Quantitation Report (QT Reviewed)

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM021924\
Data File : BM044207.D
Acq On : 20 Feb 2024 15:32
Operator : MA/JU
Sample : P1455-14
Misc :
ALS Vial : 31 Sample Multiplier: 1

Quant Time: Feb 20 16:14:07 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM021524.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Feb 15 13:40:24 2024
Response via : Initial Calibration

Instrument :
BNA_M
ClientSampleId :
C0AG0

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 02/21/2024
Supervised By :mohammad ahmed 02/21/2024

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.896	152	5119	0.400	ng/ul	0.00
4) Naphthalene-d8	10.694	136	15644	0.400	ng/ul	-0.01
9) Acenaphthene-d10	14.538	164	7858	0.400	ng/ul	# 0.00
13) Phenanthrene-d10	17.292	188	16154m	0.400	ng/ul	0.00
17) Chrysene-d12	21.497	240	11124m	0.400	ng/ul	0.00
23) Perylene-d12	23.882	264	18508	0.400	ng/ul	-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.348	96	32562	4.492	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.288	152	9609	0.439	ng/ul	-0.01
18) Fluoranthene-d10	19.326	212	18429	0.555	ng/ul	0.00
Target Compounds						Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quantitation Report (QT Reviewed)

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM021924\
Data File : BM044207.D
Acq On : 20 Feb 2024 15:32
Operator : MA/JU
Sample : P1455-14
Misc :
ALS Vial : 31 Sample Multiplier: 1

Quant Time: Feb 20 16:14:07 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM021524.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Feb 15 13:40:24 2024
Response via : Initial Calibration

Instrument :
BNA_M
ClientSampleId :
C0AG0

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 02/21/2024
Supervised By :mohammad ahmed 02/21/2024