

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM021924\
 Data File : BM044203.D
 Acq On : 20 Feb 2024 13:07
 Operator : MA/JU
 Sample : P1455-10
 Misc :
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 C0AE3

Manual Integrations
 APPROVED


Reviewed By :Yogesh Patel 02/21/2024
 Supervised By :mohammad ahmed 02/21/2024

Quant Time: Feb 20 13:47:01 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM021524.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Feb 15 13:40:24 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.896	152	4986	0.400	ng/ul	0.00
4) Naphthalene-d8	10.694	136	14749	0.400	ng/ul	-0.01
9) Acenaphthene-d10	14.539	164	7028	0.400	ng/ul	# 0.00
13) Phenanthrene-d10	17.289	188	14387m	0.400	ng/ul	-0.01
17) Chrysene-d12	21.496	240	10092m	0.400	ng/ul	0.00
23) Perylene-d12	23.881	264	17894	0.400	ng/ul	-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.348	96	29708	4.207	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.288	152	8581	0.416	ng/ul	-0.01
18) Fluoranthene-d10	19.327	212	16820	0.559	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.386	88	8937	1.266	ng/ul#	1

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
 1,4-Dioxane
 Concen: 1.266 ng/ul
 RT: 3.386 min Scan# 71
 Delta R.T. -0.004 min
 Lab File: BM044203.D
 Acq: 20 Feb 2024 13:07

Instrument :
 BNA_M
 ClientSampleId :
 C0AE3



Tgt Ion: 88 Resp: 893
 Ion Ratio Lower Upper
 88 100
 43 159.3 32.2 48.2
 58 108.3 40.6 60.8

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 02/21/2024
 Supervised By :mohammad ahmed 02/21/2024

