

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM022019\
 Data File : BM018887.D
 Acq On : 20 Feb 2019 17:15
 Operator : JU/SJ
 Sample : K1520-01
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 EFF-WW

Manual Integrations
 APPROVED

Sohil
 2/21/2019 10:08:56 AM

Quant Time: Feb 21 01:16:06 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\8270-BM021119.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Feb 11 16:02:12 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.70	152	276568	20.00	ng	-0.02
21) Naphthalene-d8	10.49	136	1185013	20.00	ng	-0.01
39) Acenaphthene-d10	14.35	164	723889	20.00	ng	-0.01
64) Phenanthrene-d10	17.10	188	1671581	20.00	ng	-0.01
76) Chrysene-d12	21.30	240	1612286	20.00	ng	-0.01
87) Perylene-d12	23.55	264	1403272	20.00	ng	-0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.30	112	1147876	68.00	ng	0.00
7) Phenol-d6	7.00	99	1001240m	42.11	ng	0.11
23) Nitrobenzene-d5	8.87	82	2382607	92.97	ng	-0.01
42) 2,4,6-Tribromophenol	15.85	330	1571827	150.64	ng	-0.01
45) 2-Fluorobiphenyl	12.97	172	4699759	86.18	ng	-0.01
79) Terphenyl-d14	19.74	244	8377655	107.19	ng	0.00
Target Compounds						
10) Phenol	7.03	94	63866m	2.553	ng	Qvalue
50) Dimethylphthalate	13.82	163	153809	2.550	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM022019\
Data File : BM018887.D
Acq On : 20 Feb 2019 17:15
Operator : JU/SJ
Sample : K1520-01
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
EFF-WW

Manual Integrations
APPROVED

Sohil
2/21/2019 10:08:56 AM

Quant Time: Feb 21 01:16:06 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\8270-BM021119.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Feb 11 16:02:12 2019
Response via : Initial Calibration

