

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM022124\
 Data File : BM044228.D
 Acq On : 21 Feb 2024 21:37
 Operator : MA/JU
 Sample : PB159095BS
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SBLK097

Quant Time: Feb 21 22:43:54 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM021524.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Feb 15 13:40:24 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.896	152	4764	0.400	ng/ul	0.00
4) Naphthalene-d8	10.699	136	12861	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.552	164	4563	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.313	188	10722	0.400	ng/ul	0.01
17) Chrysene-d12	21.512	240	8374	0.400	ng/ul	0.01
23) Perylene-d12	23.906	264	9452	0.400	ng/ul	# 0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.352	96	4739	0.702	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.299	152	5136	0.285	ng/ul	0.00
18) Fluoranthene-d10	19.336	212	9976	0.399	ng/ul	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM022124\
Data File : BM044228.D
Acq On : 21 Feb 2024 21:37
Operator : MA/JU
Sample : PB159095BS
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SBLK097

Quant Time: Feb 21 22:43:54 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM021524.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Feb 15 13:40:24 2024
Response via : Initial Calibration

