

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM022124\
 Data File : BM044240.D
 Acq On : 22 Feb 2024 15:06
 Operator : MA/JU
 Sample : P1494-04
 Misc :
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 BH3Y6

Manual Integrations
 APPROVED

Reviewed By :Jagrut Upadhyay 02/22/2024
 Supervised By :mohammad ahmed 02/23/2024

Quant Time: Feb 22 15:59:10 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM021524.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Feb 15 13:40:24 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.896	152	4348	0.400	ng/ul	0.00
4) Naphthalene-d8	10.693	136	13402	0.400	ng/ul	-0.01
9) Acenaphthene-d10	14.539	164	6746	0.400	ng/ul	# 0.00
13) Phenanthrene-d10	17.293	188	14436m	0.400	ng/ul	0.00
17) Chrysene-d12	21.499	240	11007m	0.400	ng/ul	0.00
23) Perylene-d12	23.887	264	19340	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.348	96	24739	4.018	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.288	152	6481	0.345	ng/ul	-0.01
18) Fluoranthene-d10	19.327	212	17278	0.526	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.386	88	421	0.068	ng/ul#	66

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM022124\
Data File : BM044240.D
Acq On : 22 Feb 2024 15:06
Operator : MA/JU
Sample : P1494-04
Misc :
ALS Vial : 28 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
BH3Y6

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 02/22/2024
Supervised By :mohammad ahmed 02/23/2024

Quant Time: Feb 22 15:59:10 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM021524.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Feb 15 13:40:24 2024
Response via : Initial Calibration

