

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM022124\
 Data File : BM044235.D
 Acq On : 22 Feb 2024 12:05
 Operator : MA/JU
 Sample : P1493-12
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 BH3X5

Manual Integrations
 APPROVED

Reviewed By :Jagrut Upadhyay 02/22/2024
 Supervised By :mohammad ahmed 02/23/2024

Quant Time: Feb 22 12:33:37 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM021524.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Feb 15 13:40:24 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.896	152	4361	0.400	ng/u1	0.00
4) Naphthalene-d8	10.693	136	13618	0.400	ng/u1	-0.01
9) Acenaphthene-d10	14.538	164	6791	0.400	ng/u1	# 0.00
13) Phenanthrene-d10	17.292	188	13855m	0.400	ng/u1	0.00
17) Chrysene-d12	21.497	240	10618m	0.400	ng/u1	0.00
23) Perylene-d12	23.885	264	18835	0.400	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.348	96	30773	4.983	ng/u1	0.00
6) 2-Methylnaphthalene-d10	12.288	152	10266	0.538	ng/u1	-0.01
18) Fluoranthene-d10	19.326	212	19683	0.621	ng/u1	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM022124\
Data File : BM044235.D
Acq On : 22 Feb 2024 12:05
Operator : MA/JU
Sample : P1493-12
Misc :
ALS Vial : 23 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
BH3X5

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 02/22/2024
Supervised By :mohammad ahmed 02/23/2024

Quant Time: Feb 22 12:33:37 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM021524.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Feb 15 13:40:24 2024
Response via : Initial Calibration

