

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM022216\
 Data File : BM004326.D
 Acq On : 23 Feb 2016 00:08
 Operator : SJ/UM
 Sample : H1499-11
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 0311-S15-0003

Quant Time: Feb 23 06:46:49 2016
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM022216.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Feb 23 00:11:03 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.62	152	28367	20.00	ng/ul	0.00
18) Naphthalene-d8	10.41	136	137868	20.00	ng/ul	0.00
36) Acenaphthene-d10	14.29	164	87216	20.00	ng/ul	0.00
62) Phenanthrene-d10	17.04	188	205086	20.00	ng/ul	0.00
78) Chrysene-d12	21.25	240	238340	20.00	ng/ul	0.00
86) Perylene-d12	23.47	264	202788	20.00	ng/ul	-0.01

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
3) 1,4-Dioxane-d8	3.08	96	3467	6.44	ng/uL	0.00
5) Phenol-d5	6.82	99	76195	32.03	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	6.96	67	43207	35.72	ng/ul	0.00
9) 2-Chlorophenol-d4	7.16	132	69522	33.89	ng/ul	0.00
13) 4-Methylphenol-d8	8.35	113	70480	33.16	ng/ul	0.00
19) Nitrobenzene-d5	8.79	128	36046	34.37	ng/ul	0.00
22) 2-Nitrophenol-d4	9.50	143	41269	34.52	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.05	165	70415	32.58	ng/ul	0.00
29) 4-Chloroaniline-d4	10.56	131	104645	38.85	ng/ul	0.00
44) Dimethylphthalate-d6	13.70	166	266410	35.62	ng/ul	0.00
47) Acenaphthylene-d8	13.97	160	322806	36.86	ng/ul	0.00
52) 4-Nitrophenol-d4	14.53	143	34809	28.95	ng/ul	0.00
58) Fluorene-d10	15.29	176	233255	36.67	ng/ul	0.00
63) 4,6-Dinitro-2-methylphenol	15.43	200	34725	27.56	ng/ul	0.00
71) Anthracene-d10	17.14	188	377382	37.55	ng/ul	0.00
79) Pyrene-d10	19.44	212	434207	37.66	ng/ul	0.00
90) Benzo(a)pyrene-d12	23.33	264	412045	38.72	ng/ul	0.00


Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
45) Dimethylphthalate	13.74	163	123213	15.80	ng/ul	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM022216\
Data File : BM004326.D
Acq On : 23 Feb 2016 00:08
Operator : SJ/UM
Sample : H1499-11
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
0311-S15-0003

Quant Time: Feb 23 06:46:49 2016
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM022216.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Feb 23 00:11:03 2016
Response via : Initial Calibration





#45
 Dimethylphthalate
 Concen: 15.80 ng/ul
 RT: 13.74 min Scan# 1880
 Delta R.T. -0.00 min
 Lab File: BM004326.D
 Acq: 23 Feb 2016 00:08

Instrument :
 BNA_M
 ClientSampleId :
 0311-S15-0003

Tgt Ion:	163	Resp:	123213
Ion	Ratio	Lower	Upper
163	100		
194	4.0	3.2	4.8
164	10.1	7.8	11.8

