

Data Path : Z:\HPCHEM1\BNA M\Data\BM022318\  
 Data File : BM014332.D  
 Acq On : 23 Feb 2018 15:56  
 Operator : SJ/JU  
 Sample : SSTDC0020  
 Misc :  
 ALS Vial : 2 Sample Multiplier: 1

Instrument :  
 BNA\_M  
 LabSampleId :  
 SSTD02001

Quant Time: Feb 23 16:30:54 2018  
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM-EPA-BM021418.M  
 Quant Title : SVOA CALIBRATION  
 QLast Update : Thu Feb 22 03:35:18 2018  
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min  
 Max. RRF Dev : 20% Max. Rel. Area : 150%

|      | Compound                    | AvqRF | CCRF  | %Dev   | Area% | Dev(min) |
|------|-----------------------------|-------|-------|--------|-------|----------|
| 1 I  | 1,4-Dichlorobenzene-d4      | 1.000 | 1.000 | 0.0    | 79    | 0.00     |
| 2    | 1,4-Dioxane                 | 0.519 | 0.579 | -11.6  | 92    | 0.00     |
| 3 S  | 1,4-Dioxane-d8              | 0.467 | 0.512 | -9.6   | 93    | 0.00     |
| 4    | Benzaldehyde                | 0.689 | 0.830 | -20.5# | 83    | 0.00     |
| 5 S  | Phenol-d5                   | 1.690 | 1.580 | 6.5    | 80    | 0.00     |
| 6    | Phenol                      | 1.774 | 1.581 | 10.9   | 76    | 0.00     |
| 7 S  | Bis-(2-Chloroethyl)ether-d8 | 1.023 | 1.022 | 0.1    | 82    | 0.00     |
| 8    | Bis(2-Chloroethyl)ether     | 1.287 | 1.254 | 2.6    | 80    | 0.00     |
| 9 S  | 2-Chlorophenol-d4           | 1.296 | 1.255 | 3.2    | 77    | 0.00     |
| 10   | 2-Chlorophenol              | 1.323 | 1.259 | 4.8    | 81    | 0.00     |
| 11   | 2-Methylphenol              | 1.357 | 1.271 | 6.3    | 81    | 0.00     |
| 12   | 2,2'-oxybis(1-Chloropropane | 1.299 | 1.208 | 7.0    | 79    | 0.00     |
| 13 S | 4-Methylphenol-d8           | 1.479 | 1.337 | 9.6    | 77    | 0.00     |
| 14   | Acetophenone                | 2.552 | 2.423 | 5.1    | 84    | 0.00     |
| 15 P | N-Nitroso-di-n-propylamine  | 1.279 | 1.334 | -4.3   | 88    | 0.00     |
| 16   | 4-Methylphenol              | 1.478 | 1.405 | 4.9    | 82    | 0.00     |
| 17   | Hexachloroethane            | 0.689 | 0.713 | -3.5   | 85    | 0.00     |
| 18 I | Naphthalene-d8              | 1.000 | 1.000 | 0.0    | 83    | 0.00     |
| 19 S | Nitrobenzene-d5             | 0.148 | 0.136 | 8.1    | 77    | 0.00     |
| 20   | Nitrobenzene                | 0.437 | 0.449 | -2.7   | 88    | 0.00     |
| 21   | Isophorone                  | 0.748 | 0.742 | 0.8    | 85    | 0.00     |
| 22 S | 2-Nitrophenol-d4            | 0.153 | 0.146 | 4.6    | 77    | 0.00     |
| 23 C | 2-Nitrophenol               | 0.166 | 0.155 | 6.6    | 77    | 0.00     |
| 24   | 2,4-Dimethylphenol          | 0.406 | 0.410 | -1.0   | 88    | 0.00     |
| 25   | Bis(2-Chloroethoxy)methane  | 0.394 | 0.389 | 1.3    | 83    | 0.00     |
| 26 S | 2,4-Dichlorophenol-d3       | 0.317 | 0.277 | 12.6   | 72    | 0.00     |
| 27 C | 2,4-Dichlorophenol          | 0.301 | 0.298 | 1.0    | 87    | 0.00     |
| 28   | Naphthalene                 | 1.027 | 0.987 | 3.9    | 83    | 0.00     |
| 29 S | 4-Chloroaniline-d4          | 0.312 | 0.340 | -9.0   | 86    | 0.00     |
| 30   | 4-Chloroaniline             | 0.318 | 0.346 | -8.8   | 85    | 0.00     |
| 31 C | Hexachlorobutadiene         | 0.239 | 0.238 | 0.4    | 87    | 0.00     |
| 32   | Caprolactam                 | 0.096 | 0.091 | 5.2    | 81    | 0.00     |
| 33 C | 4-Chloro-3-methylphenol     | 0.363 | 0.355 | 2.2    | 83    | 0.00     |
| 34   | 2-Methylnaphthalene         | 0.782 | 0.773 | 1.2    | 84    | 0.00     |
| 35 I | Acenaphthene-d10            | 1.000 | 1.000 | 0.0    | 86    | 0.00     |
| 36   | 1,2,4,5-Tetrachlorobenzene  | 0.642 | 0.606 | 5.6    | 84    | 0.00     |
| 37   | Hexachlorocyclopentadiene   | 0.349 | 0.316 | 9.5    | 92    | 0.00     |
| 38 C | 2,4,6-Trichlorophenol       | 0.401 | 0.373 | 7.0    | 82    | 0.00     |
| 39   | 2,4,5-Trichlorophenol       | 0.409 | 0.412 | -0.7   | 83    | 0.00     |
| 40   | 1,1'-Biphenyl               | 1.541 | 1.512 | 1.9    | 84    | 0.00     |
| 41   | 2-Chloronaphthalene         | 1.231 | 1.200 | 2.5    | 84    | 0.00     |
| 42   | 2-Nitroaniline              | 0.416 | 0.397 | 4.6    | 82    | 0.00     |
| 43 S | Dimethylphthalate-d6        | 1.652 | 1.594 | 3.5    | 84    | 0.00     |
| 44   | Dimethylphthalate           | 1.629 | 1.585 | 2.7    | 87    | 0.00     |

Data Path : Z:\HPCHEM1\BNA M\Data\BM022318\  
 Data File : BM014332.D  
 Acq On : 23 Feb 2018 15:56  
 Operator : SJ/JU  
 Sample : SSTDC0020  
 Misc :  
 ALS Vial : 2 Sample Multiplier: 1

Instrument :  
 BNA\_M  
 LabSampleId :  
 SSTD02001

Quant Time: Feb 23 16:30:54 2018  
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM-EPA-BM021418.M  
 Quant Title : SVOA CALIBRATION  
 QLast Update : Thu Feb 22 03:35:18 2018  
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min  
 Max. RRF Dev : 20% Max. Rel. Area : 150%

|      | Compound                    | AvgRF | CCRF  | %Dev  | Area% | Dev(min) |
|------|-----------------------------|-------|-------|-------|-------|----------|
| 45   | 2,6-Dinitrotoluene          | 0.309 | 0.295 | 4.5   | 83    | 0.00     |
| 46 S | Acenaphthylene-d8           | 2.037 | 1.994 | 2.1   | 86    | 0.00     |
| 47   | Acenaphthylene              | 1.925 | 1.889 | 1.9   | 85    | 0.00     |
| 48   | 3-Nitroaniline              | 0.262 | 0.227 | 13.4  | 85    | 0.00     |
| 49 C | Acenaphthene                | 1.355 | 1.354 | 0.1   | 89    | 0.00     |
| 50   | 2,4-Dinitrophenol           | 0.166 | 0.118 | 28.9# | 85    | 0.00     |
| 51 S | 4-Nitrophenol-d4            | 0.228 | 0.169 | 25.9# | 79    | 0.00     |
| 52   | 4-Nitrophenol               | 0.299 | 0.252 | 15.7  | 84    | 0.00     |
| 53   | Dibenzofuran                | 2.032 | 2.045 | -0.6  | 88    | 0.00     |
| 54   | 2,4-Dinitrotoluene          | 0.496 | 0.473 | 4.6   | 82    | 0.00     |
| 55   | 2,3,4,6-Tetrachlorophenol   | 0.414 | 0.405 | 2.2   | 86    | 0.00     |
| 56   | Diethylphthalate            | 1.815 | 1.800 | 0.8   | 87    | 0.00     |
| 57 S | Fluorene-d10                | 1.480 | 1.502 | -1.5  | 89    | 0.00     |
| 58   | Fluorene                    | 1.755 | 1.716 | 2.2   | 87    | 0.00     |
| 59   | 4-Chlorophenyl-phenylether  | 0.918 | 0.914 | 0.4   | 89    | 0.00     |
| 60   | 4-Nitroaniline              | 0.294 | 0.209 | 28.9# | 75    | 0.00     |
| 61 I | Phenanthrene-d10            | 1.000 | 1.000 | 0.0   | 87    | 0.00     |
| 62 S | 4,6-Dinitro-2-methylphenol- | 0.120 | 0.101 | 15.8  | 85    | 0.00     |
| 63   | 4,6-Dinitro-2-methylphenol  | 0.122 | 0.106 | 13.1  | 85    | 0.00     |
| 64   | N-Nitrosodiphenylamine      | 0.575 | 0.559 | 2.8   | 88    | 0.00     |
| 65   | 4-Bromophenyl-phenylether   | 0.223 | 0.215 | 3.6   | 89    | 0.00     |
| 66   | Hexachlorobenzene           | 0.235 | 0.220 | 6.4   | 87    | 0.00     |
| 67   | Atrazine                    | 0.227 | 0.207 | 8.8   | 84    | 0.00     |
| 68 C | Pentachlorophenol           | 0.123 | 0.105 | 14.6  | 89    | 0.00     |
| 69   | Phenanthrene                | 1.152 | 1.088 | 5.6   | 86    | 0.00     |
| 70 S | Anthracene-d10              | 0.965 | 0.926 | 4.0   | 86    | 0.00     |
| 71   | Anthracene                  | 1.159 | 1.126 | 2.8   | 86    | 0.00     |
| 72   | Carbazole                   | 0.974 | 0.883 | 9.3   | 85    | 0.00     |
| 73   | Di-n-butylphthalate         | 1.257 | 1.186 | 5.6   | 83    | 0.00     |
| 74 C | Fluoranthene                | 1.377 | 1.288 | 6.5   | 85    | 0.00     |
| 75 I | Chrysene-d12                | 1.000 | 1.000 | 0.0   | 82    | 0.00     |
| 76 S | Pyrene-d10                  | 1.032 | 1.032 | 0.0   | 85    | 0.00     |
| 77   | Pyrene                      | 1.329 | 1.332 | -0.2  | 85    | 0.00     |
| 78   | Butylbenzylphthalate        | 0.550 | 0.521 | 5.3   | 79    | 0.00     |
| 79   | 3,3'-Dichlorobenzidine      | 0.352 | 0.304 | 13.6  | 84    | 0.00     |
| 80   | Benzo(a)anthracene          | 1.248 | 1.201 | 3.8   | 82    | 0.00     |
| 81   | Bis(2-ethylhexyl)phthalate  | 0.787 | 0.707 | 10.2  | 75    | 0.00     |
| 82   | Chrysene                    | 1.164 | 1.147 | 1.5   | 83    | 0.00     |
| 83 I | Perylene-d12                | 1.000 | 1.000 | 0.0   | 86    | 0.00     |
| 84   | Di-n-octyl phthalate        | 1.716 | 1.327 | 22.7# | 74    | 0.00     |
| 85   | Benzo(b)fluoranthene        | 1.339 | 1.301 | 2.8   | 84    | 0.00     |
| 86   | Benzo(k)fluoranthene        | 1.273 | 1.219 | 4.2   | 85    | 0.00     |
| 87 S | Benzo(a)pyrene-d12          | 0.973 | 0.948 | 2.6   | 85    | 0.00     |

Data Path : Z:\HPCHEM1\BNA M\Data\BM022318\  
Data File : BM014332.D  
Acq On : 23 Feb 2018 15:56  
Operator : SJ/JU  
Sample : SSTDCCC020  
Misc :  
ALS Vial : 2 Sample Multiplier: 1

Instrument :  
BNA\_M  
LabSampleId :  
SSTD02001

Quant Time: Feb 23 16:30:54 2018  
Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM-EPA-BM021418.M  
Quant Title : SVOA CALIBRATION  
QLast Update : Thu Feb 22 03:35:18 2018  
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min  
Max. RRF Dev : 20% Max. Rel. Area : 150%

|      | Compound               | AvgRF | CCRF  | %Dev | Area% | Dev(min) |
|------|------------------------|-------|-------|------|-------|----------|
| 88 C | Benzo(a)pyrene         | 1.197 | 1.166 | 2.6  | 88    | 0.00     |
| 89   | Indeno(1,2,3-cd)pyrene | 1.161 | 1.152 | 0.8  | 90    | 0.00     |
| 90   | Dibenzo(a,h)anthracene | 0.965 | 0.931 | 3.5  | 88    | 0.00     |
| 91   | Benzo(a,h,i)perylene   | 0.941 | 0.921 | 2.1  | 90    | 0.00     |

(#) = Out of Range

SPCC's out = 0 CCC's out = 0