


Data Path : Z:\HPCHEM1\BNA\_M\Data\BM022318\  
 Data File : BM014325.D  
 Acq On : 23 Feb 2018 11:44  
 Operator : SJ/JU  
 Sample : DFTPP02  
 Misc :  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 BNA\_M  
 ClientSampleId :  
 DFTPP02

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA\_M\METHODS\SOM-EPA-BM021418.M  
 Title : SVOA CALIBRATION  
 Last Update : Fri Feb 23 23:12:42 2018



AutoFind: Scans 2448, 2449, 2450; Background Corrected with Scan 2442

| Target Mass | Rel. to Mass | Lower Limit% | Upper Limit% | Rel. Abn% | Raw Abn | Result Pass/Fail |
|-------------|--------------|--------------|--------------|-----------|---------|------------------|
| 51          | 198          | 10           | 80           | 55.4      | 135592  | PASS             |
| 68          | 69           | 0.00         | 2            | 1.3       | 1650    | PASS             |
| 69          | 198          | 0.00         | 100          | 53.5      | 130970  | PASS             |
| 70          | 69           | 0.00         | 2            | 0.6       | 843     | PASS             |
| 127         | 198          | 10           | 80           | 52.5      | 128685  | PASS             |
| 197         | 198          | 0.00         | 2            | 0.7       | 1620    | PASS             |
| 198         | 198          | 100          | 100          | 100.0     | 244928  | PASS             |
| 199         | 198          | 5            | 9            | 7.0       | 17196   | PASS             |
| 275         | 198          | 10           | 60           | 27.4      | 67093   | PASS             |
| 365         | 198          | 1            | 100          | 5.6       | 13730   | PASS             |
| 441         | 443          | 0.01         | 100          | 82.1      | 28874   | PASS             |
| 442         | 198          | 50           | 100          | 72.1      | 176576  | PASS             |
| 443         | 442          | 15           | 24           | 19.9      | 35165   | PASS             |

| m/z    | Abundance |
|--------|-----------|
| 37.10  | 1049.0    |
| 38.00  | 2800.0    |
| 39.10  | 14132.0   |
| 40.00  | 1027.0    |
| 43.90  | 1038.0    |
| 44.90  | 397.0     |
| 49.00  | 1866.0    |
| 50.10  | 53304.0   |
| 51.00  | 178816.0  |
| 52.00  | 9371.0    |
| 52.80  | 334.0     |
| 54.90  | 1165.0    |
| 56.00  | 4881.0    |
| 57.00  | 9264.0    |
| 58.20  | 354.0     |
| 61.10  | 3225.0    |
| 62.10  | 3103.0    |
| 63.00  | 7087.0    |
| 64.00  | 2107.0    |
| 65.00  | 3309.0    |
| 67.90  | 2626.0    |
| 69.00  | 169792.0  |
| 70.00  | 691.0     |
| 73.10  | 2149.0    |
| 74.00  | 24272.0   |
| 75.00  | 27864.0   |
| 76.10  | 10917.0   |
| 77.10  | 191296.0  |
| 78.00  | 13693.0   |
| 79.00  | 16175.0   |
| 80.00  | 11398.0   |
| 81.00  | 14793.0   |
| 82.10  | 3854.0    |
| 83.10  | 2171.0    |
| 85.00  | 2098.0    |
| 86.00  | 4781.0    |
| 86.90  | 1950.0    |
| 88.00  | 702.0     |
| 89.00  | 497.0     |
| 91.00  | 3910.0    |
| 91.90  | 3302.0    |
| 93.00  | 33000.0   |
| 93.90  | 2186.0    |
| 94.90  | 447.0     |
| 96.10  | 766.0     |
| 97.10  | 709.0     |
| 98.00  | 25160.0   |
| 99.00  | 14991.0   |
| 100.00 | 1183.0    |
| 101.00 | 7408.0    |
| 101.90 | 402.0     |
| 103.00 | 2530.0    |
| 104.00 | 7300.0    |
| 105.00 | 5370.0    |
| 106.00 | 2175.0    |
| 107.00 | 64168.0   |
| 108.00 | 8459.0    |
| 109.00 | 1556.0    |
| 110.00 | 86976.0   |
| 111.00 | 13497.0   |
| 112.00 | 2006.0    |
| 112.90 | 570.0     |
| 114.90 | 328.0     |
| 116.00 | 5174.0    |
| 117.00 | 92288.0   |
| 117.90 | 5485.0    |
| 119.10 | 499.0     |
| 120.00 | 1300.0    |
| 121.00 | 487.0     |
| 122.00 | 4939.0    |
| 122.90 | 7168.0    |
| 123.90 | 2592.0    |
| 125.00 | 2588.0    |
| 126.00 | 462.0     |
| 127.00 | 156160.0  |
| 128.00 | 11724.0   |
| 129.00 | 72832.0   |
| 130.00 | 6872.0    |
| 131.10 | 941.0     |
| 134.00 | 1553.0    |
| 135.00 | 5591.0    |
| 135.90 | 2579.0    |
| 137.10 | 4208.0    |
| 137.70 | 418.0     |
| 138.20 | 375.0     |
| 138.90 | 851.0     |
| 140.00 | 777.0     |
| 141.00 | 12383.0   |
| 141.90 | 4055.0    |
| 143.00 | 2057.0    |
| 143.90 | 1082.0    |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP02

|        |          |
|--------|----------|
| 145.10 | 598.0    |
| 146.00 | 2507.0   |
| 147.00 | 7293.0   |
| 148.00 | 18976.0  |
| 149.00 | 3687.0   |
| 149.90 | 645.0    |
| 151.00 | 1239.0   |
| 151.70 | 637.0    |
| 152.90 | 3995.0   |
| 154.10 | 2731.0   |
| 155.00 | 5756.0   |
| 156.00 | 7355.0   |
| 157.00 | 1563.0   |
| 158.00 | 2071.0   |
| 158.90 | 908.0    |
| 160.00 | 3390.0   |
| 160.90 | 5044.0   |
| 161.90 | 1700.0   |
| 163.00 | 772.0    |
| 164.00 | 473.0    |
| 164.90 | 4348.0   |
| 166.10 | 2871.0   |
| 167.00 | 26384.0  |
| 167.90 | 11511.0  |
| 169.00 | 1363.0   |
| 170.00 | 1285.0   |
| 170.90 | 1018.0   |
| 172.00 | 2278.0   |
| 173.10 | 2373.0   |
| 174.10 | 4393.0   |
| 175.00 | 8490.0   |
| 176.00 | 2036.0   |
| 177.00 | 3660.0   |
| 178.10 | 1512.0   |
| 179.00 | 14428.0  |
| 180.00 | 9164.0   |
| 181.10 | 4974.0   |
| 181.90 | 790.0    |
| 183.10 | 488.0    |
| 184.00 | 1632.0   |
| 185.00 | 7797.0   |
| 186.00 | 47552.0  |
| 187.00 | 15806.0  |
| 188.10 | 1016.0   |
| 188.90 | 4773.0   |
| 189.80 | 1388.0   |
| 191.00 | 3121.0   |
| 192.00 | 6447.0   |
| 193.00 | 5290.0   |
| 194.00 | 1777.0   |
| 195.00 | 630.0    |
| 196.00 | 8154.0   |
| 197.00 | 2323.0   |
| 197.90 | 283904.0 |
| 199.00 | 21864.0  |
| 200.00 | 1887.0   |
| 201.20 | 1368.0   |
| 203.00 | 2984.0   |
| 204.00 | 15447.0  |
| 205.00 | 23712.0  |
| 206.00 | 90544.0  |
| 207.10 | 12867.0  |
| 208.00 | 4452.0   |
| 209.00 | 1490.0   |
| 209.90 | 1965.0   |
| 210.90 | 3301.0   |
| 211.70 | 556.0    |
| 213.10 | 380.0    |
| 214.90 | 1689.0   |
| 215.90 | 2969.0   |
| 217.00 | 29896.0  |
| 218.00 | 4333.0   |
| 221.00 | 7242.0   |
| 221.80 | 4776.0   |
| 223.00 | 5878.0   |
| 224.00 | 54168.0  |
| 225.00 | 13091.0  |
| 226.00 | 1984.0   |
| 227.00 | 26808.0  |
| 227.90 | 3472.0   |
| 229.00 | 5984.0   |
| 229.90 | 661.0    |
| 230.80 | 1248.0   |
| 231.20 | 1423.0   |
| 232.00 | 584.0    |
| 232.90 | 609.0    |
| 234.00 | 2015.0   |
| 235.00 | 1974.0   |
| 236.00 | 2293.0   |
| 237.10 | 1256.0   |
| 237.80 | 379.0    |
| 238.90 | 623.0    |
| 240.20 | 588.0    |
| 241.10 | 1210.0   |
| 242.00 | 2843.0   |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP02

|        |          |
|--------|----------|
| 243.10 | 3359.0   |
| 244.00 | 38552.0  |
| 245.00 | 5172.0   |
| 246.00 | 10695.0  |
| 247.00 | 1394.0   |
| 248.00 | 525.0    |
| 249.00 | 1768.0   |
| 252.90 | 667.0    |
| 254.20 | 1746.0   |
| 255.00 | 179200.0 |
| 256.00 | 30400.0  |
| 257.00 | 1926.0   |
| 258.00 | 15683.0  |
| 259.00 | 2360.0   |
| 259.90 | 588.0    |
| 261.00 | 418.0    |
| 263.90 | 808.0    |
| 265.10 | 6418.0   |
| 265.90 | 1240.0   |
| 271.10 | 847.0    |
| 272.00 | 790.0    |
| 273.00 | 4631.0   |
| 274.00 | 17464.0  |
| 275.00 | 70320.0  |
| 276.00 | 8662.0   |
| 277.00 | 8898.0   |
| 277.90 | 1227.0   |
| 281.90 | 317.0    |
| 283.00 | 1137.0   |
| 284.00 | 776.0    |
| 284.90 | 1639.0   |
| 285.90 | 426.0    |
| 290.10 | 370.0    |
| 290.90 | 350.0    |
| 292.00 | 301.0    |
| 293.00 | 2519.0   |
| 294.00 | 505.0    |
| 296.00 | 30456.0  |
| 297.00 | 4401.0   |
| 302.00 | 471.0    |
| 303.10 | 2500.0   |
| 304.00 | 821.0    |
| 307.90 | 405.0    |
| 314.00 | 1408.0   |
| 315.00 | 3752.0   |
| 316.10 | 2149.0   |
| 317.10 | 334.0    |
| 320.90 | 1149.0   |
| 323.00 | 5844.0   |
| 324.10 | 1357.0   |
| 325.90 | 446.0    |
| 327.00 | 2341.0   |
| 327.80 | 653.0    |
| 332.90 | 891.0    |
| 334.00 | 5410.0   |
| 335.00 | 1267.0   |
| 341.00 | 415.0    |
| 345.80 | 2065.0   |
| 346.80 | 636.0    |
| 352.10 | 1740.0   |
| 353.10 | 1197.0   |
| 354.10 | 2621.0   |
| 355.00 | 403.0    |
| 365.00 | 13637.0  |
| 366.00 | 1273.0   |
| 369.80 | 368.0    |
| 370.90 | 747.0    |
| 372.10 | 3080.0   |
| 373.00 | 686.0    |
| 383.00 | 791.0    |
| 401.80 | 1093.0   |
| 403.10 | 1490.0   |
| 404.10 | 549.0    |
| 421.10 | 1679.0   |
| 421.90 | 1752.0   |
| 423.10 | 8299.0   |
| 424.10 | 2507.0   |
| 441.10 | 20152.0  |
| 442.10 | 150528.0 |
| 443.10 | 30440.0  |
| 444.20 | 2979.0   |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP02

| m/z    | Abundance |
|--------|-----------|
| 37.10  | 1149.0    |
| 37.90  | 2145.0    |
| 39.00  | 10580.0   |
| 39.90  | 1135.0    |
| 43.90  | 647.0     |
| 49.10  | 1280.0    |
| 50.00  | 50296.0   |
| 51.00  | 160832.0  |
| 52.00  | 8592.0    |
| 55.10  | 505.0     |
| 56.00  | 4753.0    |
| 57.00  | 9007.0    |
| 58.00  | 615.0     |
| 61.00  | 1557.0    |
| 62.10  | 3335.0    |
| 63.00  | 6555.0    |
| 64.00  | 871.0     |
| 65.00  | 4089.0    |
| 67.90  | 2324.0    |
| 69.00  | 153472.0  |
| 70.00  | 691.0     |
| 73.00  | 2383.0    |
| 74.10  | 21016.0   |
| 75.00  | 25864.0   |
| 76.10  | 11394.0   |
| 77.10  | 169664.0  |
| 78.00  | 11103.0   |
| 79.00  | 16380.0   |
| 80.00  | 10252.0   |
| 81.00  | 13935.0   |
| 82.00  | 4082.0    |
| 83.00  | 2631.0    |
| 85.00  | 2583.0    |
| 86.00  | 5408.0    |
| 87.00  | 1758.0    |
| 87.80  | 566.0     |
| 89.00  | 398.0     |
| 91.00  | 2516.0    |
| 92.00  | 3252.0    |
| 93.00  | 31840.0   |
| 94.00  | 2124.0    |
| 96.00  | 1270.0    |
| 97.20  | 398.0     |
| 98.00  | 25424.0   |
| 99.00  | 16448.0   |
| 100.10 | 900.0     |
| 100.90 | 7987.0    |
| 102.00 | 477.0     |
| 103.00 | 1844.0    |
| 104.00 | 5076.0    |
| 105.00 | 5291.0    |
| 106.10 | 1374.0    |
| 107.00 | 58352.0   |
| 108.00 | 9864.0    |
| 108.90 | 1761.0    |
| 110.00 | 88608.0   |
| 111.00 | 15314.0   |
| 112.00 | 2346.0    |
| 112.80 | 582.0     |
| 114.80 | 300.0     |
| 115.90 | 3192.0    |
| 117.00 | 86984.0   |
| 118.00 | 5456.0    |
| 119.80 | 390.0     |
| 120.90 | 354.0     |
| 121.90 | 4836.0    |
| 123.00 | 9222.0    |
| 124.00 | 3564.0    |
| 124.90 | 2716.0    |
| 126.00 | 915.0     |
| 127.00 | 157888.0  |
| 128.00 | 14599.0   |
| 129.00 | 73464.0   |
| 130.00 | 5281.0    |
| 131.00 | 1380.0    |
| 132.00 | 819.0     |
| 132.70 | 568.0     |
| 134.00 | 2220.0    |
| 135.00 | 5980.0    |
| 135.90 | 2579.0    |
| 137.00 | 3200.0    |
| 138.10 | 662.0     |
| 138.90 | 306.0     |
| 140.00 | 1647.0    |
| 141.00 | 11405.0   |
| 142.00 | 2481.0    |
| 142.90 | 2917.0    |
| 144.10 | 900.0     |
| 144.90 | 639.0     |
| 146.00 | 2503.0    |
| 146.90 | 5233.0    |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP02

|        |          |
|--------|----------|
| 148.00 | 19192.0  |
| 149.00 | 2618.0   |
| 150.00 | 832.0    |
| 151.20 | 716.0    |
| 152.00 | 621.0    |
| 153.00 | 4611.0   |
| 154.10 | 2139.0   |
| 155.00 | 5892.0   |
| 156.00 | 9857.0   |
| 157.00 | 1848.0   |
| 158.10 | 1866.0   |
| 159.00 | 2196.0   |
| 160.00 | 3907.0   |
| 161.00 | 4500.0   |
| 162.00 | 1441.0   |
| 163.00 | 351.0    |
| 164.00 | 466.0    |
| 164.90 | 5318.0   |
| 166.00 | 3800.0   |
| 167.00 | 28768.0  |
| 168.00 | 14401.0  |
| 169.10 | 1624.0   |
| 170.00 | 454.0    |
| 171.00 | 698.0    |
| 171.90 | 3010.0   |
| 173.00 | 2807.0   |
| 174.00 | 5362.0   |
| 175.00 | 8484.0   |
| 176.10 | 2964.0   |
| 177.00 | 4570.0   |
| 177.90 | 2285.0   |
| 178.90 | 17128.0  |
| 180.00 | 10871.0  |
| 181.10 | 4383.0   |
| 182.20 | 1071.0   |
| 183.00 | 730.0    |
| 184.10 | 1732.0   |
| 185.00 | 7510.0   |
| 186.00 | 55952.0  |
| 187.00 | 15539.0  |
| 187.80 | 1449.0   |
| 189.00 | 5321.0   |
| 189.90 | 880.0    |
| 190.90 | 2400.0   |
| 191.90 | 4566.0   |
| 193.00 | 6859.0   |
| 194.10 | 1952.0   |
| 194.70 | 816.0    |
| 195.90 | 8667.0   |
| 197.10 | 2538.0   |
| 198.00 | 294272.0 |
| 199.00 | 20328.0  |
| 200.00 | 1657.0   |
| 200.90 | 711.0    |
| 201.70 | 1241.0   |
| 202.90 | 3805.0   |
| 204.00 | 14345.0  |
| 205.00 | 26544.0  |
| 206.10 | 97216.0  |
| 207.00 | 13618.0  |
| 207.90 | 4766.0   |
| 208.90 | 948.0    |
| 210.00 | 1889.0   |
| 211.10 | 3970.0   |
| 212.10 | 596.0    |
| 215.10 | 1752.0   |
| 216.00 | 2274.0   |
| 217.00 | 32816.0  |
| 217.90 | 4083.0   |
| 220.20 | 960.0    |
| 221.00 | 9930.0   |
| 222.00 | 2901.0   |
| 223.10 | 6893.0   |
| 224.00 | 60680.0  |
| 225.00 | 12481.0  |
| 225.90 | 1327.0   |
| 227.00 | 35584.0  |
| 228.00 | 4648.0   |
| 229.00 | 6163.0   |
| 229.90 | 832.0    |
| 231.10 | 2701.0   |
| 232.00 | 378.0    |
| 233.00 | 330.0    |
| 234.00 | 2134.0   |
| 235.00 | 2738.0   |
| 235.90 | 1234.0   |
| 237.00 | 2451.0   |
| 238.90 | 1434.0   |
| 240.00 | 626.0    |
| 240.90 | 1794.0   |
| 242.00 | 3229.0   |
| 243.00 | 3836.0   |
| 244.10 | 42648.0  |
| 245.10 | 5712.0   |
| 246.00 | 11199.0  |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP02

|        |          |
|--------|----------|
| 247.00 | 1271.0   |
| 248.10 | 487.0    |
| 249.00 | 1171.0   |
| 251.20 | 424.0    |
| 251.90 | 538.0    |
| 252.90 | 1205.0   |
| 254.10 | 1544.0   |
| 255.00 | 217984.0 |
| 256.00 | 32464.0  |
| 257.00 | 3569.0   |
| 258.00 | 18104.0  |
| 259.00 | 3385.0   |
| 263.00 | 566.0    |
| 264.20 | 341.0    |
| 265.00 | 7579.0   |
| 266.00 | 1185.0   |
| 271.20 | 485.0    |
| 272.20 | 759.0    |
| 272.90 | 5906.0   |
| 274.00 | 21936.0  |
| 275.00 | 82072.0  |
| 276.10 | 9756.0   |
| 277.00 | 11449.0  |
| 278.10 | 2045.0   |
| 278.90 | 556.0    |
| 280.90 | 488.0    |
| 283.10 | 772.0    |
| 283.80 | 655.0    |
| 284.90 | 1941.0   |
| 286.00 | 557.0    |
| 289.90 | 406.0    |
| 291.00 | 558.0    |
| 292.10 | 369.0    |
| 292.80 | 1852.0   |
| 293.90 | 774.0    |
| 295.10 | 870.0    |
| 296.00 | 40048.0  |
| 297.00 | 6488.0   |
| 297.90 | 615.0    |
| 300.90 | 352.0    |
| 302.10 | 653.0    |
| 303.00 | 3923.0   |
| 304.00 | 875.0    |
| 309.00 | 501.0    |
| 309.80 | 543.0    |
| 313.10 | 670.0    |
| 314.00 | 1394.0   |
| 315.00 | 3965.0   |
| 316.10 | 2112.0   |
| 317.10 | 505.0    |
| 321.10 | 885.0    |
| 323.10 | 8832.0   |
| 324.10 | 2415.0   |
| 325.20 | 308.0    |
| 326.00 | 357.0    |
| 327.00 | 1839.0   |
| 328.00 | 1181.0   |
| 328.90 | 405.0    |
| 332.20 | 672.0    |
| 332.80 | 752.0    |
| 334.00 | 6632.0   |
| 335.10 | 1469.0   |
| 340.90 | 856.0    |
| 346.00 | 3345.0   |
| 346.90 | 661.0    |
| 352.10 | 3441.0   |
| 353.00 | 1939.0   |
| 354.10 | 2233.0   |
| 355.00 | 1061.0   |
| 361.20 | 321.0    |
| 365.00 | 16792.0  |
| 366.00 | 2062.0   |
| 370.00 | 310.0    |
| 371.20 | 640.0    |
| 372.10 | 5345.0   |
| 372.90 | 1346.0   |
| 376.90 | 366.0    |
| 383.00 | 1382.0   |
| 384.00 | 371.0    |
| 390.20 | 340.0    |
| 391.00 | 596.0    |
| 402.00 | 1379.0   |
| 403.00 | 1960.0   |
| 404.00 | 783.0    |
| 421.20 | 1673.0   |
| 422.00 | 1572.0   |
| 423.00 | 11422.0  |
| 424.10 | 3173.0   |
| 438.70 | 471.0    |
| 439.80 | 379.0    |
| 441.00 | 37296.0  |
| 442.10 | 219840.0 |
| 443.10 | 45008.0  |
| 444.10 | 4813.0   |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP02

| m/z    | Abundance |
|--------|-----------|
| 37.00  | 499.0     |
| 37.90  | 1027.0    |
| 39.10  | 4104.0    |
| 40.00  | 333.0     |
| 44.00  | 420.0     |
| 49.10  | 619.0     |
| 50.00  | 19904.0   |
| 51.10  | 67128.0   |
| 52.10  | 3520.0    |
| 54.90  | 484.0     |
| 56.10  | 2257.0    |
| 56.90  | 2783.0    |
| 61.10  | 978.0     |
| 61.90  | 912.0     |
| 63.00  | 3071.0    |
| 64.20  | 528.0     |
| 65.00  | 1350.0    |
| 68.10  | 1683.0    |
| 69.00  | 69648.0   |
| 69.90  | 1149.0    |
| 72.90  | 948.0     |
| 74.00  | 10418.0   |
| 75.00  | 12413.0   |
| 76.10  | 4263.0    |
| 77.10  | 74776.0   |
| 78.00  | 5295.0    |
| 79.10  | 6564.0    |
| 79.90  | 3753.0    |
| 81.00  | 6801.0    |
| 81.90  | 924.0     |
| 82.90  | 782.0     |
| 84.90  | 694.0     |
| 86.00  | 3093.0    |
| 86.90  | 898.0     |
| 90.90  | 1905.0    |
| 92.10  | 1478.0    |
| 93.00  | 15487.0   |
| 94.00  | 704.0     |
| 96.00  | 711.0     |
| 96.80  | 392.0     |
| 98.00  | 12668.0   |
| 99.00  | 9132.0    |
| 99.80  | 416.0     |
| 100.30 | 580.0     |
| 101.00 | 3303.0    |
| 103.00 | 910.0     |
| 104.00 | 2708.0    |
| 105.00 | 1728.0    |
| 106.10 | 873.0     |
| 107.00 | 27200.0   |
| 108.00 | 3272.0    |
| 109.00 | 843.0     |
| 110.00 | 39592.0   |
| 110.90 | 7834.0    |
| 112.10 | 1002.0    |
| 113.00 | 740.0     |
| 116.20 | 1773.0    |
| 117.00 | 45864.0   |
| 118.00 | 3346.0    |
| 118.90 | 402.0     |
| 119.90 | 307.0     |
| 122.00 | 2329.0    |
| 123.00 | 3749.0    |
| 123.90 | 1455.0    |
| 124.90 | 1104.0    |
| 127.00 | 72008.0   |
| 128.00 | 7072.0    |
| 129.00 | 37904.0   |
| 130.00 | 3312.0    |
| 130.80 | 753.0     |
| 133.10 | 330.0     |
| 134.10 | 855.0     |
| 135.00 | 3897.0    |
| 136.00 | 1400.0    |
| 137.10 | 2298.0    |
| 139.90 | 1040.0    |
| 141.00 | 6492.0    |
| 142.10 | 1911.0    |
| 143.00 | 1921.0    |
| 146.10 | 1534.0    |
| 147.00 | 3119.0    |
| 148.00 | 9277.0    |
| 148.90 | 1138.0    |
| 150.10 | 668.0     |
| 150.80 | 420.0     |
| 151.60 | 631.0     |
| 152.90 | 2001.0    |
| 154.00 | 1257.0    |
| 155.00 | 3958.0    |
| 156.10 | 4149.0    |
| 157.00 | 829.0     |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP02

|        |          |
|--------|----------|
| 157.90 | 1258.0   |
| 158.80 | 950.0    |
| 159.90 | 2418.0   |
| 160.90 | 3846.0   |
| 161.90 | 409.0    |
| 162.90 | 337.0    |
| 165.00 | 2200.0   |
| 166.00 | 1759.0   |
| 167.00 | 14877.0  |
| 168.00 | 8060.0   |
| 169.00 | 1611.0   |
| 170.00 | 460.0    |
| 171.00 | 1216.0   |
| 172.00 | 1056.0   |
| 172.80 | 1151.0   |
| 173.90 | 1893.0   |
| 175.00 | 5043.0   |
| 175.80 | 596.0    |
| 176.10 | 654.0    |
| 177.00 | 2442.0   |
| 178.00 | 1319.0   |
| 179.00 | 8073.0   |
| 180.00 | 5526.0   |
| 180.90 | 2805.0   |
| 183.10 | 403.0    |
| 183.90 | 762.0    |
| 185.00 | 4868.0   |
| 186.00 | 27232.0  |
| 187.00 | 9236.0   |
| 188.10 | 954.0    |
| 188.90 | 2827.0   |
| 191.00 | 830.0    |
| 192.10 | 2644.0   |
| 193.10 | 3156.0   |
| 194.00 | 994.0    |
| 195.90 | 3705.0   |
| 196.70 | 880.0    |
| 198.00 | 156608.0 |
| 199.00 | 9398.0   |
| 199.90 | 974.0    |
| 201.30 | 648.0    |
| 201.80 | 807.0    |
| 202.30 | 490.0    |
| 203.00 | 2028.0   |
| 204.00 | 9092.0   |
| 205.00 | 15930.0  |
| 206.00 | 53136.0  |
| 207.00 | 6816.0   |
| 208.00 | 1648.0   |
| 208.80 | 943.0    |
| 209.90 | 804.0    |
| 210.90 | 3011.0   |
| 212.00 | 475.0    |
| 215.10 | 645.0    |
| 215.90 | 1173.0   |
| 217.00 | 15989.0  |
| 217.90 | 2225.0   |
| 220.10 | 933.0    |
| 220.90 | 4690.0   |
| 221.90 | 2514.0   |
| 223.00 | 3563.0   |
| 224.00 | 30752.0  |
| 225.00 | 8935.0   |
| 226.00 | 1182.0   |
| 227.00 | 18920.0  |
| 227.90 | 2716.0   |
| 229.00 | 2554.0   |
| 230.00 | 477.0    |
| 231.20 | 1069.0   |
| 234.10 | 1177.0   |
| 234.80 | 668.0    |
| 236.00 | 946.0    |
| 237.00 | 967.0    |
| 237.80 | 590.0    |
| 239.10 | 1124.0   |
| 239.80 | 704.0    |
| 240.90 | 702.0    |
| 242.00 | 2917.0   |
| 243.00 | 2148.0   |
| 244.00 | 22960.0  |
| 245.00 | 2740.0   |
| 246.00 | 6786.0   |
| 247.00 | 1452.0   |
| 248.90 | 541.0    |
| 250.00 | 320.0    |
| 250.80 | 370.0    |
| 253.20 | 337.0    |
| 254.10 | 1134.0   |
| 255.00 | 122472.0 |
| 256.00 | 21848.0  |
| 257.00 | 696.0    |
| 258.00 | 10091.0  |
| 258.90 | 1451.0   |
| 263.90 | 321.0    |
| 265.00 | 4773.0   |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP02

|        |          |
|--------|----------|
| 265.80 | 765.0    |
| 270.90 | 324.0    |
| 272.00 | 429.0    |
| 273.00 | 3659.0   |
| 274.00 | 11165.0  |
| 275.00 | 48888.0  |
| 276.00 | 6890.0   |
| 277.00 | 6511.0   |
| 278.00 | 1039.0   |
| 281.00 | 513.0    |
| 282.20 | 385.0    |
| 282.90 | 496.0    |
| 284.00 | 875.0    |
| 285.10 | 978.0    |
| 292.90 | 872.0    |
| 294.00 | 343.0    |
| 295.00 | 549.0    |
| 296.00 | 28128.0  |
| 297.00 | 3137.0   |
| 297.80 | 358.0    |
| 300.90 | 534.0    |
| 302.00 | 465.0    |
| 303.10 | 1731.0   |
| 304.00 | 364.0    |
| 309.00 | 499.0    |
| 314.00 | 1203.0   |
| 315.10 | 2835.0   |
| 315.90 | 716.0    |
| 316.90 | 332.0    |
| 321.10 | 730.0    |
| 323.00 | 5071.0   |
| 324.10 | 900.0    |
| 326.00 | 330.0    |
| 327.00 | 2316.0   |
| 328.10 | 437.0    |
| 331.90 | 570.0    |
| 333.00 | 565.0    |
| 334.00 | 3961.0   |
| 335.10 | 974.0    |
| 341.10 | 552.0    |
| 346.10 | 1322.0   |
| 352.00 | 1373.0   |
| 353.10 | 935.0    |
| 354.10 | 1730.0   |
| 365.10 | 10762.0  |
| 366.00 | 1863.0   |
| 372.00 | 2709.0   |
| 373.20 | 407.0    |
| 383.00 | 1031.0   |
| 391.80 | 373.0    |
| 401.00 | 797.0    |
| 401.90 | 898.0    |
| 403.00 | 1485.0   |
| 403.90 | 736.0    |
| 421.10 | 1513.0   |
| 421.90 | 1212.0   |
| 423.00 | 5109.0   |
| 424.20 | 1864.0   |
| 425.10 | 467.0    |
| 438.00 | 370.0    |
| 438.80 | 372.0    |
| 439.30 | 435.0    |
| 441.10 | 29176.0  |
| 442.10 | 159360.0 |
| 443.10 | 30048.0  |
| 444.10 | 2477.0   |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP02

Scan 2442 (17.351 min): BM014325.D  
DFTPP02

| m/z    | Abundance |
|--------|-----------|
| 73.00  | 340.0     |
| 207.00 | 1122.0    |
| 208.10 | 528.0     |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP02