

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM022324\
 Data File : BM044269.D
 Acq On : 23 Feb 2024 13:56
 Operator : MA/JU
 Sample : SSTDICV020
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SICV019

Quant Time: Feb 23 22:59:46 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM022324.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Feb 23 22:49:27 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.898	152	220753	20.000	ng/u1	0.00
20) Naphthalene-d8	10.698	136	971889	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.545	164	571150	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.298	188	1184756	20.000	ng/u1	0.00
79) Chrysene-d12	21.503	240	1059263	20.000	ng/u1	0.00
88) Perylene-d12	23.897	264	1217799	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.346	96	48879	6.986	ng/uL	0.00
4) Pyridine-d5	3.757	84	371692	19.452	ng/u1	0.00
7) Phenol-d5	7.063	99	408224	19.304	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.239	67	287229	20.974	ng/u1	0.00
11) 2-Chlorophenol-d4	7.428	132	331260	20.798	ng/u1	0.00
15) 4-Methylphenol-d8	8.610	113	315495	19.717	ng/u1	0.00
21) Nitrobenzene-d5	9.069	128	160976	20.371	ng/u1	0.00
24) 2-Nitrophenol-d4	9.786	143	170030	21.090	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.322	165	292620	20.576	ng/u1	0.00
31) 4-Chloroaniline-d4	10.845	131	466167	19.517	ng/u1	0.00
46) Dimethylphthalate-d6	13.968	166	893726	20.289	ng/u1	0.00
49) Acenaphthylene-d8	14.239	160	1037267	20.467	ng/u1	0.00
54) 4-Nitrophenol-d4	14.745	143	170828	19.489	ng/u1	0.00
60) Fluorene-d10	15.539	176	749977	20.116	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.662	200	139790	19.016	ng/u1	0.00
73) Anthracene-d10	17.398	188	1122463	19.870	ng/u1	0.00
81) Pyrene-d10	19.697	212	1268667	19.908	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.744	264	1240732	19.883	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.381	88	51563	6.992	ng/uL	97
5) Pyridine	3.775	79	323375	16.177	ng/u1	97
6) Benzaldehyde	7.045	77	237469	25.869	ng/u1	97
8) Phenol	7.087	94	393213	17.888	ng/u1	99
10) Bis(2-Chloroethyl)ether	7.334	93	325479	17.690	ng/u1	99
12) 2-Chlorophenol	7.457	128	312319	18.954	ng/u1	97
13) 2-Methylphenol	8.345	108	287604	17.759	ng/u1	98
14) 2,2'-oxybis(1-Chloropr...	8.439	45	414769	16.488	ng/u1	98
16) Acetophenone	8.728	105	498215	19.489	ng/u1	99
17) N-Nitroso-di-n-propyla...	8.716	70	244167	19.066	ng/u1	99
18) 4-Methylphenol	8.675	108	314024	18.203	ng/u1	99
19) Hexachloroethane	8.969	117	126125	18.043	ng/u1	95
22) Nitrobenzene	9.110	77	340998	17.843	ng/u1	99
23) Isophorone	9.633	82	668136	18.039	ng/u1	100
25) 2-Nitrophenol	9.816	139	171029	19.127	ng/u1	96
26) 2,4-Dimethylphenol	9.880	107	252010	14.760	ng/u1	99
27) Bis(2-Chloroethoxy)met...	10.128	93	425267	17.662	ng/u1	98
29) 2,4-Dichlorophenol	10.345	162	272638	18.942	ng/u1	100
30) Naphthalene	10.751	128	982050	17.767	ng/u1	99
32) 4-Chloroaniline	10.869	127	418449	17.978	ng/u1	99
33) Hexachlorobutadiene	11.022	225	155583	17.942	ng/u1	98
34) Caprolactam	11.651	113	91719	18.435	ng/u1	92
35) 4-Chloro-3-methylphenol	11.992	107	293971	19.171	ng/u1	96
36) 2-Methylnaphthalene	12.363	142	641062	18.208	ng/u1	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM022324\
 Data File : BM044269.D
 Acq On : 23 Feb 2024 13:56
 Operator : MA/JU
 Sample : SSTDICV020
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SICV019

Quant Time: Feb 23 22:59:46 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM022324.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Feb 23 22:49:27 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.580	142	629290	18.081	ng/ul	99
39) 1,2,4,5-Tetrachloroben...	12.727	216	313712	18.554	ng/ul	98
40) Hexachlorocyclopentadiene	12.698	237	161650	14.210	ng/ul	99
41) 2,4,6-Trichlorophenol	12.974	196	199902	18.549	ng/ul	99
42) 2,4,5-Trichlorophenol	13.039	196	218181	18.727	ng/ul	100
43) 1,1'-Biphenyl	13.380	154	849833	18.277	ng/ul	99
44) 2-Chloronaphthalene	13.421	162	646227	17.749	ng/ul	99
45) 2-Nitroaniline	13.633	65	188462	18.696	ng/ul	94
47) Dimethylphthalate	14.015	163	803893	17.856	ng/ul	99
48) 2,6-Dinitrotoluene	14.133	165	167293	19.870	ng/ul	99
50) Acenaphthylene	14.268	152	1086584	18.101	ng/ul	100
51) 3-Nitroaniline	14.463	138	191285	21.536	ng/ul	97
52) Acenaphthene	14.610	153	687554	17.349	ng/ul	99
53) 2,4-Dinitrophenol	14.668	184	88490	16.938	ng/ul	94
55) 4-Nitrophenol	14.762	109	105288	17.485	ng/ul	95
56) Dibenzofuran	14.945	168	980597	18.432	ng/ul	100
57) 2,4-Dinitrotoluene	14.921	165	230079	19.022	ng/ul#	97
58) 2,3,4,6-Tetrachlorophenol	15.168	232	187212	19.967	ng/ul	99
59) Diethylphthalate	15.380	149	825343	17.810	ng/ul	100
61) Fluorene	15.592	166	772724	18.084	ng/ul	98
62) 4-Chlorophenyl-phenyle...	15.598	204	362111	17.982	ng/ul	98
63) 4-Nitroaniline	15.627	138	203803	23.723	ng/ul	99
66) 4,6-Dinitro-2-methylph...	15.674	198	129222	16.069	ng/ul	96
67) N-Nitrosodiphenylamine	15.809	169	668838	18.629	ng/ul	99
68) 4-Bromophenyl-phenylether	16.492	248	214542	18.226	ng/ul	97
69) Hexachlorobenzene	16.592	284	242964	18.033	ng/ul	97
70) Atrazine	16.774	200	211210	17.958	ng/ul	99
71) Pentachlorophenol	16.939	266	144706	16.298	ng/ul	98
72) Phenanthrene	17.339	178	1193595	17.562	ng/ul	100
74) Anthracene	17.433	178	1204445	17.601	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.333	216	293989	17.727	ng/uL	99
76) Pentachlorobenzene	14.857	250	289527	18.287	ng/uL	99
77) Carbazole	17.709	167	1154517	17.943	ng/ul	99
78) Di-n-butylphthalate	18.286	149	1491196	18.304	ng/ul	99
80) Fluoranthene	19.362	202	1326784	17.360	ng/ul	98
82) Pyrene	19.727	202	1412196	17.619	ng/ul	99
83) Butylbenzylphthalate	20.644	149	653474	18.151	ng/ul	99
84) 3,3'-Dichlorobenzidine	21.427	252	494342	20.864	ng/ul	99
85) Benzo(a)anthracene	21.486	228	1393585	17.640	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.433	149	1049337	19.038	ng/ul	99
87) Chrysene	21.539	228	1282376	17.269	ng/ul	99
89) Di-n-octyl phthalate	22.374	149	1771197	16.975	ng/ul	100
90) Benzo(b)fluoranthene	23.168	252	1374220	17.602	ng/ul	99
91) Benzo(k)fluoranthene	23.221	252	1360343	16.965	ng/ul	99
93) Benzo(a)pyrene	23.791	252	1239569	16.468	ng/ul	100
94) Indeno(1,2,3-cd)pyrene	26.379	276	1623198	17.832	ng/ul	98
95) Dibenzo(a,h)anthracene	26.415	278	1362209	17.908	ng/ul	99
96) Benzo(g,h,i)perylene	27.144	276	1293546	17.568	ng/ul	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed


BNA M

ClientSampleId :

SICV019

Quant Time: Feb 23 22:59:46 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM022324.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Feb 23 22:49:27 2024
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 6.992 ng/uL
RT: 3.381 min Scan# 50
Delta R.T. -0.000 min
Lab File: BM044269.D
Acq: 23 Feb 2024 13:56




Instrument :
BNA_M
ClientSampleId :
SICV019


Tgt Ion: 88 Resp: 51563
Ion Ratio Lower Upper
88 100
43 25.6 21.2 31.8
58 76.0 58.3 87.5



#5
Pyridine
Concen: 16.177 ng/uL
RT: 3.775 min Scan# 117
Delta R.T. -0.000 min
Lab File: BM044269.D
Acq: 23 Feb 2024 13:56

Tgt Ion: 79 Resp: 323375
Ion Ratio Lower Upper
79 100
52 76.5 58.9 88.3
51 33.9 27.8 41.8









#6
Benzaldehyde
Concen: 25.869 ng/ul
RT: 7.045 min Scan# 673
Delta R.T. -0.000 min
Lab File: BM044269.D
Acq: 23 Feb 2024 13:56




Instrument :
BNA_M
ClientSampleId :
SICV019


Tgt Ion: 77 Resp: 237469
Ion Ratio Lower Upper
77 100
105 102.3 79.0 118.6
106 98.6 77.4 116.2



#8
Phenol
Concen: 17.888 ng/ul
RT: 7.087 min Scan# 680
Delta R.T. -0.000 min
Lab File: BM044269.D
Acq: 23 Feb 2024 13:56

Tgt Ion: 94 Resp: 393213
Ion Ratio Lower Upper
94 100
65 26.5 20.6 31.0
66 31.3 25.4 38.0









#10
Bis(2-Chloroethyl)ether
Concen: 17.690 ng/ul
RT: 7.334 min Scan# 721
Delta R.T. 0.006 min
Lab File: BM044269.D
Acq: 23 Feb 2024 13:56




Instrument :
BNA_M
ClientSampleId :
SICV019


Tgt Ion: 93 Resp: 325479
Ion Ratio Lower Upper
93 100
63 75.4 61.0 91.6
95 32.6 25.1 37.7



#12
2-Chlorophenol
Concen: 18.954 ng/ul
RT: 7.457 min Scan# 743
Delta R.T. 0.006 min
Lab File: BM044269.D
Acq: 23 Feb 2024 13:56

Tgt Ion:128 Resp: 312319
Ion Ratio Lower Upper
128 100
64 46.4 39.0 58.6
130 32.6 25.4 38.0









#13
2-Methylphenol
Concen: 17.759 ng/ul
RT: 8.345 min Scan# 893
Delta R.T. 0.006 min
Lab File: BM044269.D
Acq: 23 Feb 2024 13:56




Instrument :
BNA_M
ClientSampleId :
SICV019


Tgt Ion:108 Resp: 287604
Ion Ratio Lower Upper
108 100
107 86.7 70.6 105.8



#14
2,2'-oxybis(1-Chloropropane)
Concen: 16.488 ng/ul
RT: 8.439 min Scan# 910
Delta R.T. 0.012 min
Lab File: BM044269.D
Acq: 23 Feb 2024 13:56

Tgt Ion: 45 Resp: 414769
Ion Ratio Lower Upper
45 100
77 14.5 11.3 16.9
79 11.8 8.5 12.7









#16
Acetophenone
Concen: 19.489 ng/ul
RT: 8.728 min Scan# 91
Delta R.T. -0.000 min
Lab File: BM044269.D
Acq: 23 Feb 2024 13:56




Instrument :
BNA_M
ClientSampleId :
SICV019


Tgt Ion:105 Resp: 498215
Ion Ratio Lower Upper
105 100
77 75.5 61.2 91.8
51 25.6 21.1 31.7



#17
N-Nitroso-di-n-propylamine
Concen: 19.066 ng/ul
RT: 8.716 min Scan# 957
Delta R.T. 0.006 min
Lab File: BM044269.D
Acq: 23 Feb 2024 13:56

Tgt Ion: 70 Resp: 244167
Ion Ratio Lower Upper
70 100
42 54.3 44.4 66.6
101 9.1 7.5 11.3
130 21.1 17.0 25.4









#18
4-Methylphenol
Concen: 18.203 ng/ul
RT: 8.675 min Scan# 91
Delta R.T. 0.006 min
Lab File: BM044269.D
Acq: 23 Feb 2024 13:56




Instrument :
BNA_M
ClientSampleId :
SICV019


Tgt Ion:108 Resp: 314024
Ion Ratio Lower Upper
108 100
107 109.9 86.8 130.2



#19
Hexachloroethane
Concen: 18.043 ng/ul
RT: 8.969 min Scan# 1000
Delta R.T. 0.006 min
Lab File: BM044269.D
Acq: 23 Feb 2024 13:56

Tgt Ion:117 Resp: 126125
Ion Ratio Lower Upper
117 100
201 89.2 69.8 104.6
199 57.7 40.8 61.2









#22
Nitrobenzene
Concen: 17.843 ng/ul
RT: 9.110 min Scan# 1024
Delta R.T. 0.006 min
Lab File: BM044269.D
Acq: 23 Feb 2024 13:56




Instrument :
BNA_M
ClientSampleId :
SICV019


Tgt Ion: 77 Resp: 340998
Ion Ratio Lower Upper
77 100
123 47.3 36.9 55.3
65 12.7 10.2 15.2



#23
Isophorone
Concen: 18.039 ng/ul
RT: 9.633 min Scan# 1113
Delta R.T. 0.006 min
Lab File: BM044269.D
Acq: 23 Feb 2024 13:56

Tgt Ion: 82 Resp: 668136
Ion Ratio Lower Upper
82 100
95 5.8 4.6 7.0
138 17.0 13.5 20.3









#25
 2-Nitrophenol
 Concen: 19.127 ng/ul
 RT: 9.816 min Scan# 1143
 Delta R.T. 0.006 min
 Lab File: BM044269.D
 Acq: 23 Feb 2024 13:56




Instrument :
 BNA_M
 ClientSampleId :
 SICV019


Tgt Ion:139 Resp: 171029
 Ion Ratio Lower Upper
 139 100
 65 48.1 41.0 61.4
 109 24.5 20.5 30.7



#26
 2,4-Dimethylphenol
 Concen: 14.760 ng/ul
 RT: 9.880 min Scan# 1155
 Delta R.T. 0.006 min
 Lab File: BM044269.D
 Acq: 23 Feb 2024 13:56

Tgt Ion:107 Resp: 252010
 Ion Ratio Lower Upper
 107 100
 121 51.7 41.1 61.7
 122 91.1 72.3 108.5









#27
 Bis(2-Chloroethoxy)methane
 Concen: 17.662 ng/ul
 RT: 10.128 min Scan# 1196
 Delta R.T. 0.006 min
 Lab File: BM044269.D
 Acq: 23 Feb 2024 13:56




Instrument :
 BNA_M
 ClientSampleId :
 SICV019


Tgt Ion: 93 Resp: 425267
 Ion Ratio Lower Upper
 93 100
 95 33.5 26.0 39.0
 123 12.3 9.0 13.6



#29
 2,4-Dichlorophenol
 Concen: 18.942 ng/ul
 RT: 10.345 min Scan# 1234
 Delta R.T. -0.000 min
 Lab File: BM044269.D
 Acq: 23 Feb 2024 13:56

Tgt Ion: 162 Resp: 272638
 Ion Ratio Lower Upper
 162 100
 164 64.5 51.8 77.8
 98 36.9 29.6 44.4









#30
Naphthalene
Concen: 17.767 ng/ul
RT: 10.751 min Scan# 1303
Delta R.T. 0.006 min
Lab File: BM044269.D
Acq: 23 Feb 2024 13:56




Instrument :
BNA_M
ClientSampleId :
SICV019


Tgt Ion:128 Resp: 982050
Ion Ratio Lower Upper
128 100
129 10.6 8.8 13.2
127 12.6 10.2 15.2



#32
4-Chloroaniline
Concen: 17.978 ng/ul
RT: 10.869 min Scan# 1323
Delta R.T. -0.000 min
Lab File: BM044269.D
Acq: 23 Feb 2024 13:56

Tgt Ion:127 Resp: 418449
Ion Ratio Lower Upper
127 100
129 32.2 25.3 37.9









#33
Hexachlorobutadiene
Concen: 17.942 ng/ul
RT: 11.022 min Scan# 11
Delta R.T. -0.000 min
Lab File: BM044269.D
Acq: 23 Feb 2024 13:56




Instrument :
BNA_M
ClientSampleId :
SICV019


Tgt Ion:225 Resp: 155583
Ion Ratio Lower Upper
225 100
223 61.9 49.7 74.5
227 61.7 52.2 78.2



#34
Caprolactam
Concen: 18.435 ng/ul
RT: 11.651 min Scan# 1456
Delta R.T. 0.006 min
Lab File: BM044269.D
Acq: 23 Feb 2024 13:56

Tgt Ion:113 Resp: 91719
Ion Ratio Lower Upper
113 100
55 162.7 140.7 211.1
56 120.7 102.4 153.6









#35
4-Chloro-3-methylphenol
Concen: 19.171 ng/ul
RT: 11.992 min Scan# 1514
Delta R.T. 0.006 min
Lab File: BM044269.D
Acq: 23 Feb 2024 13:56




Instrument :
BNA_M
ClientSampleId :
SICV019


Tgt Ion:107 Resp: 293971
Ion Ratio Lower Upper
107 100
144 26.5 20.6 30.8
142 84.3 64.3 96.5



#36
2-Methylnaphthalene
Concen: 18.208 ng/ul
RT: 12.363 min Scan# 1577
Delta R.T. -0.000 min
Lab File: BM044269.D
Acq: 23 Feb 2024 13:56

Tgt Ion:142 Resp: 641062
Ion Ratio Lower Upper
142 100
141 89.4 70.5 105.7









#37
1-Methylnaphthalene
Concen: 18.081 ng/ul
RT: 12.580 min Scan# 1
Delta R.T. -0.000 min
Lab File: BM044269.D
Acq: 23 Feb 2024 13:56




Instrument :
BNA_M
ClientSampleId :
SICV019


Tgt Ion:142 Resp: 629290
Ion Ratio Lower Upper
142 100
141 90.0 73.0 109.4
116 3.5 2.7 4.1



#39
1,2,4,5-Tetrachlorobenzene
Concen: 18.554 ng/ul
RT: 12.727 min Scan# 1639
Delta R.T. 0.006 min
Lab File: BM044269.D
Acq: 23 Feb 2024 13:56

Tgt Ion:216 Resp: 313712
Ion Ratio Lower Upper
216 100
214 77.9 63.9 95.9
179 21.3 17.5 26.3
108 19.2 16.2 24.4









#40
Hexachlorocyclopentadiene
Concen: 14.210 ng/ul
RT: 12.698 min Scan# 1634
Delta R.T. -0.000 min
Lab File: BM044269.D
Acq: 23 Feb 2024 13:56




Instrument :
BNA_M
ClientSampleId :
SICV019


Tgt Ion:237 Resp: 161650
Ion Ratio Lower Upper
237 100
235 60.7 49.4 74.2
272 13.5 10.6 16.0



#41
2,4,6-Trichlorophenol
Concen: 18.549 ng/ul
RT: 12.974 min Scan# 1681
Delta R.T. 0.006 min
Lab File: BM044269.D
Acq: 23 Feb 2024 13:56

Tgt Ion:196 Resp: 199902
Ion Ratio Lower Upper
196 100
198 96.8 76.1 114.1
200 31.0 25.1 37.7









#42
2,4,5-Trichlorophenol
Concen: 18.727 ng/ul
RT: 13.039 min Scan# 1
Delta R.T. -0.000 min
Lab File: BM044269.D
Acq: 23 Feb 2024 13:56




Instrument :
BNA_M
ClientSampleId :
SICV019


Tgt Ion:196 Resp: 218181
Ion Ratio Lower Upper
196 100
198 96.8 77.4 116.2
200 31.0 25.6 38.4



#43
1,1'-Biphenyl
Concen: 18.277 ng/ul
RT: 13.380 min Scan# 1750
Delta R.T. 0.006 min
Lab File: BM044269.D
Acq: 23 Feb 2024 13:56

Tgt Ion:154 Resp: 849833
Ion Ratio Lower Upper
154 100
153 40.0 32.4 48.6
76 14.0 11.6 17.4





#44

2-Chloronaphthalene

Concen: 17.749 ng/ul

RT: 13.421 min Scan# 1

Delta R.T. 0.006 min

Lab File: BM044269.D

Acq: 23 Feb 2024 13:56

Instrument :

BNA_M

ClientSampleId :

SICV019





Tgt Ion:162 Resp: 646227

Ion Ratio Lower Upper

162 100

164 32.3 26.1 39.1

127 39.0 30.9 46.3



#45

2-Nitroaniline

Concen: 18.696 ng/ul

RT: 13.633 min Scan# 1793

Delta R.T. -0.000 min

Lab File: BM044269.D

Acq: 23 Feb 2024 13:56




Tgt Ion: 65 Resp: 188462


Ion Ratio Lower Upper

65 100

92 71.8 52.5 78.7

138 112.4 86.4 129.6









#47
Dimethylphthalate
Concen: 17.856 ng/ul
RT: 14.015 min Scan# 1858
Delta R.T. 0.006 min
Lab File: BM044269.D
Acq: 23 Feb 2024 13:56




Instrument :
BNA_M
ClientSampleId :
SICV019


Tgt Ion:163 Resp: 803893
Ion Ratio Lower Upper
163 100
194 3.8 3.0 4.6
164 9.8 8.1 12.1



#48
2,6-Dinitrotoluene
Concen: 19.870 ng/ul
RT: 14.133 min Scan# 1878
Delta R.T. -0.000 min
Lab File: BM044269.D
Acq: 23 Feb 2024 13:56

Tgt Ion:165 Resp: 167293
Ion Ratio Lower Upper
165 100
89 59.9 48.2 72.2
121 22.6 18.7 28.1





#50

Acenaphthylene

Concen: 18.101 ng/ul

RT: 14.268 min Scan# 1900

Delta R.T. 0.006 min

Lab File: BM044269.D

Acq: 23 Feb 2024 13:56

Instrument :

BNA_M

ClientSampleId :

SICV019





Tgt Ion:152 Resp: 1086584

Ion Ratio Lower Upper

152 100

151 19.8 15.8 23.6

153 12.9 10.4 15.6



#51

3-Nitroaniline

Concen: 21.536 ng/ul

RT: 14.463 min Scan# 1934

Delta R.T. 0.006 min

Lab File: BM044269.D

Acq: 23 Feb 2024 13:56




Tgt Ion:138 Resp: 191285


Ion Ratio Lower Upper

138 100

108 8.6 8.6 12.8

92 114.1 93.8 140.6





#52

Acenaphthene

Concen: 17.349 ng/ul

RT: 14.610 min Scan# 1958

Delta R.T. 0.006 min

Lab File: BM044269.D

Acq: 23 Feb 2024 13:56

Instrument :

BNA_M

ClientSampleId :

SICV019





Tgt Ion:153 Resp: 687554

Ion Ratio Lower Upper

153 100

152 47.3 38.5 57.7

154 88.3 70.8 106.2



#53

2,4-Dinitrophenol

Concen: 16.938 ng/ul

RT: 14.668 min Scan# 1969

Delta R.T. 0.006 min

Lab File: BM044269.D

Acq: 23 Feb 2024 13:56




Tgt Ion:184 Resp: 88490


Ion Ratio Lower Upper

184 100

63 67.2 58.4 87.6

154 61.2 46.4 69.6









#55
4-Nitrophenol
Concen: 17.485 ng/ul
RT: 14.762 min Scan# 1984
Delta R.T. 0.006 min
Lab File: BM044269.D
Acq: 23 Feb 2024 13:56




Instrument :
BNA_M
ClientSampleId :
SICV019


Tgt Ion:109 Resp: 105288
Ion Ratio Lower Upper
109 100
139 158.0 123.0 184.6
65 145.3 122.2 183.2



#56
Dibenzofuran
Concen: 18.432 ng/ul
RT: 14.945 min Scan# 2016
Delta R.T. 0.006 min
Lab File: BM044269.D
Acq: 23 Feb 2024 13:56

Tgt Ion:168 Resp: 980597
Ion Ratio Lower Upper
168 100
139 37.2 29.9 44.9





#57

2,4-Dinitrotoluene

Concen: 19.022 ng/ul

RT: 14.921 min Scan# 2011

Delta R.T. 0.006 min

Lab File: BM044269.D

Acq: 23 Feb 2024 13:56

Instrument :

BNA_M

ClientSampleId :

SICV019





Tgt Ion:165 Resp: 230079

Ion Ratio Lower Upper

165 100

63 45.6 37.9 56.9

182 2.7 1.8 2.6#



#58

2,3,4,6-Tetrachlorophenol

Concen: 19.967 ng/ul

RT: 15.168 min Scan# 2054

Delta R.T. 0.006 min

Lab File: BM044269.D

Acq: 23 Feb 2024 13:56

Tgt Ion:232 Resp: 187212




Ion Ratio Lower Upper


232 100

131 51.8 41.9 62.9

130 2.5 1.8 2.8

166 34.5 26.9 40.3





#59

Diethylphthalate

Concen: 17.810 ng/ul

RT: 15.380 min Scan# 2089

Delta R.T. 0.006 min

Lab File: BM044269.D

Acq: 23 Feb 2024 13:56

Instrument :

BNA_M

ClientSampleId :

SICV019





Tgt Ion:149 Resp: 825343

Ion Ratio Lower Upper

149 100

177 21.5 17.0 25.6

150 12.2 9.6 14.4



#61

Fluorene

Concen: 18.084 ng/ul

RT: 15.592 min Scan# 2126

Delta R.T. -0.000 min

Lab File: BM044269.D

Acq: 23 Feb 2024 13:56




Tgt Ion:166 Resp: 772724


Ion Ratio Lower Upper

166 100

165 98.6 77.4 116.0

167 13.5 10.8 16.2









#62
4-Chlorophenyl-phenylether
Concen: 17.982 ng/ul
RT: 15.598 min Scan# 2127
Delta R.T. 0.006 min
Lab File: BM044269.D
Acq: 23 Feb 2024 13:56




Instrument :
BNA_M
ClientSampleId :
SICV019


Tgt Ion:204 Resp: 362111
Ion Ratio Lower Upper
204 100
206 32.6 26.9 40.3
141 58.3 48.3 72.5



#63
4-Nitroaniline
Concen: 23.723 ng/ul
RT: 15.627 min Scan# 2132
Delta R.T. 0.006 min
Lab File: BM044269.D
Acq: 23 Feb 2024 13:56

Tgt Ion:138 Resp: 203803
Ion Ratio Lower Upper
138 100
92 50.6 40.9 61.3
108 67.5 55.1 82.7









#66
4,6-Dinitro-2-methylphenol
Concen: 16.069 ng/ul
RT: 15.674 min Scan# 2140
Delta R.T. -0.000 min
Lab File: BM044269.D
Acq: 23 Feb 2024 13:56




Instrument :
BNA_M
ClientSampleId :
SICV019


Tgt Ion:198 Resp: 129222
Ion Ratio Lower Upper
198 100
182 4.6 3.8 5.8
77 28.3 20.6 30.8



#67
N-Nitrosodiphenylamine
Concen: 18.629 ng/ul
RT: 15.809 min Scan# 2163
Delta R.T. -0.000 min
Lab File: BM044269.D
Acq: 23 Feb 2024 13:56

Tgt Ion:169 Resp: 668838
Ion Ratio Lower Upper
169 100
168 66.6 53.6 80.4
167 36.1 29.2 43.8









#68
4-Bromophenyl-phenylether
Concen: 18.226 ng/ul
RT: 16.492 min Scan# 21
Delta R.T. 0.006 min
Lab File: BM044269.D
Acq: 23 Feb 2024 13:56




Instrument :
BNA_M
ClientSampleId :
SICV019


Tgt Ion:248 Resp: 214542
Ion Ratio Lower Upper
248 100
250 98.3 79.8 119.8
141 78.5 65.8 98.8



#69
Hexachlorobenzene
Concen: 18.033 ng/ul
RT: 16.592 min Scan# 2296
Delta R.T. 0.006 min
Lab File: BM044269.D
Acq: 23 Feb 2024 13:56

Tgt Ion:284 Resp: 242964
Ion Ratio Lower Upper
284 100
142 39.2 33.4 50.2
249 31.0 24.7 37.1





#70

Atrazine

Concen: 17.958 ng/ul

RT: 16.774 min Scan# 21

Delta R.T. 0.006 min

Lab File: BM044269.D

Acq: 23 Feb 2024 13:56

Instrument :

BNA_M

ClientSampleId :

SICV019





Tgt Ion:200 Resp: 211210

Ion Ratio Lower Upper

200 100

173 27.7 22.5 33.7

215 48.1 37.8 56.6



#71

Pentachlorophenol

Concen: 16.298 ng/ul

RT: 16.939 min Scan# 2355

Delta R.T. -0.000 min

Lab File: BM044269.D

Acq: 23 Feb 2024 13:56




Tgt Ion:266 Resp: 144706


Ion Ratio Lower Upper

266 100

264 62.7 50.9 76.3

268 65.6 51.1 76.7





#72

Phenanthrene

Concen: 17.562 ng/ul

RT: 17.339 min Scan# 2423

Delta R.T. -0.000 min

Lab File: BM044269.D

Acq: 23 Feb 2024 13:56

Instrument :

BNA_M

ClientSampleId :

SICV019





Tgt Ion:178 Resp: 1193595

Ion Ratio Lower Upper

178 100

179 15.6 12.2 18.4

176 19.5 15.6 23.4



#74

Anthracene

Concen: 17.601 ng/ul

RT: 17.433 min Scan# 2439

Delta R.T. 0.006 min

Lab File: BM044269.D

Acq: 23 Feb 2024 13:56




Tgt Ion:178 Resp: 1204445


Ion Ratio Lower Upper

178 100

179 15.6 12.2 18.2

176 18.9 15.4 23.0









#75
1,2,3,4-Tetrachlorobenzene
Concen: 17.727 ng/uL
RT: 13.333 min Scan# 11
Delta R.T. -0.000 min
Lab File: BM044269.D
Acq: 23 Feb 2024 13:56




Instrument :
BNA_M
ClientSampleId :
SICV019


Tgt Ion:216 Resp: 293989
Ion Ratio Lower Upper
216 100
214 79.0 61.9 92.9
218 47.2 37.9 56.9



#76
Pentachlorobenzene
Concen: 18.287 ng/uL
RT: 14.857 min Scan# 2001
Delta R.T. 0.006 min
Lab File: BM044269.D
Acq: 23 Feb 2024 13:56

Tgt Ion:250 Resp: 289527
Ion Ratio Lower Upper
250 100
248 61.7 48.6 73.0
252 62.2 50.0 75.0









#77
 Carbazole
 Concen: 17.943 ng/ul
 RT: 17.709 min Scan# 2485
 Delta R.T. 0.006 min
 Lab File: BM044269.D
 Acq: 23 Feb 2024 13:56




Instrument :
 BNA_M
 ClientSampleId :
 SICV019


Tgt Ion:167 Resp: 1154517
 Ion Ratio Lower Upper
 167 100
 166 21.0 17.1 25.7
 139 12.9 10.7 16.1



#78
 Di-n-butylphthalate
 Concen: 18.304 ng/ul
 RT: 18.286 min Scan# 2584
 Delta R.T. 0.006 min
 Lab File: BM044269.D
 Acq: 23 Feb 2024 13:56

Tgt Ion:149 Resp: 1491196
 Ion Ratio Lower Upper
 149 100
 150 8.9 7.3 10.9
 104 4.6 3.8 5.8









#80
Fluoranthene
Concen: 17.360 ng/uI
RT: 19.362 min Scan# 21
Delta R.T. 0.006 min
Lab File: BM044269.D
Acq: 23 Feb 2024 13:56




Instrument :
BNA_M
ClientSampleId :
SICV019


Tgt Ion:202 Resp: 1326784
Ion Ratio Lower Upper
202 100
101 12.7 11.1 16.7
100 9.8 8.2 12.2



#82
Pyrene
Concen: 17.619 ng/uI
RT: 19.727 min Scan# 2829
Delta R.T. 0.006 min
Lab File: BM044269.D
Acq: 23 Feb 2024 13:56

Tgt Ion:202 Resp: 1412196
Ion Ratio Lower Upper
202 100
101 15.6 12.6 18.8
100 12.1 10.2 15.2









#83
Butylbenzylphthalate
Concen: 18.151 ng/ul
RT: 20.644 min Scan# 2985
Delta R.T. -0.000 min
Lab File: BM044269.D
Acq: 23 Feb 2024 13:56




Instrument :
BNA_M
ClientSampleId :
SICV019


Tgt Ion:149 Resp: 653474
Ion Ratio Lower Upper
149 100
91 74.0 58.1 87.1
206 18.5 15.0 22.4



#84
3,3'-Dichlorobenzidine
Concen: 20.864 ng/ul
RT: 21.427 min Scan# 3118
Delta R.T. 0.006 min
Lab File: BM044269.D
Acq: 23 Feb 2024 13:56

Tgt Ion:252 Resp: 494342
Ion Ratio Lower Upper
252 100
254 64.0 51.6 77.4
126 13.6 11.2 16.8





#85

Benzo(a)anthracene

Concen: 17.640 ng/ul

RT: 21.486 min Scan# 3127

Delta R.T. 0.006 min

Lab File: BM044269.D

Acq: 23 Feb 2024 13:56

Instrument :

BNA_M

ClientSampleId :

SICV019





Tgt Ion:228 Resp: 1393585

Ion Ratio Lower Upper

228 100

229 19.8 15.4 23.0

226 27.0 21.2 31.8



#86

Bis(2-ethylhexyl)phthalate

Concen: 19.038 ng/ul

RT: 21.433 min Scan# 3119

Delta R.T. 0.006 min

Lab File: BM044269.D

Acq: 23 Feb 2024 13:56




Tgt Ion:149 Resp: 1049337


Ion Ratio Lower Upper

149 100

167 26.8 22.1 33.1

279 3.3 2.4 3.6





#87

Chrysene

Concen: 17.269 ng/ul

RT: 21.539 min Scan# 31

Delta R.T. 0.006 min

Lab File: BM044269.D

Acq: 23 Feb 2024 13:56

Instrument :

BNA_M

ClientSampleId :

SICV019





Tgt Ion:228 Resp: 1282376

Ion Ratio Lower Upper

228 100

226 29.9 23.8 35.6

229 19.6 16.0 24.0



#89

Di-n-octyl phthalate

Concen: 16.975 ng/ul




RT: 22.374 min Scan# 3279


Delta R.T. 0.006 min

Lab File: BM044269.D

Acq: 23 Feb 2024 13:56

Tgt Ion:149 Resp: 1771197









#90
Benzo(b)fluoranthene
Concen: 17.602 ng/ul
RT: 23.168 min Scan# 3413
Delta R.T. 0.006 min
Lab File: BM044269.D
Acq: 23 Feb 2024 13:56




Instrument :
BNA_M
ClientSampleId :
SICV019


Tgt Ion:252 Resp: 1374220
Ion Ratio Lower Upper
252 100
253 22.3 17.6 26.4
125 11.9 9.8 14.8



#91
Benzo(k)fluoranthene
Concen: 16.965 ng/ul
RT: 23.221 min Scan# 3423
Delta R.T. 0.006 min
Lab File: BM044269.D
Acq: 23 Feb 2024 13:56

Tgt Ion:252 Resp: 1360343
Ion Ratio Lower Upper
252 100
253 22.0 17.4 26.0
125 12.2 9.4 14.2





#93

Benzo(a)pyrene

Concen: 16.468 ng/ul

RT: 23.791 min Scan# 3

Delta R.T. 0.006 min

Lab File: BM044269.D

Acq: 23 Feb 2024 13:56

Instrument :

BNA_M

ClientSampleId :

SICV019





Tgt Ion:252 Resp: 1239569

Ion Ratio Lower Upper

252 100

253 22.0 17.4 26.2

125 13.6 10.8 16.2



#94

Indeno(1,2,3-cd)pyrene

Concen: 17.832 ng/ul

RT: 26.379 min Scan# 3960

Delta R.T. 0.006 min

Lab File: BM044269.D

Acq: 23 Feb 2024 13:56




Tgt Ion:276 Resp: 1623198


Ion Ratio Lower Upper

276 100

138 26.6 22.1 33.1

277 24.0 19.9 29.9





#95

Dibenzo(a,h)anthracene

Concen: 17.908 ng/ul

RT: 26.415 min Scan# 3965

Delta R.T. 0.006 min

Lab File: BM044269.D

Acq: 23 Feb 2024 13:56

Instrument :

BNA_M

ClientSampleId :

SICV019





Tgt Ion:278 Resp: 1362209

Ion Ratio Lower Upper

278 100

139 19.0 14.9 22.3

279 23.7 19.5 29.3



#96

Benzo(g,h,i)perylene

Concen: 17.568 ng/ul

RT: 27.144 min Scan# 4090

Delta R.T. 0.006 min

Lab File: BM044269.D

Acq: 23 Feb 2024 13:56

Tgt Ion:276 Resp: 1293546

Ion Ratio Lower Upper

276 100

138 26.4 21.6 32.4

277 23.7 18.5 27.7

