

Data Path : Z:\HPCHEM1\BNA M\DATA\BM022416\
 Data File : BM004389.D
 Acq On : 24 Feb 2016 17:31
 Operator : SJ/UM
 Sample : SSTDC0020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD02055

Quant Time: Feb 25 15:21:01 2016
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM02.2-EPA-BM022216.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Feb 25 02:09:59 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	135	0.00
2	1,4-Dioxane	0.437	0.397	9.2	115	0.00
3 S	1,4-Dioxane-d8	0.379	0.376	0.8	125	0.00
4	Benzaldehyde	0.907	1.011	-11.5	125	0.00
5 S	Phenol-d5	1.677	1.608	4.1	126	0.00
6	Phenol	1.788	1.697	5.1	124	0.00
7 S	Bis-(2-Chloroethyl)ether-d8	0.853	0.826	3.2	124	0.00
8	Bis(2-Chloroethyl)ether	1.324	1.284	3.0	126	0.00
9 S	2-Chlorophenol-d4	1.446	1.404	2.9	126	0.00
10	2-Chlorophenol	1.503	1.475	1.9	128	0.00
11	2-Methylphenol	1.458	1.389	4.7	126	0.00
12	2,2'-oxybis(1-Chloropropane	1.271	1.203	5.4	121	0.00
13 S	4-Methylphenol-d8	1.499	1.436	4.2	126	0.00
14	Acetophenone	2.348	2.310	1.6	124	0.00
15 P	N-Nitroso-di-n-propylamine	1.128	1.071	5.1	122	0.00
16	4-Methylphenol	1.599	1.549	3.1	125	0.00
17	Hexachloroethane	0.571	0.553	3.2	127	0.00
18 I	Naphthalene-d8	1.000	1.000	0.0	137	0.00
19 S	Nitrobenzene-d5	0.152	0.149	2.0	128	0.00
20	Nitrobenzene	0.333	0.322	3.3	126	0.00
21	Isophorone	0.689	0.636	7.7	120	0.00
22 S	2-Nitrophenol-d4	0.173	0.167	3.5	125	0.00
23 C	2-Nitrophenol	0.195	0.186	4.6	123	0.00
24	2,4-Dimethylphenol	0.390	0.371	4.9	125	0.00
25	Bis(2-Chloroethoxy)methane	0.415	0.390	6.0	123	0.00
26 S	2,4-Dichlorophenol-d3	0.314	0.305	2.9	127	0.00
27 C	2,4-Dichlorophenol	0.330	0.319	3.3	126	0.00
28	Naphthalene	1.115	1.075	3.6	126	0.00
29 S	4-Chloroaniline-d4	0.391	0.415	-6.1	126	0.00
30	4-Chloroaniline	0.412	0.444	-7.8	127	0.00
31 C	Hexachlorobutadiene	0.207	0.202	2.4	129	0.00
32	Caprolactam	0.115	0.102	11.3	117	0.00
33 C	4-Chloro-3-methylphenol	0.387	0.368	4.9	124	0.00
34	2-Methylnaphthalene	0.834	0.800	4.1	125	0.00
35	1-Methylnaphthalene	0.794	0.764	3.8	125	0.00
36 I	Acenaphthene-d10	1.000	1.000	0.0	133	0.00
37	1,2,4,5-Tetrachlorobenzene	0.665	0.663	0.3	127	0.00
38	Hexachlorocyclopentadiene	0.248	0.164	33.9#	108	0.00
39 C	2,4,6-Trichlorophenol	0.432	0.410	5.1	121	0.00
40	2,4,5-Trichlorophenol	0.469	0.458	2.3	122	0.00
41	1,1'-Biphenyl	1.719	1.696	1.3	124	0.00
42	2-Chloronaphthalene	1.308	1.305	0.2	127	0.00
43	2-Nitroaniline	0.330	0.319	3.3	120	0.00
44 S	Dimethylphthalate-d6	1.715	1.625	5.2	120	0.00

Data Path : Z:\HPCHEM1\BNA M\DATA\BM022416\
 Data File : BM004389.D
 Acq On : 24 Feb 2016 17:31
 Operator : SJ/UM
 Sample : SSTDC0020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD02055

Quant Time: Feb 25 15:21:01 2016
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM02.2-EPA-BM022216.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Feb 25 02:09:59 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
45	Dimethylphthalate	1.789	1.707	4.6	121	0.00
46	2,6-Dinitrotoluene	0.359	0.360	-0.3	126	0.00
47 S	Acenaphthylene-d8	2.008	2.004	0.2	126	0.00
48	Acenaphthylene	2.185	2.171	0.6	125	0.00
49	3-Nitroaniline	0.356	0.370	-3.9	125	0.00
50 C	Acenaphthene	1.492	1.462	2.0	124	0.00
51	2,4-Dinitrophenol	0.163	0.120	26.4#	119	0.00
52 S	4-Nitrophenol-d4	0.276	0.240	13.0	115	0.00
53	4-Nitrophenol	0.206	0.174	15.5	112	0.00
54	Dibenzofuran	2.084	2.046	1.8	123	0.00
55	2,4-Dinitrotoluene	0.536	0.526	1.9	122	0.00
56	2,3,4,6-Tetrachlorophenol	0.389	0.372	4.4	121	0.00
57	Diethylphthalate	1.851	1.763	4.8	120	0.00
58 S	Fluorene-d10	1.459	1.426	2.3	124	0.00
59	Fluorene	1.727	1.682	2.6	122	0.00
60	4-Chlorophenyl-phenylether	0.847	0.829	2.1	122	0.00
61	4-Nitroaniline	0.384	0.372	3.1	120	0.00
62 I	Phenanthrene-d10	1.000	1.000	0.0	132	0.00
63 S	4,6-Dinitro-2-methylphenol-	0.123	0.108	12.2	118	0.00
64	4,6-Dinitro-2-methylphenol	0.132	0.119	9.8	121	0.00
65	N-Nitrosodiphenylamine	0.640	0.626	2.2	122	0.00
66	4-Bromophenyl-phenylether	0.218	0.215	1.4	124	0.00
67	Hexachlorobenzene	0.242	0.237	2.1	122	0.00
68	Atrazine	0.237	0.230	3.0	118	0.00
69 C	Pentachlorophenol	0.096	0.082	14.6	120	0.00
70	Phenanthrene	1.201	1.182	1.6	123	0.00
71 S	Anthracene-d10	0.980	0.955	2.6	122	0.00
72	Anthracene	1.220	1.198	1.8	122	0.00
73	1,2,3,4-Tetrachlorobenzene	0.266	0.269	-1.1	126	0.00
74	Pentachlorobenzene	0.274	0.271	1.1	124	0.00
75	Carbazole	1.064	1.055	0.8	121	0.00
76	Di-n-butylphthalate	1.476	1.324	10.3	117	0.00
77 C	Fluoranthene	1.361	1.367	-0.4	121	0.00
78 I	Chrysene-d12	1.000	1.000	0.0	123	0.00
79 S	Pyrene-d10	0.967	0.991	-2.5	121	0.00
80	Pyrene	1.307	1.344	-2.8	120	0.00
81	Butylbenzylphthalate	0.575	0.572	0.5	117	0.00
82	3,3'-Dichlorobenzidine	0.403	0.418	-3.7	113	0.00
83	Benzo(a)anthracene	1.287	1.265	1.7	116	0.00
84	Bis(2-ethylhexyl)phthalate	0.850	0.847	0.4	115	0.00
85	Chrysene	1.238	1.185	4.3	112	0.00
86 I	Perylene-d12	1.000	1.000	0.0	97	0.00
87	Di-n-octyl phthalate	1.646	1.969	-19.6	111	0.00

Data Path : Z:\HPCHEM1\BNA M\DATA\BM022416\
Data File : BM004389.D
Acq On : 24 Feb 2016 17:31
Operator : SJ/UM
Sample : SSTDCCC020
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
LabSampleId :
SSTD02055

Quant Time: Feb 25 15:21:01 2016
Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM02.2-EPA-BM022216.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Feb 25 02:09:59 2016
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(b)fluoranthene	1.351	1.417	-4.9	95	0.00
89	Benzo(k)fluoranthene	1.273	1.351	-6.1	98	0.00
90 S	Benzo(a)pyrene-d12	1.050	1.063	-1.2	92	0.00
91 C	Benzo(a)pyrene	1.256	1.255	0.1	91	0.00
92	Indeno(1,2,3-cd)pyrene	1.175	1.126	4.2	85	0.00
93	Dibenzo(a,h)anthracene	0.969	0.940	3.0	86	0.00
94	Benzo(a,h,i)perylene	0.966	0.939	2.8	87	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0