

Data Path : Z:\HPCHEM1\BNA\_M\Data\BM022417\  
 Data File : BM009012.D  
 Acq On : 24 Feb 2017 11:37  
 Operator : SJ/MA  
 Sample : I1940-09  
 Misc :  
 ALS Vial : 4 Sample Multiplier: 1

Instrument :  
 BNA\_M  
 ClientSampleId :  
 DABP4

Quant Time: Feb 24 13:13:50 2017  
 Quant Method : Z:\HPCHEM1\BNA\_M\METHODS\SOM02.2-EPA-BM022317.M  
 Quant Title : SVOA CALIBRATION  
 QLast Update : Fri Feb 24 13:06:22 2017  
 Response via : Initial Calibration

| Internal Standards        | R.T.  | QIon | Response | Conc  | Units | Dev(Min) |
|---------------------------|-------|------|----------|-------|-------|----------|
| 1) 1,4-Dichlorobenzene-d4 | 7.90  | 152  | 177631   | 20.00 | ng/ul | 0.00     |
| 18) Naphthalene-d8        | 10.70 | 136  | 725282   | 20.00 | ng/ul | 0.00     |
| 35) Acenaphthene-d10      | 14.54 | 164  | 442447   | 20.00 | ng/ul | 0.00     |
| 61) Phenanthrene-d10      | 17.29 | 188  | 1089844  | 20.00 | ng/ul | 0.00     |
| 75) Chrysene-d12          | 21.46 | 240  | 1341619  | 20.00 | ng/ul | 0.00     |
| 83) Perylene-d12          | 23.80 | 264  | 1220779  | 20.00 | ng/ul | 0.00     |

## System Monitoring Compounds

|                                |       |     |         |       |       |      |
|--------------------------------|-------|-----|---------|-------|-------|------|
| 3) 1,4-Dioxane-d8              | 3.36  | 96  | 19130   | 4.33  | ng/uL | 0.00 |
| 5) Phenol-d5                   | 7.05  | 99  | 398775  | 24.77 | ng/ul | 0.00 |
| 7) Bis-(2-Chloroethyl)ether-d  | 7.23  | 67  | 298088  | 25.85 | ng/ul | 0.00 |
| 9) 2-Chlorophenol-d4           | 7.43  | 132 | 256913  | 23.95 | ng/ul | 0.00 |
| 13) 4-Methylphenol-d8          | 8.59  | 113 | 299837  | 25.36 | ng/ul | 0.00 |
| 19) Nitrobenzene-d5            | 9.06  | 128 | 125961  | 24.27 | ng/ul | 0.00 |
| 22) 2-Nitrophenol-d4           | 9.79  | 143 | 145813  | 26.04 | ng/ul | 0.00 |
| 26) 2,4-Dichlorophenol-d3      | 10.31 | 165 | 282341  | 24.01 | ng/ul | 0.00 |
| 29) 4-Chloroaniline-d4         | 10.84 | 131 | 366850  | 29.60 | ng/ul | 0.00 |
| 43) Dimethylphthalate-d6       | 13.95 | 166 | 985769  | 33.09 | ng/ul | 0.00 |
| 46) Acenaphthylene-d8          | 14.23 | 160 | 1186802 | 32.44 | ng/ul | 0.00 |
| 51) 4-Nitrophenol-d4           | 14.73 | 143 | 153083  | 29.43 | ng/ul | 0.00 |
| 57) Fluorene-d10               | 15.53 | 176 | 902436  | 32.45 | ng/ul | 0.00 |
| 62) 4,6-Dinitro-2-methylphenol | 15.65 | 200 | 140773  | 21.46 | ng/ul | 0.00 |
| 70) Anthracene-d10             | 17.38 | 188 | 1489451 | 28.86 | ng/ul | 0.00 |
| 76) Pyrene-d10                 | 19.67 | 212 | 1757592 | 30.05 | ng/ul | 0.00 |
| 87) Benzo(a)pyrene-d12         | 23.65 | 264 | 1624051 | 29.30 | ng/ul | 0.00 |

## Target Compounds

|                                |       |     |         |       |        | Qvalue |
|--------------------------------|-------|-----|---------|-------|--------|--------|
| 4) Benzaldehyde                | 7.05  | 77  | 475238  | 39.26 | ng/ul  | 88     |
| 10) 2-Chlorophenol             | 7.46  | 128 | 50022   | 4.48  | ng/ul# | 66     |
| 23) 2-Nitrophenol              | 9.82  | 139 | 224386  | 37.21 | ng/ul# | 65     |
| 28) Naphthalene                | 10.75 | 128 | 260983  | 7.18  | ng/ul  | 97     |
| 30) 4-Chloroaniline            | 10.75 | 127 | 35163   | 2.68  | ng/ul# | 33     |
| 31) Hexachlorobutadiene        | 11.02 | 225 | 165554  | 17.20 | ng/ul  | 98     |
| 33) 4-Chloro-3-methylphenol    | 11.97 | 107 | 183200  | 13.57 | ng/ul# | 93     |
| 36) 1,2,4,5-Tetrachlorobenzene | 12.72 | 216 | 124269  | 8.63  | ng/ul# | 92     |
| 38) 2,4,6-Trichlorophenol      | 12.96 | 196 | 168716  | 20.54 | ng/ul  | 98     |
| 39) 2,4,5-Trichlorophenol      | 12.96 | 196 | 168716  | 18.83 | ng/ul  | 97     |
| 40) 1,1'-Biphenyl              | 13.37 | 154 | 829036  | 27.17 | ng/ul  | 98     |
| 44) Dimethylphthalate          | 13.99 | 163 | 941837  | 31.60 | ng/ul  | 97     |
| 45) 2,6-Dinitrotoluene         | 14.12 | 165 | 104341  | 18.37 | ng/ul  | 83     |
| 52) 4-Nitrophenol              | 14.74 | 109 | 325868  | 48.90 | ng/ul# | 79     |
| 53) Dibenzofuran               | 14.93 | 168 | 430409  | 11.48 | ng/ul  | 98     |
| 63) 4,6-Dinitro-2-methylphenol | 15.66 | 198 | 126845  | 18.41 | ng/ul# | 77     |
| 66) Hexachlorobenzene          | 16.58 | 284 | 110281  | 8.09  | ng/ul  | 100    |
| 68) Pentachlorophenol          | 16.93 | 266 | 136298  | 16.15 | ng/ul  | 95     |
| 69) Phenanthrene               | 17.33 | 178 | 1484328 | 24.64 | ng/ul  | 98     |
| 71) Anthracene                 | 17.42 | 178 | 629082  | 10.24 | ng/ul  | 96     |
| 73) Di-n-butylphthalate        | 18.24 | 149 | 976039  | 15.84 | ng/ul  | 97     |
| 74) Fluoranthene               | 19.34 | 202 | 1486623 | 20.41 | ng/ul# | 97     |
| 77) Pyrene                     | 19.70 | 202 | 1617064 | 21.58 | ng/ul# | 94     |

Data Path : Z:\HPCHEM1\BNA\_M\Data\BM022417\  
Data File : BM009012.D  
Acq On : 24 Feb 2017 11:37  
Operator : SJ/MA  
Sample : I1940-09  
Misc :  
ALS Vial : 4 Sample Multiplier: 1

Instrument :  
BNA\_M  
ClientSampleId :  
DABP4

Quant Time: Feb 24 13:13:50 2017  
Quant Method : Z:\HPCHEM1\BNA\_M\METHODS\SOM02.2-EPA-BM022317.M  
Quant Title : SVOA CALIBRATION  
QLast Update : Fri Feb 24 13:06:22 2017  
Response via : Initial Calibration


| Internal Standards             | R.T.  | QIon | Response | Conc  | Units  | Dev(Min) |
|--------------------------------|-------|------|----------|-------|--------|----------|
| 80) Benzo(a)anthracene         | 21.49 | 228  | 1295820  | 16.93 | ng/ul  | 96       |
| 81) Bis(2-ethylhexyl)phthalate | 21.35 | 149  | 819897   | 20.98 | ng/ul  | 94       |
| 82) Chrysene                   | 21.49 | 228  | 1295820  | 17.81 | ng/ul  | 99       |
| 88) Benzo(a)pyrene             | 23.70 | 252  | 1119779  | 16.08 | ng/ul# | 99       |
| 89) Indeno(1,2,3-cd)pyrene     | 26.19 | 276  | 1402353  | 17.73 | ng/ul# | 92       |
| 90) Dibenzo(a,h)anthracene     | 26.21 | 278  | 751044   | 11.15 | ng/ul# | 95       |


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA\_M\Data\BM022417\  
Data File : BM009012.D  
Acq On : 24 Feb 2017 11:37  
Operator : SJ/MA  
Sample : I1940-09  
Misc :  
ALS Vial : 4 Sample Multiplier: 1

Instrument :  
BNA\_M  
ClientSampleId :  
DABP4

Quant Time: Feb 24 13:13:50 2017  
Quant Method : Z:\HPCHEM1\BNA\_M\METHODS\SOM02.2-EPA-BM022317.M  
Quant Title : SVOA CALIBRATION  
QLast Update : Fri Feb 24 13:06:22 2017  
Response via : Initial Calibration





#4

Benzaldehyde

Concen: 39.26 ng/ul

RT: 7.05 min Scan# 673

Delta R.T. 0.00 min

Lab File: BM009012.D

Acq: 24 Feb 2017 11:37

Instrument :

BNA\_M

ClientSampleId :

DABP4


Tgt Ion: 77 Resp: 475238

Ion Ratio Lower Upper

77 100

105 77.3 54.8 82.2




106 73.7 50.2 75.4



Abundance Ion 77.00 (76.70 to 77.70): BM009012.D (-639) (-)

Ion 105.00 (104.70 to 105.70): BM009012.D (-639) (-)

Ion 106.00 (105.70 to 106.70): BM009012.D (-639) (-)



#10

2-Chlorophenol

Concen: 4.48 ng/ul

RT: 7.46 min Scan# 744

Delta R.T. 0.00 min

Lab File: BM009012.D

Acq: 24 Feb 2017 11:37


Tgt Ion: 128 Resp: 50022

Ion Ratio Lower Upper

128 100

64 59.6 76.6 115.0#



130 19.1 26.5 39.7#




Abundance Ion 128.00 (127.70 to 128.70): BM009012.D (-710) (-)

Ion 64.00 (63.70 to 64.70): BM009012.D (-710) (-)

Ion 130.00 (129.70 to 130.70): BM009012.D (-710) (-)









#23  
2-Nitrophenol  
Concen: 37.21 ng/ul  
RT: 9.82 min Scan# 1144  
Delta R.T. 0.00 min  
Lab File: BM009012.D  
Acq: 24 Feb 2017 11:37

Instrument :  
BNA\_M  
ClientSampleId :  
DABP4

Tgt Ion:139 Resp: 224386  
Ion Ratio Lower Upper  
139 100  
65 88.0 102.7 154.1#  
109 55.2 71.1 106.7#




Abundance Ion 139.00 (138.70 to 139.70): E  
Ion 65.00 (64.70 to 65.70): BM009012.D (-1137) (-)  
Ion 109.00 (108.70 to 109.70): E






#28  
Naphthalene  
Concen: 7.18 ng/ul  
RT: 10.75 min Scan# 1303  
Delta R.T. 0.00 min  
Lab File: BM009012.D  
Acq: 24 Feb 2017 11:37

Tgt Ion:128 Resp: 260983  
Ion Ratio Lower Upper  
128 100  
129 9.9 9.7 14.5  
127 13.0 10.2 15.4



Abundance Ion 128.00 (127.70 to 128.70): E  
Ion 129.00 (128.70 to 129.70): E  
Ion 127.00 (126.70 to 127.70): E









#30  
4-Chloroaniline  
Concen: 2.68 ng/ul  
RT: 10.75 min Scan# 1302  
Delta R.T. -0.12 min  
Lab File: BM009012.D  
Acq: 24 Feb 2017 11:37




Instrument :  
BNA\_M  
ClientSampleId :  
DABP4


Tgt Ion:127 Resp: 35163  
Ion Ratio Lower Upper  
127 100  
129 66.4 23.9 35.9#



#31  
Hexachlorobutadiene  
Concen: 17.20 ng/ul  
RT: 11.02 min Scan# 1349  
Delta R.T. 0.00 min  
Lab File: BM009012.D  
Acq: 24 Feb 2017 11:37

Tgt Ion:225 Resp: 165554  
Ion Ratio Lower Upper  
225 100  
223 63.9 50.4 75.6  
227 66.3 51.0 76.4









#33  
 4-Chloro-3-methylphenol  
 Concen: 13.57 ng/ul  
 RT: 11.97 min Scan# 1510  
 Delta R.T. -0.01 min  
 Lab File: BM009012.D  
 Acq: 24 Feb 2017 11:37




Instrument :  
 BNA\_M  
 ClientSampleId :  
 DABP4


Tgt Ion:107 Resp: 183200  
 Ion Ratio Lower Upper  
 107 100  
 144 22.4 14.4 21.6#  
 142 68.3 50.7 76.1



#36  
 1,2,4,5-Tetrachlorobenzene  
 Concen: 8.63 ng/ul  
 RT: 12.72 min Scan# 1638  
 Delta R.T. -0.01 min  
 Lab File: BM009012.D  
 Acq: 24 Feb 2017 11:37

Tgt Ion:216 Resp: 124269  
 Ion Ratio Lower Upper  
 216 100  
 214 82.5 61.1 91.7  
 179 20.2 17.7 26.5  
 108 16.6 19.4 29.2#









#38  
 2,4,6-Trichlorophenol  
 Concen: 20.54 ng/ul  
 RT: 12.96 min Scan# 1679  
 Delta R.T. 0.00 min  
 Lab File: BM009012.D  
 Acq: 24 Feb 2017 11:37




Instrument :  
 BNA\_M  
 ClientSampleId :  
 DABP4


Tgt Ion:196 Resp: 168716  
 Ion Ratio Lower Upper  
 196 100  
 198 86.6 70.2 105.2  
 200 28.4 23.8 35.6



#39  
 2,4,5-Trichlorophenol  
 Concen: 18.83 ng/ul  
 RT: 12.96 min Scan# 1679  
 Delta R.T. -0.07 min  
 Lab File: BM009012.D  
 Acq: 24 Feb 2017 11:37

Tgt Ion:196 Resp: 168716  
 Ion Ratio Lower Upper  
 196 100  
 198 86.6 72.2 108.4  
 200 28.4 22.5 33.7









#40  
 1,1'-Biphenyl  
 Concen: 27.17 ng/ul  
 RT: 13.37 min Scan# 1748  
 Delta R.T. 0.00 min  
 Lab File: BM009012.D  
 Acq: 24 Feb 2017 11:37




Instrument :  
 BNA\_M  
 ClientSampled :  
 DABP4


| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 154     | 100   |       |       |
| 153     | 43.7  | 35.7  | 53.5  |
| 76      | 16.5  | 12.4  | 18.6  |



#44  
 Dimethylphthalate  
 Concen: 31.60 ng/ul  
 RT: 13.99 min Scan# 1854  
 Delta R.T. 0.00 min  
 Lab File: BM009012.D  
 Acq: 24 Feb 2017 11:37

| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 163     | 100   |       |       |
| 194     | 4.3   | 3.0   | 4.6   |
| 164     | 9.5   | 8.6   | 12.8  |









#45  
2,6-Dinitrotoluene  
Concen: 18.37 ng/ul  
RT: 14.12 min Scan# 1875  
Delta R.T. 0.00 min  
Lab File: BM009012.D  
Acq: 24 Feb 2017 11:37




Instrument :  
BNA\_M  
ClientSampled :  
DABP4


Tgt Ion:165 Resp: 104341  
Ion Ratio Lower Upper  
165 100  
89 91.0 89.4 134.0  
121 24.2 21.6 32.4



#52  
4-Nitrophenol  
Concen: 48.90 ng/ul  
RT: 14.74 min Scan# 1981  
Delta R.T. 0.00 min  
Lab File: BM009012.D  
Acq: 24 Feb 2017 11:37

Tgt Ion:109 Resp: 325868  
Ion Ratio Lower Upper  
109 100  
139 77.0 33.9 50.9#  
65 124.8 92.2 138.2





#53

Dibenzofuran

Concen: 11.48 ng/ul

RT: 14.93 min Scan# 2014

Delta R.T. -0.01 min

Lab File: BM009012.D

Acq: 24 Feb 2017 11:37

Instrument :

BNA\_M

ClientSampleId :


DABP4

Tgt Ion:168 Resp: 430409

Ion Ratio Lower Upper




168 100

139 43.3 35.6 53.4



Abundance Ion 168.00 (167.70 to 168.70): E

Ion 139.00 (138.70 to 139.70): E



#63

4,6-Dinitro-2-methylphenol

Concen: 18.41 ng/ul

RT: 15.66 min Scan# 2138

Delta R.T. 0.00 min

Lab File: BM009012.D

Acq: 24 Feb 2017 11:37


Tgt Ion:198 Resp: 126845

Ion Ratio Lower Upper

198 100

182 3.2 3.0 4.4



77 30.3 37.6 56.4#




Abundance Ion 198.00 (197.70 to 198.70): E

Ion 182.00 (181.70 to 182.70): E

Ion 77.00 (76.70 to 77.70): BM









#66  
Hexachlorobenzene  
Concen: 8.09 ng/ul  
RT: 16.58 min Scan# 2294  
Delta R.T. 0.00 min  
Lab File: BM009012.D  
Acq: 24 Feb 2017 11:37




Instrument :  
BNA\_M  
ClientSampled :  
DABP4


| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 284     | 100   |       |       |
| 142     | 39.4  | 31.6  | 47.4  |
| 249     | 36.3  | 29.0  | 43.6  |



#68  
Pentachlorophenol  
Concen: 16.15 ng/ul  
RT: 16.93 min Scan# 2353  
Delta R.T. 0.00 min  
Lab File: BM009012.D  
Acq: 24 Feb 2017 11:37

| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 266     | 100   |       |       |
| 264     | 66.5  | 54.8  | 82.2  |
| 268     | 70.5  | 62.1  | 93.1  |





#69

Phenanthrene

Concen: 24.64 ng/ul

RT: 17.33 min Scan# 2421

Delta R.T. 0.00 min

Lab File: BM009012.D

Acq: 24 Feb 2017 11:37

Instrument :

BNA\_M

ClientSampled :

DABP4


Tgt Ion:178 Resp: 1484328

Ion Ratio Lower Upper




178 100

179 15.0 12.6 19.0

176 18.8 15.8 23.6



Abundance Ion 178.00 (177.70 to 178.70): E  
Ion 179.00 (178.70 to 179.70): E  
Ion 176.00 (175.70 to 176.70): E



#71

Anthracene

Concen: 10.24 ng/ul

RT: 17.42 min Scan# 2436

Delta R.T. 0.00 min

Lab File: BM009012.D

Acq: 24 Feb 2017 11:37


Tgt Ion:178 Resp: 629082

Ion Ratio Lower Upper



178 100


179 15.8 13.1 19.7

176 17.0 15.6 23.4



Abundance Ion 178.00 (177.70 to 178.70): E  
Ion 179.00 (178.70 to 179.70): E  
Ion 176.00 (175.70 to 176.70): E





#73

Di-n-butylphthalate

Concen: 15.84 ng/ul

RT: 18.24 min Scan# 2576

Delta R.T. 0.00 min

Lab File: BM009012.D

Acq: 24 Feb 2017 11:37

Instrument :

BNA\_M

ClientSampleId :

DABP4


Tgt Ion:149 Resp: 976039

Ion Ratio Lower Upper

149 100

150 9.4 7.1 10.7




104 7.0 6.9 10.3



Abundance Ion 149.00 (148.70 to 149.70): E

Ion 150.00 (149.70 to 150.70): E

Ion 104.00 (103.70 to 104.70): E



#74

Fluoranthene

Concen: 20.41 ng/ul

RT: 19.34 min Scan# 2763

Delta R.T. 0.00 min

Lab File: BM009012.D

Acq: 24 Feb 2017 11:37


Tgt Ion:202 Resp: 1486623

Ion Ratio Lower Upper

202 100

101 8.0 5.7 8.5



100 7.1 4.6 7.0#




Abundance Ion 202.00 (201.70 to 202.70): E

Ion 101.00 (100.70 to 101.70): E

Ion 100.00 (99.70 to 100.70): B





#77

Pyrene

Concen: 21.58 ng/ul

RT: 19.70 min Scan# 2825

Delta R.T. 0.00 min

Lab File: BM009012.D

Acq: 24 Feb 2017 11:37

Instrument :

BNA\_M

ClientSampleId :

DABP4


Tgt Ion:202 Resp: 1617064

Ion Ratio Lower Upper

202 100

101 10.6 6.2 9.4#




100 9.1 6.4 9.6



Abundance Ion 202.00 (201.70 to 202.70): E

Ion 101.00 (100.70 to 101.70): E

Ion 100.00 (99.70 to 100.70): B



#80

Benzo(a)anthracene

Concen: 16.93 ng/ul

RT: 21.49 min Scan# 3129

Delta R.T. 0.05 min

Lab File: BM009012.D

Acq: 24 Feb 2017 11:37


Tgt Ion:228 Resp: 1295820

Ion Ratio Lower Upper

228 100

229 18.6 15.4 23.2



226 29.6 21.2 31.8




Abundance Ion 228.00 (227.70 to 228.70): E

Ion 229.00 (228.70 to 229.70): E

Ion 226.00 (225.70 to 226.70): E









#81  
 Bis(2-ethylhexyl)phthalate  
 Concen: 20.98 ng/ul  
 RT: 21.35 min Scan# 3105  
 Delta R.T. 0.00 min  
 Lab File: BM009012.D  
 Acq: 24 Feb 2017 11:37




Instrument :  
 BNA\_M  
 ClientSampleId :  
 DABP4


| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 149     | 100   |       |       |
| 167     | 28.8  | 20.2  | 30.2  |
| 279     | 4.4   | 4.0   | 6.0   |



#82  
 Chrysene  
 Concen: 17.81 ng/ul  
 RT: 21.49 min Scan# 3129  
 Delta R.T. 0.00 min  
 Lab File: BM009012.D  
 Acq: 24 Feb 2017 11:37

| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 228     | 100   |       |       |
| 226     | 29.6  | 23.9  | 35.9  |
| 229     | 18.6  | 15.4  | 23.2  |





#88

Benzo(a)pyrene

Concen: 16.08 ng/ul

RT: 23.70 min Scan# 3504

Delta R.T. 0.00 min

Lab File: BM009012.D

Acq: 24 Feb 2017 11:37

Instrument :

BNA\_M

ClientSampleId :

DABP4


Tgt Ion:252 Resp: 1119779

Ion Ratio Lower Upper

252 100

253 21.6 17.4 26.0




125 6.8 3.7 5.5#



Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E



#89

Indeno(1,2,3-cd)pyrene

Concen: 17.73 ng/ul

RT: 26.19 min Scan# 3928

Delta R.T. 0.00 min

Lab File: BM009012.D

Acq: 24 Feb 2017 11:37


Tgt Ion:276 Resp: 1402353

Ion Ratio Lower Upper

276 100

138 15.7 8.2 12.2#



277 24.0 21.5 32.3




Abundance Ion 276.00 (275.70 to 276.70): E

Ion 138.00 (137.70 to 138.70): E

Ion 277.00 (276.70 to 277.70): E





#90

Dibenzo(a,h)anthracene

Concen: 11.15 ng/ul

RT: 26.21 min Scan# 3931

Delta R.T. 0.01 min

Lab File: BM009012.D

Acq: 24 Feb 2017 11:37

Instrument :

BNA\_M

ClientSampleId :

DABP4


Tgt Ion: 278 Resp: 751044

Ion Ratio Lower Upper

278 100

139 14.0 5.8 8.8#

279 22.6 17.9 26.9



Abundance Ion 278.00 (277.70 to 278.70): E

Ion 139.00 (138.70 to 139.70): E

Ion 279.00 (278.70 to 279.70): E

