

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM022421\
 Data File : BM028877.D
 Acq On : 24 Feb 2021 13:05
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC040

Quant Time: Feb 24 13:43:19 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM022321.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Feb 23 16:39:41 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	105	0.00
2	1,4-Dioxane	40.000	41.204	-3.0	106	0.00
3	Pyridine	40.000	40.587	-1.5	106	0.00
4	n-Nitrosodimethylamine	40.000	42.068	-5.2	107	0.00
5 S	2-Fluorophenol	80.000	79.149	1.1	102	0.00
6	Aniline	40.000	38.797	3.0	102	0.00
7 S	Phenol-d6	80.000	78.539	1.8	103	0.00
8	2-Chlorophenol	40.000	39.088	2.3	103	0.00
9	Benzaldehyde	40.000	35.409	11.5	105	0.00
10 C	Phenol	40.000	38.823	2.9	102	0.00
11	bis(2-Chloroethyl)ether	40.000	38.907	2.7	102	0.00
12	1,3-Dichlorobenzene	40.000	38.934	2.7	102	0.00
13 C	1,4-Dichlorobenzene	40.000	39.093	2.3	102	0.00
14	1,2-Dichlorobenzene	40.000	39.122	2.2	103	0.00
15	Benzyl Alcohol	40.000	38.910	2.7	101	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	39.471	1.3	105	0.00
17	2-Methylphenol	40.000	38.552	3.6	100	0.00
18	Hexachloroethane	40.000	39.324	1.7	103	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	38.878	2.8	100	0.00
20	3+4-Methylphenols	40.000	39.037	2.4	102	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	100	0.00
22	Acetophenone	40.000	40.583	-1.5	101	0.00
23 S	Nitrobenzene-d5	80.000	81.877	-2.3	102	0.00
24	Nitrobenzene	40.000	41.150	-2.9	102	0.00
25	Isophorone	40.000	39.738	0.7	99	0.00
26 C	2-Nitrophenol	40.000	41.436	-3.6	101	0.00
27	2,4-Dimethylphenol	40.000	39.895	0.3	100	0.00
28	bis(2-Chloroethoxy)methane	40.000	40.194	-0.5	100	0.00
29 C	2,4-Dichlorophenol	40.000	40.334	-0.8	100	0.00
30	1,2,4-Trichlorobenzene	40.000	40.462	-1.2	101	0.00
31	Naphthalene	40.000	39.931	0.2	101	0.00
32	Benzoic acid	40.000	40.827	-2.1	96	0.00
33	4-Chloroaniline	40.000	39.333	1.7	98	0.00
34 C	Hexachlorobutadiene	40.000	40.917	-2.3	103	0.00
35	Caprolactam	40.000	37.834	5.4	93	0.00
36 C	4-Chloro-3-methylphenol	40.000	39.526	1.2	99	0.00
37	2-Methylnaphthalene	40.000	39.737	0.7	100	0.00
38	1-Methylnaphthalene	40.000	39.127	2.2	99	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	96	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	40.908	-2.3	99	0.00
41 P	Hexachlorocyclopentadiene	40.000	43.461	-8.7	101	0.00
42 S	2,4,6-Tribromophenol	80.000	78.869	1.4	91	0.00
43 C	2,4,6-Trichlorophenol	40.000	41.378	-3.4	97	0.00
44	2,4,5-Trichlorophenol	40.000	41.097	-2.7	98	0.00
45 S	2-Fluorobiphenyl	80.000	81.766	-2.2	99	0.00
46	1,1'-Biphenyl	40.000	40.542	-1.4	98	0.00
47	2-Chloronaphthalene	40.000	40.586	-1.5	98	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM022421\
 Data File : BM028877.D
 Acq On : 24 Feb 2021 13:05
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC040

Quant Time: Feb 24 13:43:19 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM022321.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Feb 23 16:39:41 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	40.969	-2.4	95	0.00
49	Acenaphthylene	40.000	40.526	-1.3	98	0.00
50	Dimethylphthalate	40.000	39.320	1.7	93	0.00
51	2,6-Dinitrotoluene	40.000	40.111	-0.3	93	0.00
52 C	Acenaphthene	40.000	40.220	-0.5	97	0.00
53	3-Nitroaniline	40.000	40.664	-1.7	93	0.00
54 P	2,4-Dinitrophenol	40.000	40.086	-0.2	91	0.00
55	Dibenzofuran	40.000	39.753	0.6	96	0.00
56 P	4-Nitrophenol	40.000	38.907	2.7	91	0.00
57	2,4-Dinitrotoluene	40.000	40.692	-1.7	93	0.00
58	Fluorene	40.000	39.640	0.9	95	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	40.436	-1.1	95	0.00
60	Diethylphthalate	40.000	38.997	2.5	92	0.00
61	4-Chlorophenyl-phenylether	40.000	39.473	1.3	95	0.00
62	4-Nitroaniline	40.000	40.310	-0.8	90	0.00
63	Azobenzene	40.000	40.401	-1.0	95	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	91	0.00
65	4,6-Dinitro-2-methylphenol	40.000	40.322	-0.8	89	0.00
66 c	n-Nitrosodiphenylamine	40.000	40.640	-1.6	93	0.00
67	4-Bromophenyl-phenylether	40.000	41.035	-2.6	94	0.00
68	Hexachlorobenzene	40.000	40.387	-1.0	94	0.00
69	Atrazine	40.000	39.656	0.9	89	0.00
70 C	Pentachlorophenol	40.000	42.122	-5.3	90	0.00
71	Phenanthrene	40.000	40.044	-0.1	92	0.00
72	Anthracene	40.000	40.441	-1.1	91	0.00
73	Carbazole	40.000	39.707	0.7	89	0.00
74	Di-n-butylphthalate	40.000	39.306	1.7	86	0.00
75 C	Fluoranthene	40.000	39.106	2.2	87	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	85	0.00
77	Benzidine	40.000	35.855	10.4	66	0.00
78	Pyrene	40.000	40.419	-1.0	87	0.00
79 S	Terphenyl-d14	80.000	80.041	-0.1	86	0.00
80	Butylbenzylphthalate	40.000	39.102	2.2	82	0.00
81	Benzo(a)anthracene	40.000	39.520	1.2	85	0.00
82	3,3'-Dichlorobenzidine	40.000	39.758	0.6	82	0.00
83	Chrysene	40.000	39.656	0.9	84	-0.01
84	Bis(2-ethylhexyl)phthalate	40.000	38.274	4.3	79	0.00
85 c	Di-n-octyl phthalate	40.000	38.533	3.7	79	0.00
86 I	Perylene-d12	20.000	20.000	0.0	82	-0.01
87	Indeno(1,2,3-cd)pyrene	40.000	40.038	-0.1	81	-0.02
88	Benzo(b)fluoranthene	40.000	39.398	1.5	80	0.00
89	Benzo(k)fluoranthene	40.000	40.710	-1.8	85	-0.01
90 C	Benzo(a)pyrene	40.000	39.934	0.2	82	-0.01
91	Dibenzo(a,h)anthracene	40.000	40.264	-0.7	82	-0.02
92	Benzo(g,h,i)perylene	40.000	40.292	-0.7	82	-0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM022421\
Data File : BM028877.D
Acq On : 24 Feb 2021 13:05
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
LabSampleId :
SSTDCCC040

Quant Time: Feb 24 13:43:19 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM022321.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Feb 23 16:39:41 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0