

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM022424\
 Data File : BM044323.D
 Acq On : 25 Feb 2024 09:13
 Operator : MA/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 47 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD020197

Quant Time: Feb 25 23:08:11 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM022324.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sun Feb 25 22:57:38 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.922	152	322766	20.000	ng/ul	0.00
20) Naphthalene-d8	10.727	136	1440747	20.000	ng/ul	0.01
38) Acenaphthene-d10	14.568	164	862928	20.000	ng/ul	0.01
64) Phenanthrene-d10	17.315	188	1860698	20.000	ng/ul	0.00
79) Chrysene-d12	21.515	240	1662650	20.000	ng/ul	0.00
88) Perylene-d12	23.927	264	1874169	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.357	96	73371	7.172	ng/uL	0.00
4) Pyridine-d5	3.769	84	521771	18.676	ng/ul	0.00
7) Phenol-d5	7.087	99	595267	19.252	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.269	67	373537	18.656	ng/ul	0.01
11) 2-Chlorophenol-d4	7.451	132	466749	20.043	ng/ul	0.01
15) 4-Methylphenol-d8	8.639	113	453787	19.396	ng/ul	0.01
21) Nitrobenzene-d5	9.092	128	225830	19.278	ng/ul	0.00
24) 2-Nitrophenol-d4	9.816	143	246326	20.610	ng/ul	0.01
28) 2,4-Dichlorophenol-d3	10.345	165	424976	20.158	ng/ul	0.00
31) 4-Chloroaniline-d4	10.874	131	677801	19.143	ng/ul	0.01
46) Dimethylphthalate-d6	13.992	166	1323301	19.883	ng/ul	0.01
49) Acenaphthylene-d8	14.257	160	1502533	19.622	ng/ul	0.00
54) 4-Nitrophenol-d4	14.774	143	266400	20.116	ng/ul	0.01
60) Fluorene-d10	15.562	176	1101521	19.555	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.686	200	216544	18.756	ng/ul	0.00
73) Anthracene-d10	17.415	188	1740576	19.619	ng/ul	0.00
81) Pyrene-d10	19.715	212	1954465	19.539	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.768	264	1863359	19.403	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.393	88	78885	7.316	ng/uL	97
5) Pyridine	3.787	79	544431	18.627	ng/ul	99
6) Benzaldehyde	7.069	77	327830	24.425	ng/ul	100
8) Phenol	7.116	94	608010	18.917	ng/ul	99
10) Bis(2-Chloroethyl)ether	7.357	93	510134	18.963	ng/ul	99
12) 2-Chlorophenol	7.481	128	473441	19.651	ng/ul	97
13) 2-Methylphenol	8.369	108	455345	19.230	ng/ul	100
14) 2,2'-oxybis(1-Chloropr...	8.451	45	679860	18.484	ng/ul	98
16) Acetophenone	8.757	105	743366	19.888	ng/ul	98
17) N-Nitroso-di-n-propyla...	8.745	70	391352	20.901	ng/ul	99
18) 4-Methylphenol	8.698	108	487917	19.344	ng/ul	100
19) Hexachloroethane	8.992	117	196967	19.271	ng/ul	96
22) Nitrobenzene	9.139	77	537579	18.975	ng/ul	98
23) Isophorone	9.663	82	1097009	19.980	ng/ul	100
25) 2-Nitrophenol	9.845	139	267398	20.173	ng/ul	99
26) 2,4-Dimethylphenol	9.904	107	494035	19.519	ng/ul	99
27) Bis(2-Chloroethoxy)met...	10.157	93	690706	19.351	ng/ul	100
29) 2,4-Dichlorophenol	10.375	162	426719	19.999	ng/ul	100
30) Naphthalene	10.780	128	1556164	18.992	ng/ul	99
32) 4-Chloroaniline	10.898	127	665205	19.279	ng/ul	99
33) Hexachlorobutadiene	11.051	225	244038	18.984	ng/ul	98
34) Caprolactam	11.680	113	151205	20.501	ng/ul	96
35) 4-Chloro-3-methylphenol	12.021	107	480021	21.117	ng/ul	99
36) 2-Methylnaphthalene	12.392	142	1029155	19.719	ng/ul	99


Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM022424\
 Data File : BM044323.D
 Acq On : 25 Feb 2024 09:13
 Operator : MA/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 47 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD020197

Quant Time: Feb 25 23:08:11 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM022324.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sun Feb 25 22:57:38 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.610	142	1017996	19.731	ng/ul	99
39) 1,2,4,5-Tetrachloroben...	12.751	216	479102	18.755	ng/ul	99
40) Hexachlorocyclopentadiene	12.727	237	297666	17.319	ng/ul	99
41) 2,4,6-Trichlorophenol	12.998	196	330156	20.276	ng/ul	97
42) 2,4,5-Trichlorophenol	13.068	196	354776	20.155	ng/ul	99
43) 1,1'-Biphenyl	13.404	154	1314226	18.707	ng/ul	98
44) 2-Chloronaphthalene	13.445	162	1041065	18.926	ng/ul	99
45) 2-Nitroaniline	13.657	65	318680	20.925	ng/ul	98
47) Dimethylphthalate	14.039	163	1353974	19.905	ng/ul	99
48) 2,6-Dinitrotoluene	14.163	165	268503	21.108	ng/ul	96
50) Acenaphthylene	14.292	152	1754659	19.346	ng/ul	99
51) 3-Nitroaniline	14.486	138	289100	21.543	ng/ul	94
52) Acenaphthene	14.627	153	1132586	18.916	ng/ul	100
53) 2,4-Dinitrophenol	14.692	184	145847	18.477	ng/ul	97
55) 4-Nitrophenol	14.786	109	185887	20.432	ng/ul	98
56) Dibenzofuran	14.968	168	1542058	19.185	ng/ul	99
57) 2,4-Dinitrotoluene	14.945	165	396978	21.723	ng/ul	97
58) 2,3,4,6-Tetrachlorophenol	15.192	232	302781	21.374	ng/ul	98
59) Diethylphthalate	15.404	149	1442377	20.600	ng/ul	99
61) Fluorene	15.615	166	1272249	19.707	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.615	204	604256	19.861	ng/ul	99
63) 4-Nitroaniline	15.645	138	298028	22.961	ng/ul	99
66) 4,6-Dinitro-2-methylph...	15.704	198	238188	18.859	ng/ul	99
67) N-Nitrosodiphenylamine	15.833	169	1096410	19.444	ng/ul	99
68) 4-Bromophenyl-phenylether	16.515	248	330867	17.897	ng/ul	97
69) Hexachlorobenzene	16.609	284	406650	19.218	ng/ul	99
70) Atrazine	16.792	200	373449	20.218	ng/ul	100
71) Pentachlorophenol	16.962	266	263577	18.902	ng/ul	98
72) Phenanthrene	17.362	178	2024907	18.970	ng/ul	100
74) Anthracene	17.451	178	2089046	19.438	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.363	216	471210	18.091	ng/ul	99
76) Pentachlorobenzene	14.880	250	455708	18.327	ng/ul	97
77) Carbazole	17.727	167	1957044	19.366	ng/ul	100
78) Di-n-butylphthalate	18.303	149	2679327	20.940	ng/ul	99
80) Fluoranthene	19.380	202	2315233	19.299	ng/ul	99
82) Pyrene	19.745	202	2426391	19.286	ng/ul	99
83) Butylbenzylphthalate	20.662	149	1208748	21.390	ng/ul	99
84) 3,3'-Dichlorobenzidine	21.439	252	735767	19.784	ng/ul	99
85) Benzo(a)anthracene	21.503	228	2422742	19.538	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.450	149	1891149	21.859	ng/ul#	99
87) Chrysene	21.556	228	2267264	19.451	ng/ul	99
89) Di-n-octyl phthalate	22.391	149	3264375	20.329	ng/ul	100
90) Benzo(b)fluoranthene	23.191	252	2354509	19.596	ng/ul	100
91) Benzo(k)fluoranthene	23.244	252	2395788	19.414	ng/ul	99
93) Benzo(a)pyrene	23.821	252	2245946	19.388	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.432	276	2672977	19.081	ng/ul	99
95) Dibenzo(a,h)anthracene	26.462	278	2244931	19.177	ng/ul	99
96) Benzo(g,h,i)perylene	27.203	276	2122910	18.734	ng/ul	98





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane
Concen: 7.316 ng/uL
RT: 3.393 min Scan# 51
Delta R.T. 0.006 min
Lab File: BM044323.D
Acq: 25 Feb 2024 09:13




Instrument :
BNA_M
ClientSampleId :
SSTD020197


Tgt Ion: 88 Resp: 78885
Ion Ratio Lower Upper
88 100
43 29.6 21.2 31.8
58 74.6 58.3 87.5



#5
Pyridine
Concen: 18.627 ng/uL
RT: 3.787 min Scan# 119
Delta R.T. -0.000 min
Lab File: BM044323.D
Acq: 25 Feb 2024 09:13

Tgt Ion: 79 Resp: 544431
Ion Ratio Lower Upper
79 100
52 74.2 58.9 88.3
51 33.5 27.8 41.8









#6
Benzaldehyde
Concen: 24.425 ng/ul
RT: 7.069 min Scan# 61
Delta R.T. 0.006 min
Lab File: BM044323.D
Acq: 25 Feb 2024 09:13




Instrument :
BNA_M
ClientSampleId :
SSTD020197


Tgt Ion: 77 Resp: 327830
Ion Ratio Lower Upper
77 100
105 98.5 79.0 118.6
106 97.3 77.4 116.2



#8
Phenol
Concen: 18.917 ng/ul
RT: 7.116 min Scan# 685
Delta R.T. 0.011 min
Lab File: BM044323.D
Acq: 25 Feb 2024 09:13

Tgt Ion: 94 Resp: 608010
Ion Ratio Lower Upper
94 100
65 27.0 20.6 31.0
66 31.9 25.4 38.0









#10
Bis(2-Chloroethyl)ether
Concen: 18.963 ng/ul
RT: 7.357 min Scan# 71
Delta R.T. 0.006 min
Lab File: BM044323.D
Acq: 25 Feb 2024 09:13




Instrument :
BNA_M
ClientSampleId :
SSTD020197


Tgt Ion: 93 Resp: 510134
Ion Ratio Lower Upper
93 100
63 75.2 61.0 91.6
95 32.5 25.1 37.7



#12
2-Chlorophenol
Concen: 19.651 ng/ul
RT: 7.481 min Scan# 747
Delta R.T. 0.006 min
Lab File: BM044323.D
Acq: 25 Feb 2024 09:13

Tgt Ion:128 Resp: 473441
Ion Ratio Lower Upper
128 100
64 46.5 39.0 58.6
130 32.4 25.4 38.0









#13
2-Methylphenol
Concen: 19.230 ng/ul
RT: 8.369 min Scan# 896
Delta R.T. 0.012 min
Lab File: BM044323.D
Acq: 25 Feb 2024 09:13




Instrument :
BNA_M
ClientSampleId :
SSTD020197


Tgt Ion:108 Resp: 455345
Ion Ratio Lower Upper
108 100
107 88.3 70.6 105.8



#14
2,2'-oxybis(1-Chloropropane)
Concen: 18.484 ng/ul
RT: 8.451 min Scan# 912
Delta R.T. 0.006 min
Lab File: BM044323.D
Acq: 25 Feb 2024 09:13

Tgt Ion: 45 Resp: 679860
Ion Ratio Lower Upper
45 100
77 15.0 11.3 16.9
79 10.9 8.5 12.7









#16
Acetophenone
Concen: 19.888 ng/ul
RT: 8.757 min Scan# 90
Delta R.T. 0.012 min
Lab File: BM044323.D
Acq: 25 Feb 2024 09:13




Instrument :
BNA_M
ClientSampleId :
SSTD020197


Tgt Ion:105 Resp: 743366
Ion Ratio Lower Upper
105 100
77 74.9 61.2 91.8
51 25.5 21.1 31.7



#17
N-Nitroso-di-n-propylamine
Concen: 20.901 ng/ul
RT: 8.745 min Scan# 962
Delta R.T. 0.012 min
Lab File: BM044323.D
Acq: 25 Feb 2024 09:13

Tgt Ion: 70 Resp: 391352
Ion Ratio Lower Upper
70 100
42 54.1 44.4 66.6
101 9.5 7.5 11.3
130 20.8 17.0 25.4









#18
4-Methylphenol
Concen: 19.344 ng/ul
RT: 8.698 min Scan# 91
Delta R.T. 0.006 min
Lab File: BM044323.D
Acq: 25 Feb 2024 09:13




Instrument :
BNA_M
ClientSampleId :
SSTD020197


Tgt Ion:108 Resp: 487917
Ion Ratio Lower Upper
108 100
107 108.9 86.8 130.2



#19
Hexachloroethane
Concen: 19.271 ng/ul
RT: 8.992 min Scan# 1004
Delta R.T. 0.006 min
Lab File: BM044323.D
Acq: 25 Feb 2024 09:13

Tgt Ion:117 Resp: 196967
Ion Ratio Lower Upper
117 100
201 89.0 69.8 104.6
199 56.1 40.8 61.2









#22
Nitrobenzene
Concen: 18.975 ng/ul
RT: 9.139 min Scan# 1027
Delta R.T. 0.011 min
Lab File: BM044323.D
Acq: 25 Feb 2024 09:13

Instrument :
BNA_M
ClientSampleId :
SSTD020197

Tgt Ion: 77 Resp: 537579




Ion	Ratio	Lower	Upper
77	100		
123	47.4	36.9	55.3
65	13.3	10.2	15.2




#23
Isophorone
Concen: 19.980 ng/ul
RT: 9.663 min Scan# 1118
Delta R.T. 0.012 min
Lab File: BM044323.D
Acq: 25 Feb 2024 09:13

Tgt Ion: 82 Resp: 1097009

Ion	Ratio	Lower	Upper
82	100		
95	6.0	4.6	7.0
138	17.0	13.5	20.3









#25
2-Nitrophenol
Concen: 20.173 ng/ul
RT: 9.845 min Scan# 1147
Delta R.T. 0.011 min
Lab File: BM044323.D
Acq: 25 Feb 2024 09:13




Instrument :
BNA_M
ClientSampleId :
SSTD020197


Tgt Ion:139 Resp: 267398
Ion Ratio Lower Upper
139 100
65 49.8 41.0 61.4
109 25.5 20.5 30.7



#26
2,4-Dimethylphenol
Concen: 19.519 ng/ul
RT: 9.904 min Scan# 1159
Delta R.T. 0.006 min
Lab File: BM044323.D
Acq: 25 Feb 2024 09:13

Tgt Ion:107 Resp: 494035
Ion Ratio Lower Upper
107 100
121 50.3 41.1 61.7
122 90.3 72.3 108.5









#27
 Bis(2-Chloroethoxy)methane
 Concen: 19.351 ng/ul
 RT: 10.157 min Scan# 1202
 Delta R.T. 0.011 min
 Lab File: BM044323.D
 Acq: 25 Feb 2024 09:13




Instrument :
 BNA_M
 ClientSampleId :
 SSTD020197


Tgt Ion: 93 Resp: 690706
 Ion Ratio Lower Upper
 93 100
 95 32.5 26.0 39.0
 123 11.8 9.0 13.6



#29
 2,4-Dichlorophenol
 Concen: 19.999 ng/ul
 RT: 10.375 min Scan# 1239
 Delta R.T. 0.011 min
 Lab File: BM044323.D
 Acq: 25 Feb 2024 09:13

Tgt Ion: 162 Resp: 426719
 Ion Ratio Lower Upper
 162 100
 164 64.8 51.8 77.8
 98 37.3 29.6 44.4









#30
Naphthalene
Concen: 18.992 ng/ul
RT: 10.780 min Scan# 1308
Delta R.T. 0.011 min
Lab File: BM044323.D
Acq: 25 Feb 2024 09:13




Instrument :
BNA_M
ClientSampleId :
SSTD020197


Tgt Ion:128 Resp: 1556164
Ion Ratio Lower Upper
128 100
129 10.8 8.8 13.2
127 13.0 10.2 15.2



#32
4-Chloroaniline
Concen: 19.279 ng/ul
RT: 10.898 min Scan# 1328
Delta R.T. 0.006 min
Lab File: BM044323.D
Acq: 25 Feb 2024 09:13

Tgt Ion:127 Resp: 665205
Ion Ratio Lower Upper
127 100
129 32.4 25.3 37.9





#33

Hexachlorobutadiene

Concen: 18.984 ng/ul

RT: 11.051 min Scan# 1

Delta R.T. 0.006 min

Lab File: BM044323.D

Acq: 25 Feb 2024 09:13

Instrument :

BNA_M

ClientSampleId :

SSTD020197





Tgt Ion:225 Resp: 244038

Ion Ratio Lower Upper

225 100

223 61.2 49.7 74.5

227 62.6 52.2 78.2



#34

Caprolactam

Concen: 20.501 ng/ul

RT: 11.680 min Scan# 1461

Delta R.T. 0.012 min

Lab File: BM044323.D

Acq: 25 Feb 2024 09:13




Tgt Ion:113 Resp: 151205


Ion Ratio Lower Upper

113 100

55 169.0 140.7 211.1

56 125.4 102.4 153.6









#35
4-Chloro-3-methylphenol
Concen: 21.117 ng/ul
RT: 12.021 min Scan# 1519
Delta R.T. 0.012 min
Lab File: BM044323.D
Acq: 25 Feb 2024 09:13




Instrument :
BNA_M
ClientSampleId :
SSTD020197


Tgt Ion:107 Resp: 480021
Ion Ratio Lower Upper
107 100
144 27.0 20.6 30.8
142 81.3 64.3 96.5



#36
2-Methylnaphthalene
Concen: 19.719 ng/ul
RT: 12.392 min Scan# 1582
Delta R.T. 0.011 min
Lab File: BM044323.D
Acq: 25 Feb 2024 09:13

Tgt Ion:142 Resp: 1029155
Ion Ratio Lower Upper
142 100
141 88.7 70.5 105.7





#37

1-Methylnaphthalene

Concen: 19.731 ng/ul

RT: 12.610 min Scan# 1

Delta R.T. 0.011 min

Lab File: BM044323.D

Acq: 25 Feb 2024 09:13

Instrument :

BNA_M

ClientSampleId :

SSTD020197





Tgt Ion:142 Resp: 1017996

Ion Ratio Lower Upper

142 100

141 90.4 73.0 109.4

116 3.4 2.7 4.1



#39

1,2,4,5-Tetrachlorobenzene

Concen: 18.755 ng/ul

RT: 12.751 min Scan# 1643

Delta R.T. 0.006 min

Lab File: BM044323.D

Acq: 25 Feb 2024 09:13

Tgt Ion:216 Resp: 479102




Ion Ratio Lower Upper


216 100

214 79.4 63.9 95.9

179 22.3 17.5 26.3

108 19.8 16.2 24.4





#40

Hexachlorocyclopentadiene

Concen: 17.319 ng/ul

RT: 12.727 min Scan# 1

Delta R.T. 0.011 min

Lab File: BM044323.D

Acq: 25 Feb 2024 09:13

Instrument :

BNA_M

ClientSampleId :

SSTD020197





Tgt Ion:237 Resp: 297666

Ion Ratio Lower Upper

237 100

235 61.2 49.4 74.2

272 13.3 10.6 16.0



#41

2,4,6-Trichlorophenol

Concen: 20.276 ng/ul

RT: 12.998 min Scan# 1685

Delta R.T. 0.006 min

Lab File: BM044323.D

Acq: 25 Feb 2024 09:13




Tgt Ion:196 Resp: 330156


Ion Ratio Lower Upper

196 100

198 92.1 76.1 114.1

200 30.0 25.1 37.7





#42

2,4,5-Trichlorophenol

Concen: 20.155 ng/ul

RT: 13.068 min Scan# 1

Delta R.T. 0.006 min

Lab File: BM044323.D

Acq: 25 Feb 2024 09:13

Instrument :

BNA_M

ClientSampleId :

SSTD020197





Tgt Ion:196 Resp: 354776

Ion Ratio Lower Upper

196 100

198 96.0 77.4 116.2

200 30.1 25.6 38.4



#43

1,1'-Biphenyl

Concen: 18.707 ng/ul

RT: 13.404 min Scan# 1754

Delta R.T. 0.006 min

Lab File: BM044323.D

Acq: 25 Feb 2024 09:13




Tgt Ion:154 Resp: 1314226


Ion Ratio Lower Upper

154 100

153 38.9 32.4 48.6

76 14.2 11.6 17.4





#44

2-Chloronaphthalene

Concen: 18.926 ng/ul

RT: 13.445 min Scan# 1

Delta R.T. 0.011 min

Lab File: BM044323.D

Acq: 25 Feb 2024 09:13

Instrument :

BNA_M

ClientSampleId :

SSTD020197





Tgt Ion:162 Resp: 1041065

Ion Ratio Lower Upper

162 100

164 33.3 26.1 39.1

127 38.1 30.9 46.3



#45

2-Nitroaniline

Concen: 20.925 ng/ul

RT: 13.657 min Scan# 1797

Delta R.T. 0.006 min

Lab File: BM044323.D

Acq: 25 Feb 2024 09:13




Tgt Ion: 65 Resp: 318680


Ion Ratio Lower Upper

65 100

92 69.0 52.5 78.7

138 108.3 86.4 129.6





#47

Dimethylphthalate

Concen: 19.905 ng/ul

RT: 14.039 min Scan# 1862

Delta R.T. 0.011 min

Lab File: BM044323.D

Acq: 25 Feb 2024 09:13

Instrument :

BNA_M

ClientSampleId :

SSTD020197





Tgt Ion:163 Resp: 1353974

Ion Ratio Lower Upper

163 100

194 3.9 3.0 4.6

164 10.3 8.1 12.1



#48

2,6-Dinitrotoluene

Concen: 21.108 ng/ul

RT: 14.163 min Scan# 1883

Delta R.T. 0.011 min

Lab File: BM044323.D

Acq: 25 Feb 2024 09:13




Tgt Ion:165 Resp: 268503


Ion Ratio Lower Upper

165 100

89 56.8 48.2 72.2

121 22.1 18.7 28.1





#50

Acenaphthylene

Concen: 19.346 ng/ul

RT: 14.292 min Scan# 1905

Delta R.T. 0.012 min

Lab File: BM044323.D

Acq: 25 Feb 2024 09:13

Instrument :

BNA_M

ClientSampleId :

SSTD020197





Tgt Ion:152 Resp: 1754659

Ion Ratio Lower Upper

152 100

151 19.1 15.8 23.6

153 12.8 10.4 15.6



#51

3-Nitroaniline

Concen: 21.543 ng/ul

RT: 14.486 min Scan# 1938

Delta R.T. 0.006 min

Lab File: BM044323.D

Acq: 25 Feb 2024 09:13




Tgt Ion:138 Resp: 289100


Ion Ratio Lower Upper

138 100

108 9.2 8.6 12.8

92 110.1 93.8 140.6





#52

Acenaphthene

Concen: 18.916 ng/ul

RT: 14.627 min Scan# 1961

Delta R.T. 0.006 min

Lab File: BM044323.D

Acq: 25 Feb 2024 09:13

Instrument :

BNA_M

ClientSampleId :

SSTD020197





Tgt Ion:153 Resp: 1132586

Ion Ratio Lower Upper

153 100

152 48.4 38.5 57.7

154 88.3 70.8 106.2



#53

2,4-Dinitrophenol

Concen: 18.477 ng/ul

RT: 14.692 min Scan# 1973

Delta R.T. 0.006 min

Lab File: BM044323.D

Acq: 25 Feb 2024 09:13




Tgt Ion:184 Resp: 145847


Ion Ratio Lower Upper

184 100

63 69.3 58.4 87.6

154 59.3 46.4 69.6









#55
4-Nitrophenol
Concen: 20.432 ng/ul
RT: 14.786 min Scan# 1988
Delta R.T. 0.006 min
Lab File: BM044323.D
Acq: 25 Feb 2024 09:13




Instrument :
BNA_M
ClientSampleId :
SSTD020197


Tgt Ion:109 Resp: 185887
Ion Ratio Lower Upper
109 100
139 154.4 123.0 184.6
65 148.1 122.2 183.2



#56
Dibenzofuran
Concen: 19.185 ng/ul
RT: 14.968 min Scan# 2020
Delta R.T. 0.011 min
Lab File: BM044323.D
Acq: 25 Feb 2024 09:13

Tgt Ion:168 Resp: 1542058
Ion Ratio Lower Upper
168 100
139 37.0 29.9 44.9





#57

2,4-Dinitrotoluene

Concen: 21.723 ng/ul

RT: 14.945 min Scan# 2016

Delta R.T. 0.012 min

Lab File: BM044323.D

Acq: 25 Feb 2024 09:13

Instrument :

BNA_M

ClientSampleId :

SSTD020197





Tgt Ion:165 Resp: 396978

Ion Ratio Lower Upper

165 100

63 45.0 37.9 56.9

182 2.6 1.8 2.6



#58

2,3,4,6-Tetrachlorophenol

Concen: 21.374 ng/ul

RT: 15.192 min Scan# 2058

Delta R.T. 0.011 min

Lab File: BM044323.D

Acq: 25 Feb 2024 09:13

Tgt Ion:232 Resp: 302781




Ion Ratio Lower Upper


232 100

131 50.6 41.9 62.9

130 2.2 1.8 2.8

166 32.9 26.9 40.3





#59

Diethylphthalate

Concen: 20.600 ng/ul

RT: 15.404 min Scan# 2093

Delta R.T. 0.006 min

Lab File: BM044323.D

Acq: 25 Feb 2024 09:13

Instrument :

BNA_M

ClientSampleId :

SSTD020197





Tgt Ion:149 Resp: 1442377

Ion Ratio Lower Upper

149 100

177 20.6 17.0 25.6

150 12.3 9.6 14.4



#61

Fluorene

Concen: 19.707 ng/ul

RT: 15.615 min Scan# 2130

Delta R.T. 0.006 min

Lab File: BM044323.D

Acq: 25 Feb 2024 09:13




Tgt Ion:166 Resp: 1272249


Ion Ratio Lower Upper

166 100

165 95.3 77.4 116.0

167 13.4 10.8 16.2









#62
4-Chlorophenyl-phenylether
Concen: 19.861 ng/ul
RT: 15.615 min Scan# 2129
Delta R.T. 0.006 min
Lab File: BM044323.D
Acq: 25 Feb 2024 09:13




Instrument :
BNA_M
ClientSampleId :
SSTD020197


Tgt Ion:204 Resp: 604256
Ion Ratio Lower Upper
204 100
206 32.3 26.9 40.3
141 60.6 48.3 72.5



#63
4-Nitroaniline
Concen: 22.961 ng/ul
RT: 15.645 min Scan# 2135
Delta R.T. 0.006 min
Lab File: BM044323.D
Acq: 25 Feb 2024 09:13

Tgt Ion:138 Resp: 298028
Ion Ratio Lower Upper
138 100
92 51.3 40.9 61.3
108 67.9 55.1 82.7









#66
4,6-Dinitro-2-methylphenol
Concen: 18.859 ng/ul
RT: 15.704 min Scan# 2145
Delta R.T. 0.011 min
Lab File: BM044323.D
Acq: 25 Feb 2024 09:13




Instrument :
BNA_M
ClientSampleId :
SSTD020197


Tgt Ion:198 Resp: 238188
Ion Ratio Lower Upper
198 100
182 4.0 3.8 5.8
77 25.4 20.6 30.8



#67
N-Nitrosodiphenylamine
Concen: 19.444 ng/ul
RT: 15.833 min Scan# 2167
Delta R.T. 0.006 min
Lab File: BM044323.D
Acq: 25 Feb 2024 09:13

Tgt Ion:169 Resp: 1096410
Ion Ratio Lower Upper
169 100
168 67.7 53.6 80.4
167 36.2 29.2 43.8









#68
4-Bromophenyl-phenylether
Concen: 17.897 ng/ul
RT: 16.515 min Scan# 21
Delta R.T. 0.006 min
Lab File: BM044323.D
Acq: 25 Feb 2024 09:13




Instrument :
BNA_M
ClientSampleId :
SSTD020197


Tgt Ion:248 Resp: 330867
Ion Ratio Lower Upper
248 100
250 99.6 79.8 119.8
141 76.8 65.8 98.8



#69
Hexachlorobenzene
Concen: 19.218 ng/ul
RT: 16.609 min Scan# 2299
Delta R.T. 0.006 min
Lab File: BM044323.D
Acq: 25 Feb 2024 09:13

Tgt Ion:284 Resp: 406650
Ion Ratio Lower Upper
284 100
142 41.0 33.4 50.2
249 31.1 24.7 37.1





#70

Atrazine

Concen: 20.218 ng/ul

RT: 16.792 min Scan# 21

Delta R.T. 0.006 min

Lab File: BM044323.D

Acq: 25 Feb 2024 09:13

Instrument :

BNA_M

ClientSampleId :

SSTD020197





Tgt Ion:200 Resp: 373449

Ion Ratio Lower Upper

200 100

173 27.6 22.5 33.7

215 47.4 37.8 56.6



#71

Pentachlorophenol

Concen: 18.902 ng/ul

RT: 16.962 min Scan# 2359

Delta R.T. 0.006 min

Lab File: BM044323.D

Acq: 25 Feb 2024 09:13




Tgt Ion:266 Resp: 263577


Ion Ratio Lower Upper

266 100

264 62.0 50.9 76.3

268 65.2 51.1 76.7









#72
Phenanthrene
Concen: 18.970 ng/ul
RT: 17.362 min Scan# 2426
Delta R.T. 0.006 min
Lab File: BM044323.D
Acq: 25 Feb 2024 09:13




Instrument :
BNA_M
ClientSampleId :
SSTD020197


Tgt Ion:178 Resp: 2024907
Ion Ratio Lower Upper
178 100
179 15.0 12.2 18.4
176 19.6 15.6 23.4



#74
Anthracene
Concen: 19.438 ng/ul
RT: 17.451 min Scan# 2442
Delta R.T. 0.006 min
Lab File: BM044323.D
Acq: 25 Feb 2024 09:13

Tgt Ion:178 Resp: 2089046
Ion Ratio Lower Upper
178 100
179 15.2 12.2 18.2
176 18.6 15.4 23.0





#75

1,2,3,4-Tetrachlorobenzene

Concen: 18.091 ng/uL

RT: 13.363 min Scan# 1

Delta R.T. 0.012 min

Lab File: BM044323.D

Acq: 25 Feb 2024 09:13

Instrument :

BNA_M

ClientSampleId :

SSTD020197





Tgt Ion:216 Resp: 471210

Ion Ratio Lower Upper

216 100

214 79.0 61.9 92.9

218 47.1 37.9 56.9



#76

Pentachlorobenzene

Concen: 18.327 ng/uL

RT: 14.880 min Scan# 2005

Delta R.T. 0.006 min

Lab File: BM044323.D

Acq: 25 Feb 2024 09:13




Tgt Ion:250 Resp: 455708


Ion Ratio Lower Upper

250 100

248 63.9 48.6 73.0

252 64.2 50.0 75.0





#77

Carbazole

Concen: 19.366 ng/ul

RT: 17.727 min Scan# 2489

Delta R.T. 0.006 min

Lab File: BM044323.D

Acq: 25 Feb 2024 09:13

Instrument :

BNA_M

ClientSampleId :

SSTD020197





Tgt Ion:167 Resp: 1957044

Ion Ratio Lower Upper

167 100

166 21.4 17.1 25.7

139 13.1 10.7 16.1



#78

Di-n-butylphthalate

Concen: 20.940 ng/ul

RT: 18.303 min Scan# 2587

Delta R.T. 0.006 min

Lab File: BM044323.D

Acq: 25 Feb 2024 09:13




Tgt Ion:149 Resp: 2679327


Ion Ratio Lower Upper

149 100

150 9.4 7.3 10.9

104 4.8 3.8 5.8





#80

Fluoranthene

Concen: 19.299 ng/uI

RT: 19.380 min Scan# 21

Delta R.T. 0.006 min

Lab File: BM044323.D

Acq: 25 Feb 2024 09:13

Instrument :

BNA_M

ClientSampleId :

SSTD020197





Tgt Ion:202 Resp: 2315233

Ion Ratio Lower Upper

202 100

101 13.4 11.1 16.7

100 9.9 8.2 12.2



#82

Pyrene

Concen: 19.286 ng/uI

RT: 19.745 min Scan# 2832

Delta R.T. 0.006 min

Lab File: BM044323.D

Acq: 25 Feb 2024 09:13




Tgt Ion:202 Resp: 2426391


Ion Ratio Lower Upper

202 100

101 15.2 12.6 18.8

100 12.6 10.2 15.2





#83

Butylbenzylphthalate

Concen: 21.390 ng/ul

RT: 20.662 min Scan# 2988

Delta R.T. -0.000 min

Lab File: BM044323.D

Acq: 25 Feb 2024 09:13

Instrument :

BNA_M

ClientSampleId :

SSTD020197





Tgt Ion:149 Resp: 1208748

Ion Ratio Lower Upper

149 100

91 71.9 58.1 87.1

206 18.3 15.0 22.4



#84

3,3'-Dichlorobenzidine

Concen: 19.784 ng/ul

RT: 21.439 min Scan# 3120

Delta R.T. -0.000 min

Lab File: BM044323.D

Acq: 25 Feb 2024 09:13




Tgt Ion:252 Resp: 735767


Ion Ratio Lower Upper

252 100

254 64.0 51.6 77.4

126 13.9 11.2 16.8





#85

Benzo(a)anthracene

Concen: 19.538 ng/ul

RT: 21.503 min Scan# 3131

Delta R.T. 0.006 min

Lab File: BM044323.D

Acq: 25 Feb 2024 09:13

Instrument :

BNA_M

ClientSampleId :

SSTD020197





Tgt Ion:228 Resp: 2422742

Ion Ratio Lower Upper

228 100

229 19.7 15.4 23.0

226 26.6 21.2 31.8



#86

Bis(2-ethylhexyl)phthalate

Concen: 21.859 ng/ul

RT: 21.450 min Scan# 3122

Delta R.T. 0.006 min

Lab File: BM044323.D

Acq: 25 Feb 2024 09:13




Tgt Ion:149 Resp: 1891149


Ion Ratio Lower Upper

149 100

167 27.8 22.1 33.1

279 3.7 2.4 3.6#









#87
Chrysene
Concen: 19.451 ng/ul
RT: 21.556 min Scan# 3139
Delta R.T. 0.006 min
Lab File: BM044323.D
Acq: 25 Feb 2024 09:13




Instrument :
BNA_M
ClientSampleId :
SSTD020197


Tgt Ion:228 Resp: 2267264
Ion Ratio Lower Upper
228 100
226 29.8 23.8 35.6
229 19.5 16.0 24.0



#89
Di-n-octyl phthalate
Concen: 20.329 ng/ul
RT: 22.391 min Scan# 3282
Delta R.T. 0.006 min
Lab File: BM044323.D
Acq: 25 Feb 2024 09:13

Tgt Ion:149 Resp: 3264375









#90
Benzo(b)fluoranthene
Concen: 19.596 ng/ul
RT: 23.191 min Scan# 3417
Delta R.T. 0.006 min
Lab File: BM044323.D
Acq: 25 Feb 2024 09:13




Instrument :
BNA_M
ClientSampleId :
SSTD020197


Tgt Ion:252 Resp: 2354509
Ion Ratio Lower Upper
252 100
253 21.8 17.6 26.4
125 12.2 9.8 14.8



#91
Benzo(k)fluoranthene
Concen: 19.414 ng/ul
RT: 23.244 min Scan# 3427
Delta R.T. 0.006 min
Lab File: BM044323.D
Acq: 25 Feb 2024 09:13

Tgt Ion:252 Resp: 2395788
Ion Ratio Lower Upper
252 100
253 22.0 17.4 26.0
125 12.0 9.4 14.2





#93

Benzo(a)pyrene

Concen: 19.388 ng/ul

RT: 23.821 min Scan# 3

Delta R.T. 0.006 min

Lab File: BM044323.D

Acq: 25 Feb 2024 09:13

Instrument :

BNA_M

ClientSampleId :

SSTD020197





Tgt Ion:252 Resp: 2245946

Ion Ratio Lower Upper

252 100

253 22.4 17.4 26.2

125 13.4 10.8 16.2



#94

Indeno(1,2,3-cd)pyrene

Concen: 19.081 ng/ul

RT: 26.432 min Scan# 3969

Delta R.T. 0.012 min

Lab File: BM044323.D

Acq: 25 Feb 2024 09:13




Tgt Ion:276 Resp: 2672977


Ion Ratio Lower Upper

276 100

138 27.0 22.1 33.1

277 24.3 19.9 29.9





#95

Dibenzo(a,h)anthracene

Concen: 19.177 ng/ul

RT: 26.462 min Scan# 3972

Delta R.T. 0.012 min

Lab File: BM044323.D

Acq: 25 Feb 2024 09:13

Instrument :

BNA_M

ClientSampleId :

SSTD020197





Tgt Ion:278 Resp: 2244931

Ion Ratio Lower Upper

278 100

139 19.0 14.9 22.3

279 23.9 19.5 29.3



#96

Benzo(g,h,i)perylene

Concen: 18.734 ng/ul

RT: 27.203 min Scan# 4100

Delta R.T. 0.017 min

Lab File: BM044323.D

Acq: 25 Feb 2024 09:13

Tgt Ion:276 Resp: 2122910

Ion Ratio Lower Upper

276 100

138 25.4 21.6 32.4

277 23.5 18.5 27.7

