


Data Path : Z:\svoasrv\HPCHEM1\BNA M\Data\BM022519\
 Data File : BM018909.D
 Acq On : 25 Feb 2019 10:59
 Operator : JU/SJ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\8270-BM021119.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Mon Feb 11 16:02:12 2019



AutoFind: Scans 2453, 2454, 2455; Background Corrected with Scan 2447

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	34.4	79482	PASS
68	69	0.00	2	2.0	2569	PASS
69	198	0.00	100	54.9	126965	PASS
70	69	0.00	2	0.6	733	PASS
127	198	10	80	53.3	123309	PASS
197	198	0.00	2	0.4	865	PASS
198	198	100	100	100.0	231210	PASS
199	198	5	9	6.9	15873	PASS
275	198	10	60	29.0	67149	PASS
365	198	1	100	4.2	9623	PASS
441	198	0.01	100	14.6	33662	PASS
442	198	50	100	94.2	217701	PASS
443	442	15	24	19.7	42810	PASS