

Data Path : Z:\HPCHEM1\BNA M\DATA\BM022616\
 Data File : BM004442.D
 Acq On : 27 Feb 2016 00:52
 Operator : SJ/UM
 Sample : H1564-16
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 D9R35

Quant Time: Feb 27 04:33:55 2016
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM02.2-EPA-BM022216.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Feb 27 00:43:30 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.62	152	27772	20.00	ng/ul	0.00
18) Naphthalene-d8	10.40	136	113310	20.00	ng/ul	0.00
36) Acenaphthene-d10	14.28	164	59938	20.00	ng/ul	0.00
62) Phenanthrene-d10	17.03	188	125527	20.00	ng/ul	0.00
78) Chrysene-d12	21.24	240	136612	20.00	ng/ul	0.00
86) Perylene-d12	23.46	264	137991	20.00	ng/ul	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
3) 1,4-Dioxane-d8	3.08	96	3273	6.21	ng/uL	0.00
5) Phenol-d5	6.80	99	35366	15.19	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	6.95	67	36264	30.62	ng/ul	0.00
9) 2-Chlorophenol-d4	7.15	132	22540	11.22	ng/ul	0.00
13) 4-Methylphenol-d8	8.34	113	46803	22.49	ng/ul	0.00
19) Nitrobenzene-d5	8.77	128	30488	35.37	ng/ul	0.00
22) 2-Nitrophenol-d4	9.50	143	8042	8.18	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.05	165	17501	9.85	ng/ul	0.00
29) 4-Chloroaniline-d4	10.55	131	80792	36.49	ng/ul	0.00
44) Dimethylphthalate-d6	13.69	166	169043	32.88	ng/ul	0.00
47) Acenaphthylene-d8	13.97	160	215515	35.81	ng/ul	0.00
52) 4-Nitrophenol-d4	0.00	143	0	0.00	ng/ul	0.00
58) Fluorene-d10	15.27	176	146453	33.50	ng/ul	0.00
63) 4,6-Dinitro-2-methylphenol	0.00	200	0	0.00	ng/ul	0.00
71) Anthracene-d10	17.13	188	222751	36.22	ng/ul	0.00
79) Pyrene-d10	19.44	212	245146	37.10	ng/ul	0.00
90) Benzo(a)pyrene-d12	23.33	264	268629	37.10	ng/ul	0.00


Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
45) Dimethylphthalate	13.74	163	53489	9.98	ng/ul	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA M\DATA\BM022616\
Data File : BM004442.D
Acq On : 27 Feb 2016 00:52
Operator : SJ/UM
Sample : H1564-16
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
D9R35

Quant Time: Feb 27 04:33:55 2016
Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM02.2-EPA-BM022216.M
Quant Title : SVOA CALIBRATION
QLast Update : Sat Feb 27 00:43:30 2016
Response via : Initial Calibration





#45

Dimethylphthalate

Concen: 9.98 ng/ul

RT: 13.74 min Scan# 1879

Delta R.T. 0.00 min

Lab File: BM004442.D

Acq: 27 Feb 2016 00:52

Instrument :

BNA_M

ClientSampleId :

D9R35


Tot Ion:163 Resp: 53489

Ion Ratio Lower Upper

163 100

194 4.3 3.2 4.8

164 9.6 7.8 11.8



Abundance Ion 163.00 (162.70 to 163.70): E

Ion 194.00 (193.70 to 194.70): E

Ion 164.00 (163.70 to 164.70): E

