

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM022624\
 Data File : BM044329.D
 Acq On : 26 Feb 2024 13:15
 Operator : MA/JU
 Sample : P1442-13
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 BH9Q6

Quant Time: Feb 26 14:06:08 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM022324.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sun Feb 25 22:57:38 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.928	152	169775	20.000	ng/u1	0.01
20) Naphthalene-d8	10.728	136	736348	20.000	ng/u1	0.01
38) Acenaphthene-d10	14.569	164	423490	20.000	ng/u1	0.01
64) Phenanthrene-d10	17.316	188	896282	20.000	ng/u1	0.00
79) Chrysene-d12	21.515	240	792535	20.000	ng/u1	0.00
88) Perylene-d12	23.927	264	902832	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.358	96	34279	6.370	ng/uL	0.00
4) Pyridine-d5	3.770	84	516754	35.164	ng/u1	0.00
7) Phenol-d5	7.087	99	585238	35.984	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.269	67	398947	37.880	ng/u1	0.01
11) 2-Chlorophenol-d4	7.452	132	464027	37.882	ng/u1	0.01
15) 4-Methylphenol-d8	8.634	113	454644	36.945	ng/u1	0.00
21) Nitrobenzene-d5	9.099	128	238822	39.890	ng/u1	0.01
24) 2-Nitrophenol-d4	9.816	143	256947	42.065	ng/u1	0.01
28) 2,4-Dichlorophenol-d3	10.351	165	429171	39.830	ng/u1	0.01
31) 4-Chloroaniline-d4	10.875	131	663487	36.664	ng/u1	0.01
46) Dimethylphthalate-d6	13.992	166	1452586	44.474	ng/u1	0.01
49) Acenaphthylene-d8	14.263	160	1705191	45.377	ng/u1	0.01
54) 4-Nitrophenol-d4	14.775	143	226820	34.899	ng/u1	0.01
60) Fluorene-d10	15.563	176	1289509	46.646	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.686	200	186547	33.544	ng/u1	0.00
73) Anthracene-d10	17.416	188	1918688	44.896	ng/u1	0.00
81) Pyrene-d10	19.716	212	2243976	47.063	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.774	264	2103306	45.465	ng/u1	0.01

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM022624\
Data File : BM044329.D
Acq On : 26 Feb 2024 13:15
Operator : MA/JU
Sample : P1442-13
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
BH9Q6

Quant Time: Feb 26 14:06:08 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM022324.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Sun Feb 25 22:57:38 2024
Response via : Initial Calibration

