

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM022624\
 Data File : BM044330.D
 Acq On : 26 Feb 2024 13:51
 Operator : MA/JU
 Sample : P1442-01
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 BH9Q7

Quant Time: Feb 26 14:57:54 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM022324.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sun Feb 25 22:57:38 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.928	152	225272	20.000	ng/u1	0.01
20) Naphthalene-d8	10.728	136	966415	20.000	ng/u1	0.01
38) Acenaphthene-d10	14.569	164	560938	20.000	ng/u1	0.01
64) Phenanthrene-d10	17.321	188	1193407	20.000	ng/u1	0.01
79) Chrysene-d12	21.515	240	1040518	20.000	ng/u1	0.00
88) Perylene-d12	23.927	264	1170894	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.357	96	29205	4.090	ng/uL	0.00
4) Pyridine-d5	3.775	84	430877	22.097	ng/u1	0.00
7) Phenol-d5	7.093	99	517005	23.957	ng/u1	0.01
9) Bis-(2-Chloroethyl)eth...	7.269	67	336813	24.102	ng/u1	0.01
11) 2-Chlorophenol-d4	7.451	132	392479	24.148	ng/u1	0.01
15) 4-Methylphenol-d8	8.640	113	417392	25.562	ng/u1	0.01
21) Nitrobenzene-d5	9.098	128	210531	26.793	ng/u1	0.01
24) 2-Nitrophenol-d4	9.816	143	228427	28.494	ng/u1	0.01
28) 2,4-Dichlorophenol-d3	10.351	165	380953	26.939	ng/u1	0.01
31) 4-Chloroaniline-d4	10.881	131	622904	26.227	ng/u1	0.02
46) Dimethylphthalate-d6	13.992	166	1427115	32.988	ng/u1	0.01
49) Acenaphthylene-d8	14.263	160	1607424	32.294	ng/u1	0.01
54) 4-Nitrophenol-d4	14.774	143	171362	19.906	ng/u1	0.01
60) Fluorene-d10	15.563	176	1243514	33.960	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.686	200	132779	17.931	ng/u1	0.00
73) Anthracene-d10	17.421	188	1908034	33.531	ng/u1	0.01
81) Pyrene-d10	19.715	212	2257483	36.063	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.774	264	2133916	35.567	ng/u1	0.01

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM022624\
Data File : BM044330.D
Acq On : 26 Feb 2024 13:51
Operator : MA/JU
Sample : P1442-01
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
BH9Q7

Quant Time: Feb 26 14:57:54 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM022324.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Sun Feb 25 22:57:38 2024
Response via : Initial Calibration

