

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM022624\
 Data File : BM044338.D
 Acq On : 26 Feb 2024 18:44
 Operator : MA/JU
 Sample : P1442-14
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 BH9S0

Quant Time: Feb 26 23:14:33 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM022324.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sun Feb 25 22:57:38 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.928	152	186489	20.000	ng/u1	0.01
20) Naphthalene-d8	10.733	136	810291	20.000	ng/u1	0.02
38) Acenaphthene-d10	14.569	164	476662	20.000	ng/u1	0.01
64) Phenanthrene-d10	17.321	188	999831	20.000	ng/u1	0.01
79) Chrysene-d12	21.521	240	872822	20.000	ng/u1	0.00
88) Perylene-d12	23.933	264	973948	20.000	ng/u1	0.01

System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.357	96	28076	4.750	ng/uL	0.00
4) Pyridine-d5	3.775	84	400205	24.792	ng/u1	0.00
7) Phenol-d5	7.093	99	480260	26.883	ng/u1	0.01
9) Bis-(2-Chloroethyl)eth...	7.269	67	317640	27.457	ng/u1	0.01
11) 2-Chlorophenol-d4	7.457	132	373198	27.736	ng/u1	0.02
15) 4-Methylphenol-d8	8.639	113	357250	26.429	ng/u1	0.01
21) Nitrobenzene-d5	9.098	128	191334	29.041	ng/u1	0.01
24) 2-Nitrophenol-d4	9.822	143	201607	29.994	ng/u1	0.02
28) 2,4-Dichlorophenol-d3	10.351	165	339602	28.641	ng/u1	0.01
31) 4-Chloroaniline-d4	10.881	131	527980	26.513	ng/u1	0.02
46) Dimethylphthalate-d6	13.992	166	1226622	33.366	ng/u1	0.01
49) Acenaphthylene-d8	14.263	160	1393664	32.950	ng/u1	0.01
54) 4-Nitrophenol-d4	14.780	143	192593	26.328	ng/u1	0.02
60) Fluorene-d10	15.563	176	1079071	34.680	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.686	200	156513	25.229	ng/u1	0.00
73) Anthracene-d10	17.421	188	1624727	34.080	ng/u1	0.01
81) Pyrene-d10	19.721	212	1905836	36.295	ng/u1	0.01
92) Benzo(a)pyrene-d12	23.774	264	1779266	35.652	ng/u1	0.01

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM022624\
Data File : BM044338.D
Acq On : 26 Feb 2024 18:44
Operator : MA/JU
Sample : P1442-14
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
BH9S0

Quant Time: Feb 26 23:14:33 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM022324.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Sun Feb 25 22:57:38 2024
Response via : Initial Calibration

