

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM022624\
 Data File : BM044401.D
 Acq On : 28 Feb 2024 13:12
 Operator : MA/JU
 Sample : PB159253BS
 Misc :
 ALS Vial : 83 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SLCS253

Quant Time: Feb 28 14:00:37 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM022324.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Feb 27 23:05:48 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.946	152	202935	20.000	ng/u1	0.00
20) Naphthalene-d8	10.757	136	881054	20.000	ng/u1	0.01
38) Acenaphthene-d10	14.586	164	499756	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.339	188	1003928	20.000	ng/u1	0.00
79) Chrysene-d12	21.539	240	816724	20.000	ng/u1	0.00
88) Perylene-d12	23.956	264	868164	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.369	96	41541	6.459	ng/uL	0.00
4) Pyridine-d5	3.781	84	592064	33.706	ng/u1	0.00
7) Phenol-d5	7.116	99	684743	35.223	ng/u1	0.01
9) Bis-(2-Chloroethyl)eth...	7.293	67	429695	34.133	ng/u1	0.01
11) 2-Chlorophenol-d4	7.475	132	540556	36.919	ng/u1	0.00
15) 4-Methylphenol-d8	8.663	113	518385	35.241	ng/u1	0.00
21) Nitrobenzene-d5	9.122	128	268316	37.455	ng/u1	0.01
24) 2-Nitrophenol-d4	9.840	143	285670	39.086	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.375	165	493612	38.287	ng/u1	0.00
31) 4-Chloroaniline-d4	10.904	131	713990	32.974	ng/u1	0.01
46) Dimethylphthalate-d6	14.016	166	1400339	36.331	ng/u1	0.01
49) Acenaphthylene-d8	14.286	160	1666367	37.577	ng/u1	0.01
54) 4-Nitrophenol-d4	14.798	143	272733	35.560	ng/u1	0.01
60) Fluorene-d10	15.580	176	1196399	36.673	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.710	200	223399	35.863	ng/u1	0.00
73) Anthracene-d10	17.439	188	1765073	36.873	ng/u1	0.00
81) Pyrene-d10	19.733	212	1933684	39.354	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.803	264	1721681	38.702	ng/u1	0.01
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.405	88	91564	13.506	ng/uL	95
5) Pyridine	3.805	79	641945	34.932	ng/u1	96
6) Benzaldehyde	7.093	77	358763	42.514	ng/u1	99
8) Phenol	7.140	94	747397	36.986	ng/u1	98
10) Bis(2-Chloroethyl)ether	7.387	93	614081	36.306	ng/u1	97
12) 2-Chlorophenol	7.504	128	585718	38.667	ng/u1	98
13) 2-Methylphenol	8.399	108	549577	36.915	ng/u1	99
14) 2,2'-oxybis(1-Chloropr...	8.481	45	815860	35.279	ng/u1	98
16) Acetophenone	8.781	105	880986	37.488	ng/u1	100
17) N-Nitroso-di-n-propyla...	8.775	70	452459	38.434	ng/u1	99
18) 4-Methylphenol	8.728	108	593674	37.435	ng/u1	99
19) Hexachloroethane	9.022	117	243080	37.827	ng/u1	94
22) Nitrobenzene	9.163	77	646326	37.306	ng/u1	98
23) Isophorone	9.693	82	1278598	38.080	ng/u1	99
25) 2-Nitrophenol	9.875	139	324941	40.087	ng/u1	98
26) 2,4-Dimethylphenol	9.934	107	493447	31.881	ng/u1	98
27) Bis(2-Chloroethoxy)met...	10.181	93	807840	37.010	ng/u1	100
29) 2,4-Dichlorophenol	10.404	162	514837	39.457	ng/u1	100
30) Naphthalene	10.804	128	1864068	37.201	ng/u1	100
32) 4-Chloroaniline	10.928	127	625805	29.659	ng/u1	98
33) Hexachlorobutadiene	11.081	225	296493	37.717	ng/u1	99
34) Caprolactam	11.716	113	173079	38.374	ng/u1	98
35) 4-Chloro-3-methylphenol	12.051	107	561169	40.368	ng/u1	97
36) 2-Methylnaphthalene	12.416	142	1212570	37.992	ng/u1	100

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM022624\
 Data File : BM044401.D
 Acq On : 28 Feb 2024 13:12
 Operator : MA/JU
 Sample : PB159253BS
 Misc :
 ALS Vial : 83 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SLCS253

Quant Time: Feb 28 14:00:37 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM022324.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Feb 27 23:05:48 2024
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.634	142	1190662	37.737	ng/ul	100
39) 1,2,4,5-Tetrachloroben...	12.781	216	569595	38.501	ng/ul	98
40) Hexachlorocyclopentadiene	12.751	237	293330	29.470	ng/ul	99
41) 2,4,6-Trichlorophenol	13.028	196	370157	39.253	ng/ul	98
42) 2,4,5-Trichlorophenol	13.098	196	402508	39.484	ng/ul	99
43) 1,1'-Biphenyl	13.428	154	1526209	37.512	ng/ul	99
44) 2-Chloronaphthalene	13.469	162	1208731	37.942	ng/ul	99
45) 2-Nitroaniline	13.686	65	365849	41.479	ng/ul	97
47) Dimethylphthalate	14.063	163	1482717	37.639	ng/ul	100
48) 2,6-Dinitrotoluene	14.186	165	303143	41.149	ng/ul	97
50) Acenaphthylene	14.310	152	1994595	37.973	ng/ul	99
51) 3-Nitroaniline	14.510	138	329737	42.426	ng/ul	94
52) Acenaphthene	14.651	153	1301382	37.530	ng/ul	100
53) 2,4-Dinitrophenol	14.716	184	155421	33.998	ng/ul	98
55) 4-Nitrophenol	14.810	109	193747	36.772	ng/ul	97
56) Dibenzofuran	14.992	168	1751384	37.624	ng/ul	100
57) 2,4-Dinitrotoluene	14.969	165	433220	40.933	ng/ul	98
58) 2,3,4,6-Tetrachlorophenol	15.216	232	309550	37.731	ng/ul	98
59) Diethylphthalate	15.427	149	1549890	38.222	ng/ul	99
61) Fluorene	15.639	166	1410075	37.715	ng/ul	100
62) 4-Chlorophenyl-phenyle...	15.639	204	672234	38.151	ng/ul	99
63) 4-Nitroaniline	15.675	138	337923	44.954	ng/ul	99
66) 4,6-Dinitro-2-methylph...	15.727	198	248193	36.423	ng/ul	98
67) N-Nitrosodiphenylamine	15.857	169	1196619	39.332	ng/ul	100
68) 4-Bromophenyl-phenylether	16.533	248	391803	39.280	ng/ul	97
69) Hexachlorobenzene	16.633	284	451510	39.548	ng/ul	98
70) Atrazine	16.816	200	236934	23.774	ng/ul	99
71) Pentachlorophenol	16.980	266	255421	33.949	ng/ul	99
72) Phenanthrene	17.380	178	2190850	38.041	ng/ul	100
74) Anthracene	17.474	178	2198880	37.921	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.386	216	564714	40.184	ng/ul	99
76) Pentachlorobenzene	14.904	250	527836	39.344	ng/ul	98
77) Carbazole	17.751	167	2063543	37.847	ng/ul	99
78) Di-n-butylphthalate	18.327	149	2784873	40.340	ng/ul	100
80) Fluoranthene	19.398	202	2417542	41.024	ng/ul	100
82) Pyrene	19.762	202	2502959	40.500	ng/ul	100
83) Butylbenzylphthalate	20.686	149	1212775	43.691	ng/ul	96
84) 3,3'-Dichlorobenzidine	21.462	252	722846	39.568	ng/ul	99
85) Benzo(a)anthracene	21.521	228	2384143	39.141	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.468	149	1867490	43.943	ng/ul	100
87) Chrysene	21.574	228	2200582	38.433	ng/ul	100
89) Di-n-octyl phthalate	22.415	149	3167921	42.589	ng/ul	100
90) Benzo(b)fluoranthene	23.221	252	2253343	40.486	ng/ul	99
91) Benzo(k)fluoranthene	23.274	252	2302412	40.277	ng/ul	99
93) Benzo(a)pyrene	23.856	252	2143757	39.950	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.486	276	2571750	39.632	ng/ul	98
95) Dibenzo(a,h)anthracene	26.515	278	2159647	39.826	ng/ul	99
96) Benzo(g,h,i)perylene	27.262	276	2061861	39.280	ng/ul	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM022624\
Data File : BM044401.D
Acq On : 28 Feb 2024 13:12
Operator : MA/JU
Sample : PB159253BS
Misc :
ALS Vial : 83 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SLCS253

Quant Time: Feb 28 14:00:37 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM022324.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Feb 27 23:05:48 2024
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 13.506 ng/uL
RT: 3.405 min Scan# 54
Delta R.T. 0.006 min
Lab File: BM044401.D
Acq: 28 Feb 2024 13:12




Instrument :
BNA_M
ClientSampleId :
SLCS253


Tgt Ion: 88 Resp: 91564
Ion Ratio Lower Upper
88 100
43 30.0 21.2 31.8
58 69.2 58.3 87.5



#5
Pyridine
Concen: 34.932 ng/uL
RT: 3.805 min Scan# 122
Delta R.T. 0.006 min
Lab File: BM044401.D
Acq: 28 Feb 2024 13:12

Tgt Ion: 79 Resp: 641945
Ion Ratio Lower Upper
79 100
52 69.9 58.9 88.3
51 32.8 27.8 41.8









#6
Benzaldehyde
Concen: 42.514 ng/ul
RT: 7.093 min Scan# 681
Delta R.T. 0.006 min
Lab File: BM044401.D
Acq: 28 Feb 2024 13:12




Instrument :
BNA_M
ClientSampleId :
SLCS253


Tgt Ion: 77 Resp: 358763
Ion Ratio Lower Upper
77 100
105 99.2 79.0 118.6
106 95.5 77.4 116.2



#8
Phenol
Concen: 36.986 ng/ul
RT: 7.140 min Scan# 689
Delta R.T. 0.006 min
Lab File: BM044401.D
Acq: 28 Feb 2024 13:12

Tgt Ion: 94 Resp: 747397
Ion Ratio Lower Upper
94 100
65 27.0 20.6 31.0
66 32.2 25.4 38.0









#10
Bis(2-Chloroethyl)ether
Concen: 36.306 ng/ul
RT: 7.387 min Scan# 710
Delta R.T. 0.012 min
Lab File: BM044401.D
Acq: 28 Feb 2024 13:12




Instrument :
BNA_M
ClientSampleId :
SLCS253


Tgt Ion: 93 Resp: 614081
Ion Ratio Lower Upper
93 100
63 74.4 61.0 91.6
95 33.3 25.1 37.7



#12
2-Chlorophenol
Concen: 38.667 ng/ul
RT: 7.504 min Scan# 751
Delta R.T. 0.006 min
Lab File: BM044401.D
Acq: 28 Feb 2024 13:12

Tgt Ion:128 Resp: 585718
Ion Ratio Lower Upper
128 100
64 47.2 39.0 58.6
130 31.4 25.4 38.0









#13
2-Methylphenol
Concen: 36.915 ng/ul
RT: 8.399 min Scan# 902
Delta R.T. 0.012 min
Lab File: BM044401.D
Acq: 28 Feb 2024 13:12




Instrument :
BNA_M
ClientSampleId :
SLCS253


Tgt Ion:108 Resp: 549577
Ion Ratio Lower Upper
108 100
107 87.5 70.6 105.8



#14
2,2'-oxybis(1-Chloropropane)
Concen: 35.279 ng/ul
RT: 8.481 min Scan# 917
Delta R.T. 0.006 min
Lab File: BM044401.D
Acq: 28 Feb 2024 13:12

Tgt Ion: 45 Resp: 815860
Ion Ratio Lower Upper
45 100
77 15.0 11.3 16.9
79 10.9 8.5 12.7









#16
Acetophenone
Concen: 37.488 ng/ul
RT: 8.781 min Scan# 968
Delta R.T. 0.006 min
Lab File: BM044401.D
Acq: 28 Feb 2024 13:12




Instrument :
BNA_M
ClientSampleId :
SLCS253


Tgt Ion:105 Resp: 880986
Ion Ratio Lower Upper
105 100
77 76.5 61.2 91.8
51 26.4 21.1 31.7



#17
N-Nitroso-di-n-propylamine
Concen: 38.434 ng/ul
RT: 8.775 min Scan# 967
Delta R.T. 0.012 min
Lab File: BM044401.D
Acq: 28 Feb 2024 13:12

Tgt Ion: 70 Resp: 452459
Ion Ratio Lower Upper
70 100
42 54.3 44.4 66.6
101 9.0 7.5 11.3
130 21.5 17.0 25.4









#18
4-Methylphenol
Concen: 37.435 ng/ul
RT: 8.728 min Scan# 91
Delta R.T. 0.006 min
Lab File: BM044401.D
Acq: 28 Feb 2024 13:12




Instrument :
BNA_M
ClientSampleId :
SLCS253


Tgt Ion:108 Resp: 593674
Ion Ratio Lower Upper
108 100
107 109.6 86.8 130.2



#19
Hexachloroethane
Concen: 37.827 ng/ul
RT: 9.022 min Scan# 1009
Delta R.T. 0.012 min
Lab File: BM044401.D
Acq: 28 Feb 2024 13:12

Tgt Ion:117 Resp: 243080
Ion Ratio Lower Upper
117 100
201 90.8 69.8 104.6
199 57.2 40.8 61.2









#22
Nitrobenzene
Concen: 37.306 ng/ul
RT: 9.163 min Scan# 1033
Delta R.T. 0.006 min
Lab File: BM044401.D
Acq: 28 Feb 2024 13:12




Instrument :
BNA_M
ClientSampleId :
SLCS253


Tgt Ion: 77 Resp: 646326
Ion Ratio Lower Upper
77 100
123 47.5 36.9 55.3
65 13.3 10.2 15.2



#23
Isophorone
Concen: 38.080 ng/ul
RT: 9.693 min Scan# 1123
Delta R.T. 0.012 min
Lab File: BM044401.D
Acq: 28 Feb 2024 13:12

Tgt Ion: 82 Resp: 1278598
Ion Ratio Lower Upper
82 100
95 6.4 4.6 7.0
138 17.4 13.5 20.3









#25
2-Nitrophenol
Concen: 40.087 ng/ul
RT: 9.875 min Scan# 1154
Delta R.T. 0.012 min
Lab File: BM044401.D
Acq: 28 Feb 2024 13:12




Instrument :
BNA_M
ClientSampleId :
SLCS253


Tgt Ion:139 Resp: 324941
Ion Ratio Lower Upper
139 100
65 49.2 41.0 61.4
109 25.7 20.5 30.7



#26
2,4-Dimethylphenol
Concen: 31.881 ng/ul
RT: 9.934 min Scan# 1164
Delta R.T. 0.006 min
Lab File: BM044401.D
Acq: 28 Feb 2024 13:12

Tgt Ion:107 Resp: 493447
Ion Ratio Lower Upper
107 100
121 51.1 41.1 61.7
122 87.9 72.3 108.5









#27
 Bis(2-Chloroethoxy)methane
 Concen: 37.010 ng/ul
 RT: 10.181 min Scan# 1206
 Delta R.T. 0.006 min
 Lab File: BM044401.D
 Acq: 28 Feb 2024 13:12




Instrument :
 BNA_M
 ClientSampleId :
 SLCS253


Tgt Ion: 93 Resp: 807840
 Ion Ratio Lower Upper
 93 100
 95 32.5 26.0 39.0
 123 10.8 9.0 13.6



#29
 2,4-Dichlorophenol
 Concen: 39.457 ng/ul
 RT: 10.404 min Scan# 1244
 Delta R.T. 0.012 min
 Lab File: BM044401.D
 Acq: 28 Feb 2024 13:12

Tgt Ion: 162 Resp: 514837
 Ion Ratio Lower Upper
 162 100
 164 64.6 51.8 77.8
 98 37.0 29.6 44.4









#30
Naphthalene
Concen: 37.201 ng/u1
RT: 10.804 min Scan# 11
Delta R.T. 0.006 min
Lab File: BM044401.D
Acq: 28 Feb 2024 13:12




Instrument :
BNA_M
ClientSampleId :
SLCS253


Tgt Ion:128 Resp: 1864068
Ion Ratio Lower Upper
128 100
129 11.2 8.8 13.2
127 12.6 10.2 15.2



#32
4-Chloroaniline
Concen: 29.659 ng/u1
RT: 10.928 min Scan# 1333
Delta R.T. 0.006 min
Lab File: BM044401.D
Acq: 28 Feb 2024 13:12

Tgt Ion:127 Resp: 625805
Ion Ratio Lower Upper
127 100
129 32.8 25.3 37.9





#33

Hexachlorobutadiene

Concen: 37.717 ng/ul

RT: 11.081 min Scan# 1

Delta R.T. 0.012 min

Lab File: BM044401.D

Acq: 28 Feb 2024 13:12

Instrument :

BNA_M

ClientSampleId :

SLCS253





Tgt Ion:225 Resp: 296493

Ion Ratio Lower Upper

225 100

223 62.0 49.7 74.5

227 64.1 52.2 78.2



#34

Caprolactam

Concen: 38.374 ng/ul

RT: 11.716 min Scan# 1467

Delta R.T. 0.018 min

Lab File: BM044401.D

Acq: 28 Feb 2024 13:12




Tgt Ion:113 Resp: 173079


Ion Ratio Lower Upper

113 100

55 170.4 140.7 211.1

56 128.4 102.4 153.6









#35
 4-Chloro-3-methylphenol
 Concen: 40.368 ng/ul
 RT: 12.051 min Scan# 1524
 Delta R.T. 0.012 min
 Lab File: BM044401.D
 Acq: 28 Feb 2024 13:12




Instrument :
 BNA_M
 ClientSampleId :
 SLCS253


Tgt Ion:107 Resp: 561169
 Ion Ratio Lower Upper
 107 100
 144 27.0 20.6 30.8
 142 82.6 64.3 96.5



#36
 2-Methylnaphthalene
 Concen: 37.992 ng/ul
 RT: 12.416 min Scan# 1586
 Delta R.T. 0.006 min
 Lab File: BM044401.D
 Acq: 28 Feb 2024 13:12

Tgt Ion:142 Resp: 1212570
 Ion Ratio Lower Upper
 142 100
 141 87.9 70.5 105.7





#37

1-Methylnaphthalene

Concen: 37.737 ng/ul

RT: 12.634 min Scan# 1

Delta R.T. 0.006 min

Lab File: BM044401.D

Acq: 28 Feb 2024 13:12

Instrument :

BNA_M

ClientSampleId :

SLCS253





Tgt Ion:142 Resp: 1190662

Ion Ratio Lower Upper

142 100

141 91.3 73.0 109.4

116 3.8 2.7 4.1



#39

1,2,4,5-Tetrachlorobenzene

Concen: 38.501 ng/ul

RT: 12.781 min Scan# 1648

Delta R.T. 0.012 min

Lab File: BM044401.D

Acq: 28 Feb 2024 13:12

Tgt Ion:216 Resp: 569595




Ion Ratio Lower Upper


216 100

214 78.3 63.9 95.9

179 21.1 17.5 26.3

108 19.5 16.2 24.4









#40
Hexachlorocyclopentadiene
Concen: 29.470 ng/ul
RT: 12.751 min Scan# 1
Delta R.T. 0.006 min
Lab File: BM044401.D
Acq: 28 Feb 2024 13:12




Instrument :
BNA_M
ClientSampleId :
SLCS253


Tgt Ion:237 Resp: 293330
Ion Ratio Lower Upper
237 100
235 61.3 49.4 74.2
272 13.6 10.6 16.0



#41
2,4,6-Trichlorophenol
Concen: 39.253 ng/ul
RT: 13.028 min Scan# 1690
Delta R.T. 0.012 min
Lab File: BM044401.D
Acq: 28 Feb 2024 13:12

Tgt Ion:196 Resp: 370157
Ion Ratio Lower Upper
196 100
198 96.8 76.1 114.1
200 30.9 25.1 37.7





#42

2,4,5-Trichlorophenol

Concen: 39.484 ng/ul

RT: 13.098 min Scan# 1

Delta R.T. 0.012 min

Lab File: BM044401.D

Acq: 28 Feb 2024 13:12

Instrument :

BNA_M

ClientSampleId :

SLCS253





Tgt Ion:196 Resp: 402508

Ion Ratio Lower Upper

196 100

198 97.8 77.4 116.2

200 31.3 25.6 38.4



#43

1,1'-Biphenyl

Concen: 37.512 ng/ul

RT: 13.428 min Scan# 1758

Delta R.T. 0.006 min

Lab File: BM044401.D

Acq: 28 Feb 2024 13:12




Tgt Ion:154 Resp: 1526209


Ion Ratio Lower Upper

154 100

153 39.8 32.4 48.6

76 14.2 11.6 17.4





#44

2-Chloronaphthalene

Concen: 37.942 ng/ul

RT: 13.469 min Scan# 1

Delta R.T. 0.006 min

Lab File: BM044401.D

Acq: 28 Feb 2024 13:12

Instrument :

BNA_M

ClientSampleId :

SLCS253





Tgt Ion:162 Resp: 1208731

Ion Ratio Lower Upper

162 100

164 32.8 26.1 39.1

127 39.0 30.9 46.3



#45

2-Nitroaniline

Concen: 41.479 ng/ul

RT: 13.686 min Scan# 1802

Delta R.T. 0.012 min

Lab File: BM044401.D

Acq: 28 Feb 2024 13:12




Tgt Ion: 65 Resp: 365849


Ion Ratio Lower Upper

65 100

92 68.6 52.5 78.7

138 109.8 86.4 129.6









#47
Dimethylphthalate
Concen: 37.639 ng/ul
RT: 14.063 min Scan# 1866
Delta R.T. 0.012 min
Lab File: BM044401.D
Acq: 28 Feb 2024 13:12




Instrument :
BNA_M
ClientSampleId :
SLCS253


Tgt Ion:163 Resp: 1482717
Ion Ratio Lower Upper
163 100
194 3.9 3.0 4.6
164 10.0 8.1 12.1



#48
2,6-Dinitrotoluene
Concen: 41.149 ng/ul
RT: 14.186 min Scan# 1887
Delta R.T. 0.012 min
Lab File: BM044401.D
Acq: 28 Feb 2024 13:12

Tgt Ion:165 Resp: 303143
Ion Ratio Lower Upper
165 100
89 58.3 48.2 72.2
121 21.5 18.7 28.1









#50
Acenaphthylene
Concen: 37.973 ng/ul
RT: 14.310 min Scan# 1908
Delta R.T. 0.006 min
Lab File: BM044401.D
Acq: 28 Feb 2024 13:12




Instrument :
BNA_M
ClientSampleId :
SLCS253


Tgt Ion:152 Resp: 1994595
Ion Ratio Lower Upper
152 100
151 20.0 15.8 23.6
153 13.2 10.4 15.6



#51
3-Nitroaniline
Concen: 42.426 ng/ul
RT: 14.510 min Scan# 1942
Delta R.T. 0.006 min
Lab File: BM044401.D
Acq: 28 Feb 2024 13:12

Tgt Ion:138 Resp: 329737
Ion Ratio Lower Upper
138 100
108 8.8 8.6 12.8
92 111.0 93.8 140.6









#52
Acenaphthene
Concen: 37.530 ng/ul
RT: 14.651 min Scan# 1966
Delta R.T. 0.006 min
Lab File: BM044401.D
Acq: 28 Feb 2024 13:12




Instrument :
BNA_M
ClientSampleId :
SLCS253


Tgt Ion:153 Resp: 1301382
Ion Ratio Lower Upper
153 100
152 47.6 38.5 57.7
154 88.7 70.8 106.2



#53
2,4-Dinitrophenol
Concen: 33.998 ng/ul
RT: 14.716 min Scan# 1977
Delta R.T. 0.006 min
Lab File: BM044401.D
Acq: 28 Feb 2024 13:12

Tgt Ion:184 Resp: 155421
Ion Ratio Lower Upper
184 100
63 69.9 58.4 87.6
154 57.7 46.4 69.6





#55

4-Nitrophenol

Concen: 36.772 ng/ul

RT: 14.810 min Scan# 1992

Delta R.T. 0.006 min

Lab File: BM044401.D

Acq: 28 Feb 2024 13:12

Instrument :

BNA_M

ClientSampleId :

SLCS253





Tgt Ion:109 Resp: 193747

Ion Ratio Lower Upper

109 100

139 150.3 123.0 184.6

65 147.6 122.2 183.2



#56

Dibenzofuran

Concen: 37.624 ng/ul

RT: 14.992 min Scan# 2024

Delta R.T. 0.012 min

Lab File: BM044401.D




Acq: 28 Feb 2024 13:12


Tgt Ion:168 Resp: 1751384

Ion Ratio Lower Upper

168 100

139 37.5 29.9 44.9









#57
2,4-Dinitrotoluene
Concen: 40.933 ng/ul
RT: 14.969 min Scan# 2019
Delta R.T. 0.012 min
Lab File: BM044401.D
Acq: 28 Feb 2024 13:12




Instrument :
BNA_M
ClientSampleId :
SLCS253


Tgt Ion:165 Resp: 433220
Ion Ratio Lower Upper
165 100
63 45.9 37.9 56.9
182 2.3 1.8 2.6



#58
2,3,4,6-Tetrachlorophenol
Concen: 37.731 ng/ul
RT: 15.216 min Scan# 2062
Delta R.T. 0.012 min
Lab File: BM044401.D
Acq: 28 Feb 2024 13:12

Tgt Ion:232 Resp: 309550
Ion Ratio Lower Upper
232 100
131 49.6 41.9 62.9
130 2.6 1.8 2.8
166 33.7 26.9 40.3









#59
Diethylphthalate
Concen: 38.222 ng/ul
RT: 15.427 min Scan# 2097
Delta R.T. 0.012 min
Lab File: BM044401.D
Acq: 28 Feb 2024 13:12




Instrument :
BNA_M
ClientSampleId :
SLCS253


Tgt Ion:149 Resp: 1549890
Ion Ratio Lower Upper
149 100
177 21.5 17.0 25.6
150 12.4 9.6 14.4



#61
Fluorene
Concen: 37.715 ng/ul
RT: 15.639 min Scan# 2134
Delta R.T. 0.006 min
Lab File: BM044401.D
Acq: 28 Feb 2024 13:12

Tgt Ion:166 Resp: 1410075
Ion Ratio Lower Upper
166 100
165 96.7 77.4 116.0
167 13.7 10.8 16.2









#62
4-Chlorophenyl-phenylether
Concen: 38.151 ng/ul
RT: 15.639 min Scan# 2133
Delta R.T. 0.006 min
Lab File: BM044401.D
Acq: 28 Feb 2024 13:12




Instrument :
BNA_M
ClientSampleId :
SLCS253


Tgt Ion:204 Resp: 672234
Ion Ratio Lower Upper
204 100
206 33.0 26.9 40.3
141 60.7 48.3 72.5



#63
4-Nitroaniline
Concen: 44.954 ng/ul
RT: 15.675 min Scan# 2140
Delta R.T. 0.012 min
Lab File: BM044401.D
Acq: 28 Feb 2024 13:12

Tgt Ion:138 Resp: 337923
Ion Ratio Lower Upper
138 100
92 49.1 40.9 61.3
108 68.9 55.1 82.7









#66
 4,6-Dinitro-2-methylphenol
 Concen: 36.423 ng/ul
 RT: 15.727 min Scan# 2149
 Delta R.T. 0.012 min
 Lab File: BM044401.D
 Acq: 28 Feb 2024 13:12




Instrument :
 BNA_M
 ClientSampleId :
 SLCS253


Tgt Ion:198 Resp: 248193
 Ion Ratio Lower Upper
 198 100
 182 4.1 3.8 5.8
 77 24.9 20.6 30.8



#67
 N-Nitrosodiphenylamine
 Concen: 39.332 ng/ul
 RT: 15.857 min Scan# 2171
 Delta R.T. 0.012 min
 Lab File: BM044401.D
 Acq: 28 Feb 2024 13:12

Tgt Ion:169 Resp: 1196619
 Ion Ratio Lower Upper
 169 100
 168 67.2 53.6 80.4
 167 36.4 29.2 43.8









#68
4-Bromophenyl-phenylether
Concen: 39.280 ng/ul
RT: 16.533 min Scan# 21
Delta R.T. 0.006 min
Lab File: BM044401.D
Acq: 28 Feb 2024 13:12




Instrument :
BNA_M
ClientSampleId :
SLCS253


Tgt Ion:248 Resp: 391803
Ion Ratio Lower Upper
248 100
250 96.6 79.8 119.8
141 80.6 65.8 98.8



#69
Hexachlorobenzene
Concen: 39.548 ng/ul
RT: 16.633 min Scan# 2303
Delta R.T. 0.006 min
Lab File: BM044401.D
Acq: 28 Feb 2024 13:12

Tgt Ion:284 Resp: 451510
Ion Ratio Lower Upper
284 100
142 39.3 33.4 50.2
249 30.8 24.7 37.1





#70

Atrazine

Concen: 23.774 ng/ul

RT: 16.816 min Scan# 21

Delta R.T. 0.006 min

Lab File: BM044401.D

Acq: 28 Feb 2024 13:12

Instrument :

BNA_M

ClientSampleId :

SLCS253





Tgt Ion:200 Resp: 236934

Ion Ratio Lower Upper

200 100

173 26.7 22.5 33.7

215 46.9 37.8 56.6



#71

Pentachlorophenol

Concen: 33.949 ng/ul

RT: 16.980 min Scan# 2362

Delta R.T. 0.006 min

Lab File: BM044401.D

Acq: 28 Feb 2024 13:12




Tgt Ion:266 Resp: 255421


Ion Ratio Lower Upper

266 100

264 64.1 50.9 76.3

268 64.8 51.1 76.7





#72

Phenanthrene

Concen: 38.041 ng/uI

RT: 17.380 min Scan# 2430

Delta R.T. 0.006 min

Lab File: BM044401.D

Acq: 28 Feb 2024 13:12

Instrument :

BNA_M

ClientSampleId :

SLCS253





Tgt Ion:178 Resp: 2190850

Ion Ratio Lower Upper

178 100

179 15.5 12.2 18.4

176 19.7 15.6 23.4



#74

Anthracene

Concen: 37.921 ng/uI

RT: 17.474 min Scan# 2446

Delta R.T. 0.012 min

Lab File: BM044401.D

Acq: 28 Feb 2024 13:12




Tgt Ion:178 Resp: 2198880


Ion Ratio Lower Upper

178 100

179 15.4 12.2 18.2

176 18.8 15.4 23.0





#75

1,2,3,4-Tetrachlorobenzene

Concen: 40.184 ng/uL

RT: 13.386 min Scan# 1

Delta R.T. 0.006 min

Lab File: BM044401.D

Acq: 28 Feb 2024 13:12

Instrument :

BNA_M

ClientSampleId :

SLCS253





Tgt Ion:216 Resp: 564714

Ion Ratio Lower Upper

216 100

214 78.3 61.9 92.9

218 47.6 37.9 56.9



#76

Pentachlorobenzene

Concen: 39.344 ng/uL

RT: 14.904 min Scan# 2009

Delta R.T. 0.012 min

Lab File: BM044401.D

Acq: 28 Feb 2024 13:12




Tgt Ion:250 Resp: 527836


Ion Ratio Lower Upper

250 100

248 62.8 48.6 73.0

252 63.7 50.0 75.0









#77
 Carbazole
 Concen: 37.847 ng/ul
 RT: 17.751 min Scan# 2493
 Delta R.T. 0.006 min
 Lab File: BM044401.D
 Acq: 28 Feb 2024 13:12




Instrument :
 BNA_M
 ClientSampleId :
 SLCS253


Tgt Ion:167 Resp: 2063543
 Ion Ratio Lower Upper
 167 100
 166 21.3 17.1 25.7
 139 13.0 10.7 16.1



#78
 Di-n-butylphthalate
 Concen: 40.340 ng/ul
 RT: 18.327 min Scan# 2591
 Delta R.T. 0.012 min
 Lab File: BM044401.D
 Acq: 28 Feb 2024 13:12

Tgt Ion:149 Resp: 2784873
 Ion Ratio Lower Upper
 149 100
 150 9.0 7.3 10.9
 104 4.8 3.8 5.8









#80
Fluoranthene
Concen: 41.024 ng/u1
RT: 19.398 min Scan# 21
Delta R.T. 0.006 min
Lab File: BM044401.D
Acq: 28 Feb 2024 13:12




Instrument :
BNA_M
ClientSampleId :
SLCS253


Tgt Ion:202 Resp: 2417542
Ion Ratio Lower Upper
202 100
101 13.6 11.1 16.7
100 10.2 8.2 12.2



#82
Pyrene
Concen: 40.500 ng/u1
RT: 19.762 min Scan# 2835
Delta R.T. 0.006 min
Lab File: BM044401.D
Acq: 28 Feb 2024 13:12

Tgt Ion:202 Resp: 2502959
Ion Ratio Lower Upper
202 100
101 15.9 12.6 18.8
100 12.8 10.2 15.2





#83

Butylbenzylphthalate

Concen: 43.691 ng/ul

RT: 20.686 min Scan# 2991

Delta R.T. 0.012 min

Lab File: BM044401.D

Acq: 28 Feb 2024 13:12

Instrument :

BNA_M

ClientSampleId :

SLCS253





Tgt Ion:149 Resp: 1212775

Ion Ratio Lower Upper

149 100

91 69.4 58.1 87.1

206 19.8 15.0 22.4



#84

3,3'-Dichlorobenzidine

Concen: 39.568 ng/ul

RT: 21.462 min Scan# 3124

Delta R.T. 0.006 min

Lab File: BM044401.D

Acq: 28 Feb 2024 13:12




Tgt Ion:252 Resp: 722846


Ion Ratio Lower Upper

252 100

254 63.5 51.6 77.4

126 13.3 11.2 16.8





#85

Benzo(a)anthracene

Concen: 39.141 ng/ul

RT: 21.521 min Scan# 3134

Delta R.T. 0.006 min

Lab File: BM044401.D

Acq: 28 Feb 2024 13:12

Instrument :

BNA_M

ClientSampleId :

SLCS253





Tgt Ion:228 Resp: 2384143

Ion Ratio Lower Upper

228 100

229 19.9 15.4 23.0

226 27.1 21.2 31.8



#86

Bis(2-ethylhexyl)phthalate

Concen: 43.943 ng/ul

RT: 21.468 min Scan# 3125

Delta R.T. 0.006 min

Lab File: BM044401.D

Acq: 28 Feb 2024 13:12




Tgt Ion:149 Resp: 1867490


Ion Ratio Lower Upper

149 100

167 27.5 22.1 33.1

279 3.4 2.4 3.6





#87

Chrysene

Concen: 38.433 ng/ul

RT: 21.574 min Scan# 3143

Delta R.T. 0.006 min

Lab File: BM044401.D

Acq: 28 Feb 2024 13:12

Instrument :

BNA_M

ClientSampleId :

SLCS253





Tgt Ion:228 Resp: 2200582

Ion Ratio Lower Upper

228 100

226 29.9 23.8 35.6

229 20.0 16.0 24.0



#89

Di-n-octyl phthalate

Concen: 42.589 ng/ul




RT: 22.415 min Scan# 3286


Delta R.T. 0.012 min

Lab File: BM044401.D

Acq: 28 Feb 2024 13:12

Tgt Ion:149 Resp: 3167921









#90
Benzo(b)fluoranthene
Concen: 40.486 ng/ul
RT: 23.221 min Scan# 3422
Delta R.T. 0.012 min
Lab File: BM044401.D
Acq: 28 Feb 2024 13:12




Instrument :
BNA_M
ClientSampleId :
SLCS253


Tgt Ion:252 Resp: 2253343
Ion Ratio Lower Upper
252 100
253 21.9 17.6 26.4
125 11.9 9.8 14.8



#91
Benzo(k)fluoranthene
Concen: 40.277 ng/ul
RT: 23.274 min Scan# 3432
Delta R.T. 0.012 min
Lab File: BM044401.D
Acq: 28 Feb 2024 13:12

Tgt Ion:252 Resp: 2302412
Ion Ratio Lower Upper
252 100
253 22.1 17.4 26.0
125 11.4 9.4 14.2





#93

Benzo(a)pyrene

Concen: 39.950 ng/ul

RT: 23.856 min Scan# 3531

Delta R.T. 0.018 min

Lab File: BM044401.D

Acq: 28 Feb 2024 13:12

Instrument :

BNA_M

ClientSampleId :

SLCS253





Tgt Ion:252 Resp: 2143757

Ion Ratio Lower Upper

252 100

253 21.8 17.4 26.2

125 12.5 10.8 16.2



#94

Indeno(1,2,3-cd)pyrene

Concen: 39.632 ng/ul

RT: 26.486 min Scan# 3978

Delta R.T. 0.018 min

Lab File: BM044401.D

Acq: 28 Feb 2024 13:12




Tgt Ion:276 Resp: 2571750


Ion Ratio Lower Upper

276 100

138 26.1 22.1 33.1

277 24.4 19.9 29.9





#95

Dibenzo(a,h)anthracene

Concen: 39.826 ng/ul

RT: 26.515 min Scan# 3983

Delta R.T. 0.018 min

Lab File: BM044401.D

Acq: 28 Feb 2024 13:12

Instrument :

BNA_M

ClientSampleId :

SLCS253





Tgt Ion:278 Resp: 2159647

Ion Ratio Lower Upper

278 100

139 18.9 14.9 22.3

279 23.6 19.5 29.3



#96

Benzo(g,h,i)perylene

Concen: 39.280 ng/ul

RT: 27.262 min Scan# 4110

Delta R.T. 0.024 min

Lab File: BM044401.D

Acq: 28 Feb 2024 13:12

Tgt Ion:276 Resp: 2061861

Ion Ratio Lower Upper

276 100

138 25.6 21.6 32.4

277 23.4 18.5 27.7

