


Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM022624\
Data File : BM044324.D
Acq On : 26 Feb 2024 10:13
Operator : MA/JU
Sample : DFTPP003
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DFTPP003

Integration File: LSCINT.P

Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM022324.MA.M
Title : SVOA CALIBRATION
Last Update : Tue Feb 27 00:07:53 2024



AutoFind: Scans 2494, 2495, 2496; Background Corrected with Scan 2489

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
51	198	10	80	42.3	131471	PASS
68	69	0.00	2	1.5	2057	PASS
69	198	0.00	100	43.1	133891	PASS
70	69	0.00	2	0.5	628	PASS
127	198	10	80	52.6	163435	PASS
197	198	0.00	2	0.3	846	PASS
198	198	100	100	100.0	310720	PASS
199	198	5	9	6.5	20185	PASS
275	198	10	60	22.0	68365	PASS
365	198	1	100	2.5	7765	PASS
441	198	0.01	100	9.6	29970	PASS
442	198	50	100	60.7	188453	PASS
443	442	15	24	19.7	37077	PASS

m/z	Abundance
39.10	234.0
40.00	2922.0
41.00	211.0
43.00	168.0
44.00	5306.0
45.10	248.0
48.90	635.0
50.90	318.0
54.90	183.0
73.00	851.0
73.90	166.0
75.00	156.0
77.00	280.0
77.90	265.0
83.80	546.0
85.90	310.0
90.90	182.0
96.10	299.0
132.90	318.0
134.80	224.0
146.90	245.0
191.00	245.0
193.00	300.0
207.00	2584.0
207.90	561.0
209.00	360.0
220.90	278.0
280.90	538.0
281.80	155.0
282.90	161.0

Instrument :
BNA_M
ClientSampleId :
DFTPP003

m/z	Abundance
35.90	273.0
36.90	361.0
38.00	1376.0
39.10	7949.0
40.00	3386.0
40.90	408.0
43.00	191.0
44.00	6424.0
44.90	325.0
46.90	197.0
48.00	205.0
49.00	1277.0
50.00	27024.0
51.00	109736.0
52.00	5934.0
52.90	317.0
54.90	598.0
56.00	3042.0
57.00	7893.0
57.90	345.0
58.90	205.0
60.90	1360.0
61.90	1224.0
63.00	3957.0
64.00	728.0
65.00	2487.0
65.80	220.0
66.90	185.0
67.90	1808.0
69.00	107976.0
70.00	510.0
73.00	1878.0
74.00	10400.0
75.00	16944.0
76.00	6475.0
77.00	128856.0
78.00	10245.0
79.00	6866.0
79.90	5789.0
81.00	8269.0
82.00	1747.0
83.00	2316.0
83.90	726.0
84.90	1247.0
86.00	2681.0
87.00	967.0
88.00	467.0
89.10	231.0
90.90	1905.0
92.00	2058.0
93.00	12005.0
93.90	704.0
94.90	293.0
96.00	1047.0
97.00	283.0
98.00	8720.0
99.00	11083.0
99.90	860.0
101.00	4425.0
101.90	308.0
103.00	1093.0
104.00	2364.0
104.90	2846.0
106.00	825.0
107.00	35208.0
108.00	5083.0
109.00	1002.0
110.00	64592.0
111.00	10192.0
112.00	1224.0
113.00	443.0
114.90	222.0
115.90	1633.0
117.00	25192.0
118.00	2393.0
118.90	340.0
119.90	403.0
121.90	2448.0
122.90	2965.0
123.90	1518.0
125.00	1326.0
126.10	543.0
127.00	123968.0
128.00	9065.0
129.00	43760.0
130.00	4316.0
130.90	984.0
132.00	346.0
132.90	337.0
133.90	1153.0
134.90	3518.0
136.00	1397.0

Instrument :
BNA_M
ClientSampleId :
DFTPP003

136.90	2711.0
138.00	457.0
139.90	513.0
141.00	4810.0
141.90	1869.0
142.90	1677.0
144.00	266.0
144.90	271.0
145.90	1119.0
147.00	2970.0
148.00	6242.0
149.00	2075.0
149.80	350.0
150.90	540.0
152.00	509.0
152.90	1907.0
153.90	1144.0
155.00	3036.0
156.00	4638.0
157.10	796.0
157.90	885.0
158.90	741.0
159.90	1463.0
161.00	2838.0
161.80	650.0
163.00	279.0
164.10	286.0
165.00	2016.0
165.90	1616.0
167.00	11231.0
168.00	11584.0
168.90	1406.0
169.90	437.0
171.00	349.0
171.90	1071.0
172.90	1296.0
174.00	2204.0
175.00	4046.0
176.00	1225.0
177.00	1788.0
177.80	485.0
178.90	7845.0
180.00	5748.0
180.90	2706.0
181.90	468.0
182.70	188.0
183.90	474.0
185.00	3626.0
186.00	28072.0
187.00	8074.0
188.00	991.0
189.00	1809.0
189.90	437.0
190.90	1119.0
191.90	2732.0
193.00	2876.0
194.00	669.0
195.00	242.0
196.00	5108.0
196.70	987.0
197.90	214976.0
198.90	13900.0
200.00	1176.0
201.50	866.0
203.00	1354.0
204.00	6758.0
205.00	11814.0
206.00	52992.0
207.00	8833.0
208.00	2051.0
208.90	870.0
209.90	824.0
211.00	1644.0
212.00	440.0
212.90	207.0
214.90	616.0
216.00	912.0
216.90	12800.0
217.90	1877.0
219.80	274.0
221.00	6523.0
222.90	3004.0
224.00	26880.0
225.00	6914.0
226.00	941.0
227.00	10448.0
227.90	1574.0
228.90	2342.0
230.00	241.0
231.00	1065.0
231.90	187.0
233.00	173.0
233.90	591.0
235.00	634.0
236.00	547.0
237.00	1249.0

Instrument :
BNA_M
ClientSampleId :
DFTPP003

237.90	262.0
238.90	262.0
239.90	284.0
241.00	692.0
242.00	1406.0
243.00	1261.0
244.00	22160.0
245.00	3212.0
245.90	3769.0
246.90	734.0
249.00	690.0
249.90	219.0
250.90	288.0
252.00	253.0
252.90	410.0
253.80	395.0
254.10	395.0
255.00	101440.0
256.00	16090.0
257.00	1375.0
258.00	4965.0
258.90	873.0
259.90	170.0
263.90	178.0
265.00	1924.0
266.00	316.0
266.80	152.0
270.90	461.0
271.90	228.0
273.00	2811.0
274.00	7690.0
275.00	42176.0
276.00	5924.0
276.90	3601.0
277.90	737.0
279.00	190.0
281.00	567.0
282.00	211.0
283.00	354.0
283.90	241.0
284.90	612.0
285.80	161.0
291.90	153.0
292.90	730.0
293.90	277.0
294.90	348.0
296.00	11073.0
296.90	1535.0
298.00	229.0
301.00	258.0
302.00	222.0
302.90	1359.0
303.90	311.0
308.00	162.0
309.90	167.0
313.90	448.0
314.90	1197.0
316.00	715.0
320.90	346.0
323.00	3723.0
324.00	735.0
327.00	696.0
327.90	289.0
332.00	349.0
332.80	503.0
334.00	1797.0
334.90	585.0
341.00	588.0
346.00	648.0
347.00	161.0
352.00	1076.0
353.00	708.0
354.00	1011.0
354.90	237.0
364.90	4448.0
365.90	581.0
370.00	249.0
370.80	215.0
372.00	1643.0
372.90	540.0
382.90	334.0
401.90	653.0
403.00	727.0
403.90	326.0
420.90	835.0
422.00	660.0
423.00	5089.0
424.00	2013.0
424.90	397.0
441.00	14821.0
442.00	90480.0
443.00	18880.0
444.00	1911.0

Instrument :
BNA_M
ClientSampleId :
DFTPP003

m/z	Abundance
35.90	223.0
36.90	723.0
38.00	1634.0
39.10	10893.0
40.00	3104.0
41.00	536.0
42.90	272.0
44.00	6061.0
44.90	556.0
46.90	166.0
49.00	1512.0
50.00	39192.0
51.00	154304.0
52.00	7846.0
53.10	269.0
55.00	712.0
56.00	5040.0
57.00	10198.0
58.00	494.0
59.00	270.0
60.00	176.0
60.90	1850.0
62.00	2164.0
63.00	6173.0
64.00	1115.0
65.00	3064.0
65.90	222.0
67.00	241.0
68.00	2359.0
69.00	159040.0
70.00	732.0
73.00	1918.0
74.00	14758.0
75.00	25072.0
76.00	8468.0
77.00	189760.0
78.00	13214.0
79.00	10779.0
80.00	8392.0
81.00	11517.0
82.00	3139.0
83.00	3148.0
83.90	971.0
85.00	1652.0
86.00	3518.0
87.00	1893.0
88.00	870.0
88.90	242.0
91.00	2498.0
91.90	3043.0
93.00	18128.0
94.00	1157.0
94.90	269.0
96.00	1127.0
96.90	246.0
98.00	12627.0
99.00	14456.0
100.00	967.0
101.00	7116.0
101.90	580.0
103.00	2298.0
104.00	4076.0
105.00	3654.0
106.00	1422.0
107.00	52952.0
108.00	8469.0
108.90	1602.0
110.00	98920.0
111.00	14474.0
112.00	1765.0
113.00	460.0
115.00	269.0
115.90	2435.0
117.00	36904.0
117.90	2891.0
119.10	480.0
119.90	593.0
122.00	2741.0
123.00	4688.0
124.00	2222.0
125.00	2367.0
126.10	744.0
127.00	193344.0
128.00	13897.0
129.00	70544.0
130.00	5671.0
131.00	1236.0
131.90	641.0
132.90	355.0
134.00	1902.0
135.00	5832.0
136.00	2229.0

Instrument :
BNA_M
ClientSampleId :
DFTPP003

137.00	3998.0
138.00	726.0
138.90	280.0
140.00	603.0
141.00	8769.0
142.00	2232.0
143.00	1872.0
143.90	666.0
144.90	425.0
146.00	1238.0
147.00	4785.0
147.90	9424.0
149.00	2576.0
150.00	591.0
151.00	1052.0
151.90	615.0
152.90	2712.0
154.00	2034.0
155.00	5050.0
156.00	7350.0
157.00	1562.0
157.90	1429.0
158.90	1191.0
159.90	2506.0
161.00	4294.0
161.90	1085.0
162.90	277.0
164.10	407.0
165.00	2834.0
166.00	2652.0
167.00	17248.0
168.00	13552.0
169.00	1720.0
169.90	536.0
170.90	567.0
171.90	1377.0
172.90	1896.0
174.00	3508.0
175.00	6971.0
175.90	2301.0
176.90	3222.0
177.90	1002.0
178.90	12180.0
180.00	9715.0
180.90	4031.0
181.90	594.0
182.70	156.0
184.00	922.0
185.00	5367.0
186.00	46736.0
187.00	12581.0
188.00	1951.0
189.00	2802.0
190.00	462.0
191.00	1393.0
192.00	4546.0
193.00	4340.0
194.00	1136.0
195.00	539.0
196.00	10113.0
197.00	1514.0
197.90	365760.0
198.90	24080.0
199.90	1803.0
201.40	1203.0
203.00	2082.0
204.00	12687.0
205.00	21504.0
206.00	90072.0
207.00	14588.0
208.00	3120.0
208.90	1455.0
209.90	1171.0
211.00	3468.0
211.90	500.0
213.00	226.0
214.90	798.0
215.90	2028.0
216.90	23472.0
217.90	3438.0
219.00	294.0
221.00	12848.0
221.90	3156.0
223.00	5721.0
224.00	49128.0
225.00	11827.0
226.00	1416.0
227.00	19000.0
228.00	2864.0
229.00	3760.0
229.90	538.0
231.00	1710.0
231.80	188.0
233.00	253.0
234.00	1184.0
234.90	1526.0

Instrument :
BNA_M
ClientSampleId :
DFTPP003

236.00	1127.0
237.00	1847.0
238.00	238.0
238.90	713.0
239.90	583.0
240.90	852.0
242.00	2573.0
243.00	2815.0
244.00	36720.0
245.00	4775.0
246.00	6329.0
246.90	1446.0
247.90	352.0
249.00	1506.0
250.00	286.0
251.00	256.0
251.80	222.0
252.90	563.0
253.90	949.0
255.00	188800.0
256.00	26544.0
256.90	1895.0
258.00	9613.0
258.90	1587.0
259.90	300.0
261.00	414.0
263.00	255.0
263.90	231.0
265.00	4629.0
265.90	637.0
269.80	160.0
270.90	600.0
271.90	682.0
273.00	5622.0
274.00	13986.0
275.00	81160.0
276.00	10972.0
277.00	6138.0
278.00	1393.0
278.90	243.0
281.00	609.0
281.80	285.0
283.00	804.0
284.00	460.0
285.00	1036.0
288.90	289.0
290.00	275.0
290.90	195.0
292.00	249.0
292.90	1687.0
294.00	488.0
295.00	429.0
296.00	20928.0
297.00	3397.0
297.90	246.0
300.80	178.0
301.90	415.0
303.00	2453.0
304.00	526.0
307.80	445.0
308.90	217.0
309.90	259.0
312.90	278.0
314.00	652.0
315.00	2200.0
316.00	1280.0
316.90	341.0
320.90	517.0
321.90	452.0
323.00	6364.0
324.00	945.0
325.10	168.0
326.90	1467.0
327.80	678.0
328.90	190.0
332.00	504.0
332.90	654.0
334.00	4052.0
335.00	1217.0
341.00	1127.0
342.10	259.0
345.90	1328.0
346.80	303.0
350.90	174.0
352.00	1790.0
353.00	1106.0
354.00	1985.0
355.00	526.0
358.80	194.0
364.90	7925.0
366.00	1310.0
369.90	172.0
371.00	495.0
372.00	3284.0
373.00	849.0
383.00	1233.0

Instrument :
BNA_M
ClientSampleId :
DFTPP003

384.00	242.0
390.00	439.0
391.00	250.0
391.90	240.0
401.00	197.0
402.00	1031.0
403.00	1873.0
403.90	598.0
421.00	2021.0
422.00	1568.0
423.00	11101.0
424.00	4065.0
425.10	498.0
441.00	33440.0
442.00	212160.0
443.00	42072.0
444.00	3860.0
444.90	246.0

Instrument :
BNA_M
ClientSampleId :
DFTPP003

137.00	3998.0
138.00	726.0
138.90	280.0
140.00	603.0
141.00	8769.0
142.00	2232.0
143.00	1872.0
143.90	666.0
144.90	425.0
146.00	1238.0
147.00	4785.0
147.90	9424.0
149.00	2576.0
150.00	591.0
151.00	1052.0
151.90	615.0
152.90	2712.0
154.00	2034.0
155.00	5050.0
156.00	7350.0
157.00	1562.0
157.90	1429.0
158.90	1191.0
159.90	2506.0
161.00	4294.0
161.90	1085.0
162.90	277.0
164.10	407.0
165.00	2834.0
166.00	2652.0
167.00	17248.0
168.00	13552.0
169.00	1720.0
169.90	536.0
170.90	567.0
171.90	1377.0
172.90	1896.0
174.00	3508.0
175.00	6971.0
175.90	2301.0
176.90	3222.0
177.90	1002.0
178.90	12180.0
180.00	9715.0
180.90	4031.0
181.90	594.0
182.70	156.0
184.00	922.0
185.00	5367.0
186.00	46736.0
187.00	12581.0
188.00	1951.0
189.00	2802.0
190.00	462.0
191.00	1393.0
192.00	4546.0
193.00	4340.0
194.00	1136.0
195.00	539.0
196.00	10113.0
197.00	1514.0
197.90	365760.0
198.90	24080.0
199.90	1803.0
201.40	1203.0
203.00	2082.0
204.00	12687.0
205.00	21504.0
206.00	90072.0
207.00	14588.0
208.00	3120.0
208.90	1455.0
209.90	1171.0
211.00	3468.0
211.90	500.0
213.00	226.0
214.90	798.0
215.90	2028.0
216.90	23472.0
217.90	3438.0
219.00	294.0
221.00	12848.0
221.90	3156.0
223.00	5721.0
224.00	49128.0
225.00	11827.0
226.00	1416.0
227.00	19000.0
228.00	2864.0
229.00	3760.0
229.90	538.0
231.00	1710.0
231.80	188.0
233.00	253.0
234.00	1184.0
234.90	1526.0

Instrument :
BNA_M
ClientSampleId :
DFTPP003

236.00	1127.0
237.00	1847.0
238.00	238.0
238.90	713.0
239.90	583.0
240.90	852.0
242.00	2573.0
243.00	2815.0
244.00	36720.0
245.00	4775.0
246.00	6329.0
246.90	1446.0
247.90	352.0
249.00	1506.0
250.00	286.0
251.00	256.0
251.80	222.0
252.90	563.0
253.90	949.0
255.00	188800.0
256.00	26544.0
256.90	1895.0
258.00	9613.0
258.90	1587.0
259.90	300.0
261.00	414.0
263.00	255.0
263.90	231.0
265.00	4629.0
265.90	637.0
269.80	160.0
270.90	600.0
271.90	682.0
273.00	5622.0
274.00	13986.0
275.00	81160.0
276.00	10972.0
277.00	6138.0
278.00	1393.0
278.90	243.0
281.00	609.0
281.80	285.0
283.00	804.0
284.00	460.0
285.00	1036.0
288.90	289.0
290.00	275.0
290.90	195.0
292.00	249.0
292.90	1687.0
294.00	488.0
295.00	429.0
296.00	20928.0
297.00	3397.0
297.90	246.0
300.80	178.0
301.90	415.0
303.00	2453.0
304.00	526.0
307.80	445.0
308.90	217.0
309.90	259.0
312.90	278.0
314.00	652.0
315.00	2200.0
316.00	1280.0
316.90	341.0
320.90	517.0
321.90	452.0
323.00	6364.0
324.00	945.0
325.10	168.0
326.90	1467.0
327.80	678.0
328.90	190.0
332.00	504.0
332.90	654.0
334.00	4052.0
335.00	1217.0
341.00	1127.0
342.10	259.0
345.90	1328.0
346.80	303.0
350.90	174.0
352.00	1790.0
353.00	1106.0
354.00	1985.0
355.00	526.0
358.80	194.0
364.90	7925.0
366.00	1310.0
369.90	172.0
371.00	495.0
372.00	3284.0
373.00	849.0
383.00	1233.0

Instrument :
BNA_M
ClientSampleId :
DFTPP003

384.00	242.0
390.00	439.0
391.00	250.0
391.90	240.0
401.00	197.0
402.00	1031.0
403.00	1873.0
403.90	598.0
421.00	2021.0
422.00	1568.0
423.00	11101.0
424.00	4065.0
425.10	498.0
441.00	33440.0
442.00	212160.0
443.00	42072.0
444.00	3860.0
444.90	246.0

Instrument :
BNA_M
ClientSampleId :
DFTPP003

m/z	Abundance
37.00	729.0
38.00	1599.0
39.00	9461.0
40.00	3289.0
41.00	248.0
43.10	270.0
44.00	6145.0
45.00	319.0
48.00	180.0
49.00	1478.0
50.00	32024.0
51.10	131328.0
52.10	6739.0
53.00	332.0
55.00	604.0
56.00	3337.0
57.00	9369.0
58.00	446.0
59.00	160.0
61.00	1661.0
62.00	1667.0
63.00	5432.0
63.90	724.0
65.00	2641.0
65.90	249.0
67.00	201.0
67.90	2003.0
69.00	134656.0
69.90	643.0
73.10	1682.0
74.00	12695.0
75.00	20672.0
76.00	7002.0
77.00	164480.0
78.00	11865.0
79.00	8617.0
80.00	6667.0
81.00	10982.0
82.00	2949.0
83.00	2218.0
83.90	567.0
85.00	2074.0
86.00	3282.0
87.00	1124.0
87.90	780.0
89.00	282.0
91.00	2086.0
92.00	2276.0
93.00	15602.0
93.90	1039.0
94.90	416.0
96.00	942.0
97.00	380.0
98.00	12017.0
99.00	10973.0
99.90	676.0
101.00	6051.0
102.10	335.0
102.90	1888.0
103.90	3229.0
105.00	3530.0
106.00	1052.0
107.00	47744.0
108.00	7387.0
109.10	1343.0
110.00	91144.0
111.00	13036.0
112.00	1333.0
112.90	427.0
115.00	316.0
115.90	2471.0
117.00	34192.0
118.00	2524.0
118.90	293.0
119.90	536.0
120.80	247.0
121.90	2696.0
122.90	4416.0
123.90	1873.0
125.00	1951.0
126.10	500.0
127.00	172992.0
128.00	12977.0
129.00	63696.0
130.00	4994.0
130.90	1118.0
132.10	376.0
132.80	558.0
133.90	1636.0
135.00	5043.0
136.00	2233.0
137.00	3082.0

Instrument :
BNA_M
ClientSampleId :
DFTPP003

138.00	477.0
138.80	262.0
139.90	683.0
141.00	6608.0
142.00	2549.0
142.90	1860.0
144.00	561.0
144.90	458.0
145.90	1381.0
147.00	3563.0
147.90	9160.0
149.00	2242.0
149.90	738.0
151.20	977.0
152.00	538.0
152.90	2507.0
154.00	1701.0
155.00	4566.0
156.00	6404.0
157.00	1187.0
157.90	1596.0
159.00	1149.0
159.90	2279.0
161.00	3458.0
161.90	1456.0
163.00	242.0
164.00	346.0
164.90	2834.0
166.00	2552.0
167.00	14998.0
168.00	10773.0
169.00	1507.0
169.90	445.0
171.00	563.0
171.90	1473.0
173.00	1716.0
173.90	3278.0
175.00	6818.0
176.00	1981.0
176.90	2725.0
177.90	1037.0
178.90	11342.0
180.00	9054.0
181.00	3975.0
182.00	499.0
182.90	207.0
184.00	956.0
185.00	5281.0
186.00	45448.0
187.00	12282.0
188.00	1381.0
189.00	2730.0
189.90	578.0
190.90	1569.0
192.00	3968.0
193.00	4573.0
193.90	1051.0
194.80	530.0
196.00	9333.0
196.70	1551.0
197.90	351424.0
198.90	22576.0
199.90	1717.0
201.50	1327.0
202.90	2236.0
204.00	12799.0
205.00	21088.0
206.00	89000.0
207.00	13139.0
208.00	3302.0
209.00	901.0
210.00	1428.0
211.00	3549.0
212.00	678.0
212.90	260.0
213.90	161.0
214.90	911.0
216.00	1719.0
217.00	22168.0
218.00	2969.0
218.90	183.0
221.00	12336.0
221.70	3080.0
222.90	4901.0
224.00	47216.0
225.00	12209.0
225.90	1109.0
227.00	18016.0
227.90	2818.0
228.90	4002.0
229.90	422.0
231.00	1852.0
231.90	421.0
232.90	232.0
233.90	1237.0
234.90	1350.0

Instrument :
BNA_M
ClientSampleId :
DFTPP003

235.90	1065.0
237.00	1623.0
237.90	219.0
239.00	759.0
240.00	689.0
240.90	1095.0
242.00	2389.0
243.00	2691.0
244.00	39344.0
245.00	4988.0
246.00	7950.0
246.90	1413.0
248.00	289.0
248.90	1107.0
250.00	151.0
251.00	473.0
251.90	299.0
252.90	737.0
254.00	906.0
255.00	187904.0
256.00	26376.0
257.00	2301.0
257.90	9929.0
259.00	1590.0
260.00	276.0
260.90	294.0
263.90	245.0
265.00	4603.0
265.90	679.0
270.00	258.0
271.00	517.0
272.00	568.0
273.00	5874.0
274.00	14174.0
275.00	81760.0
276.00	10832.0
277.00	6547.0
278.00	1117.0
279.10	281.0
281.00	827.0
281.90	357.0
282.90	956.0
284.00	602.0
284.90	1020.0
286.00	229.0
288.90	319.0
290.00	263.0
290.90	214.0
291.90	351.0
292.90	1694.0
294.00	375.0
295.00	414.0
296.00	22640.0
296.90	3064.0
298.00	190.0
301.00	314.0
301.90	502.0
303.00	2597.0
304.00	860.0
308.00	218.0
308.90	198.0
309.90	223.0
313.00	280.0
314.00	1122.0
314.90	2104.0
315.90	1227.0
316.90	373.0
320.90	810.0
322.10	369.0
323.00	7892.0
324.00	1192.0
324.90	166.0
325.90	172.0
327.00	1397.0
327.90	672.0
331.90	514.0
332.90	857.0
334.00	4899.0
335.00	1135.0
341.00	971.0
341.90	175.0
345.90	1951.0
346.90	242.0
350.90	151.0
351.90	2111.0
353.00	1691.0
354.00	2165.0
355.00	362.0
359.10	193.0
364.90	10922.0
366.00	1099.0
370.00	335.0
371.00	714.0
372.00	3867.0
373.00	644.0
382.90	1119.0

Instrument :
BNA_M
ClientSampleId :
DFTPP003

384.00	326.0
389.90	488.0
391.00	242.0
392.00	329.0
401.00	445.0
402.00	2044.0
402.90	2033.0
403.90	686.0
404.90	203.0
421.00	1723.0
421.90	1786.0
423.00	12100.0
424.00	3182.0
425.00	612.0
441.00	41648.0
442.00	262720.0
443.00	50280.0
444.00	4354.0
445.00	349.0

Instrument :
BNA_M
ClientSampleId :
DFTPP003