

Data Path : Z:\HPCHEM1\BNA M\DATA\BM022816\
 Data File : BM004460.D
 Acq On : 28 Feb 2016 20:56
 Operator : SJ/UM
 Sample : SSTDICV001
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 ICVBM022816

Quant Time: Feb 29 06:36:15 2016
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\8270-SIM-BM022816.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Feb 29 05:57:18 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	86	0.00
2	1,4-Dioxane	0.475	0.453	4.6	104	0.00
3	n-Nitrosodimethylamine	0.371	0.376	-1.3	101	0.00
4 S	2-Fluorophenol	1.100	1.087	1.2	97	0.00
5 S	Phenol-d6	1.248	1.141	8.6	97	0.00
6	bis(2-Chloroethyl)ether	0.992	1.024	-3.2	100	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	91	0.00
8 S	Nitrobenzene-d5	0.276	0.258	6.5	97	0.00
9	Nitrobenzene	0.307	0.296	3.6	98	0.00
10	Hexachlorobutadiene	0.204	0.201	1.5	98	0.00
11	2-Methylnaphthalene	0.761	0.778	-2.2	103	0.00
12 I	Acenaphthene-d10	1.000	1.000	0.0	95	0.00
13 S	2,4,6-Tribromophenol	0.198	0.150	24.2	89	0.00
14 S	2-Fluorobiphenyl	1.525	1.430	6.2	99	0.00
15 I	Phenanthrene-d10	1.000	1.000	0.0	94	0.00
16	Hexachlorobenzene	0.278	0.268	3.6	104	0.00
17	Pentachlorophenol	0.103	0.065	36.9#	94	0.00
18	Fluoranthene	1.981	1.765	10.9	100	0.00
19 I	Chrysene-d12	1.000	1.000	0.0	93	0.00
20	Benzidine	0.257	0.361	-40.5#	158#	0.00
21	Pvrene	1.489	1.400	6.0	99	0.00
22 S	Terphenyl-d14	0.637	0.616	3.3	103	0.00
23	Benzo(a)anthracene	1.428	1.268	11.2	105	0.00
24	3,3'-Dichlorobenzidine	0.490	0.512	-4.5	101	0.00
25	Chrysene	2.306	1.355	41.2#	99	0.00
26	Bis(2-ethylhexyl)phthalate	0.906	0.696	23.2	86	0.00
27	Indeno(1,2,3-cd)pyrene	1.301	1.234	5.1	102	-0.01
28 I	Perylene-d12	1.000	1.000	0.0	91	-0.01
29	Benzo(b)fluoranthene	1.885	1.794	4.8	96	0.00
30	Benzo(k)fluoranthene	1.625	1.448	10.9	100	0.00
31 C	Benzo(a)pyrene	1.466	1.453	0.9	102	0.00
32	Dibenzo(a,h)anthracene	1.309	1.250	4.5	102	0.00
33	Benzo(g,h,i)perylene	1.438	1.380	4.0	103	-0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0