

Data Path : Z:\HPCHEM1\BNA M\DATA\BM022816\
 Data File : BM004462.D
 Acq On : 28 Feb 2016 23:21
 Operator : SJ/UM
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC001

Quant Time: Feb 29 06:41:25 2016
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\8270-SIM-BM022816.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Feb 29 05:57:18 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	97	0.00
2	1,4-Dioxane	1.000	0.933	6.7	100	0.00
3	n-Nitrosodimethylamine	1.000	0.894	10.6	100	0.00
4 S	2-Fluorophenol	1.000	0.829	17.1	91	0.00
5 S	Phenol-d6	1.000	0.776	22.4	93	0.00
6	bis(2-Chloroethyl)ether	1.000	0.907	9.3	98	0.00
7 I	Naphthalene-d8	5.000	5.000	0.0	100	0.00
8 S	Nitrobenzene-d5	1.000	0.786	21.4	90	0.00
9	Nitrobenzene	1.000	0.815	18.5	92	0.00
10	Hexachlorobutadiene	1.000	0.863	13.7	95	0.00
11	2-Methylnaphthalene	1.000	0.915	8.5	102	0.00
12 I	Acenaphthene-d10	5.000	5.000	0.0	107	0.00
13 S	2,4,6-Tribromophenol	1.000	0.671	32.9#	79	0.00
14 S	2-Fluorobiphenyl	1.000	0.830	17.0	99	0.00
15 I	Phenanthrene-d10	5.000	5.000	0.0	97	0.00
16	Hexachlorobenzene	1.000	0.969	3.1	101	0.00
17	Pentachlorophenol	1.000	0.781	21.9	91	0.00
18	Fluoranthene	1.000	0.856	14.4	99	0.00
19 I	Chrysene-d12	5.000	5.000	0.0	100	0.00
20	Benzidine	1.000	1.109	-10.9	158	0.00
21	Pvrene	1.000	0.866	13.4	98	0.00
22 S	Terphenyl-d14	1.000	0.875	12.5	100	0.00
23	Benzo(a)anthracene	1.000	0.856	14.4	109	0.00
24	3,3'-Dichlorobenzidine	1.000	0.951	4.9	99	0.00
25	Chrysene	1.000	0.762	23.8	100	0.00
26	Bis(2-ethylhexyl)phthalate	1.000	0.621	37.9#	75	0.00
27	Indeno(1,2,3-cd)pyrene	1.000	0.859	14.1	99	-0.01
28 I	Perylene-d12	5.000	5.000	0.0	101	-0.01
29	Benzo(b)fluoranthene	1.000	0.757	24.3	84	-0.01
30	Benzo(k)fluoranthene	1.000	0.894	10.6	111	0.00
31 C	Benzo(a)pyrene	1.000	0.881	11.9	100	0.00
32	Dibenzo(a,h)anthracene	1.000	0.833	16.7	99	0.00
33	Benzo(g,h,i)perylene	1.000	0.850	15.0	101	-0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0