

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM022822\
 Data File : BM034092.D
 Acq On : 28 Feb 2022 11:35
 Operator : CG/JU
 Sample : SSTD00532
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD005032

Quant Time: Feb 28 17:42:11 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM022822.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Feb 28 15:19:40 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.995	152	147769	20.000	ng/u1	0.00
20) Naphthalene-d8	10.813	136	567035	20.000	ng/u1	# 0.00
38) Acenaphthene-d10	14.636	164	398141	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.371	188	872547	20.000	ng/u1	# 0.00
79) Chrysene-d12	21.542	240	878455	20.000	ng/u1	# 0.00
88) Perylene-d12	23.994	264	867436	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.372	96	6918	2.159	ng/uL	0.00
4) Pyridine-d5	3.796	84	39237	4.659	ng/u1	0.00
7) Phenol-d5	7.143	99	45576	4.318	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.319	67	27763	4.739	ng/u1	0.00
11) 2-Chlorophenol-d4	7.525	132	41703	4.584	ng/u1	0.00
15) 4-Methylphenol-d8	8.695	113	37680	4.195	ng/u1	0.00
21) Nitrobenzene-d5	9.160	128	19802	4.623	ng/u1	0.00
24) 2-Nitrophenol-d4	9.884	143	20805	4.273	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.431	165	44214	4.488	ng/u1	0.00
31) 4-Chloroaniline-d4	10.942	131	55208	4.443	ng/u1	0.00
46) Dimethylphthalate-d6	14.036	166	146403	4.860	ng/u1	0.00
49) Acenaphthylene-d8	14.324	160	167712	4.571	ng/u1	0.00
54) 4-Nitrophenol-d4	14.807	143	18862	3.656	ng/u1	0.00
60) Fluorene-d10	15.624	176	126117	4.844	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.730	200	19998	3.511	ng/u1	0.00
73) Anthracene-d10	17.471	188	197661	4.685	ng/u1	0.00
81) Pyrene-d10	19.753	212	226630	4.786	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.830	264	201832	4.566	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.407	88	6701	2.130	ng/uL	88
5) Pyridine	3.819	79	40596	4.723	ng/u1#	79
6) Benzaldehyde	7.125	77	32540	5.429	ng/u1#	80
8) Phenol	7.172	94	45480	4.414	ng/u1	89
10) Bis(2-Chloroethyl)ether	7.413	93	38427	4.735	ng/u1	85
12) 2-Chlorophenol	7.554	128	41688	4.617	ng/u1#	79
13) 2-Methylphenol	8.437	108	35411	4.373	ng/u1	93
14) 2,2'-oxybis(1-Chloropr...	8.525	45	42537	4.803	ng/u1#	88
16) Acetophenone	8.819	105	65224	4.669	ng/u1#	83
17) N-Nitroso-di-n-propyla...	8.801	70	29432	4.619	ng/u1#	87
18) 4-Methylphenol	8.760	108	38554	4.390	ng/u1	96
19) Hexachloroethane	9.095	117	17828	4.715	ng/u1#	61
22) Nitrobenzene	9.201	77	42745	4.653	ng/u1#	88
23) Isophorone	9.731	82	77049	4.549	ng/u1#	89
25) 2-Nitrophenol	9.913	139	21321	4.242	ng/u1#	77
26) 2,4-Dimethylphenol	9.978	107	46266	4.596	ng/u1	91
27) Bis(2-Chloroethoxy)met...	10.213	93	48803	4.752	ng/u1#	97
29) 2,4-Dichlorophenol	10.454	162	42491	4.490	ng/u1#	90
30) Naphthalene	10.866	128	135888	4.782	ng/u1	98
32) 4-Chloroaniline	10.966	127	53477	4.354	ng/u1	99
33) Hexachlorobutadiene	11.160	225	35974	4.950	ng/u1	94
34) Caprolactam	11.719	113	10370	4.035	ng/u1#	52
35) 4-Chloro-3-methylphenol	12.083	107	39509	4.459	ng/u1#	77
36) 2-Methylnaphthalene	12.472	142	96752	4.664	ng/u1	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM022822\
 Data File : BM034092.D
 Acq On : 28 Feb 2022 11:35
 Operator : CG/JU
 Sample : SST00532
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SST005032

Quant Time: Feb 28 17:42:11 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM022822.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Feb 28 15:19:40 2022
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.689	142	98452	4.682	ng/ul#	96
39) 1,2,4,5-Tetrachloroben...	12.836	216	64541	4.879	ng/ul#	96
40) Hexachlorocyclopentadiene	12.825	237	34981	3.878	ng/ul	97
41) 2,4,6-Trichlorophenol	13.072	196	35967	4.451	ng/ul	99
42) 2,4,5-Trichlorophenol	13.136	196	40116	4.514	ng/ul	95
43) 1,1'-Biphenyl	13.472	154	141524	4.822	ng/ul#	97
44) 2-Chloronaphthalene	13.513	162	108022	4.738	ng/ul	95
45) 2-Nitroaniline	13.707	65	20944	4.080	ng/ul#	64
47) Dimethylphthalate	14.083	163	141174	4.841	ng/ul	98
48) 2,6-Dinitrotoluene	14.195	165	24284	4.255	ng/ul	88
50) Acenaphthylene	14.354	152	168240	4.669	ng/ul	99
51) 3-Nitroaniline	14.524	138	22706	4.090	ng/ul#	74
52) Acenaphthene	14.695	153	114473	4.787	ng/ul	98
53) 2,4-Dinitrophenol	14.724	184	11688	3.087	ng/ul#	76
55) 4-Nitrophenol	14.819	109	19386	4.128	ng/ul#	80
56) Dibenzofuran	15.030	168	171888	4.895	ng/ul	92
57) 2,4-Dinitrotoluene	14.983	165	34674	4.181	ng/ul#	70
58) 2,3,4,6-Tetrachlorophenol	15.254	232	34608	4.501	ng/ul#	83
59) Diethylphthalate	15.448	149	137238	4.714	ng/ul	96
61) Fluorene	15.677	166	140656	4.803	ng/ul	98
62) 4-Chlorophenyl-phenyle...	15.671	204	75556	4.824	ng/ul	90
63) 4-Nitroaniline	15.677	138	22847	4.033	ng/ul#	85
66) 4,6-Dinitro-2-methylph...	15.742	198	19723	3.543	ng/ul#	84
67) N-Nitrosodiphenylamine	15.883	169	118058	4.660	ng/ul	98
68) 4-Bromophenyl-phenylether	16.565	248	43202	4.538	ng/ul#	90
69) Hexachlorobenzene	16.683	284	55026	4.766	ng/ul#	91
70) Atrazine	16.824	200	44921	4.320	ng/ul	97
71) Pentachlorophenol	17.024	266	28824	3.834	ng/ul	95
72) Phenanthrene	17.412	178	227498	4.807	ng/ul	99
74) Anthracene	17.507	178	224908	4.693	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.436	216	65247	4.737	ng/ul	97
76) Pentachlorobenzene	14.954	250	65098	4.778	ng/ul	98
77) Carbazole	17.771	167	184914	4.427	ng/ul	99
78) Di-n-butylphthalate	18.342	149	212960	4.373	ng/ul	99
80) Fluoranthene	19.424	202	257749	4.711	ng/ul#	90
82) Pyrene	19.783	202	274474	4.856	ng/ul#	87
83) Butylbenzylphthalate	20.683	149	83627	4.196	ng/ul#	75
84) 3,3'-Dichlorobenzidine	21.453	252	75859	4.074	ng/ul	94
85) Benzo(a)anthracene	21.530	228	254879	4.746	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.459	149	125918	4.098	ng/ul	99
87) Chrysene	21.583	228	255273	4.841	ng/ul	97
89) Di-n-octyl phthalate	22.400	149	188783	3.536	ng/ul	100
90) Benzo(b)fluoranthene	23.241	252	244984	4.454	ng/ul#	97
91) Benzo(k)fluoranthene	23.294	252	240986	4.654	ng/ul#	96
93) Benzo(a)pyrene	23.883	252	244436	4.641	ng/ul#	97
94) Indeno(1,2,3-cd)pyrene	26.547	276	281963	4.811	ng/ul	97
95) Dibenzo(a,h)anthracene	26.565	278	248129	4.890	ng/ul	98
96) Benzo(g,h,i)perylene	27.329	276	242551	4.900	ng/ul#	93


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM022822\
 Data File : BM034092.D
 Acq On : 28 Feb 2022 11:35
 Operator : CG/JU
 Sample : SSTD00532
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD005032

Quant Time: Feb 28 17:42:11 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM022822.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Feb 28 15:19:40 2022
 Response via : Initial Calibration









#2
 1,4-Dioxane
 Concen: 2.130 ng/uL
 RT: 3.407 min Scan# 29
 Delta R.T. 0.000 min
 Lab File: BM034092.D
 Acq: 28 Feb 2022 11:35




Instrument :
 BNA_M
 ClientSampleId :
 SSTD005032


Tgt Ion: 88 Resp: 6701
 Ion Ratio Lower Upper
 88 100
 43 36.9 36.2 54.4
 58 74.5 68.5 102.7



#5
 Pyridine
 Concen: 4.723 ng/uL
 RT: 3.819 min Scan# 99
 Delta R.T. 0.006 min
 Lab File: BM034092.D
 Acq: 28 Feb 2022 11:35

Tgt Ion: 79 Resp: 40596
 Ion Ratio Lower Upper
 79 100
 52 70.3 78.1 117.1#
 51 40.4 36.0 54.0









#6
Benzaldehyde
Concen: 5.429 ng/ul
RT: 7.125 min Scan# 661
Delta R.T. 0.000 min
Lab File: BM034092.D
Acq: 28 Feb 2022 11:35




Instrument :
BNA_M
ClientSampleId :
SSTD005032


Tgt Ion: 77 Resp: 32540
Ion Ratio Lower Upper
77 100
105 96.8 65.6 98.4
106 95.8 60.6 90.8#



#8
Phenol
Concen: 4.414 ng/ul
RT: 7.172 min Scan# 669
Delta R.T. 0.000 min
Lab File: BM034092.D
Acq: 28 Feb 2022 11:35

Tgt Ion: 94 Resp: 45480
Ion Ratio Lower Upper
94 100
65 33.9 31.7 47.5
66 47.5 45.4 68.0









#10
Bis(2-Chloroethyl)ether
Concen: 4.735 ng/ul
RT: 7.413 min Scan# 710
Delta R.T. 0.000 min
Lab File: BM034092.D
Acq: 28 Feb 2022 11:35




Instrument :
BNA_M
ClientSampleId :
SSTD005032


Tgt Ion: 93 Resp: 38427
Ion Ratio Lower Upper
93 100
63 72.2 71.7 107.5
95 30.9 26.3 39.5



#12
2-Chlorophenol
Concen: 4.617 ng/ul
RT: 7.554 min Scan# 734
Delta R.T. -0.006 min
Lab File: BM034092.D
Acq: 28 Feb 2022 11:35

Tgt Ion: 128 Resp: 41688
Ion Ratio Lower Upper
128 100
64 38.6 49.4 74.2#
130 30.5 26.8 40.2









#13
2-Methylphenol
Concen: 4.373 ng/ul
RT: 8.437 min Scan# 884
Delta R.T. 0.000 min
Lab File: BM034092.D
Acq: 28 Feb 2022 11:35




Instrument :
BNA_M
ClientSampleId :
SSTD005032


Tgt Ion:108 Resp: 35411
Ion Ratio Lower Upper
108 100
107 81.3 70.2 105.4



#14
2,2'-oxybis(1-Chloropropane)
Concen: 4.803 ng/ul
RT: 8.525 min Scan# 899
Delta R.T. -0.012 min
Lab File: BM034092.D
Acq: 28 Feb 2022 11:35

Tgt Ion: 45 Resp: 42537
Ion Ratio Lower Upper
45 100
77 18.0 9.9 14.9#
79 14.0 8.3 12.5#









#16
Acetophenone
Concen: 4.669 ng/ul
RT: 8.819 min Scan# 949
Delta R.T. 0.000 min
Lab File: BM034092.D
Acq: 28 Feb 2022 11:35




Instrument :
BNA_M
ClientSampleId :
SSTD005032


Tgt Ion:105 Resp: 65224
Ion Ratio Lower Upper
105 100
77 75.5 73.1 109.7
51 26.9 30.5 45.7#



#17
N-Nitroso-di-n-propylamine
Concen: 4.619 ng/ul
RT: 8.801 min Scan# 946
Delta R.T. -0.012 min
Lab File: BM034092.D
Acq: 28 Feb 2022 11:35

Tgt Ion: 70 Resp: 29432
Ion Ratio Lower Upper
70 100
42 58.6 56.9 85.3
101 9.8 6.9 10.3
130 22.2 13.8 20.8#









#18
4-Methylphenol
Concen: 4.390 ng/ul
RT: 8.760 min Scan# 911
Delta R.T. -0.006 min
Lab File: BM034092.D
Acq: 28 Feb 2022 11:35




Instrument :
BNA_M
ClientSampleId :
SSTD005032


Tgt Ion:108 Resp: 38554
Ion Ratio Lower Upper
108 100
107 113.5 87.4 131.2



#19
Hexachloroethane
Concen: 4.715 ng/ul
RT: 9.095 min Scan# 996
Delta R.T. 0.006 min
Lab File: BM034092.D
Acq: 28 Feb 2022 11:35

Tgt Ion:117 Resp: 17828
Ion Ratio Lower Upper
117 100
201 125.8 69.6 104.4#
199 73.5 40.6 60.8#









#22
Nitrobenzene
Concen: 4.653 ng/ul
RT: 9.201 min Scan# 1014
Delta R.T. 0.000 min
Lab File: BM034092.D
Acq: 28 Feb 2022 11:35




Instrument :
BNA_M
ClientSampleId :
SSTD005032


Tgt Ion: 77 Resp: 42745
Ion Ratio Lower Upper
77 100
123 45.8 29.0 43.6#
65 12.9 11.4 17.2



#23
Isophorone
Concen: 4.549 ng/ul
RT: 9.731 min Scan# 1104
Delta R.T. 0.000 min
Lab File: BM034092.D
Acq: 28 Feb 2022 11:35

Tgt Ion: 82 Resp: 77049
Ion Ratio Lower Upper
82 100
95 10.1 6.5 9.7#
138 20.8 11.8 17.8#









#25
2-Nitrophenol
Concen: 4.242 ng/ul
RT: 9.913 min Scan# 1135
Delta R.T. -0.006 min
Lab File: BM034092.D
Acq: 28 Feb 2022 11:35




Instrument :
BNA_M
ClientSampleId :
SSTD005032


Tgt Ion:139 Resp: 21321
Ion Ratio Lower Upper
139 100
65 51.4 59.0 88.6#
109 34.5 37.2 55.8#



#26
2,4-Dimethylphenol
Concen: 4.596 ng/ul
RT: 9.978 min Scan# 1146
Delta R.T. 0.000 min
Lab File: BM034092.D
Acq: 28 Feb 2022 11:35

Tgt Ion:107 Resp: 46266
Ion Ratio Lower Upper
107 100
121 49.1 34.8 52.2
122 83.1 60.0 90.0





#27

Bis(2-Chloroethoxy)methane

Concen: 4.752 ng/ul

RT: 10.213 min Scan# 1186

Delta R.T. -0.006 min

Lab File: BM034092.D

Acq: 28 Feb 2022 11:35

Instrument :

BNA_M

ClientSampleId :

SSTD005032





Tgt Ion: 93 Resp: 48803

Ion Ratio Lower Upper

93 100

95 32.7 25.3 37.9

123 12.8 7.9 11.9#



#29

2,4-Dichlorophenol

Concen: 4.490 ng/ul

RT: 10.454 min Scan# 1227

Delta R.T. 0.000 min

Lab File: BM034092.D

Acq: 28 Feb 2022 11:35




Tgt Ion:162 Resp: 42491


Ion Ratio Lower Upper

162 100

164 68.5 50.0 75.0

98 29.5 30.6 45.8#









#30
Naphthalene
Concen: 4.782 ng/ul
RT: 10.866 min Scan# 1297
Delta R.T. 0.000 min
Lab File: BM034092.D
Acq: 28 Feb 2022 11:35




Instrument :
BNA_M
ClientSampleId :
SSTD005032


Tgt Ion:128 Resp: 135888
Ion Ratio Lower Upper
128 100
129 10.3 9.1 13.7
127 13.3 10.2 15.4



#32
4-Chloroaniline
Concen: 4.354 ng/ul
RT: 10.966 min Scan# 1314
Delta R.T. 0.000 min
Lab File: BM034092.D
Acq: 28 Feb 2022 11:35

Tgt Ion:127 Resp: 53477
Ion Ratio Lower Upper
127 100
129 31.1 25.4 38.0









#33
Hexachlorobutadiene
Concen: 4.950 ng/ul
RT: 11.160 min Scan# 11
Delta R.T. 0.000 min
Lab File: BM034092.D
Acq: 28 Feb 2022 11:35




Instrument :
BNA_M
ClientSampleId :
SSTD005032


Tgt Ion:225 Resp: 35974
Ion Ratio Lower Upper
225 100
223 62.1 52.2 78.2
227 60.1 52.8 79.2



#34
Caprolactam
Concen: 4.035 ng/ul
RT: 11.719 min Scan# 1442
Delta R.T. -0.006 min
Lab File: BM034092.D
Acq: 28 Feb 2022 11:35

Tgt Ion:113 Resp: 10370
Ion Ratio Lower Upper
113 100
55 119.6 157.9 236.9#
56 105.3 131.8 197.6#









#35
4-Chloro-3-methylphenol
Concen: 4.459 ng/ul
RT: 12.083 min Scan# 1504
Delta R.T. 0.000 min
Lab File: BM034092.D
Acq: 28 Feb 2022 11:35




Instrument :
BNA_M
ClientSampleId :
SSTD005032


Tgt Ion:107 Resp: 39509
Ion Ratio Lower Upper
107 100
144 26.4 17.1 25.7#
142 83.0 49.4 74.2#



#36
2-Methylnaphthalene
Concen: 4.664 ng/ul
RT: 12.472 min Scan# 1570
Delta R.T. 0.000 min
Lab File: BM034092.D
Acq: 28 Feb 2022 11:35

Tgt Ion:142 Resp: 96752
Ion Ratio Lower Upper
142 100
141 89.7 71.1 106.7





#37

1-Methylnaphthalene

Concen: 4.682 ng/ul

RT: 12.689 min Scan# 1

Delta R.T. 0.000 min

Lab File: BM034092.D

Acq: 28 Feb 2022 11:35

Instrument :

BNA_M

ClientSampleId :

SSTD005032





Tgt Ion:142 Resp: 98452

Ion Ratio Lower Upper

142 100

141 90.5 75.5 113.3

116 3.9 4.0 6.0#



#39

1,2,4,5-Tetrachlorobenzene

Concen: 4.879 ng/ul

RT: 12.836 min Scan# 1632

Delta R.T. 0.000 min

Lab File: BM034092.D

Acq: 28 Feb 2022 11:35

Tgt Ion:216 Resp: 64541




Ion Ratio Lower Upper


216 100

214 73.7 61.3 91.9

179 21.8 18.0 27.0

108 15.4 16.6 25.0#









#40
Hexachlorocyclopentadiene
Concen: 3.878 ng/ul
RT: 12.825 min Scan# 1630
Delta R.T. 0.000 min
Lab File: BM034092.D
Acq: 28 Feb 2022 11:35




Instrument :
BNA_M
ClientSampleId :
SSTD005032


Tgt Ion:237 Resp: 34981
Ion Ratio Lower Upper
237 100
235 60.6 50.6 76.0
272 12.2 9.6 14.4



#41
2,4,6-Trichlorophenol
Concen: 4.451 ng/ul
RT: 13.072 min Scan# 1672
Delta R.T. 0.000 min
Lab File: BM034092.D
Acq: 28 Feb 2022 11:35

Tgt Ion:196 Resp: 35967
Ion Ratio Lower Upper
196 100
198 99.1 78.9 118.3
200 31.1 23.7 35.5









#42
2,4,5-Trichlorophenol
Concen: 4.514 ng/ul
RT: 13.136 min Scan# 1
Delta R.T. 0.000 min
Lab File: BM034092.D
Acq: 28 Feb 2022 11:35




Instrument :
BNA_M
ClientSampleId :
SSTD005032


Tgt Ion:196 Resp: 40116
Ion Ratio Lower Upper
196 100
198 98.7 82.2 123.2
200 29.5 27.9 41.9



#43
1,1'-Biphenyl
Concen: 4.822 ng/ul
RT: 13.472 min Scan# 1740
Delta R.T. 0.000 min
Lab File: BM034092.D
Acq: 28 Feb 2022 11:35

Tgt Ion:154 Resp: 141524
Ion Ratio Lower Upper
154 100
153 41.1 33.1 49.7
76 8.8 10.7 16.1#





#44

2-Chloronaphthalene

Concen: 4.738 ng/ul

RT: 13.513 min Scan# 1

Delta R.T. 0.000 min

Lab File: BM034092.D

Acq: 28 Feb 2022 11:35

Instrument :

BNA_M

ClientSampleId :

SSTD005032





Tgt Ion:162 Resp: 108022

Ion Ratio Lower Upper

162 100

164 32.5 26.2 39.4

127 36.2 33.7 50.5



#45

2-Nitroaniline

Concen: 4.080 ng/ul

RT: 13.707 min Scan# 1780

Delta R.T. 0.000 min

Lab File: BM034092.D

Acq: 28 Feb 2022 11:35




Tgt Ion: 65 Resp: 20944


Ion Ratio Lower Upper

65 100

92 73.8 47.8 71.6#

138 118.8 61.0 91.6#





#47

Dimethylphthalate

Concen: 4.841 ng/ul

RT: 14.083 min Scan# 1844

Delta R.T. -0.006 min

Lab File: BM034092.D

Acq: 28 Feb 2022 11:35

Instrument :

BNA_M

ClientSampleId :

SSTD005032





Tgt Ion:163 Resp: 141174

Ion Ratio Lower Upper

163 100

194 4.7 3.3 4.9

164 11.0 8.2 12.2



#48

2,6-Dinitrotoluene

Concen: 4.255 ng/ul

RT: 14.195 min Scan# 1863

Delta R.T. -0.006 min

Lab File: BM034092.D

Acq: 28 Feb 2022 11:35




Tgt Ion:165 Resp: 24284


Ion Ratio Lower Upper

165 100

89 57.3 55.7 83.5

121 22.9 20.0 30.0









#50
Acenaphthylene
Concen: 4.669 ng/ul
RT: 14.354 min Scan# 1891
Delta R.T. -0.006 min
Lab File: BM034092.D
Acq: 28 Feb 2022 11:35




Instrument :
BNA_M
ClientSampleId :
SSTD005032


Tgt Ion:152 Resp: 168240
Ion Ratio Lower Upper
152 100
151 19.7 15.8 23.8
153 13.5 10.1 15.1



#51
3-Nitroaniline
Concen: 4.090 ng/ul
RT: 14.524 min Scan# 1919
Delta R.T. 0.000 min
Lab File: BM034092.D
Acq: 28 Feb 2022 11:35

Tgt Ion:138 Resp: 22706
Ion Ratio Lower Upper
138 100
108 11.9 10.8 16.2
92 104.8 110.6 166.0#









#52
Acenaphthene
Concen: 4.787 ng/ul
RT: 14.695 min Scan# 1948
Delta R.T. -0.006 min
Lab File: BM034092.D
Acq: 28 Feb 2022 11:35




Instrument :
BNA_M
ClientSampleId :
SSTD005032


Tgt Ion:153 Resp: 114473
Ion Ratio Lower Upper
153 100
152 46.1 39.4 59.2
154 92.0 72.7 109.1



#53
2,4-Dinitrophenol
Concen: 3.087 ng/ul
RT: 14.724 min Scan# 1953
Delta R.T. -0.006 min
Lab File: BM034092.D
Acq: 28 Feb 2022 11:35

Tgt Ion:184 Resp: 11688
Ion Ratio Lower Upper
184 100
63 54.9 73.8 110.8#
154 85.0 63.5 95.3









#55
4-Nitrophenol
Concen: 4.128 ng/ul
RT: 14.819 min Scan# 1970
Delta R.T. -0.006 min
Lab File: BM034092.D
Acq: 28 Feb 2022 11:35




Instrument :
BNA_M
ClientSampleId :
SSTD005032


Tgt Ion:109 Resp: 19386
Ion Ratio Lower Upper
109 100
139 99.9 77.6 116.4
65 100.9 112.2 168.4#



#56
Dibenzofuran
Concen: 4.895 ng/ul
RT: 15.030 min Scan# 2005
Delta R.T. 0.000 min
Lab File: BM034092.D
Acq: 28 Feb 2022 11:35

Tgt Ion:168 Resp: 171888
Ion Ratio Lower Upper
168 100
139 37.3 34.0 51.0









#57
2,4-Dinitrotoluene
Concen: 4.181 ng/ul
RT: 14.983 min Scan# 1997
Delta R.T. 0.000 min
Lab File: BM034092.D
Acq: 28 Feb 2022 11:35




Instrument :
BNA_M
ClientSampleId :
SSTD005032


Tgt Ion:165 Resp: 34674
Ion Ratio Lower Upper
165 100
63 39.7 51.8 77.8#
182 3.6 2.6 4.0



#58
2,3,4,6-Tetrachlorophenol
Concen: 4.501 ng/ul
RT: 15.254 min Scan# 2043
Delta R.T. 0.000 min
Lab File: BM034092.D
Acq: 28 Feb 2022 11:35

Tgt Ion:232 Resp: 34608
Ion Ratio Lower Upper
232 100
131 36.8 41.4 62.2#
130 3.0 2.0 3.0
166 27.4 27.4 41.0#





#59

Diethylphthalate

Concen: 4.714 ng/ul

RT: 15.448 min Scan# 2076

Delta R.T. 0.000 min

Lab File: BM034092.D

Acq: 28 Feb 2022 11:35

Instrument :

BNA_M

ClientSampleId :

SSTD005032





Tgt Ion:149 Resp: 137238

Ion Ratio Lower Upper

149 100

177 23.8 17.1 25.7

150 12.8 9.9 14.9



#61

Fluorene

Concen: 4.803 ng/ul

RT: 15.677 min Scan# 2115

Delta R.T. 0.000 min

Lab File: BM034092.D

Acq: 28 Feb 2022 11:35




Tgt Ion:166 Resp: 140656


Ion Ratio Lower Upper

166 100

165 99.0 77.9 116.9

167 13.1 10.3 15.5









#62
4-Chlorophenyl-phenylether
Concen: 4.824 ng/ul
RT: 15.671 min Scan# 2114
Delta R.T. 0.000 min
Lab File: BM034092.D
Acq: 28 Feb 2022 11:35




Instrument :
BNA_M
ClientSampleId :
SSTD005032


Tgt Ion:204 Resp: 75556
Ion Ratio Lower Upper
204 100
206 32.9 27.8 41.8
141 57.4 54.9 82.3



#63
4-Nitroaniline
Concen: 4.033 ng/ul
RT: 15.677 min Scan# 2115
Delta R.T. -0.006 min
Lab File: BM034092.D
Acq: 28 Feb 2022 11:35

Tgt Ion:138 Resp: 22847
Ion Ratio Lower Upper
138 100
92 46.9 48.0 72.0#
108 98.0 89.5 134.3









#66
4,6-Dinitro-2-methylphenol
Concen: 3.543 ng/ul
RT: 15.742 min Scan# 2127
Delta R.T. -0.006 min
Lab File: BM034092.D
Acq: 28 Feb 2022 11:35




Instrument :
BNA_M
ClientSampleId :
SSTD005032


Tgt Ion:198 Resp: 19723
Ion Ratio Lower Upper
198 100
182 5.3 4.2 6.2
77 22.4 26.3 39.5#



#67
N-Nitrosodiphenylamine
Concen: 4.660 ng/ul
RT: 15.883 min Scan# 2150
Delta R.T. 0.000 min
Lab File: BM034092.D
Acq: 28 Feb 2022 11:35

Tgt Ion:169 Resp: 118058
Ion Ratio Lower Upper
169 100
168 66.3 52.9 79.3
167 33.9 29.9 44.9









#68
4-Bromophenyl-phenylether
Concen: 4.538 ng/ul
RT: 16.565 min Scan# 21
Delta R.T. 0.000 min
Lab File: BM034092.D
Acq: 28 Feb 2022 11:35




Instrument :
BNA_M
ClientSampleId :
SSTD005032


Tgt Ion:248 Resp: 43202
Ion Ratio Lower Upper
248 100
250 99.6 79.0 118.4
141 69.1 69.7 104.5#



#69
Hexachlorobenzene
Concen: 4.766 ng/ul
RT: 16.683 min Scan# 2286
Delta R.T. 0.000 min
Lab File: BM034092.D
Acq: 28 Feb 2022 11:35

Tgt Ion:284 Resp: 55026
Ion Ratio Lower Upper
284 100
142 34.3 34.3 51.5#
249 34.1 26.1 39.1









#70
Atrazine
Concen: 4.320 ng/ul
RT: 16.824 min Scan# 21
Delta R.T. -0.006 min
Lab File: BM034092.D
Acq: 28 Feb 2022 11:35




Instrument :
BNA_M
ClientSampleId :
SSTD005032


Tgt Ion:200 Resp: 44921
Ion Ratio Lower Upper
200 100
173 24.0 20.6 30.8
215 47.6 36.9 55.3



#71
Pentachlorophenol
Concen: 3.834 ng/ul
RT: 17.024 min Scan# 2344
Delta R.T. 0.000 min
Lab File: BM034092.D
Acq: 28 Feb 2022 11:35

Tgt Ion:266 Resp: 28824
Ion Ratio Lower Upper
266 100
264 63.6 49.1 73.7
268 65.4 47.8 71.8









#72
Phenanthrene
Concen: 4.807 ng/ul
RT: 17.412 min Scan# 2411
Delta R.T. -0.006 min
Lab File: BM034092.D
Acq: 28 Feb 2022 11:35




Instrument :
BNA_M
ClientSampleId :
SSTD005032


Tgt Ion:178 Resp: 227498
Ion Ratio Lower Upper
178 100
179 15.4 11.7 17.5
176 18.5 15.1 22.7



#74
Anthracene
Concen: 4.693 ng/ul
RT: 17.507 min Scan# 2426
Delta R.T. 0.000 min
Lab File: BM034092.D
Acq: 28 Feb 2022 11:35

Tgt Ion:178 Resp: 224908
Ion Ratio Lower Upper
178 100
179 15.9 12.3 18.5
176 18.2 14.5 21.7





#75

1,2,3,4-Tetrachlorobenzene

Concen: 4.737 ng/uL

RT: 13.436 min Scan# 1

Delta R.T. -0.006 min

Lab File: BM034092.D

Acq: 28 Feb 2022 11:35

Instrument :

BNA_M

ClientSampleId :

SSTD005032





Tgt Ion:216 Resp: 65247

Ion Ratio Lower Upper

216 100

214 77.9 65.4 98.0

218 47.2 38.8 58.2



#76

Pentachlorobenzene

Concen: 4.778 ng/uL

RT: 14.954 min Scan# 1992

Delta R.T. 0.000 min

Lab File: BM034092.D

Acq: 28 Feb 2022 11:35




Tgt Ion:250 Resp: 65098


Ion Ratio Lower Upper

250 100

248 66.3 51.6 77.4

252 65.9 52.0 78.0





#77

Carbazole

Concen: 4.427 ng/ul

RT: 17.771 min Scan# 2471

Delta R.T. 0.000 min

Lab File: BM034092.D

Acq: 28 Feb 2022 11:35

Instrument :

BNA_M

ClientSampleId :

SSTD005032





Tgt Ion:167 Resp: 184914

Ion Ratio Lower Upper

167 100

166 21.3 16.9 25.3

139 12.5 10.5 15.7



#78

Di-n-butylphthalate

Concen: 4.373 ng/ul

RT: 18.342 min Scan# 2568

Delta R.T. 0.000 min

Lab File: BM034092.D

Acq: 28 Feb 2022 11:35




Tgt Ion:149 Resp: 212960


Ion Ratio Lower Upper

149 100

150 9.0 7.0 10.6

104 5.7 5.1 7.7









#80
Fluoranthene
Concen: 4.711 ng/ul
RT: 19.424 min Scan# 21
Delta R.T. 0.000 min
Lab File: BM034092.D
Acq: 28 Feb 2022 11:35




Instrument :
BNA_M
ClientSampleId :
SSTD005032


Tgt Ion:202 Resp: 257749
Ion Ratio Lower Upper
202 100
101 6.4 8.2 12.4#
100 4.9 6.9 10.3#



#82
Pyrene
Concen: 4.856 ng/ul
RT: 19.783 min Scan# 2813
Delta R.T. -0.006 min
Lab File: BM034092.D
Acq: 28 Feb 2022 11:35

Tgt Ion:202 Resp: 274474
Ion Ratio Lower Upper
202 100
101 6.9 9.8 14.6#
100 6.2 8.6 13.0#









#83
 Butylbenzylphthalate
 Concen: 4.196 ng/ul
 RT: 20.683 min Scan# 2966
 Delta R.T. 0.000 min
 Lab File: BM034092.D
 Acq: 28 Feb 2022 11:35




Instrument :
 BNA_M
 ClientSampleId :
 SSTD005032


Tgt Ion:149 Resp: 83627
 Ion Ratio Lower Upper
 149 100
 91 60.0 69.3 103.9#
 206 24.2 16.1 24.1#



#84
 3,3'-Dichlorobenzidine
 Concen: 4.074 ng/ul
 RT: 21.453 min Scan# 3097
 Delta R.T. -0.006 min
 Lab File: BM034092.D
 Acq: 28 Feb 2022 11:35

Tgt Ion:252 Resp: 75859
 Ion Ratio Lower Upper
 252 100
 254 66.2 49.4 74.2
 126 8.5 8.2 12.4









#85
 Benzo(a)anthracene
 Concen: 4.746 ng/ul
 RT: 21.530 min Scan# 3110
 Delta R.T. 0.000 min
 Lab File: BM034092.D
 Acq: 28 Feb 2022 11:35




Instrument :
 BNA_M
 ClientSampleId :
 SSTD005032


Tgt Ion:228 Resp: 254879
 Ion Ratio Lower Upper
 228 100
 229 19.2 15.0 22.4
 226 26.5 21.4 32.0



#86
 Bis(2-ethylhexyl)phthalate
 Concen: 4.098 ng/ul
 RT: 21.459 min Scan# 3098
 Delta R.T. 0.000 min
 Lab File: BM034092.D
 Acq: 28 Feb 2022 11:35

Tgt Ion:149 Resp: 125918
 Ion Ratio Lower Upper
 149 100
 167 27.7 21.8 32.8
 279 4.7 3.2 4.8





#87

Chrysene

Concen: 4.841 ng/ul

RT: 21.583 min Scan# 3119

Delta R.T. 0.000 min

Lab File: BM034092.D

Acq: 28 Feb 2022 11:35

Instrument :

BNA_M

ClientSampleId :

SSTD005032





Tgt Ion:228 Resp: 255273

Ion Ratio Lower Upper

228 100

226 28.5 24.5 36.7

229 19.1 15.9 23.9



#89

Di-n-octyl phthalate

Concen: 3.536 ng/ul




RT: 22.400 min Scan# 3258


Delta R.T. -0.006 min

Lab File: BM034092.D

Acq: 28 Feb 2022 11:35

Tgt Ion:149 Resp: 188783









#90
Benzo(b)fluoranthene
Concen: 4.454 ng/ul
RT: 23.241 min Scan# 3402
Delta R.T. -0.006 min
Lab File: BM034092.D
Acq: 28 Feb 2022 11:35




Instrument :
BNA_M
ClientSampleId :
SSTD005032


Tgt Ion:252 Resp: 244984
Ion Ratio Lower Upper
252 100
253 21.8 17.7 26.5
125 6.4 7.8 11.6#



#91
Benzo(k)fluoranthene
Concen: 4.654 ng/ul
RT: 23.294 min Scan# 3410
Delta R.T. 0.000 min
Lab File: BM034092.D
Acq: 28 Feb 2022 11:35

Tgt Ion:252 Resp: 240986
Ion Ratio Lower Upper
252 100
253 21.0 17.6 26.4
125 6.3 7.4 11.2#









#93
Benzo(a)pyrene
Concen: 4.641 ng/ul
RT: 23.883 min Scan# 31
Delta R.T. -0.006 min
Lab File: BM034092.D
Acq: 28 Feb 2022 11:35




Instrument :
BNA_M
ClientSampleId :
SSTD005032


Tgt Ion:252 Resp: 244436
Ion Ratio Lower Upper
252 100
253 22.0 17.4 26.0
125 6.8 7.8 11.6#



#94
Indeno(1,2,3-cd)pyrene
Concen: 4.811 ng/ul
RT: 26.547 min Scan# 3963
Delta R.T. -0.006 min
Lab File: BM034092.D
Acq: 28 Feb 2022 11:35

Tgt Ion:276 Resp: 281963
Ion Ratio Lower Upper
276 100
138 17.4 16.5 24.7
277 25.7 20.6 31.0









#95
 Dibenzo(a,h)anthracene
 Concen: 4.890 ng/ul
 RT: 26.565 min Scan# 3966
 Delta R.T. -0.006 min
 Lab File: BM034092.D
 Acq: 28 Feb 2022 11:35

Instrument :
 BNA_M
 ClientSampleId :
 SSTD005032

Tgt Ion:278 Resp: 248129
 Ion Ratio Lower Upper
 278 100
 139 11.4 11.0 16.4
 279 22.5 18.1 27.1



#96
 Benzo(g,h,i)perylene
 Concen: 4.900 ng/ul
 RT: 27.329 min Scan# 4096
 Delta R.T. -0.006 min
 Lab File: BM034092.D
 Acq: 28 Feb 2022 11:35

Tgt Ion:276 Resp: 242551
 Ion Ratio Lower Upper
 276 100
 138 13.6 15.6 23.4#
 277 24.3 18.7 28.1

