

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM022822\
 Data File : BM034094.D
 Acq On : 28 Feb 2022 12:47
 Operator : CG/JU
 Sample : SST02034
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SST020034

Quant Time: Feb 28 17:43:53 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM022822.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Feb 28 15:19:40 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.995	152	167673	20.000	ng/ul	0.00
20) Naphthalene-d8	10.813	136	663827	20.000	ng/ul	# 0.00
38) Acenaphthene-d10	14.636	164	482703	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.371	188	1054343	20.000	ng/ul	# 0.00
79) Chrysene-d12	21.547	240	1084092	20.000	ng/ul	# 0.00
88) Perylene-d12	23.994	264	986812	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.372	96	28518	7.842	ng/uL	0.00
4) Pyridine-d5	3.796	84	184996	19.359	ng/ul	0.00
7) Phenol-d5	7.148	99	225656	18.844	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.319	67	128858	19.383	ng/ul	0.00
11) 2-Chlorophenol-d4	7.525	132	202985	19.662	ng/ul	0.00
15) 4-Methylphenol-d8	8.701	113	195741	19.204	ng/ul	0.00
21) Nitrobenzene-d5	9.160	128	100004	19.944	ng/ul	0.00
24) 2-Nitrophenol-d4	9.889	143	113455	19.906	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.425	165	230864	20.015	ng/ul	0.00
31) 4-Chloroaniline-d4	10.942	131	283428	19.485	ng/ul	0.00
46) Dimethylphthalate-d6	14.042	166	722276	19.776	ng/ul	0.00
49) Acenaphthylene-d8	14.330	160	883038	19.852	ng/ul	0.00
54) 4-Nitrophenol-d4	14.807	143	119771	19.146	ng/ul	0.00
60) Fluorene-d10	15.624	176	623518	19.752	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.730	200	126192	18.333	ng/ul	0.00
73) Anthracene-d10	17.471	188	1014041	19.889	ng/ul	0.00
81) Pyrene-d10	19.759	212	1173635	20.083	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.835	264	990134	19.688	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.407	88	27599	7.732	ng/uL#	79
5) Pyridine	3.813	79	188662	19.346	ng/ul#	79
6) Benzaldehyde	7.125	77	153661	22.596	ng/ul#	78
8) Phenol	7.172	94	220498	18.860	ng/ul#	85
10) Bis(2-Chloroethyl)ether	7.413	93	177406	19.264	ng/ul#	85
12) 2-Chlorophenol	7.560	128	202654	19.779	ng/ul#	81
13) 2-Methylphenol	8.436	108	175359	19.086	ng/ul	99
14) 2,2'-oxybis(1-Chloropr...	8.536	45	195727	19.477	ng/ul#	88
16) Acetophenone	8.819	105	306739	19.351	ng/ul#	85
17) N-Nitroso-di-n-propyla...	8.813	70	146087	20.205	ng/ul#	88
18) 4-Methylphenol	8.766	108	191279	19.195	ng/ul	94
19) Hexachloroethane	9.089	117	83515	19.465	ng/ul#	65
22) Nitrobenzene	9.201	77	213290	19.830	ng/ul#	86
23) Isophorone	9.730	82	399826	20.163	ng/ul#	91
25) 2-Nitrophenol	9.919	139	117871	20.031	ng/ul#	75
26) 2,4-Dimethylphenol	9.978	107	235814	20.009	ng/ul	91
27) Bis(2-Chloroethoxy)met...	10.219	93	240855	20.033	ng/ul#	98
29) 2,4-Dichlorophenol	10.454	162	219661	19.829	ng/ul#	91
30) Naphthalene	10.866	128	656983	19.749	ng/ul	99
32) 4-Chloroaniline	10.966	127	280517	19.508	ng/ul	99
33) Hexachlorobutadiene	11.160	225	166978	19.625	ng/ul	99
34) Caprolactam	11.724	113	58759	19.532	ng/ul#	60
35) 4-Chloro-3-methylphenol	12.083	107	210680	20.311	ng/ul#	74
36) 2-Methylnaphthalene	12.472	142	485537	19.994	ng/ul	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM022822\
 Data File : BM034094.D
 Acq On : 28 Feb 2022 12:47
 Operator : CG/JU
 Sample : SST02034
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SST020034


Quant Time: Feb 28 17:43:53 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM022822.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Feb 28 15:19:40 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.689	142	493080	20.030	ng/ul	97
39) 1,2,4,5-Tetrachloroben...	12.836	216	308157	19.216	ng/ul#	97
40) Hexachlorocyclopentadiene	12.824	237	194703	17.804	ng/ul	98
41) 2,4,6-Trichlorophenol	13.071	196	194066	19.811	ng/ul	99
42) 2,4,5-Trichlorophenol	13.136	196	212489	19.723	ng/ul	93
43) 1,1'-Biphenyl	13.471	154	696280	19.569	ng/ul#	96
44) 2-Chloronaphthalene	13.513	162	540425	19.549	ng/ul	95
45) 2-Nitroaniline	13.707	65	124261	19.967	ng/ul#	66
47) Dimethylphthalate	14.089	163	697561	19.730	ng/ul	99
48) 2,6-Dinitrotoluene	14.201	165	139168	20.113	ng/ul#	84
50) Acenaphthylene	14.360	152	866630	19.838	ng/ul	100
51) 3-Nitroaniline	14.524	138	133575	19.848	ng/ul#	74
52) Acenaphthene	14.701	153	571871	19.726	ng/ul	98
53) 2,4-Dinitrophenol	14.730	184	81428	17.740	ng/ul#	71
55) 4-Nitrophenol	14.824	109	111539	19.588	ng/ul#	78
56) Dibenzofuran	15.030	168	841926	19.775	ng/ul	93
57) 2,4-Dinitrotoluene	14.983	165	204472	20.337	ng/ul#	72
58) 2,3,4,6-Tetrachlorophenol	15.254	232	185150	19.860	ng/ul#	88
59) Diethylphthalate	15.448	149	708979	20.086	ng/ul	99
61) Fluorene	15.677	166	704830	19.850	ng/ul	100
62) 4-Chlorophenyl-phenyle...	15.671	204	377214	19.863	ng/ul	89
63) 4-Nitroaniline	15.683	138	143605	20.908	ng/ul#	81
66) 4,6-Dinitro-2-methylph...	15.748	198	124439	18.498	ng/ul#	80
67) N-Nitrosodiphenylamine	15.883	169	607452	19.842	ng/ul	98
68) 4-Bromophenyl-phenylether	16.565	248	227939	19.815	ng/ul#	91
69) Hexachlorobenzene	16.683	284	274234	19.657	ng/ul#	92
70) Atrazine	16.830	200	243846	19.408	ng/ul	96
71) Pentachlorophenol	17.024	266	168033	18.496	ng/ul	96
72) Phenanthrene	17.418	178	1119033	19.567	ng/ul	99
74) Anthracene	17.506	178	1147600	19.815	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.442	216	323181	19.419	ng/ul	96
76) Pentachlorobenzene	14.954	250	321422	19.524	ng/ul	99
77) Carbazole	17.771	167	992142	19.657	ng/ul	99
78) Di-n-butylphthalate	18.342	149	1190817	20.237	ng/ul	98
80) Fluoranthene	19.424	202	1360283	20.146	ng/ul#	91
82) Pyrene	19.789	202	1396233	20.017	ng/ul#	88
83) Butylbenzylphthalate	20.683	149	495302	20.138	ng/ul#	76
84) 3,3'-Dichlorobenzidine	21.459	252	449069	19.541	ng/ul#	95
85) Benzo(a)anthracene	21.530	228	1309341	19.758	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.459	149	750619	19.795	ng/ul	100
87) Chrysene	21.583	228	1274924	19.593	ng/ul	99
89) Di-n-octyl phthalate	22.406	149	1160730	19.113	ng/ul	100
90) Benzo(b)fluoranthene	23.247	252	1259887	20.133	ng/ul#	96
91) Benzo(k)fluoranthene	23.294	252	1178132	19.999	ng/ul#	97
93) Benzo(a)pyrene	23.888	252	1190885	19.877	ng/ul#	98
94) Indeno(1,2,3-cd)pyrene	26.553	276	1305567	19.581	ng/ul	95
95) Dibenzo(a,h)anthracene	26.571	278	1134746	19.660	ng/ul	96
96) Benzo(g,h,i)perylene	27.335	276	1095255	19.450	ng/ul#	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Feb 28 17:43:53 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM022822.M
Quant Title : SVOA CALIBRATION
QLast Update : Mon Feb 28 15:19:40 2022
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 7.732 ng/uL
RT: 3.407 min Scan# 29
Delta R.T. 0.000 min
Lab File: BM034094.D
Acq: 28 Feb 2022 12:47




Instrument :
BNA_M
ClientSampleId :
SSTD020034


Tgt Ion: 88 Resp: 27599
Ion Ratio Lower Upper
88 100
43 32.2 36.2 54.4#
58 65.5 68.5 102.7#



#5
Pyridine
Concen: 19.346 ng/uL
RT: 3.813 min Scan# 98
Delta R.T. 0.000 min
Lab File: BM034094.D
Acq: 28 Feb 2022 12:47

Tgt Ion: 79 Resp: 188662
Ion Ratio Lower Upper
79 100
52 71.6 78.1 117.1#
51 38.9 36.0 54.0









#6
Benzaldehyde
Concen: 22.596 ng/ul
RT: 7.125 min Scan# 661
Delta R.T. 0.000 min
Lab File: BM034094.D
Acq: 28 Feb 2022 12:47




Instrument :
BNA_M
ClientSampleId :
SSTD020034


Tgt Ion: 77 Resp: 153661
Ion Ratio Lower Upper
77 100
105 102.0 65.6 98.4#
106 94.8 60.6 90.8#



#8
Phenol
Concen: 18.860 ng/ul
RT: 7.172 min Scan# 669
Delta R.T. 0.000 min
Lab File: BM034094.D
Acq: 28 Feb 2022 12:47

Tgt Ion: 94 Resp: 220498
Ion Ratio Lower Upper
94 100
65 31.4 31.7 47.5#
66 45.1 45.4 68.0#









#10
Bis(2-Chloroethyl)ether
Concen: 19.264 ng/ul
RT: 7.413 min Scan# 710
Delta R.T. 0.000 min
Lab File: BM034094.D
Acq: 28 Feb 2022 12:47




Instrument :
BNA_M
ClientSampleId :
SSTD020034


Tgt Ion: 93 Resp: 177406
Ion Ratio Lower Upper
93 100
63 71.5 71.7 107.5#
95 31.4 26.3 39.5



#12
2-Chlorophenol
Concen: 19.779 ng/ul
RT: 7.560 min Scan# 735
Delta R.T. 0.000 min
Lab File: BM034094.D
Acq: 28 Feb 2022 12:47

Tgt Ion:128 Resp: 202654
Ion Ratio Lower Upper
128 100
64 40.3 49.4 74.2#
130 31.8 26.8 40.2









#13
2-Methylphenol
Concen: 19.086 ng/ul
RT: 8.436 min Scan# 884
Delta R.T. 0.000 min
Lab File: BM034094.D
Acq: 28 Feb 2022 12:47




Instrument :
BNA_M
ClientSampleId :
SSTD020034


Tgt Ion:108 Resp: 175359
Ion Ratio Lower Upper
108 100
107 86.9 70.2 105.4



#14
2,2'-oxybis(1-Chloropropane)
Concen: 19.477 ng/ul
RT: 8.536 min Scan# 901
Delta R.T. 0.000 min
Lab File: BM034094.D
Acq: 28 Feb 2022 12:47

Tgt Ion: 45 Resp: 195727
Ion Ratio Lower Upper
45 100
77 18.7 9.9 14.9#
79 13.4 8.3 12.5#









#16
Acetophenone
Concen: 19.351 ng/ul
RT: 8.819 min Scan# 949
Delta R.T. 0.000 min
Lab File: BM034094.D
Acq: 28 Feb 2022 12:47




Instrument :
BNA_M
ClientSampleId :
SSTD020034


Tgt Ion:105 Resp: 306739
Ion Ratio Lower Upper
105 100
77 77.3 73.1 109.7
51 28.6 30.5 45.7#



#17
N-Nitroso-di-n-propylamine
Concen: 20.205 ng/ul
RT: 8.813 min Scan# 948
Delta R.T. 0.000 min
Lab File: BM034094.D
Acq: 28 Feb 2022 12:47

Tgt Ion: 70 Resp: 146087
Ion Ratio Lower Upper
70 100
42 61.4 56.9 85.3
101 11.2 6.9 10.3#
130 24.3 13.8 20.8#









#18
4-Methylphenol
Concen: 19.195 ng/ul
RT: 8.766 min Scan# 940
Delta R.T. 0.000 min
Lab File: BM034094.D
Acq: 28 Feb 2022 12:47




Instrument :
BNA_M
ClientSampleId :
SSTD020034


Tgt Ion:108 Resp: 191279
Ion Ratio Lower Upper
108 100
107 115.5 87.4 131.2



#19
Hexachloroethane
Concen: 19.465 ng/ul
RT: 9.089 min Scan# 995
Delta R.T. 0.000 min
Lab File: BM034094.D
Acq: 28 Feb 2022 12:47

Tgt Ion:117 Resp: 83515
Ion Ratio Lower Upper
117 100
201 119.8 69.6 104.4#
199 73.7 40.6 60.8#









#22
Nitrobenzene
Concen: 19.830 ng/ul
RT: 9.201 min Scan# 1014
Delta R.T. 0.000 min
Lab File: BM034094.D
Acq: 28 Feb 2022 12:47




Instrument :
BNA_M
ClientSampleId :
SSTD020034


Tgt Ion: 77 Resp: 213290
Ion Ratio Lower Upper
77 100
123 47.6 29.0 43.6#
65 15.1 11.4 17.2



#23
Isophorone
Concen: 20.163 ng/ul
RT: 9.730 min Scan# 1104
Delta R.T. 0.000 min
Lab File: BM034094.D
Acq: 28 Feb 2022 12:47

Tgt Ion: 82 Resp: 399826
Ion Ratio Lower Upper
82 100
95 9.2 6.5 9.7
138 19.8 11.8 17.8#









#25
2-Nitrophenol
Concen: 20.031 ng/ul
RT: 9.919 min Scan# 1136
Delta R.T. 0.000 min
Lab File: BM034094.D
Acq: 28 Feb 2022 12:47




Instrument :
BNA_M
ClientSampleId :
SSTD020034


Tgt Ion:139 Resp: 117871
Ion Ratio Lower Upper
139 100
65 47.1 59.0 88.6#
109 37.1 37.2 55.8#



#26
2,4-Dimethylphenol
Concen: 20.009 ng/ul
RT: 9.978 min Scan# 1146
Delta R.T. 0.000 min
Lab File: BM034094.D
Acq: 28 Feb 2022 12:47

Tgt Ion:107 Resp: 235814
Ion Ratio Lower Upper
107 100
121 49.7 34.8 52.2
122 82.4 60.0 90.0





#27

Bis(2-Chloroethoxy)methane

Concen: 20.033 ng/ul

RT: 10.219 min Scan# 1187

Delta R.T. 0.000 min

Lab File: BM034094.D

Acq: 28 Feb 2022 12:47

Instrument :

BNA_M

ClientSampleId :

SSTD020034





Tgt Ion: 93 Resp: 240855

Ion Ratio Lower Upper

93 100

95 31.4 25.3 37.9

123 12.4 7.9 11.9#



#29

2,4-Dichlorophenol

Concen: 19.829 ng/ul

RT: 10.454 min Scan# 1227

Delta R.T. 0.000 min

Lab File: BM034094.D

Acq: 28 Feb 2022 12:47




Tgt Ion:162 Resp: 219661


Ion Ratio Lower Upper

162 100

164 66.7 50.0 75.0

98 30.0 30.6 45.8#









#30
Naphthalene
Concen: 19.749 ng/ul
RT: 10.866 min Scan# 1297
Delta R.T. 0.000 min
Lab File: BM034094.D
Acq: 28 Feb 2022 12:47




Instrument :
BNA_M
ClientSampleId :
SSTD020034


Tgt Ion:128 Resp: 656983
Ion Ratio Lower Upper
128 100
129 10.7 9.1 13.7
127 13.3 10.2 15.4



#32
4-Chloroaniline
Concen: 19.508 ng/ul
RT: 10.966 min Scan# 1314
Delta R.T. 0.000 min
Lab File: BM034094.D
Acq: 28 Feb 2022 12:47

Tgt Ion:127 Resp: 280517
Ion Ratio Lower Upper
127 100
129 31.1 25.4 38.0









#33
Hexachlorobutadiene
Concen: 19.625 ng/ul
RT: 11.160 min Scan# 1347
Delta R.T. 0.000 min
Lab File: BM034094.D
Acq: 28 Feb 2022 12:47




Instrument :
BNA_M
ClientSampleId :
SSTD020034


Tgt Ion:225 Resp: 166978
Ion Ratio Lower Upper
225 100
223 66.3 52.2 78.2
227 64.9 52.8 79.2



#34
Caprolactam
Concen: 19.532 ng/ul
RT: 11.724 min Scan# 1443
Delta R.T. 0.000 min
Lab File: BM034094.D
Acq: 28 Feb 2022 12:47

Tgt Ion:113 Resp: 58759
Ion Ratio Lower Upper
113 100
55 135.9 157.9 236.9#
56 111.8 131.8 197.6#









#35
4-Chloro-3-methylphenol
Concen: 20.311 ng/ul
RT: 12.083 min Scan# 1504
Delta R.T. 0.000 min
Lab File: BM034094.D
Acq: 28 Feb 2022 12:47




Instrument :
BNA_M
ClientSampleId :
SSTD020034


Tgt Ion:107 Resp: 210680
Ion Ratio Lower Upper
107 100
144 27.5 17.1 25.7#
142 85.6 49.4 74.2#



#36
2-Methylnaphthalene
Concen: 19.994 ng/ul
RT: 12.472 min Scan# 1570
Delta R.T. 0.000 min
Lab File: BM034094.D
Acq: 28 Feb 2022 12:47

Tgt Ion:142 Resp: 485537
Ion Ratio Lower Upper
142 100
141 89.5 71.1 106.7









#37
1-Methylnaphthalene
Concen: 20.030 ng/ul
RT: 12.689 min Scan# 1607
Delta R.T. 0.000 min
Lab File: BM034094.D
Acq: 28 Feb 2022 12:47




Instrument :
BNA_M
ClientSampleId :
SSTD020034


Tgt Ion:142 Resp: 493080
Ion Ratio Lower Upper
142 100
141 91.8 75.5 113.3
116 4.2 4.0 6.0



#39
1,2,4,5-Tetrachlorobenzene
Concen: 19.216 ng/ul
RT: 12.836 min Scan# 1632
Delta R.T. 0.000 min
Lab File: BM034094.D
Acq: 28 Feb 2022 12:47

Tgt Ion:216 Resp: 308157
Ion Ratio Lower Upper
216 100
214 77.2 61.3 91.9
179 20.1 18.0 27.0
108 15.9 16.6 25.0#









#40
Hexachlorocyclopentadiene
Concen: 17.804 ng/ul
RT: 12.824 min Scan# 1630
Delta R.T. 0.000 min
Lab File: BM034094.D
Acq: 28 Feb 2022 12:47




Instrument :
BNA_M
ClientSampleId :
SSTD020034


Tgt Ion:237 Resp: 194703
Ion Ratio Lower Upper
237 100
235 61.8 50.6 76.0
272 13.8 9.6 14.4



#41
2,4,6-Trichlorophenol
Concen: 19.811 ng/ul
RT: 13.071 min Scan# 1672
Delta R.T. 0.000 min
Lab File: BM034094.D
Acq: 28 Feb 2022 12:47

Tgt Ion:196 Resp: 194066
Ion Ratio Lower Upper
196 100
198 97.4 78.9 118.3
200 30.7 23.7 35.5









#42
2,4,5-Trichlorophenol
Concen: 19.723 ng/ul
RT: 13.136 min Scan# 1
Delta R.T. 0.000 min
Lab File: BM034094.D
Acq: 28 Feb 2022 12:47




Instrument :
BNA_M
ClientSampleId :
SSTD020034


Tgt Ion:196 Resp: 212489
Ion Ratio Lower Upper
196 100
198 95.1 82.2 123.2
200 31.2 27.9 41.9



#43
1,1'-Biphenyl
Concen: 19.569 ng/ul
RT: 13.471 min Scan# 1740
Delta R.T. 0.000 min
Lab File: BM034094.D
Acq: 28 Feb 2022 12:47

Tgt Ion:154 Resp: 696280
Ion Ratio Lower Upper
154 100
153 40.5 33.1 49.7
76 8.9 10.7 16.1#





#44

2-Chloronaphthalene

Concen: 19.549 ng/ul

RT: 13.513 min Scan# 1747

Delta R.T. 0.000 min

Lab File: BM034094.D

Acq: 28 Feb 2022 12:47

Instrument :

BNA_M

ClientSampleId :

SSTD020034





Tgt Ion:162 Resp: 540425

Ion Ratio Lower Upper

162 100

164 32.9 26.2 39.4

127 36.4 33.7 50.5



#45

2-Nitroaniline

Concen: 19.967 ng/ul

RT: 13.707 min Scan# 1780

Delta R.T. 0.000 min

Lab File: BM034094.D

Acq: 28 Feb 2022 12:47




Tgt Ion: 65 Resp: 124261


Ion Ratio Lower Upper

65 100

92 71.0 47.8 71.6

138 118.4 61.0 91.6#









#47
Dimethylphthalate
Concen: 19.730 ng/ul
RT: 14.089 min Scan# 1845
Delta R.T. 0.000 min
Lab File: BM034094.D
Acq: 28 Feb 2022 12:47




Instrument :
BNA_M
ClientSampleId :
SSTD020034


Tgt Ion:163 Resp: 697561
Ion Ratio Lower Upper
163 100
194 4.2 3.3 4.9
164 9.9 8.2 12.2



#48
2,6-Dinitrotoluene
Concen: 20.113 ng/ul
RT: 14.201 min Scan# 1864
Delta R.T. 0.000 min
Lab File: BM034094.D
Acq: 28 Feb 2022 12:47

Tgt Ion:165 Resp: 139168
Ion Ratio Lower Upper
165 100
89 54.4 55.7 83.5#
121 19.8 20.0 30.0#









#50
Acenaphthylene
Concen: 19.838 ng/ul
RT: 14.360 min Scan# 1891
Delta R.T. 0.000 min
Lab File: BM034094.D
Acq: 28 Feb 2022 12:47




Instrument :
BNA_M
ClientSampleId :
SSTD020034


Tgt Ion:152 Resp: 866630
Ion Ratio Lower Upper
152 100
151 19.7 15.8 23.8
153 12.6 10.1 15.1



#51
3-Nitroaniline
Concen: 19.848 ng/ul
RT: 14.524 min Scan# 1919
Delta R.T. 0.000 min
Lab File: BM034094.D
Acq: 28 Feb 2022 12:47

Tgt Ion:138 Resp: 133575
Ion Ratio Lower Upper
138 100
108 9.9 10.8 16.2#
92 105.4 110.6 166.0#









#52
Acenaphthene
Concen: 19.726 ng/ul
RT: 14.701 min Scan# 1949
Delta R.T. 0.000 min
Lab File: BM034094.D
Acq: 28 Feb 2022 12:47




Instrument :
BNA_M
ClientSampleId :
SSTD020034


Tgt Ion:153 Resp: 571871
Ion Ratio Lower Upper
153 100
152 47.8 39.4 59.2
154 88.6 72.7 109.1



#53
2,4-Dinitrophenol
Concen: 17.740 ng/ul
RT: 14.730 min Scan# 1954
Delta R.T. 0.000 min
Lab File: BM034094.D
Acq: 28 Feb 2022 12:47

Tgt Ion:184 Resp: 81428
Ion Ratio Lower Upper
184 100
63 56.6 73.8 110.8#
154 62.6 63.5 95.3#









#55
4-Nitrophenol
Concen: 19.588 ng/ul
RT: 14.824 min Scan# 1970
Delta R.T. 0.000 min
Lab File: BM034094.D
Acq: 28 Feb 2022 12:47




Instrument :
BNA_M
ClientSampleId :
SSTD020034


Tgt Ion:109 Resp: 111539
Ion Ratio Lower Upper
109 100
139 108.9 77.6 116.4
65 104.4 112.2 168.4#



#56
Dibenzofuran
Concen: 19.775 ng/ul
RT: 15.030 min Scan# 2005
Delta R.T. 0.000 min
Lab File: BM034094.D
Acq: 28 Feb 2022 12:47

Tgt Ion:168 Resp: 841926
Ion Ratio Lower Upper
168 100
139 37.9 34.0 51.0





#57

2,4-Dinitrotoluene

Concen: 20.337 ng/ul

RT: 14.983 min Scan# 1997

Delta R.T. 0.000 min

Lab File: BM034094.D

Acq: 28 Feb 2022 12:47

Instrument :

BNA_M

ClientSampleId :

SSTD020034





Tgt Ion:165 Resp: 204472

Ion Ratio Lower Upper

165 100

63 41.3 51.8 77.8#

182 2.9 2.6 4.0



#58

2,3,4,6-Tetrachlorophenol

Concen: 19.860 ng/ul

RT: 15.254 min Scan# 2043

Delta R.T. 0.000 min

Lab File: BM034094.D

Acq: 28 Feb 2022 12:47

Tgt Ion:232 Resp: 185150




Ion Ratio Lower Upper


232 100

131 41.0 41.4 62.2#

130 2.5 2.0 3.0

166 29.3 27.4 41.0









#59
Diethylphthalate
Concen: 20.086 ng/ul
RT: 15.448 min Scan# 2076
Delta R.T. 0.000 min
Lab File: BM034094.D
Acq: 28 Feb 2022 12:47




Instrument :
BNA_M
ClientSampleId :
SSTD020034


Tgt Ion:149 Resp: 708979
Ion Ratio Lower Upper
149 100
177 22.0 17.1 25.7
150 12.4 9.9 14.9



#61
Fluorene
Concen: 19.850 ng/ul
RT: 15.677 min Scan# 2115
Delta R.T. 0.000 min
Lab File: BM034094.D
Acq: 28 Feb 2022 12:47

Tgt Ion:166 Resp: 704830
Ion Ratio Lower Upper
166 100
165 97.6 77.9 116.9
167 13.5 10.3 15.5









#62
4-Chlorophenyl-phenylether
Concen: 19.863 ng/ul
RT: 15.671 min Scan# 2114
Delta R.T. 0.000 min
Lab File: BM034094.D
Acq: 28 Feb 2022 12:47




Instrument :
BNA_M
ClientSampleId :
SSTD020034


Tgt Ion:204 Resp: 377214
Ion Ratio Lower Upper
204 100
206 32.9 27.8 41.8
141 57.0 54.9 82.3



#63
4-Nitroaniline
Concen: 20.908 ng/ul
RT: 15.683 min Scan# 2116
Delta R.T. 0.000 min
Lab File: BM034094.D
Acq: 28 Feb 2022 12:47

Tgt Ion:138 Resp: 143605
Ion Ratio Lower Upper
138 100
92 46.7 48.0 72.0#
108 90.1 89.5 134.3









#66
 4,6-Dinitro-2-methylphenol
 Concen: 18.498 ng/ul
 RT: 15.748 min Scan# 2127
 Delta R.T. 0.000 min
 Lab File: BM034094.D
 Acq: 28 Feb 2022 12:47




Instrument :
 BNA_M
 ClientSampleId :
 SSTD020034


Tgt Ion:198 Resp: 124439
 Ion Ratio Lower Upper
 198 100
 182 4.1 4.2 6.2#
 77 19.8 26.3 39.5#



#67
 N-Nitrosodiphenylamine
 Concen: 19.842 ng/ul
 RT: 15.883 min Scan# 2150
 Delta R.T. 0.000 min
 Lab File: BM034094.D
 Acq: 28 Feb 2022 12:47

Tgt Ion:169 Resp: 607452
 Ion Ratio Lower Upper
 169 100
 168 67.4 52.9 79.3
 167 36.1 29.9 44.9









#68
4-Bromophenyl-phenylether
Concen: 19.815 ng/ul
RT: 16.565 min Scan# 2266
Delta R.T. 0.000 min
Lab File: BM034094.D
Acq: 28 Feb 2022 12:47




Instrument :
BNA_M
ClientSampleId :
SSTD020034


Tgt Ion:248 Resp: 227939
Ion Ratio Lower Upper
248 100
250 98.6 79.0 118.4
141 68.7 69.7 104.5#



#69
Hexachlorobenzene
Concen: 19.657 ng/ul
RT: 16.683 min Scan# 2286
Delta R.T. 0.000 min
Lab File: BM034094.D
Acq: 28 Feb 2022 12:47

Tgt Ion:284 Resp: 274234
Ion Ratio Lower Upper
284 100
142 33.9 34.3 51.5#
249 32.2 26.1 39.1





#70

Atrazine

Concen: 19.408 ng/ul

RT: 16.830 min Scan# 2311

Delta R.T. 0.000 min

Lab File: BM034094.D

Acq: 28 Feb 2022 12:47

Instrument :

BNA_M

ClientSampleId :

SSTD020034





Tgt Ion:200 Resp: 243846

Ion Ratio Lower Upper

200 100

173 24.6 20.6 30.8

215 49.4 36.9 55.3



#71

Pentachlorophenol

Concen: 18.496 ng/ul

RT: 17.024 min Scan# 2344

Delta R.T. 0.000 min

Lab File: BM034094.D

Acq: 28 Feb 2022 12:47




Tgt Ion:266 Resp: 168033


Ion Ratio Lower Upper

266 100

264 63.1 49.1 73.7

268 65.0 47.8 71.8









#72
Phenanthrene
Concen: 19.567 ng/ul
RT: 17.418 min Scan# 2411
Delta R.T. 0.000 min
Lab File: BM034094.D
Acq: 28 Feb 2022 12:47




Instrument :
BNA_M
ClientSampleId :
SSTD020034


Tgt Ion:178 Resp: 1119033
Ion Ratio Lower Upper
178 100
179 15.4 11.7 17.5
176 19.3 15.1 22.7



#74
Anthracene
Concen: 19.815 ng/ul
RT: 17.506 min Scan# 2426
Delta R.T. 0.000 min
Lab File: BM034094.D
Acq: 28 Feb 2022 12:47

Tgt Ion:178 Resp: 1147600
Ion Ratio Lower Upper
178 100
179 15.4 12.3 18.5
176 18.5 14.5 21.7









#75
1,2,3,4-Tetrachlorobenzene
Concen: 19.419 ng/uL
RT: 13.442 min Scan# 1735
Delta R.T. 0.000 min
Lab File: BM034094.D
Acq: 28 Feb 2022 12:47




Instrument :
BNA_M
ClientSampleId :
SSTD020034


Tgt Ion:216 Resp: 323181
Ion Ratio Lower Upper
216 100
214 76.3 65.4 98.0
218 48.1 38.8 58.2



#76
Pentachlorobenzene
Concen: 19.524 ng/uL
RT: 14.954 min Scan# 1992
Delta R.T. 0.000 min
Lab File: BM034094.D
Acq: 28 Feb 2022 12:47

Tgt Ion:250 Resp: 321422
Ion Ratio Lower Upper
250 100
248 64.0 51.6 77.4
252 64.7 52.0 78.0





#77

Carbazole

Concen: 19.657 ng/ul

RT: 17.771 min Scan# 2471

Delta R.T. 0.000 min

Lab File: BM034094.D

Acq: 28 Feb 2022 12:47

Instrument :

BNA_M

ClientSampleId :

SSTD020034





Tgt Ion:167 Resp: 992142

Ion Ratio Lower Upper

167 100

166 21.8 16.9 25.3

139 12.5 10.5 15.7



#78

Di-n-butylphthalate

Concen: 20.237 ng/ul

RT: 18.342 min Scan# 2568

Delta R.T. 0.000 min

Lab File: BM034094.D

Acq: 28 Feb 2022 12:47




Tgt Ion:149 Resp: 1190817


Ion Ratio Lower Upper

149 100

150 9.2 7.0 10.6

104 5.5 5.1 7.7









#80
Fluoranthene
Concen: 20.146 ng/ul
RT: 19.424 min Scan# 21
Delta R.T. 0.000 min
Lab File: BM034094.D
Acq: 28 Feb 2022 12:47




Instrument :
BNA_M
ClientSampleId :
SSTD020034


Tgt Ion:202 Resp: 1360283
Ion Ratio Lower Upper
202 100
101 6.9 8.2 12.4#
100 5.4 6.9 10.3#



#82
Pyrene
Concen: 20.017 ng/ul
RT: 19.789 min Scan# 2814
Delta R.T. 0.000 min
Lab File: BM034094.D
Acq: 28 Feb 2022 12:47

Tgt Ion:202 Resp: 1396233
Ion Ratio Lower Upper
202 100
101 7.4 9.8 14.6#
100 6.2 8.6 13.0#





#83

Butylbenzylphthalate

Concen: 20.138 ng/ul

RT: 20.683 min Scan# 2966

Delta R.T. 0.000 min

Lab File: BM034094.D

Acq: 28 Feb 2022 12:47

Instrument :

BNA_M

ClientSampleId :

SSTD020034





Tgt Ion:149 Resp: 495302

Ion Ratio Lower Upper

149 100

91 61.1 69.3 103.9#

206 23.8 16.1 24.1



#84

3,3'-Dichlorobenzidine

Concen: 19.541 ng/ul

RT: 21.459 min Scan# 3098

Delta R.T. 0.000 min

Lab File: BM034094.D

Acq: 28 Feb 2022 12:47




Tgt Ion:252 Resp: 449069


Ion Ratio Lower Upper

252 100

254 65.8 49.4 74.2

126 8.2 8.2 12.4#









#85
Benzo(a)anthracene
Concen: 19.758 ng/ul
RT: 21.530 min Scan# 3110
Delta R.T. 0.000 min
Lab File: BM034094.D
Acq: 28 Feb 2022 12:47




Instrument :
BNA_M
ClientSampleId :
SSTD020034


Tgt Ion:228 Resp: 1309341
Ion Ratio Lower Upper
228 100
229 19.6 15.0 22.4
226 27.2 21.4 32.0



#86
Bis(2-ethylhexyl)phthalate
Concen: 19.795 ng/ul
RT: 21.459 min Scan# 3098
Delta R.T. 0.000 min
Lab File: BM034094.D
Acq: 28 Feb 2022 12:47

Tgt Ion:149 Resp: 750619
Ion Ratio Lower Upper
149 100
167 27.4 21.8 32.8
279 4.8 3.2 4.8









#87
Chrysene
Concen: 19.593 ng/ul
RT: 21.583 min Scan# 3119
Delta R.T. 0.000 min
Lab File: BM034094.D
Acq: 28 Feb 2022 12:47




Instrument :
BNA_M
ClientSampleId :
SSTD020034


Tgt Ion:228 Resp: 1274924
Ion Ratio Lower Upper
228 100
226 29.9 24.5 36.7
229 20.1 15.9 23.9



#89
Di-n-octyl phthalate
Concen: 19.113 ng/ul
RT: 22.406 min Scan# 3259
Delta R.T. 0.000 min
Lab File: BM034094.D
Acq: 28 Feb 2022 12:47

Tgt Ion:149 Resp: 1160730









#90
Benzo(b)fluoranthene
Concen: 20.133 ng/ul
RT: 23.247 min Scan# 3402
Delta R.T. 0.000 min
Lab File: BM034094.D
Acq: 28 Feb 2022 12:47




Instrument :
BNA_M
ClientSampleId :
SSTD020034


Tgt Ion:252 Resp: 1259887
Ion Ratio Lower Upper
252 100
253 21.4 17.7 26.5
125 6.3 7.8 11.6#



#91
Benzo(k)fluoranthene
Concen: 19.999 ng/ul
RT: 23.294 min Scan# 3410
Delta R.T. 0.000 min
Lab File: BM034094.D
Acq: 28 Feb 2022 12:47

Tgt Ion:252 Resp: 1178132
Ion Ratio Lower Upper
252 100
253 21.6 17.6 26.4
125 6.8 7.4 11.2#





#93

Benzo(a)pyrene

Concen: 19.877 ng/ul

RT: 23.888 min Scan# 3511

Delta R.T. 0.000 min

Lab File: BM034094.D

Acq: 28 Feb 2022 12:47

Instrument :

BNA_M

ClientSampleId :

SSTD020034





Tgt Ion:252 Resp: 1190885

Ion Ratio Lower Upper

252 100

253 21.9 17.4 26.0

125 7.1 7.8 11.6#



#94

Indeno(1,2,3-cd)pyrene

Concen: 19.581 ng/ul

RT: 26.553 min Scan# 3964

Delta R.T. 0.000 min

Lab File: BM034094.D

Acq: 28 Feb 2022 12:47




Tgt Ion:276 Resp: 1305567


Ion Ratio Lower Upper

276 100

138 16.7 16.5 24.7

277 24.6 20.6 31.0









#95
 Dibenzo(a,h)anthracene
 Concen: 19.660 ng/ul
 RT: 26.571 min Scan# 3967
 Delta R.T. 0.000 min
 Lab File: BM034094.D
 Acq: 28 Feb 2022 12:47

Instrument :
 BNA_M
 ClientSampleId :
 SSTD020034

Tgt Ion:278 Resp: 1134746
 Ion Ratio Lower Upper
 278 100
 139 11.4 11.0 16.4
 279 23.6 18.1 27.1



#96
 Benzo(g,h,i)perylene
 Concen: 19.450 ng/ul
 RT: 27.335 min Scan# 4097
 Delta R.T. 0.000 min
 Lab File: BM034094.D
 Acq: 28 Feb 2022 12:47

Tgt Ion:276 Resp: 1095255
 Ion Ratio Lower Upper
 276 100
 138 15.1 15.6 23.4#
 277 24.0 18.7 28.1

