

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM022822\
 Data File : BM034098.D
 Acq On : 28 Feb 2022 15:32
 Operator : CG/JU
 Sample : SSTDICV020
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SICV038

Quant Time: Mar 01 03:50:58 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM022822.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Mar 01 03:48:18 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.001	152	181758	20.000	ng/u1	0.00
20) Naphthalene-d8	10.819	136	731863	20.000	ng/u1	# 0.00
38) Acenaphthene-d10	14.636	164	512052	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.377	188	1089113	20.000	ng/u1	# 0.00
79) Chrysene-d12	21.553	240	1013125	20.000	ng/u1	# 0.00
88) Perylene-d12	24.000	264	947156	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.372	96	30783	7.809	ng/uL	0.00
4) Pyridine-d5	3.796	84	199402	19.250	ng/u1	0.00
7) Phenol-d5	7.148	99	250139	19.269	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.319	67	141256	19.601	ng/u1	0.00
11) 2-Chlorophenol-d4	7.525	132	222299	19.864	ng/u1	0.00
15) 4-Methylphenol-d8	8.701	113	214683	19.430	ng/u1	0.00
21) Nitrobenzene-d5	9.160	128	110257	19.945	ng/u1	0.00
24) 2-Nitrophenol-d4	9.889	143	127779	20.335	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.430	165	249503	19.620	ng/u1	0.00
31) 4-Chloroaniline-d4	10.942	131	307295	19.162	ng/u1	0.00
46) Dimethylphthalate-d6	14.042	166	776146	20.033	ng/u1	0.00
49) Acenaphthylene-d8	14.330	160	949653	20.126	ng/u1	0.00
54) 4-Nitrophenol-d4	14.813	143	128804	19.410	ng/u1	0.00
60) Fluorene-d10	15.624	176	657700	19.641	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.730	200	131147	18.444	ng/u1	0.00
73) Anthracene-d10	17.477	188	1040640	19.759	ng/u1	0.00
81) Pyrene-d10	19.759	212	1145400	20.972	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.847	264	952518	19.734	ng/u1	0.01
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.407	88	29327	7.580	ng/uL#	81
5) Pyridine	3.813	79	207459	19.624	ng/u1#	78
6) Benzaldehyde	7.125	77	169765	23.029	ng/u1#	71
8) Phenol	7.172	94	245679	19.386	ng/u1#	85
10) Bis(2-Chloroethyl)ether	7.413	93	199062	19.941	ng/u1	86
12) 2-Chlorophenol	7.560	128	225481	20.301	ng/u1#	80
13) 2-Methylphenol	8.436	108	194144	19.493	ng/u1	95
14) 2,2'-oxybis(1-Chloropr...	8.531	45	221572	20.340	ng/u1#	91
16) Acetophenone	8.825	105	340335	19.807	ng/u1#	83
17) N-Nitroso-di-n-propyla...	8.813	70	160164	20.435	ng/u1#	86
18) 4-Methylphenol	8.766	108	214581	19.865	ng/u1	98
19) Hexachloroethane	9.095	117	91690	19.714	ng/u1#	62
22) Nitrobenzene	9.201	77	231982	19.563	ng/u1#	84
23) Isophorone	9.736	82	444560	20.335	ng/u1#	90
25) 2-Nitrophenol	9.919	139	131234	20.229	ng/u1#	75
26) 2,4-Dimethylphenol	9.978	107	253450	19.506	ng/u1	89
27) Bis(2-Chloroethoxy)met...	10.219	93	266574	20.111	ng/u1	97
29) 2,4-Dichlorophenol	10.454	162	238161	19.501	ng/u1	94
30) Naphthalene	10.866	128	719837	19.626	ng/u1	99
32) 4-Chloroaniline	10.966	127	302655	19.091	ng/u1	99
33) Hexachlorobutadiene	11.166	225	179727	19.160	ng/u1	98
34) Caprolactam	11.825	113	917	0.276	ng/u1	81
35) 4-Chloro-3-methylphenol	12.089	107	228201	19.955	ng/u1#	72
36) 2-Methylnaphthalene	12.472	142	524299	19.583	ng/u1	100

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM022822\
 Data File : BM034098.D
 Acq On : 28 Feb 2022 15:32
 Operator : CG/JU
 Sample : SSTDICV020
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SICV038

Quant Time: Mar 01 03:50:58 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM022822.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Mar 01 03:48:18 2022
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.689	142	533137	19.644	ng/ul#	98
39) 1,2,4,5-Tetrachloroben...	12.842	216	330851	19.449	ng/ul#	94
40) Hexachlorocyclopentadiene	12.830	237	217499	18.748	ng/ul	99
41) 2,4,6-Trichlorophenol	13.071	196	207829	20.000	ng/ul	98
42) 2,4,5-Trichlorophenol	13.142	196	230168	20.139	ng/ul	94
43) 1,1'-Biphenyl	13.477	154	748753	19.838	ng/ul#	96
44) 2-Chloronaphthalene	13.519	162	581553	19.831	ng/ul	94
45) 2-Nitroaniline	13.707	65	137066	20.763	ng/ul#	68
47) Dimethylphthalate	14.089	163	747872	19.940	ng/ul	100
48) 2,6-Dinitrotoluene	14.201	165	149220	20.330	ng/ul	87
50) Acenaphthylene	14.360	152	921530	19.886	ng/ul	99
51) 3-Nitroaniline	14.530	138	144740	20.274	ng/ul#	68
52) Acenaphthene	14.701	153	606753	19.729	ng/ul	97
53) 2,4-Dinitrophenol	14.730	184	87606	17.992	ng/ul#	75
55) 4-Nitrophenol	14.824	109	115872	19.182	ng/ul#	79
56) Dibenzofuran	15.036	168	892439	19.760	ng/ul	92
57) 2,4-Dinitrotoluene	14.983	165	216833	20.331	ng/ul#	71
58) 2,3,4,6-Tetrachlorophenol	15.254	232	197740	19.994	ng/ul	89
59) Diethylphthalate	15.454	149	741337	19.799	ng/ul	98
61) Fluorene	15.683	166	741281	19.680	ng/ul	100
62) 4-Chlorophenyl-phenyle...	15.677	204	399099	19.811	ng/ul	89
63) 4-Nitroaniline	15.683	138	150003	20.536	ng/ul#	81
66) 4,6-Dinitro-2-methylph...	15.748	198	130432	18.770	ng/ul#	85
67) N-Nitrosodiphenylamine	15.883	169	630106	19.925	ng/ul	98
68) 4-Bromophenyl-phenylether	16.565	248	240248	20.219	ng/ul	90
69) Hexachlorobenzene	16.689	284	285912	19.840	ng/ul#	87
70) Atrazine	16.830	200	250262	19.283	ng/ul	100
71) Pentachlorophenol	17.024	266	174156	18.558	ng/ul	96
72) Phenanthrene	17.418	178	1164757	19.716	ng/ul	99
74) Anthracene	17.512	178	1189517	19.884	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.442	216	345711	20.109	ng/ul	96
76) Pentachlorobenzene	14.960	250	340095	19.999	ng/ul	97
77) Carbazole	17.771	167	1010362	19.378	ng/ul	99
78) Di-n-butylphthalate	18.348	149	1189229	19.564	ng/ul	98
80) Fluoranthene	19.430	202	1338691	21.215	ng/ul#	91
82) Pyrene	19.789	202	1378310	21.144	ng/ul#	90
83) Butylbenzylphthalate	20.689	149	477891	20.791	ng/ul#	75
84) 3,3'-Dichlorobenzidine	21.465	252	443798	20.664	ng/ul#	96
85) Benzo(a)anthracene	21.536	228	1239669	20.017	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.465	149	721971	20.373	ng/ul	99
87) Chrysene	21.589	228	1208880	19.879	ng/ul	99
89) Di-n-octyl phthalate	22.412	149	1081311	18.550	ng/ul	100
90) Benzo(b)fluoranthene	23.253	252	1182451	19.687	ng/ul#	97
91) Benzo(k)fluoranthene	23.300	252	1127341	19.938	ng/ul#	98
93) Benzo(a)pyrene	23.894	252	1136146	19.758	ng/ul#	98
94) Indeno(1,2,3-cd)pyrene	26.565	276	1249264	19.521	ng/ul	96
95) Dibenzo(a,h)anthracene	26.582	278	1086925	19.620	ng/ul#	96
96) Benzo(g,h,i)perylene	27.353	276	1046077	19.354	ng/ul#	95


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM022822\
 Data File : BM034098.D
 Acq On : 28 Feb 2022 15:32
 Operator : CG/JU
 Sample : SSTDICV020
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SICV038

Quant Time: Mar 01 03:50:58 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM022822.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Mar 01 03:48:18 2022
 Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 7.580 ng/uL
RT: 3.407 min Scan# 29
Delta R.T. 0.000 min
Lab File: BM034098.D
Acq: 28 Feb 2022 15:32




Instrument :
BNA_M
ClientSampleId :
SICV038


Tgt Ion: 88 Resp: 29327
Ion Ratio Lower Upper
88 100
43 30.5 36.2 54.4#
58 70.5 68.5 102.7



#5
Pyridine
Concen: 19.624 ng/uL
RT: 3.813 min Scan# 98
Delta R.T. 0.000 min
Lab File: BM034098.D
Acq: 28 Feb 2022 15:32

Tgt Ion: 79 Resp: 207459
Ion Ratio Lower Upper
79 100
52 70.9 78.1 117.1#
51 38.2 36.0 54.0









#6
Benzaldehyde
Concen: 23.029 ng/ul
RT: 7.125 min Scan# 661
Delta R.T. 0.000 min
Lab File: BM034098.D
Acq: 28 Feb 2022 15:32




Instrument :
BNA_M
ClientSampleId :
SICV038


Tgt Ion: 77 Resp: 169765
Ion Ratio Lower Upper
77 100
105 107.7 65.6 98.4#
106 99.9 60.6 90.8#



#8
Phenol
Concen: 19.386 ng/ul
RT: 7.172 min Scan# 669
Delta R.T. 0.000 min
Lab File: BM034098.D
Acq: 28 Feb 2022 15:32

Tgt Ion: 94 Resp: 245679
Ion Ratio Lower Upper
94 100
65 32.0 31.7 47.5
66 43.8 45.4 68.0#





#10

Bis(2-Chloroethyl)ether

Concen: 19.941 ng/ul

RT: 7.413 min Scan# 710

Delta R.T. 0.000 min

Lab File: BM034098.D

Acq: 28 Feb 2022 15:32

Instrument :

BNA_M

ClientSampleId :

SICV038





Tgt Ion: 93 Resp: 199062

Ion Ratio Lower Upper

93 100

63 72.2 71.7 107.5

95 32.1 26.3 39.5



#12

2-Chlorophenol

Concen: 20.301 ng/ul

RT: 7.560 min Scan# 735

Delta R.T. 0.000 min

Lab File: BM034098.D

Acq: 28 Feb 2022 15:32




Tgt Ion:128 Resp: 225481


Ion Ratio Lower Upper

128 100

64 39.0 49.4 74.2#

130 32.0 26.8 40.2









#13
2-Methylphenol
Concen: 19.493 ng/ul
RT: 8.436 min Scan# 884
Delta R.T. 0.000 min
Lab File: BM034098.D
Acq: 28 Feb 2022 15:32




Instrument :
BNA_M
ClientSampleId :
SICV038


Tgt Ion:108 Resp: 194144
Ion Ratio Lower Upper
108 100
107 92.5 70.2 105.4



#14
2,2'-oxybis(1-Chloropropane)
Concen: 20.340 ng/ul
RT: 8.531 min Scan# 900
Delta R.T. -0.006 min
Lab File: BM034098.D
Acq: 28 Feb 2022 15:32

Tgt Ion: 45 Resp: 221572
Ion Ratio Lower Upper
45 100
77 17.6 9.9 14.9#
79 12.1 8.3 12.5









#16
Acetophenone
Concen: 19.807 ng/ul
RT: 8.825 min Scan# 91
Delta R.T. 0.006 min
Lab File: BM034098.D
Acq: 28 Feb 2022 15:32




Instrument :
BNA_M
ClientSampleId :
SICV038


Tgt Ion:105 Resp: 340335
Ion Ratio Lower Upper
105 100
77 75.5 73.1 109.7
51 27.7 30.5 45.7#



#17
N-Nitroso-di-n-propylamine
Concen: 20.435 ng/ul
RT: 8.813 min Scan# 948
Delta R.T. 0.000 min
Lab File: BM034098.D
Acq: 28 Feb 2022 15:32

Tgt Ion: 70 Resp: 160164
Ion Ratio Lower Upper
70 100
42 58.6 56.9 85.3
101 10.1 6.9 10.3
130 24.2 13.8 20.8#









#18
4-Methylphenol
Concen: 19.865 ng/ul
RT: 8.766 min Scan# 940
Delta R.T. 0.000 min
Lab File: BM034098.D
Acq: 28 Feb 2022 15:32




Instrument :
BNA_M
ClientSampleId :
SICV038


Tgt Ion:108 Resp: 214581
Ion Ratio Lower Upper
108 100
107 111.0 87.4 131.2



#19
Hexachloroethane
Concen: 19.714 ng/ul
RT: 9.095 min Scan# 996
Delta R.T. 0.006 min
Lab File: BM034098.D
Acq: 28 Feb 2022 15:32

Tgt Ion:117 Resp: 91690
Ion Ratio Lower Upper
117 100
201 124.0 69.6 104.4#
199 75.3 40.6 60.8#









#22
Nitrobenzene
Concen: 19.563 ng/ul
RT: 9.201 min Scan# 1104
Delta R.T. 0.000 min
Lab File: BM034098.D
Acq: 28 Feb 2022 15:32




Instrument :
BNA_M
ClientSampleId :
SICV038


Tgt Ion: 77 Resp: 231982
Ion Ratio Lower Upper
77 100
123 49.3 29.0 43.6#
65 15.1 11.4 17.2



#23
Isophorone
Concen: 20.335 ng/ul
RT: 9.736 min Scan# 1105
Delta R.T. 0.006 min
Lab File: BM034098.D
Acq: 28 Feb 2022 15:32

Tgt Ion: 82 Resp: 444560
Ion Ratio Lower Upper
82 100
95 8.6 6.5 9.7
138 20.9 11.8 17.8#









#25
2-Nitrophenol
Concen: 20.229 ng/ul
RT: 9.919 min Scan# 1136
Delta R.T. 0.000 min
Lab File: BM034098.D
Acq: 28 Feb 2022 15:32




Instrument :
BNA_M
ClientSampleId :
SICV038


Tgt Ion:139 Resp: 131234
Ion Ratio Lower Upper
139 100
65 48.2 59.0 88.6#
109 35.3 37.2 55.8#



#26
2,4-Dimethylphenol
Concen: 19.506 ng/ul
RT: 9.978 min Scan# 1146
Delta R.T. 0.000 min
Lab File: BM034098.D
Acq: 28 Feb 2022 15:32

Tgt Ion:107 Resp: 253450
Ion Ratio Lower Upper
107 100
121 51.3 34.8 52.2
122 84.2 60.0 90.0





#27

Bis(2-Chloroethoxy)methane

Concen: 20.111 ng/ul

RT: 10.219 min Scan# 1187

Delta R.T. 0.000 min

Lab File: BM034098.D

Acq: 28 Feb 2022 15:32

Instrument :

BNA_M

ClientSampleId :

SICV038





Tgt Ion: 93 Resp: 266574

Ion Ratio Lower Upper

93 100

95 33.0 25.3 37.9

123 11.8 7.9 11.9



#29

2,4-Dichlorophenol

Concen: 19.501 ng/ul

RT: 10.454 min Scan# 1227

Delta R.T. 0.000 min

Lab File: BM034098.D

Acq: 28 Feb 2022 15:32




Tgt Ion:162 Resp: 238161


Ion Ratio Lower Upper

162 100

164 66.4 50.0 75.0

98 33.0 30.6 45.8









#30
Naphthalene
Concen: 19.626 ng/ul
RT: 10.866 min Scan# 1297
Delta R.T. 0.000 min
Lab File: BM034098.D
Acq: 28 Feb 2022 15:32




Instrument :
BNA_M
ClientSampleId :
SICV038


Tgt Ion:128 Resp: 719837
Ion Ratio Lower Upper
128 100
129 11.3 9.1 13.7
127 13.2 10.2 15.4



#32
4-Chloroaniline
Concen: 19.091 ng/ul
RT: 10.966 min Scan# 1314
Delta R.T. 0.000 min
Lab File: BM034098.D
Acq: 28 Feb 2022 15:32

Tgt Ion:127 Resp: 302655
Ion Ratio Lower Upper
127 100
129 32.2 25.4 38.0









#33
Hexachlorobutadiene
Concen: 19.160 ng/ul
RT: 11.166 min Scan# 11
Delta R.T. 0.006 min
Lab File: BM034098.D
Acq: 28 Feb 2022 15:32




Instrument :
BNA_M
ClientSampleId :
SICV038


Tgt Ion:225 Resp: 179727
Ion Ratio Lower Upper
225 100
223 63.1 52.2 78.2
227 64.7 52.8 79.2



#34
Caprolactam
Concen: 0.276 ng/ul
RT: 11.825 min Scan# 1460
Delta R.T. 0.100 min
Lab File: BM034098.D
Acq: 28 Feb 2022 15:32

Tgt Ion:113 Resp: 917
Ion Ratio Lower Upper
113 100
55 161.7 157.9 236.9
56 147.8 131.8 197.6









#35
4-Chloro-3-methylphenol
Concen: 19.955 ng/ul
RT: 12.089 min Scan# 1504
Delta R.T. 0.006 min
Lab File: BM034098.D
Acq: 28 Feb 2022 15:32




Instrument :
BNA_M
ClientSampleId :
SICV038


Tgt Ion:107 Resp: 228201
Ion Ratio Lower Upper
107 100
144 27.6 17.1 25.7#
142 87.1 49.4 74.2#



#36
2-Methylnaphthalene
Concen: 19.583 ng/ul
RT: 12.472 min Scan# 1570
Delta R.T. 0.000 min
Lab File: BM034098.D
Acq: 28 Feb 2022 15:32

Tgt Ion:142 Resp: 524299
Ion Ratio Lower Upper
142 100
141 88.9 71.1 106.7





#37

1-Methylnaphthalene

Concen: 19.644 ng/ul

RT: 12.689 min Scan# 1

Delta R.T. 0.000 min

Lab File: BM034098.D

Acq: 28 Feb 2022 15:32

Instrument :

BNA_M

ClientSampleId :

SICV038





Tgt Ion:142 Resp: 533137

Ion Ratio Lower Upper

142 100

141 92.5 75.5 113.3

116 3.9 4.0 6.0#



#39

1,2,4,5-Tetrachlorobenzene

Concen: 19.449 ng/ul

RT: 12.842 min Scan# 1633

Delta R.T. 0.006 min

Lab File: BM034098.D

Acq: 28 Feb 2022 15:32

Tgt Ion:216 Resp: 330851




Ion Ratio Lower Upper


216 100

214 80.9 61.3 91.9

179 21.0 18.0 27.0

108 14.6 16.6 25.0#









#40
Hexachlorocyclopentadiene
Concen: 18.748 ng/ul
RT: 12.830 min Scan# 1631
Delta R.T. 0.006 min
Lab File: BM034098.D
Acq: 28 Feb 2022 15:32




Instrument :
BNA_M
ClientSampleId :
SICV038


Tgt Ion:237 Resp: 217499
Ion Ratio Lower Upper
237 100
235 62.5 50.6 76.0
272 12.8 9.6 14.4



#41
2,4,6-Trichlorophenol
Concen: 20.000 ng/ul
RT: 13.071 min Scan# 1672
Delta R.T. 0.000 min
Lab File: BM034098.D
Acq: 28 Feb 2022 15:32

Tgt Ion:196 Resp: 207829
Ion Ratio Lower Upper
196 100
198 96.4 78.9 118.3
200 30.5 23.7 35.5









#42
2,4,5-Trichlorophenol
Concen: 20.139 ng/ul
RT: 13.142 min Scan# 1
Delta R.T. 0.006 min
Lab File: BM034098.D
Acq: 28 Feb 2022 15:32




Instrument :
BNA_M
ClientSampleId :
SICV038


Tgt Ion:196 Resp: 230168
Ion Ratio Lower Upper
196 100
198 96.5 82.2 123.2
200 32.0 27.9 41.9



#43
1,1'-Biphenyl
Concen: 19.838 ng/ul
RT: 13.477 min Scan# 1741
Delta R.T. 0.006 min
Lab File: BM034098.D
Acq: 28 Feb 2022 15:32

Tgt Ion:154 Resp: 748753
Ion Ratio Lower Upper
154 100
153 40.8 33.1 49.7
76 8.8 10.7 16.1#





#44

2-Chloronaphthalene

Concen: 19.831 ng/ul

RT: 13.519 min Scan# 1

Delta R.T. 0.006 min

Lab File: BM034098.D

Acq: 28 Feb 2022 15:32

Instrument :

BNA_M

ClientSampleId :

SICV038





Tgt Ion:162 Resp: 581553

Ion Ratio Lower Upper

162 100

164 32.0 26.2 39.4

127 35.9 33.7 50.5



#45

2-Nitroaniline

Concen: 20.763 ng/ul

RT: 13.707 min Scan# 1780

Delta R.T. 0.000 min

Lab File: BM034098.D

Acq: 28 Feb 2022 15:32




Tgt Ion: 65 Resp: 137066


Ion Ratio Lower Upper

65 100

92 69.5 47.8 71.6

138 116.7 61.0 91.6#





#47

Dimethylphthalate

Concen: 19.940 ng/ul

RT: 14.089 min Scan# 1845

Delta R.T. 0.000 min

Lab File: BM034098.D

Acq: 28 Feb 2022 15:32

Instrument :

BNA_M

ClientSampleId :

SICV038





Tgt Ion:163 Resp: 747872

Ion Ratio Lower Upper

163 100

194 4.3 3.3 4.9

164 10.1 8.2 12.2



#48

2,6-Dinitrotoluene

Concen: 20.330 ng/ul

RT: 14.201 min Scan# 1864

Delta R.T. 0.000 min

Lab File: BM034098.D

Acq: 28 Feb 2022 15:32




Tgt Ion:165 Resp: 149220


Ion Ratio Lower Upper

165 100

89 56.8 55.7 83.5

121 21.6 20.0 30.0









#50
Acenaphthylene
Concen: 19.886 ng/ul
RT: 14.360 min Scan# 1891
Delta R.T. 0.000 min
Lab File: BM034098.D
Acq: 28 Feb 2022 15:32




Instrument :
BNA_M
ClientSampleId :
SICV038


Tgt Ion:152 Resp: 921530
Ion Ratio Lower Upper
152 100
151 20.0 15.8 23.8
153 13.0 10.1 15.1



#51
3-Nitroaniline
Concen: 20.274 ng/ul
RT: 14.530 min Scan# 1920
Delta R.T. 0.006 min
Lab File: BM034098.D
Acq: 28 Feb 2022 15:32

Tgt Ion:138 Resp: 144740
Ion Ratio Lower Upper
138 100
108 9.9 10.8 16.2#
92 97.5 110.6 166.0#









#52
Acenaphthene
Concen: 19.729 ng/ul
RT: 14.701 min Scan# 1949
Delta R.T. 0.000 min
Lab File: BM034098.D
Acq: 28 Feb 2022 15:32




Instrument :
BNA_M
ClientSampleId :
SICV038


Tgt Ion:153 Resp: 606753
Ion Ratio Lower Upper
153 100
152 46.8 39.4 59.2
154 87.9 72.7 109.1



#53
2,4-Dinitrophenol
Concen: 17.992 ng/ul
RT: 14.730 min Scan# 1954
Delta R.T. 0.000 min
Lab File: BM034098.D
Acq: 28 Feb 2022 15:32

Tgt Ion:184 Resp: 87606
Ion Ratio Lower Upper
184 100
63 59.0 73.8 110.8#
154 66.9 63.5 95.3









#55
4-Nitrophenol
Concen: 19.182 ng/ul
RT: 14.824 min Scan# 1970
Delta R.T. 0.000 min
Lab File: BM034098.D
Acq: 28 Feb 2022 15:32




Instrument :
BNA_M
ClientSampleId :
SICV038


Tgt Ion:109 Resp: 115872
Ion Ratio Lower Upper
109 100
139 111.5 77.6 116.4
65 110.4 112.2 168.4#



#56
Dibenzofuran
Concen: 19.760 ng/ul
RT: 15.036 min Scan# 2006
Delta R.T. 0.006 min
Lab File: BM034098.D
Acq: 28 Feb 2022 15:32

Tgt Ion:168 Resp: 892439
Ion Ratio Lower Upper
168 100
139 37.5 34.0 51.0









#57
2,4-Dinitrotoluene
Concen: 20.331 ng/ul
RT: 14.983 min Scan# 1997
Delta R.T. 0.000 min
Lab File: BM034098.D
Acq: 28 Feb 2022 15:32




Instrument :
BNA_M
ClientSampleId :
SICV038


Tgt Ion:165 Resp: 216833
Ion Ratio Lower Upper
165 100
63 40.5 51.8 77.8#
182 2.9 2.6 4.0



#58
2,3,4,6-Tetrachlorophenol
Concen: 19.994 ng/ul
RT: 15.254 min Scan# 2043
Delta R.T. 0.000 min
Lab File: BM034098.D
Acq: 28 Feb 2022 15:32

Tgt Ion:232 Resp: 197740
Ion Ratio Lower Upper
232 100
131 41.6 41.4 62.2
130 2.3 2.0 3.0
166 30.3 27.4 41.0





#59

Diethylphthalate

Concen: 19.799 ng/ul

RT: 15.454 min Scan# 2076

Delta R.T. 0.006 min

Lab File: BM034098.D

Acq: 28 Feb 2022 15:32

Instrument :

BNA_M

ClientSampleId :

SICV038





Tgt Ion:149 Resp: 741337

Ion Ratio Lower Upper

149 100

177 22.9 17.1 25.7

150 12.7 9.9 14.9



#61

Fluorene

Concen: 19.680 ng/ul

RT: 15.683 min Scan# 2116

Delta R.T. 0.006 min

Lab File: BM034098.D

Acq: 28 Feb 2022 15:32




Tgt Ion:166 Resp: 741281


Ion Ratio Lower Upper

166 100

165 97.6 77.9 116.9

167 13.4 10.3 15.5









#62
4-Chlorophenyl-phenylether
Concen: 19.811 ng/ul
RT: 15.677 min Scan# 2115
Delta R.T. 0.006 min
Lab File: BM034098.D
Acq: 28 Feb 2022 15:32




Instrument :
BNA_M
ClientSampleId :
SICV038


Tgt Ion:204 Resp: 399099
Ion Ratio Lower Upper
204 100
206 32.5 27.8 41.8
141 56.8 54.9 82.3



#63
4-Nitroaniline
Concen: 20.536 ng/ul
RT: 15.683 min Scan# 2116
Delta R.T. 0.000 min
Lab File: BM034098.D
Acq: 28 Feb 2022 15:32

Tgt Ion:138 Resp: 150003
Ion Ratio Lower Upper
138 100
92 46.4 48.0 72.0#
108 91.2 89.5 134.3









#66
 4,6-Dinitro-2-methylphenol
 Concen: 18.770 ng/ul
 RT: 15.748 min Scan# 2127
 Delta R.T. 0.000 min
 Lab File: BM034098.D
 Acq: 28 Feb 2022 15:32




Instrument :
 BNA_M
 ClientSampleId :
 SICV038


Tgt Ion:198 Resp: 130432
 Ion Ratio Lower Upper
 198 100
 182 3.9 4.2 6.2#
 77 23.6 26.3 39.5#



#67
 N-Nitrosodiphenylamine
 Concen: 19.925 ng/ul
 RT: 15.883 min Scan# 2150
 Delta R.T. 0.000 min
 Lab File: BM034098.D
 Acq: 28 Feb 2022 15:32

Tgt Ion:169 Resp: 630106
 Ion Ratio Lower Upper
 169 100
 168 67.4 52.9 79.3
 167 36.1 29.9 44.9









#68
4-Bromophenyl-phenylether
Concen: 20.219 ng/ul
RT: 16.565 min Scan# 21
Delta R.T. 0.000 min
Lab File: BM034098.D
Acq: 28 Feb 2022 15:32




Instrument :
BNA_M
ClientSampleId :
SICV038


Tgt Ion:248 Resp: 240248
Ion Ratio Lower Upper
248 100
250 96.5 79.0 118.4
141 70.0 69.7 104.5



#69
Hexachlorobenzene
Concen: 19.840 ng/ul
RT: 16.689 min Scan# 2287
Delta R.T. 0.006 min
Lab File: BM034098.D
Acq: 28 Feb 2022 15:32

Tgt Ion:284 Resp: 285912
Ion Ratio Lower Upper
284 100
142 29.9 34.3 51.5#
249 31.2 26.1 39.1





#70

Atrazine

Concen: 19.283 ng/ul

RT: 16.830 min Scan# 21

Delta R.T. 0.000 min

Lab File: BM034098.D

Acq: 28 Feb 2022 15:32

Instrument :

BNA_M

ClientSampleId :

SICV038





Tgt Ion:200 Resp: 250262

Ion Ratio Lower Upper

200 100

173 25.7 20.6 30.8

215 46.2 36.9 55.3



#71

Pentachlorophenol

Concen: 18.558 ng/ul

RT: 17.024 min Scan# 2344

Delta R.T. 0.000 min

Lab File: BM034098.D

Acq: 28 Feb 2022 15:32




Tgt Ion:266 Resp: 174156


Ion Ratio Lower Upper

266 100

264 62.6 49.1 73.7

268 64.8 47.8 71.8









#72
Phenanthrene
Concen: 19.716 ng/ul
RT: 17.418 min Scan# 2411
Delta R.T. 0.000 min
Lab File: BM034098.D
Acq: 28 Feb 2022 15:32




Instrument :
BNA_M
ClientSampleId :
SICV038


Tgt Ion:178 Resp: 1164757
Ion Ratio Lower Upper
178 100
179 15.1 11.7 17.5
176 18.9 15.1 22.7



#74
Anthracene
Concen: 19.884 ng/ul
RT: 17.512 min Scan# 2427
Delta R.T. 0.006 min
Lab File: BM034098.D
Acq: 28 Feb 2022 15:32

Tgt Ion:178 Resp: 1189517
Ion Ratio Lower Upper
178 100
179 15.2 12.3 18.5
176 18.2 14.5 21.7









#75
1,2,3,4-Tetrachlorobenzene
Concen: 20.109 ng/uL
RT: 13.442 min Scan# 11
Delta R.T. 0.000 min
Lab File: BM034098.D
Acq: 28 Feb 2022 15:32




Instrument :
BNA_M
ClientSampleId :
SICV038


Tgt Ion:216 Resp: 345711
Ion Ratio Lower Upper
216 100
214 77.0 65.4 98.0
218 46.8 38.8 58.2



#76
Pentachlorobenzene
Concen: 19.999 ng/uL
RT: 14.960 min Scan# 1993
Delta R.T. 0.006 min
Lab File: BM034098.D
Acq: 28 Feb 2022 15:32

Tgt Ion:250 Resp: 340095
Ion Ratio Lower Upper
250 100
248 62.2 51.6 77.4
252 62.7 52.0 78.0





#77

Carbazole

Concen: 19.378 ng/ul

RT: 17.771 min Scan# 2471

Delta R.T. 0.000 min

Lab File: BM034098.D

Acq: 28 Feb 2022 15:32

Instrument :

BNA_M

ClientSampleId :

SICV038





Tgt Ion:167 Resp: 1010362

Ion Ratio Lower Upper

167 100

166 21.3 16.9 25.3

139 12.7 10.5 15.7



#78

Di-n-butylphthalate

Concen: 19.564 ng/ul

RT: 18.348 min Scan# 2569

Delta R.T. 0.006 min

Lab File: BM034098.D

Acq: 28 Feb 2022 15:32




Tgt Ion:149 Resp: 1189229


Ion Ratio Lower Upper

149 100

150 9.0 7.0 10.6

104 5.3 5.1 7.7









#80
 Fluoranthene
 Concen: 21.215 ng/ul
 RT: 19.430 min Scan# 21
 Delta R.T. 0.006 min
 Lab File: BM034098.D
 Acq: 28 Feb 2022 15:32




Instrument :
 BNA_M
 ClientSampleId :
 SICV038


Tgt Ion:202 Resp: 1338691
 Ion Ratio Lower Upper
 202 100
 101 6.9 8.2 12.4#
 100 5.6 6.9 10.3#



#82
 Pyrene
 Concen: 21.144 ng/ul
 RT: 19.789 min Scan# 2814
 Delta R.T. 0.000 min
 Lab File: BM034098.D
 Acq: 28 Feb 2022 15:32

Tgt Ion:202 Resp: 1378310
 Ion Ratio Lower Upper
 202 100
 101 8.3 9.8 14.6#
 100 7.1 8.6 13.0#









#83
 Butylbenzylphthalate
 Concen: 20.791 ng/ul
 RT: 20.689 min Scan# 2966
 Delta R.T. 0.006 min
 Lab File: BM034098.D
 Acq: 28 Feb 2022 15:32




Instrument :
 BNA_M
 ClientSampleId :
 SICV038


Tgt Ion:149 Resp: 477891
 Ion Ratio Lower Upper
 149 100
 91 59.9 69.3 103.9#
 206 24.1 16.1 24.1



#84
 3,3'-Dichlorobenzidine
 Concen: 20.664 ng/ul
 RT: 21.465 min Scan# 3099
 Delta R.T. 0.006 min
 Lab File: BM034098.D
 Acq: 28 Feb 2022 15:32

Tgt Ion:252 Resp: 443798
 Ion Ratio Lower Upper
 252 100
 254 64.2 49.4 74.2
 126 8.0 8.2 12.4#









#85
Benzo(a)anthracene
Concen: 20.017 ng/ul
RT: 21.536 min Scan# 3111
Delta R.T. 0.006 min
Lab File: BM034098.D
Acq: 28 Feb 2022 15:32




Instrument :
BNA_M
ClientSampleId :
SICV038


Tgt Ion:228 Resp: 1239669
Ion Ratio Lower Upper
228 100
229 19.2 15.0 22.4
226 27.0 21.4 32.0



#86
Bis(2-ethylhexyl)phthalate
Concen: 20.373 ng/ul
RT: 21.465 min Scan# 3099
Delta R.T. 0.006 min
Lab File: BM034098.D
Acq: 28 Feb 2022 15:32

Tgt Ion:149 Resp: 721971
Ion Ratio Lower Upper
149 100
167 27.5 21.8 32.8
279 4.7 3.2 4.8





#87

Chrysene

Concen: 19.879 ng/ul

RT: 21.589 min Scan# 3119

Delta R.T. 0.006 min

Lab File: BM034098.D

Acq: 28 Feb 2022 15:32

Instrument :

BNA_M

ClientSampleId :

SICV038





Tgt Ion:228 Resp: 1208880

Ion Ratio Lower Upper

228 100

226 30.1 24.5 36.7

229 19.6 15.9 23.9



#89

Di-n-octyl phthalate

Concen: 18.550 ng/ul




RT: 22.412 min Scan# 3260


Delta R.T. 0.006 min

Lab File: BM034098.D

Acq: 28 Feb 2022 15:32

Tgt Ion:149 Resp: 1081311





#90

Benzo(b)fluoranthene

Concen: 19.687 ng/ul

RT: 23.253 min Scan# 3402

Delta R.T. 0.006 min

Lab File: BM034098.D

Acq: 28 Feb 2022 15:32

Instrument :

BNA_M

ClientSampleId :

SICV038





Tgt Ion:252 Resp: 1182451

Ion Ratio Lower Upper

252 100

253 22.2 17.7 26.5

125 6.7 7.8 11.6#



#91

Benzo(k)fluoranthene

Concen: 19.938 ng/ul

RT: 23.300 min Scan# 3411

Delta R.T. 0.006 min

Lab File: BM034098.D

Acq: 28 Feb 2022 15:32




Tgt Ion:252 Resp: 1127341


Ion Ratio Lower Upper

252 100

253 22.1 17.6 26.4

125 6.7 7.4 11.2#





#93

Benzo(a)pyrene

Concen: 19.758 ng/ul

RT: 23.894 min Scan# 3

Delta R.T. 0.006 min

Lab File: BM034098.D

Acq: 28 Feb 2022 15:32

Instrument :

BNA_M

ClientSampleId :

SICV038





Tgt Ion:252 Resp: 1136146

Ion Ratio Lower Upper

252 100

253 21.7 17.4 26.0

125 7.2 7.8 11.6#



#94

Indeno(1,2,3-cd)pyrene

Concen: 19.521 ng/ul

RT: 26.565 min Scan# 3966

Delta R.T. 0.012 min

Lab File: BM034098.D

Acq: 28 Feb 2022 15:32




Tgt Ion:276 Resp: 1249264


Ion Ratio Lower Upper

276 100

138 17.0 16.5 24.7

277 25.6 20.6 31.0





#95

Dibenzo(a,h)anthracene

Concen: 19.620 ng/ul

RT: 26.582 min Scan# 3969

Delta R.T. 0.012 min

Lab File: BM034098.D

Acq: 28 Feb 2022 15:32

Instrument :

BNA_M

ClientSampleId :

SICV038





Tgt Ion:278 Resp: 1086925

Ion Ratio Lower Upper

278 100

139 10.5 11.0 16.4#

279 23.4 18.1 27.1



#96

Benzo(g,h,i)perylene

Concen: 19.354 ng/ul

RT: 27.353 min Scan# 4100

Delta R.T. 0.018 min

Lab File: BM034098.D

Acq: 28 Feb 2022 15:32

Tgt Ion:276 Resp: 1046077

Ion Ratio Lower Upper

276 100

138 14.2 15.6 23.4#

277 23.5 18.7 28.1

