

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM022824\
 Data File : BM044438.D
 Acq On : 29 Feb 2024 14:14
 Operator : MA/JU
 Sample : PB159263BL
 Misc :
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SBLK263

Quant Time: Feb 29 23:08:17 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM022324.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Feb 29 03:37:20 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.957	152	204900	20.000	ng/u1	0.00
20) Naphthalene-d8	10.763	136	846643	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.598	164	497686	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.345	188	1004657	20.000	ng/u1	0.00
79) Chrysene-d12	21.539	240	817185	20.000	ng/u1	0.00
88) Perylene-d12	23.962	264	916320	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.369	96	35018	5.392	ng/uL	0.00
4) Pyridine-d5	3.787	84	486408	27.425	ng/u1	0.00
7) Phenol-d5	7.122	99	519291	26.456	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.298	67	338273	26.613	ng/u1	0.00
11) 2-Chlorophenol-d4	7.481	132	418810	28.330	ng/u1	0.00
15) 4-Methylphenol-d8	8.669	113	389940	26.255	ng/u1	0.00
21) Nitrobenzene-d5	9.128	128	198564	28.845	ng/u1	0.00
24) 2-Nitrophenol-d4	9.851	143	206956	29.467	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.386	165	339672	27.417	ng/u1	0.01
31) 4-Chloroaniline-d4	10.916	131	549074	26.389	ng/u1	0.01
46) Dimethylphthalate-d6	14.021	166	1069079	27.852	ng/u1	0.00
49) Acenaphthylene-d8	14.292	160	1259061	28.510	ng/u1	0.00
54) 4-Nitrophenol-d4	14.810	143	147125	19.262	ng/u1	0.01
60) Fluorene-d10	15.592	176	936035	28.812	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.715	200	121221	19.446	ng/u1	0.00
73) Anthracene-d10	17.445	188	1477496	30.843	ng/u1	0.00
81) Pyrene-d10	19.739	212	1640217	33.363	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.803	264	1467735	31.260	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM022824\
Data File : BM044438.D
Acq On : 29 Feb 2024 14:14
Operator : MA/JU
Sample : PB159263BL
Misc :
ALS Vial : 38 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SBLK263

Quant Time: Feb 29 23:08:17 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM022324.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Feb 29 03:37:20 2024
Response via : Initial Calibration

