

Data Path : S:\HPCHEM1\BNA_M\DATA\BM022916\
 Data File : BM004493.D
 Acq On : 01 Mar 2016 06:18
 Operator : SJ/UM
 Sample : H1509-05
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 MW-5

Quant Time: Mar 01 07:41:27 2016
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\8270-SIM-BM022916.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Mar 01 00:52:28 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.62	152	17885	5.00	ng	0.00
7) Naphthalene-d8	10.39	136	70768	5.00	ng	0.00
13) Acenaphthene-d10	14.27	164	39181	5.00	ng	0.00
19) Phenanthrene-d10	17.02	188	93792	5.00	ng	0.00
26) Chrysene-d12	21.24	240	81538	5.00	ng	0.00
35) Perylene-d12	23.45	264	68291	5.00	ng	-0.01
System Monitoring Compounds						
4) 2-Fluorophenol	5.22	112	8885	2.14	ng	0.00
5) Phenol-d6	6.83	99	6051	1.29	ng	0.02
8) Nitrobenzene-d5	8.79	82	13054	3.62	ng	0.00
14) 2,4,6-Tribromophenol	15.78	330	12645	8.40	ng	-0.02
15) 2-Fluorobiphenyl	12.89	172	51835	4.32	ng	0.00
29) Terphenyl-d14	19.66	244	50307	4.79	ng	-0.02
Target Compounds						Qvalue
25) Fluoranthene	19.10	202	1087	0.03	ng	97
28) Pyrene	19.47	202	1868	0.06	ng	# 88
33) Bis(2-ethylhexyl)phthalate	21.13	149	2432	0.10	ng	# 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : S:\HPCHEM1\BNA_M\DATA\BM022916\
Data File : BM004493.D
Acq On : 01 Mar 2016 06:18
Operator : SJ/UM
Sample : H1509-05
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
MW-5

Quant Time: Mar 01 07:41:27 2016
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\8270-SIM-BM022916.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Mar 01 00:52:28 2016
Response via : Initial Calibration

