

Data Path : S:\HPCHEM1\BNA_M\DATA\BM022916\
 Data File : BM004498.D
 Acq On : 01 Mar 2016 09:17
 Operator : SJ/UM
 Sample : H1519-04
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 KY038PS038-160217

Quant Time: Mar 01 13:51:21 2016
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\8270-SIM-BM022916.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Mar 01 00:52:28 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.62	152	20533	5.00	ng	0.00
7) Naphthalene-d8	10.39	136	78394	5.00	ng	0.00
13) Acenaphthene-d10	14.26	164	42329	5.00	ng	0.00
19) Phenanthrene-d10	17.01	188	90520	5.00	ng	0.00
26) Chrysene-d12	21.25	240	82667	5.00	ng	0.01
35) Perylene-d12	23.45	264	81247	5.00	ng	-0.01

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
4) 2-Fluorophenol	5.22	112	8431	1.77	ng	0.00
5) Phenol-d6	6.83	99	5420	1.01	ng	0.02
8) Nitrobenzene-d5	8.78	82	13668	3.43	ng	0.00
14) 2,4,6-Tribromophenol	15.79	330	11575	7.12	ng	0.00
15) 2-Fluorobiphenyl	12.89	172	53104	4.10	ng	0.00
29) Terphenyl-d14	19.67	244	59207	5.56	ng	0.00

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : S:\HPCHEM1\BNA_M\DATA\BM022916\
Data File : BM004498.D
Acq On : 01 Mar 2016 09:17
Operator : SJ/UM
Sample : H1519-04
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
KY038PS038-160217

Quant Time: Mar 01 13:51:21 2016
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\8270-SIM-BM022916.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Mar 01 00:52:28 2016
Response via : Initial Calibration

