

Data Path : S:\HPCHEM1\BNA_M\DATA\BM022916\
 Data File : BM004490.D
 Acq On : 01 Mar 2016 03:47
 Operator : SJ/UM
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC001

Quant Time: Mar 01 07:29:49 2016
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\8270-SIM-BM022916.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Mar 01 00:52:28 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	84	0.00
2	1,4-Dioxane	1.000	0.939	6.1	80	0.00
3	n-Nitrosodimethylamine	1.000	0.871	12.9	78	0.00
4 S	2-Fluorophenol	1.000	0.877	12.3	79	0.00
5 S	Phenol-d6	1.000	0.942	5.8	83	0.02
6	bis(2-Chloroethyl)ether	1.000	0.936	6.4	81	0.00
7 I	Naphthalene-d8	5.000	5.000	0.0	88	0.00
8 S	Nitrobenzene-d5	1.000	0.911	8.9	85	0.00
9	Nitrobenzene	1.000	0.904	9.6	83	0.00
10	Naphthalene	1.000	0.938	6.2	86	0.00
11	Hexachlorobutadiene	1.000	0.966	3.4	87	-0.01
12	2-Methylnaphthalene	1.000	0.995	0.5	92	0.00
13 I	Acenaphthene-d10	5.000	5.000	0.0	92	0.00
14 S	2,4,6-Tribromophenol	1.000	0.862	13.8	89	0.00
15 S	2-Fluorobiphenyl	1.000	0.963	3.7	93	0.00
16	Acenaphthylene	1.000	0.944	5.6	91	0.00
17	Acenaphthene	1.000	0.972	2.8	100	0.00
18	Fluorene	1.000	1.005	-0.5	98	0.00
19 I	Phenanthrene-d10	5.000	5.000	0.0	87	0.00
20	4-Bromophenyl-phenylether	1.000	1.078	-7.8	109	0.00
21	Hexachlorobenzene	1.000	1.088	-8.8	114	0.00
22	Pentachlorophenol	1.000	1.101	-10.1	122	0.00
23	Phenanthrene	1.000	1.031	-3.1	106	0.00
24	Anthracene	1.000	0.887	11.3	88	0.02
25	Fluoranthene	1.000	0.957	4.3	98	0.00
26 I	Chrysene-d12	5.000	5.000	0.0	79	-0.01
27	Benzydine	1.000	0.797	20.3	72	0.00
28	Pyrene	1.000	1.078	-7.8	98	0.00
29 S	Terphenyl-d14	1.000	1.041	-4.1	95	0.00
30	Benzo(a)anthracene	1.000	0.965	3.5	86	-0.01
31	3,3'-Dichlorobenzidine	1.000	1.034	-3.4	99	0.00
32	Chrysene	1.000	0.834	16.6	75	0.00
33	Bis(2-ethylhexyl)phthalate	1.000	0.909	9.1	69	-0.01
34	Indeno(1,2,3-cd)pyrene	1.000	1.042	-4.2	92	-0.01
35 I	Perylene-d12	5.000	5.000	0.0	93	-0.01
36	Benzo(b)fluoranthene	1.000	1.137	-13.7	107	-0.01
37	Benzo(k)fluoranthene	1.000	1.010	-1.0	99	0.00
38 C	Benzo(a)pyrene	1.000	0.989	1.1	95	0.00
39	Dibenzo(a,h)anthracene	1.000	0.956	4.4	93	-0.01
40	Benzo(g,h,i)perylene	1.000	0.960	4.0	93	-0.01

Data Path : S:\HPCHEM1\BNA_M\DATA\BM022916\
Data File : BM004490.D
Acq On : 01 Mar 2016 03:47
Operator : SJ/UM
Sample : SSTDCCC001
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_M
LabSampleId :
SSTDCCC001

Quant Time: Mar 01 07:29:49 2016
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\8270-SIM-BM022916.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Mar 01 00:52:28 2016
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev Area	% Dev(min)

(#) = Out of Range				
SPCC's out = 0 CCC's out = 0				