

Data Path : Z:\HPCHEM1\BNA M\DATA\BM030116\
 Data File : BM004518.D
 Acq On : 02 Mar 2016 13:56
 Operator : SJ/UM
 Sample : H1509-04
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 MW-4

Quant Time: Mar 02 16:38:13 2016
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\8270-SIM-BM030116.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 02 14:46:57 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.60	152	18107	5.00	ng	0.00
7) Naphthalene-d8	10.39	136	71627	5.00	ng	0.00
13) Acenaphthene-d10	14.26	164	38740	5.00	ng	0.00
19) Phenanthrene-d10	17.01	188	91455	5.00	ng	0.00
26) Chrysene-d12	21.23	240	70039	5.00	ng	0.00
35) Perylene-d12	23.45	264	66007	5.00	ng	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
4) 2-Fluorophenol	5.22	112	11051	2.92	ng	-0.02
5) Phenol-d6	6.81	99	7775	1.65	ng	-0.02
8) Nitrobenzene-d5	8.78	82	15336	4.03	ng	-0.01
14) 2,4,6-Tribromophenol	15.77	330	12785	8.94	ng	-0.02
15) 2-Fluorobiphenyl	12.88	172	57376	4.76	ng	0.00
29) Terphenyl-d14	19.67	244	52824	5.09	ng	0.00

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
10) Naphthalene	10.44	128	931	0.05	ng	# 37
12) 2-Methylnaphthalene	12.08	142	220	0.02	ng	# 56
16) Acenaphthylene	13.97	152	192	0.11	ng	# 79
22) Pentachlorophenol	16.73	266	101	0.33	ng	# 87
24) Anthracene	17.16	178	583	0.21	ng	# 92
25) Fluoranthene	19.11	202	4152	0.13	ng	# 95
28) Pyrene	19.47	202	3887	0.13	ng	# 97
33) Bis(2-ethylhexyl)phthalate	21.12	149	7192	0.60	ng	# 94
34) Indeno(1,2,3-cd)pyrene	25.72	276	881	0.04	ng	# 98
40) Benzo(q,h,i)perylene	26.39	276	1142	0.06	ng	# 56

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA M\DATA\BM030116\
Data File : BM004518.D
Acq On : 02 Mar 2016 13:56
Operator : SJ/UM
Sample : H1509-04
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
MW-4

Quant Time: Mar 02 16:38:13 2016
Quant Method : Z:\HPCHEM1\BNA M\METHODS\8270-SIM-BM030116.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Mar 02 14:46:57 2016
Response via : Initial Calibration

