

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM030117\
 Data File : BM009059.D
 Acq On : 01 Mar 2017 20:11
 Operator : SJ/MA
 Sample : I1988-15
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DABS9

Quant Time: Mar 02 03:56:00 2017
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM022317.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Mar 02 03:54:18 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.90	152	175309	20.00	ng/ul	0.00
18) Naphthalene-d8	10.70	136	691583	20.00	ng/ul	0.00
35) Acenaphthene-d10	14.53	164	412640	20.00	ng/ul	0.00
61) Phenanthrene-d10	17.28	188	987093	20.00	ng/ul	0.00
75) Chrysene-d12	21.45	240	1228099	20.00	ng/ul	0.00
83) Perylene-d12	23.80	264	1200321	20.00	ng/ul	0.00

System Monitoring Compounds

3) 1,4-Dioxane-d8	3.36	96	20663	4.73	ng/uL	0.00
5) Phenol-d5	7.05	99	386324	24.31	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	7.23	67	290693	25.54	ng/ul	0.00
9) 2-Chlorophenol-d4	7.42	132	266362	25.16	ng/ul	0.00
13) 4-Methylphenol-d8	8.59	113	261169	22.38	ng/ul	0.00
19) Nitrobenzene-d5	9.06	128	132360	26.75	ng/ul	0.00
22) 2-Nitrophenol-d4	9.78	143	136912	25.64	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.31	165	274080	24.45	ng/ul	0.00
29) 4-Chloroaniline-d4	10.83	131	317750	26.88	ng/ul	0.00
43) Dimethylphthalate-d6	13.94	166	831677	29.94	ng/ul	0.00
46) Acenaphthylene-d8	14.23	160	1026319	30.08	ng/ul	0.00
51) 4-Nitrophenol-d4	14.72	143	113650	23.42	ng/ul	0.00
57) Fluorene-d10	15.53	176	781371	30.12	ng/ul	0.00
62) 4,6-Dinitro-2-methylphenol	15.64	200	105555	17.77	ng/ul	0.00
70) Anthracene-d10	17.38	188	1254045	26.83	ng/ul	0.00
76) Pyrene-d10	19.67	212	1475895	27.57	ng/ul	0.00
87) Benzo(a)pyrene-d12	23.65	264	1467833	26.93	ng/ul	0.00


Target Compounds					Qvalue
44) Dimethylphthalate	13.99	163	500049	17.99 ng/ul	97


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM030117\
Data File : BM009059.D
Acq On : 01 Mar 2017 20:11
Operator : SJ/MA
Sample : I1988-15
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DABS9

Quant Time: Mar 02 03:56:00 2017
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM022317.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Mar 02 03:54:18 2017
Response via : Initial Calibration





#44

Dimethylphthalate

Concen: 17.99 ng/ul

RT: 13.99 min Scan# 1853

Delta R.T. 0.00 min

Lab File: BM009059.D


Acq: 01 Mar 2017 20:11

Instrument :

BNA_M

ClientSampleId :

DABS9



Tgt Ion:163 Resp: 500049

Ion Ratio Lower Upper

163 100

194 4.3 3.0 4.6

164 9.3 8.6 12.8

