

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM030124\
 Data File : BM044480.D
 Acq On : 01 Mar 2024 20:54
 Operator : MA/JU
 Sample : P1496-12
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 BHA05

Quant Time: Mar 01 22:07:09 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM022324.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Mar 01 16:56:24 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.116	152	172513	20.000	ng/u1	0.06
20) Naphthalene-d8	10.951	136	741968	20.000	ng/u1	0.07
38) Acenaphthene-d10	14.780	164	437262	20.000	ng/u1	0.06
64) Phenanthrene-d10	17.527	188	931761	20.000	ng/u1	0.05
79) Chrysene-d12	21.715	240	855701	20.000	ng/u1	0.05
88) Perylene-d12	24.279	264	951349	20.000	ng/u1	0.08
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.428	96	27443	5.019	ng/uL	0.03
4) Pyridine-d5	3.869	84	112184	7.513	ng/u1	0.04
7) Phenol-d5	7.275	99	206630	12.503	ng/u1	0.06
9) Bis-(2-Chloroethyl)eth...	7.451	67	362651	33.887	ng/u1	0.06
11) 2-Chlorophenol-d4	7.633	132	403828	32.444	ng/u1	0.06
15) 4-Methylphenol-d8	8.845	113	300844	24.059	ng/u1	0.07
21) Nitrobenzene-d5	9.304	128	218046	36.144	ng/u1	0.07
24) 2-Nitrophenol-d4	10.033	143	228257	37.085	ng/u1	0.07
28) 2,4-Dichlorophenol-d3	10.574	165	402772	37.097	ng/u1	0.08
31) 4-Chloroaniline-d4	11.110	131	235927	12.938	ng/u1	0.08
46) Dimethylphthalate-d6	14.204	166	1266119	37.544	ng/u1	0.06
49) Acenaphthylene-d8	14.474	160	1521135	39.204	ng/u1	0.06
54) 4-Nitrophenol-d4	14.992	143	66992	9.983	ng/u1	0.06
60) Fluorene-d10	15.768	176	1157734	40.560	ng/u1	0.06
65) 4,6-Dinitro-2-methylph...	15.898	200	170546	29.499	ng/u1	0.06
73) Anthracene-d10	17.627	188	1826033	41.101	ng/u1	0.05
81) Pyrene-d10	19.915	212	2204087	42.814	ng/u1	0.05
92) Benzo(a)pyrene-d12	24.115	264	2037503	41.797	ng/u1	0.08

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM030124\
Data File : BM044480.D
Acq On : 01 Mar 2024 20:54
Operator : MA/JU
Sample : P1496-12
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
BHA05

Quant Time: Mar 01 22:07:09 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM022324.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Mar 01 16:56:24 2024
Response via : Initial Calibration

