

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM030124\
 Data File : BM044493.D
 Acq On : 02 Mar 2024 05:27
 Operator : MA/JU
 Sample : PB159374BL
 Misc :
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SBLK374

Quant Time: Mar 04 01:39:06 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM022324.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Mar 04 01:18:48 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.469	152	227977	20.000	ng/u1	-0.12
20) Naphthalene-d8	11.339	136	881005	20.000	ng/u1	-0.14
38) Acenaphthene-d10	15.092	164	442444	20.000	ng/u1	-0.11
64) Phenanthrene-d10	17.821	188	777653	20.000	ng/u1	-0.09
79) Chrysene-d12	21.956	240	719932	20.000	ng/u1	-0.07
88) Perylene-d12	24.685	264	880672	20.000	ng/u1	-0.12
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.581	96	42341	5.860	ng/uL	-0.05
4) Pyridine-d5	4.046	84	569503	28.860	ng/u1	-0.06
7) Phenol-d5	7.622	99	598373	27.399	ng/u1	-0.12
9) Bis-(2-Chloroethyl)eth...	7.804	67	405934	28.703	ng/u1	-0.12
11) 2-Chlorophenol-d4	7.981	132	492475	29.940	ng/u1	-0.12
15) 4-Methylphenol-d8	9.216	113	432951	26.200	ng/u1	-0.14
21) Nitrobenzene-d5	9.675	128	224821	31.385	ng/u1	-0.13
24) 2-Nitrophenol-d4	10.416	143	231442	31.668	ng/u1	-0.12
28) 2,4-Dichlorophenol-d3	10.957	165	366668	28.442	ng/u1	-0.14
31) 4-Chloroaniline-d4	11.498	131	559331	25.833	ng/u1	-0.13
46) Dimethylphthalate-d6	14.533	166	986512	28.910	ng/u1	-0.11
49) Acenaphthylene-d8	14.792	160	1220754	31.094	ng/u1	-0.11
54) 4-Nitrophenol-d4	15.310	143	126940	18.695	ng/u1	-0.09
60) Fluorene-d10	16.074	176	832912	28.839	ng/u1	-0.10
65) 4,6-Dinitro-2-methylph...	16.198	200	89780	18.606	ng/u1	-0.10
73) Anthracene-d10	17.915	188	1204336	32.480	ng/u1	-0.09
81) Pyrene-d10	20.162	212	1319446	30.464	ng/u1	-0.08
92) Benzo(a)pyrene-d12	24.503	264	1498801	33.213	ng/u1	-0.12

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM030124\
Data File : BM044493.D
Acq On : 02 Mar 2024 05:27
Operator : MA/JU
Sample : PB159374BL
Misc :
ALS Vial : 24 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SBLK374

Quant Time: Mar 04 01:39:06 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM022324.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Mon Mar 04 01:18:48 2024
Response via : Initial Calibration

