

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM030222\
 Data File : BM034114.D
 Acq On : 02 Mar 2022 12:50
 Operator : CG/JU
 Sample : PB142784BL
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SBLK784

Quant Time: Mar 02 13:50:59 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM022822.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Mar 01 03:48:18 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.996	152	132053	20.000	ng/u1	0.00
20) Naphthalene-d8	10.813	136	477574	20.000	ng/u1	# 0.00
38) Acenaphthene-d10	14.637	164	305735	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.372	188	630588	20.000	ng/u1	# 0.00
79) Chrysene-d12	21.542	240	638179	20.000	ng/u1	# 0.00
88) Perylene-d12	23.995	264	699140	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.373	96	16745	5.847	ng/uL	0.00
4) Pyridine-d5	3.790	84	216384	28.752	ng/u1	0.00
7) Phenol-d5	7.143	99	251117	26.626	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.313	67	144231	27.547	ng/u1	0.00
11) 2-Chlorophenol-d4	7.525	132	230539	28.355	ng/u1	0.00
15) 4-Methylphenol-d8	8.696	113	207747	25.880	ng/u1	0.00
21) Nitrobenzene-d5	9.155	128	104132	28.867	ng/u1	0.00
24) 2-Nitrophenol-d4	9.884	143	113431	27.663	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.425	165	220730	26.600	ng/u1	0.00
31) 4-Chloroaniline-d4	10.937	131	279554	26.714	ng/u1	0.00
46) Dimethylphthalate-d6	14.037	166	682462	29.502	ng/u1	0.00
49) Acenaphthylene-d8	14.325	160	807736	28.670	ng/u1	0.00
54) 4-Nitrophenol-d4	14.807	143	95720	24.158	ng/u1	0.00
60) Fluorene-d10	15.625	176	575411	28.779	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.731	200	96117	23.347	ng/u1	0.00
73) Anthracene-d10	17.472	188	910303	29.853	ng/u1	0.00
81) Pyrene-d10	19.760	212	1056260	30.703	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.836	264	1101673	30.920	ng/u1	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM030222\
Data File : BM034114.D
Acq On : 02 Mar 2022 12:50
Operator : CG/JU
Sample : PB142784BL
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SBLK784

Quant Time: Mar 02 13:50:59 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM022822.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Mar 01 03:48:18 2022
Response via : Initial Calibration

