

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM030222\
 Data File : BM034112.D
 Acq On : 02 Mar 2022 10:01
 Operator : CG/JU
 Sample : PB142853BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SBLK853

Quant Time: Mar 02 11:13:23 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM022822.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Mar 01 03:48:18 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.996	152	154967	20.000	ng/u1	0.00
20) Naphthalene-d8	10.807	136	597233	20.000	ng/u1	# 0.00
38) Acenaphthene-d10	14.630	164	405188	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.371	188	827774	20.000	ng/u1	# 0.00
79) Chrysene-d12	21.542	240	728630	20.000	ng/u1	# 0.00
88) Perylene-d12	23.989	264	748086	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.372	96	19275	5.735	ng/uL	0.00
4) Pyridine-d5	3.790	84	254643	28.833	ng/u1	0.00
7) Phenol-d5	7.143	99	298677	26.986	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.313	67	174704	28.433	ng/u1	0.00
11) 2-Chlorophenol-d4	7.519	132	272756	28.587	ng/u1	0.00
15) 4-Methylphenol-d8	8.696	113	251381	26.685	ng/u1	0.00
21) Nitrobenzene-d5	9.154	128	128897	28.573	ng/u1	0.00
24) 2-Nitrophenol-d4	9.884	143	142996	27.886	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.425	165	272848	26.293	ng/u1	0.00
31) 4-Chloroaniline-d4	10.937	131	354267	27.070	ng/u1	0.00
46) Dimethylphthalate-d6	14.036	166	919608	29.996	ng/u1	0.00
49) Acenaphthylene-d8	14.325	160	1063600	28.486	ng/u1	0.00
54) 4-Nitrophenol-d4	14.801	143	125766	23.950	ng/u1	0.00
60) Fluorene-d10	15.619	176	759804	28.674	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.725	200	128877	23.847	ng/u1	0.00
73) Anthracene-d10	17.471	188	1194868	29.851	ng/u1	0.00
81) Pyrene-d10	19.754	212	1312355	33.412	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.836	264	1176749	30.866	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM030222\
Data File : BM034112.D
Acq On : 02 Mar 2022 10:01
Operator : CG/JU
Sample : PB142853BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SBLK853

Quant Time: Mar 02 11:13:23 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM022822.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Mar 01 03:48:18 2022
Response via : Initial Calibration

