

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM030315\
 Data File : BM000621.D
 Acq On : 03 Mar 2015 18:00
 Operator : TP/IZ
 Sample : SSTD00141
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD00141

Quant Time: Mar 04 02:46:22 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM01.2-EPA-SIM-BM030315.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 04 02:34:44 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.67	152	8260	0.40	ng/ul	0.00
4) Naphthalene-d8	10.44	136	29679	0.40	ng/ul	-0.01
8) Acenaphthene-d10	14.31	164	13126	0.40	ng/ul	0.00
12) Phenanthrene-d10	17.05	188	28704	0.40	ng/ul	0.00
18) Chrysene-d12	21.26	240	18058	0.40	ng/ul	0.00
22) Perylene-d12	23.46	264	13666	0.40	ng/ul	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
2) 1,4-Dioxane-d8	3.14	96	39111	6.75	ng/uL	0.00
6) 2-Methylnaphthalene-d10	12.04	152	35848	0.97	ng/ul	0.00
16) Fluoranthene-d10	19.09	212	48017	0.88	ng/ul	0.00

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
3) 1,4-Dioxane	3.17	88	54577	6.01	ng/ul#	82
5) Naphthalene	10.49	128	72897	0.96	ng/ul	99
7) 2-Methylnaphthalene	12.12	142	50275	0.99	ng/ul	100
9) Acenaphthylene	14.03	152	70276	1.20	ng/ul	99
10) Acenaphthene	14.38	153	53275	1.16	ng/ul	98
11) Fluorene	15.36	166	58897	1.10	ng/ul	99
13) Pentachlorophenol	16.71	266	3895	1.37	ng/ul	97
14) Phenanthrene	17.10	178	80460	0.96	ng/ul	99
15) Anthracene	17.19	178	73869	1.02	ng/ul	99
17) Fluoranthene	19.12	202	82461	0.95	ng/ul	99
19) Pyrene	19.49	202	77346	1.00	ng/ul	100
20) Benzo(a)anthracene	21.24	228	51138	0.98	ng/ul	99
21) Chrysene	21.29	228	55426	0.89	ng/ul	99
23) Benzo(b)fluoranthene	22.79	252	47638	1.00	ng/ul	93
24) Benzo(k)fluoranthene	22.83	252	49219	1.00	ng/ul	93
25) Benzo(a)pyrene	23.35	252	45484	1.00	ng/ul#	90
26) Indeno(1,2,3-cd)pyrene	25.67	276	47735	0.95	ng/ul	98
27) Dibenzo(a,h)anthracene	25.69	278	38287	0.97	ng/ul	96
28) Benzo(g,h,i)perylene	26.35	276	41449	0.93	ng/ul	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM030315\
Data File : BM000621.D
Acq On : 03 Mar 2015 18:00
Operator : TP/IZ
Sample : SSTD00141
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSTD00141

Quant Time: Mar 04 02:46:22 2015
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM01.2-EPA-SIM-BM030315.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Mar 04 02:34:44 2015
Response via : Initial Calibration

