

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM030316\
 Data File : BM004540.D
 Acq On : 03 Mar 2016 20:13
 Operator : SJ/UM
 Sample : H1616-01
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 BLDG06-P001-SS001-01

Quant Time: Mar 04 04:42:53 2016
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM030316.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Mar 04 04:25:59 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.59	152	32301	20.00	ng/ul	0.00
18) Naphthalene-d8	10.37	136	130052	20.00	ng/ul	0.00
36) Acenaphthene-d10	14.26	164	65135	20.00	ng/ul	0.00
62) Phenanthrene-d10	17.01	188	133864	20.00	ng/ul	0.00
78) Chrysene-d12	21.23	240	119593	20.00	ng/ul	0.00
86) Perylene-d12	23.48	264	140283	20.00	ng/ul	0.05

System Monitoring Compounds

3) 1,4-Dioxane-d8	3.05	96	2318	3.86	ng/uL	0.00
5) Phenol-d5	6.79	99	50321	19.34	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	6.93	67	27976	21.33	ng/ul	0.00
9) 2-Chlorophenol-d4	7.13	132	48221	20.78	ng/ul	0.00
13) 4-Methylphenol-d8	8.32	113	39185	17.08	ng/ul	0.00
19) Nitrobenzene-d5	8.76	128	22246	22.89	ng/ul	0.00
22) 2-Nitrophenol-d4	9.47	143	24166	21.72	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.02	165	35978	17.87	ng/ul	0.00
29) 4-Chloroaniline-d4	10.54	131	19038	7.69	ng/ul	0.01
44) Dimethylphthalate-d6	13.67	166	119514	22.06	ng/ul	0.00
47) Acenaphthylene-d8	13.94	160	145933	22.28	ng/ul	0.00
52) 4-Nitrophenol-d4	14.54	143	6498	8.13	ng/ul	0.03
58) Fluorene-d10	15.26	176	101469	21.59	ng/ul	0.00
63) 4,6-Dinitro-2-methylphenol	15.42	200	3756	4.32	ng/ul	0.01
71) Anthracene-d10	17.11	188	136641	20.86	ng/ul	0.00
79) Pyrene-d10	19.42	212	143598	22.33	ng/ul	0.00
90) Benzo(a)pyrene-d12	23.35	264	114493	15.34	ng/ul	0.06

Target Compounds

						Qvalue
14) Acetophenone	8.43	105	9438	2.61	ng/ul	94
28) Naphthalene	10.43	128	454891	63.59	ng/ul	100
45) Dimethylphthalate	13.72	163	50522	8.87	ng/ul	99
54) Dibenzofuran	14.66	168	9216	1.37	ng/ul#	67
70) Phenanthrene	17.06	178	16683	2.08	ng/ul	97
77) Fluoranthene	19.09	202	18485	2.14	ng/ul	97
80) Pyrene	19.45	202	23832	2.75	ng/ul	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM030316\
Data File : BM004540.D
Acq On : 03 Mar 2016 20:13
Operator : SJ/UM
Sample : H1616-01
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
BLDG06-P001-SS001-01

Quant Time: Mar 04 04:42:53 2016
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM030316.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Mar 04 04:25:59 2016
Response via : Initial Calibration

