

Data Path : Z:\HPCHEM1\BNA M\DATA\BM030317\
 Data File : BM009071.D
 Acq On : 03 Mar 2017 16:59
 Operator : SJ/MA
 Sample : SSTD3.239
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD3.239

Quant Time: Mar 03 17:30:43 2017
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM-EPA-SIM-BM030317.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 03 16:43:52 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.90	152	127	0.40	ng/ul	0.00
2) Naphthalene-d8	10.69	136	370	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.54	164	218	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.28	188	537	0.40	ng/ul	0.00
16) Chrysene-d12	21.45	240	715	0.40	ng/ul	0.00
20) Perylene-d12	23.78	264	699	0.40	ng/ul	0.00
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	0.00	152	0d	0.00	ng/ul	
14) Fluoranthene-d10	0.00	212	0d	0.00	ng/ul	
Target Compounds						
11) Pentachlorophenol	16.92	266	845	3.25	ng/ul	Ovalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA M\DATA\BM030317\
Data File : BM009071.D
Acq On : 03 Mar 2017 16:59
Operator : SJ/MA
Sample : SSTD3.239
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSTD3.239

Quant Time: Mar 03 17:30:43 2017
Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM-EPA-SIM-BM030317.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 03 16:43:52 2017
Response via : Initial Calibration

