

Data Path : Z:\SVOASRV\HPCHEM1\BNA\_M\DATA\BM030320\  
Quantitation Report (QT Reviewed)

Data File : BM025222.D  
Acq On : 03 Mar 2020 21:13  
Operator : CG/JU  
Sample : L1542-09  
Misc :  
ALS Vial : 18 Sample Multiplier: 1

Instrument :  
BNA\_M  
ClientSampleId :

Quant Time: Mar 04 01:06:32 2020  
Quant Method : Z:\SVOASRV\HPCHEM1\BNA\_M\METHODS\SOM-EPA-SIM-BM030320.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Wed Mar 04 00:52:51 2020  
Response via : Initial Calibration

| Internal Standards        | R.T.  | QIon | Response | Conc | Units | Dev(Min) |
|---------------------------|-------|------|----------|------|-------|----------|
| 1) 1,4-Dichlorobenzene-d4 | 7.68  | 152  | 3750     | 0.40 | ng/ul | 0.00     |
| 2) Naphthalene-d8         | 10.45 | 136  | 15879    | 0.40 | ng/ul | 0.00     |
| 6) Acenaphthene-d10       | 14.31 | 164  | 11691    | 0.40 | ng/ul | 0.00     |
| 10) Phenanthrene-d10      | 17.05 | 188  | 27194m   | 0.40 | ng/ul | 0.00     |
| 16) Chrysene-d12          | 21.23 | 240  | 24326    | 0.40 | ng/ul | 0.00     |
| 20) Perylene-d12          | 23.43 | 264  | 22240    | 0.40 | ng/ul | 0.00     |

| System Monitoring Compounds | R.T.  | QIon | Response | Conc | Units | Dev(Min) |
|-----------------------------|-------|------|----------|------|-------|----------|
| 4) 2-Methylnaphthalene-d10  | 12.04 | 152  | 7395     | 0.25 | ng/ul | 0.00     |
| 14) Fluoranthene-d10        | 19.08 | 212  | 20364    | 0.23 | ng/ul | 0.00     |

| Target Compounds | Qvalue |
|------------------|--------|
|------------------|--------|

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : BM025222.D  
Acq On : 03 Mar 2020 21:13  
Operator : CG/JU  
Sample : L1542-09  
Misc :  
ALS Vial : 18 Sample Multiplier: 1

Instrument :  
BNA\_M  
ClientSampleId :

Quant Time: Mar 04 01:06:32 2020  
Quant Method : Z:\SVOASRV\HPCHEM1\BNA\_M\METHODS\SOM-EPA-SIM-BM030320.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Wed Mar 04 00:52:51 2020  
Response via : Initial Calibration

