

Data File : BM025228.D
Acq On : 04 Mar 2020 00:51
Operator : CG/JU
Sample : L1542-11
Misc :
ALS Vial : 24 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DBD47

Quant Time: Mar 04 04:07:59 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-SIM-BM030320.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Mar 04 00:52:51 2020
Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.68	152	4319	0.40	ng/ul	0.00
2) Naphthalene-d8	10.44	136	18476	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.31	164	13774	0.40	ng/ul	0.00
10) Phenanthrene-d10	0.00	188	0	0.00	ng/ul	-17.05
16) Chrysene-d12	21.23	240	30918	0.40	ng/ul	0.00
20) Perylene-d12	23.42	264	29646	0.40	ng/ul	0.00
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	12.04	152	12686	0.37	ng/ul	0.00
14) Fluoranthene-d10	19.08	212	36532	0.00	ng/ul	0.00
Target Compounds						
17) Pyrene	19.44	202	5731	0.049	ng/ul#	1
23) Benzo(a)pyrene	23.29	252	9814	0.099	ng/ul#	86


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : BM025228.D
Acq On : 04 Mar 2020 00:51
Operator : CG/JU
Sample : L1542-11
Misc :
ALS Vial : 24 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DBD47

Quant Time: Mar 04 04:07:59 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-SIM-BM030320.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Mar 04 00:52:51 2020
Response via : Initial Calibration









#17
Pyrene
Concen: 0.049 ng/ul
RT: 19.44 min Scan# 5253
Delta R.T. -0.03 min
Lab File: BM025228.D
Acq: 04 Mar 2020 00:51

Instrument :
BNA_M
ClientSampleId :
DBD47

Tgt Ion:202 Resp: 5731
Ion Ratio Lower Upper
202 100
101 142.0 8.7 13.1#
100 1004.4 7.3 10.9#



#23
Benzo(a)pyrene
Concen: 0.099 ng/ul
RT: 23.29 min Scan# 6682
Delta R.T. -0.04 min
Lab File: BM025228.D
Acq: 04 Mar 2020 00:51

Tgt Ion:252 Resp: 9814
Ion Ratio Lower Upper
252 100
253 33.1 20.3 30.5#
125 21.7 13.0 19.4#

