

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM030320\
 Data File : BM025233.D
 Acq On : 04 Mar 2020 03:52
 Operator : CG/JU
 Sample : L1542-16
 Misc :
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DBD52

Quant Time: Mar 04 04:50:42 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\SOM-EPA-SIM-BM030320.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 04 00:52:51 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.68	152	6789	0.40	ng/ul	0.00
2) Naphthalene-d8	10.44	136	29576	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.31	164	21436	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.05	188	51650	0.40	ng/ul	0.00
16) Chrysene-d12	21.23	240	49490	0.40	ng/ul	0.00
20) Perylene-d12	23.42	264	47702	0.40	ng/ul	0.00
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	12.04	152	11933	0.22	ng/ul	0.00
14) Fluoranthene-d10	19.08	212	36920	0.22	ng/ul	0.00
Target Compounds				Ovalue		
12) Phenanthrene	17.09	178	4270	0.026	ng/ul	99
15) Fluoranthene	19.11	202	6685	0.032	ng/ul	98
17) Pyrene	19.47	202	5351	0.029	ng/ul	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM030320\
Data File : BM025233.D
Acq On : 04 Mar 2020 03:52
Operator : CG/JU
Sample : L1542-16
Misc :
ALS Vial : 29 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DBD52

Quant Time: Mar 04 04:50:42 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\SOM-EPA-SIM-BM030320.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Mar 04 00:52:51 2020
Response via : Initial Calibration

