

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM030320\
 Data File : BM025246.D
 Acq On : 04 Mar 2020 12:24
 Operator : CG/JU
 Sample : L1542-08
 Misc :
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DBD44

Manual Integrations
 APPROVED

mohammad
 3/5/2020 7:57:04 AM

Quant Time: Mar 04 13:11:22 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\SOM-EPA-SIM-BM030320.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 04 08:02:03 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.68	152	3233	0.40	ng/ul	0.00
2) Naphthalene-d8	10.45	136	14168	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.30	164	10833	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.04	188	25944m	0.40	ng/ul	0.00
16) Chrysene-d12	21.23	240	26015	0.40	ng/ul	0.00
20) Perylene-d12	23.43	264	25848	0.40	ng/ul	0.00
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	12.04	152	11542	0.44	ng/ul	0.00
14) Fluoranthene-d10	19.07	212	34367	0.41	ng/ul	0.00
Target Compounds						
15) Fluoranthene	19.10	202	2475	0.024	ng/ul	93
17) Pyrene	19.47	202	2297	0.023	ng/ul#	91
19) Chrysene	21.27	228	2115	0.020	ng/ul	95
21) Benzo(b)fluoranthene	22.77	252	3016m	0.029	ng/ul	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM030320\
Data File : BM025246.D
Acq On : 04 Mar 2020 12:24
Operator : CG/JU
Sample : L1542-08
Misc :
ALS Vial : 42 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DBD44

Manual Integrations
APPROVED

mohammad
3/5/2020 7:57:04 AM

Quant Time: Mar 04 13:11:22 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\SOM-EPA-SIM-BM030320.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Mar 04 08:02:03 2020
Response via : Initial Calibration

