

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM030516\
 Data File : BM004571.D
 Acq On : 06 Mar 2016 00:07
 Operator : SJ/UM
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC001

Quant Time: Mar 07 08:04:39 2016
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\8270-SIM-BM030516.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Mar 07 07:54:56 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	110	0.00
2	1,4-Dioxane	1.000	0.946	5.4	109	0.00
3	n-Nitrosodimethylamine	1.000	0.866	13.4	106	0.00
4 S	2-Fluorophenol	1.000	0.829	17.1	101	0.02
5 S	Phenol-d6	1.000	0.841	15.9	102	0.00
6	bis(2-Chloroethyl)ether	1.000	0.859	14.1	102	0.02
7 I	Naphthalene-d8	5.000	5.000	0.0	106	0.00
8 S	Nitrobenzene-d5	1.000	0.853	14.7	101	0.00
9	Nitrobenzene	1.000	0.835	16.5	101	0.00
10	Naphthalene	1.000	0.997	0.3	118	0.01
11	Hexachlorobutadiene	1.000	0.917	8.3	108	0.00
12	2-Methylnaphthalene	1.000	0.884	11.6	105	0.00
13 I	Acenaphthene-d10	5.000	5.000	0.0	100	-0.02
14 S	2,4,6-Tribromophenol	1.000	0.864	13.6	96	0.00
15 S	2-Fluorobiphenyl	1.000	0.991	0.9	105	0.00
16	Acenaphthylene	1.000	0.942	5.8	101	0.00
17	Acenaphthene	1.000	1.046	-4.6	112	-0.02
18	Fluorene	1.000	0.923	7.7	99	0.00
19 I	Phenanthrene-d10	5.000	5.000	0.0	90	0.00
20	Hexachlorobenzene	1.000	1.051	-5.1	98	-0.02
21	Pentachlorophenol	1.000	0.611	38.9#	68	0.00
22	Phenanthrene	1.000	1.171	-17.1	111	-0.02
23	Anthracene	1.000	1.072	-7.2	112	-0.02
24	Fluoranthene	1.000	1.069	-6.9	103	0.00
25 I	Chrysene-d12	5.000	5.000	0.0	106	-0.01
26	Benzidine	1.000	1.105	-10.5	161	0.00
27	Pyrene	1.000	0.893	10.7	103	0.00
28 S	Terphenyl-d14	1.000	0.882	11.8	104	0.00
29	Benzo(a)anthracene	1.000	0.887	11.3	104	0.00
30	3,3'-Dichlorobenzidine	1.000	0.884	11.6	99	0.01
31	Chrysene	1.000	0.918	8.2	103	-0.01
32	Bis(2-ethylhexyl)phthalate	1.000	0.862	13.8	105	-0.01
33	Indeno(1,2,3-cd)pyrene	1.000	0.899	10.1	104	0.00
34 I	Perylene-d12	5.000	5.000	0.0	107	0.00
35 C	Benzo(a)pyrene	1.000	0.905	9.5	106	0.00
36	Dibenzo(a,h)anthracene	1.000	0.857	14.3	105	0.00
37	Benzo(g,h,i)perylene	1.000	0.867	13.3	105	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0