


Data Path : Z:\HPCHEM1\BNA_M\Data\BM030518\
 Data File : BM014467.D
 Acq On : 05 Mar 2018 09:18
 Operator : SJ/JU
 Sample : DFTPP08
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP08

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM021418.M
 Title : SVOA CALIBRATION
 Last Update : Mon Mar 05 23:57:03 2018



AutoFind: Scans 2447, 2448, 2449; Background Corrected with Scan 2442

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	51.4	86050	PASS
68	69	0.00	2	1.7	1410	PASS
69	198	0.00	100	50.5	84474	PASS
70	69	0.00	2	0.9	784	PASS
127	198	10	80	54.6	91333	PASS
197	198	0.00	2	0.9	1532	PASS
198	198	100	100	100.0	167402	PASS
199	198	5	9	6.7	11296	PASS
275	198	10	60	29.1	48760	PASS
365	198	1	100	4.8	7956	PASS
441	443	0.01	100	83.4	18374	PASS
442	198	50	100	69.1	115594	PASS
443	442	15	24	19.1	22042	PASS

m/z	Abundance
37.00	378.0
37.90	2135.0
39.10	8634.0
40.00	688.0
41.10	950.0
43.10	784.0
43.90	1722.0
45.00	382.0
49.10	833.0
50.00	30312.0
51.00	93808.0
52.10	4017.0
52.90	457.0
55.00	1563.0
56.10	3041.0
57.00	4249.0
58.00	384.0
61.10	1197.0
62.00	1438.0
63.00	3866.0
64.00	842.0
65.10	2466.0
67.00	1150.0
68.00	2273.0
69.00	88392.0
70.10	1371.0
73.10	1314.0
74.00	13023.0
75.00	15821.0
76.00	4902.0
77.10	97264.0
78.00	7932.0
79.00	9788.0
80.00	6150.0
81.00	9648.0
82.10	1720.0
83.00	1724.0
85.00	1359.0
86.00	2972.0
87.00	1393.0
87.70	504.0
91.10	1795.0
92.10	1779.0
93.00	16600.0
94.10	1301.0
95.00	978.0
96.10	889.0
97.10	1099.0
98.00	14371.0
99.00	8747.0
99.90	836.0
100.90	4080.0
103.00	1160.0
104.00	2550.0
105.00	3020.0
105.80	588.0
107.00	30128.0
108.10	5585.0
109.20	1854.0
110.00	46592.0
111.00	8317.0
111.90	1363.0
115.20	436.0
116.00	2104.0
117.00	43760.0
118.00	2678.0
119.10	930.0
120.10	574.0
121.00	391.0
121.90	1890.0
123.00	4114.0
124.00	1887.0
124.90	936.0
126.10	552.0
127.00	91480.0
128.00	7073.0
129.00	42664.0
130.00	3504.0
130.90	461.0
132.20	307.0
133.20	762.0
133.90	1973.0
135.00	3322.0
136.10	1592.0
137.10	1890.0
140.00	421.0
141.00	5310.0
141.90	2878.0
143.00	1825.0
144.00	462.0
145.00	549.0

Instrument :
BNA_M
ClientSampleId :
DFTPP08

145.90	1913.0
147.00	3653.0
148.00	10664.0
149.00	1615.0
150.10	304.0
151.00	1144.0
151.70	393.0
152.90	2386.0
153.90	1103.0
155.00	3049.0
155.90	3943.0
156.90	901.0
157.80	588.0
158.10	522.0
159.00	1662.0
160.00	2405.0
161.10	3698.0
161.90	1120.0
163.70	300.0
164.90	2407.0
166.10	1734.0
167.00	12724.0
168.00	6324.0
169.10	918.0
171.10	825.0
172.00	1346.0
172.90	1439.0
173.90	2741.0
175.00	4233.0
175.80	1154.0
177.00	2338.0
177.70	689.0
178.90	9200.0
180.00	5283.0
181.20	2197.0
181.90	1142.0
182.80	492.0
183.90	421.0
185.00	4252.0
186.00	24584.0
187.00	9156.0
187.90	950.0
189.00	2537.0
190.10	470.0
191.00	936.0
192.00	3706.0
193.00	3396.0
194.10	970.0
196.00	3592.0
197.10	1515.0
198.00	155456.0
198.90	11500.0
200.00	602.0
202.00	430.0
202.90	1938.0
204.10	8746.0
205.00	13597.0
206.10	46160.0
207.00	8280.0
208.00	2963.0
209.10	675.0
209.90	575.0
210.90	1875.0
211.10	1796.0
211.80	711.0
214.80	1398.0
215.90	1039.0
216.90	15610.0
218.10	2172.0
221.00	6382.0
221.90	2036.0
223.00	3608.0
224.00	28656.0
225.00	8476.0
226.10	775.0
227.00	18920.0
227.90	2847.0
229.10	3311.0
230.00	763.0
230.90	1052.0
233.10	402.0
234.10	845.0
234.90	643.0
235.90	705.0
237.00	1052.0
238.90	501.0
240.90	733.0
242.00	1427.0
243.20	1299.0
244.00	19672.0
244.90	2640.0
245.90	5556.0
247.00	1368.0
249.00	439.0
250.70	394.0

Instrument :
BNA_M
ClientSampleId :
DFTPP08

252.90	774.0
254.10	659.0
255.00	109592.0
256.10	16060.0
256.80	1006.0
258.00	8043.0
259.00	1618.0
261.10	306.0
265.00	3871.0
265.90	394.0
272.20	397.0
273.00	3056.0
274.00	8690.0
275.00	45536.0
276.00	6154.0
277.00	4411.0
277.90	903.0
281.10	540.0
282.00	313.0
283.00	495.0
284.10	415.0
285.10	661.0
288.90	310.0
291.90	375.0
293.00	963.0
293.90	377.0
295.20	409.0
296.00	14519.0
297.10	2543.0
300.90	337.0
303.00	2856.0
314.10	473.0
315.00	1784.0
316.00	417.0
323.10	4314.0
324.10	828.0
326.90	718.0
333.10	661.0
334.10	2446.0
334.90	1001.0
341.00	553.0
345.90	681.0
352.00	1806.0
353.00	1275.0
354.20	1675.0
355.10	610.0
365.00	5970.0
366.10	1008.0
370.90	494.0
372.10	1358.0
383.00	512.0
390.00	472.0
392.10	405.0
403.10	897.0
421.00	819.0
422.00	784.0
423.10	4376.0
423.90	833.0
441.10	12979.0
442.10	84560.0
443.10	16760.0
444.10	2068.0

Instrument :

BNA_M

ClientSampleId :

DFTPP08

m/z	Abundance
37.00	481.0
38.00	2557.0
39.10	9136.0
40.10	1525.0
41.10	792.0
43.00	376.0
44.00	766.0
45.10	442.0
49.10	902.0
50.00	31936.0
51.10	104520.0
52.00	4502.0
53.20	528.0
55.00	1919.0
56.00	2489.0
57.00	5745.0
57.80	335.0
61.10	1402.0
62.10	2137.0
63.00	4797.0
63.80	571.0
65.10	1672.0
67.00	787.0
68.00	1822.0
69.00	104808.0
70.00	427.0
71.00	356.0
72.90	2085.0
74.00	12755.0
75.00	19072.0
76.10	6726.0
77.00	110352.0
78.10	7033.0
79.00	11486.0
80.00	5667.0
81.00	9594.0
81.90	2142.0
82.90	1358.0
85.00	1738.0
86.00	3452.0
87.00	2422.0
88.00	645.0
91.00	2403.0
92.10	2047.0
93.00	16896.0
93.90	1129.0
95.10	642.0
95.80	300.0
98.00	15191.0
99.00	9449.0
99.90	953.0
101.10	5072.0
103.00	1450.0
104.00	3306.0
104.90	2934.0
106.20	879.0
107.00	37936.0
107.90	4950.0
109.10	1288.0
110.00	56016.0
111.00	8344.0
112.10	1703.0
114.80	498.0
116.00	3074.0
117.00	51648.0
118.00	3535.0
119.10	919.0
122.00	3010.0
123.00	4146.0
124.00	2006.0
124.90	1434.0
126.20	436.0
127.00	111256.0
128.00	7500.0
129.00	51520.0
130.00	4403.0
130.90	1002.0
133.00	911.0
133.90	2047.0
135.00	5135.0
135.90	2960.0
137.00	3105.0
137.90	383.0
140.10	1207.0
141.00	8426.0
142.00	2787.0
143.00	2176.0
144.00	363.0
144.90	485.0
145.90	1975.0
147.00	4468.0

Instrument :
BNA_M
ClientSampleId :
DFTPP08

148.00	11256.0
149.00	2507.0
150.00	599.0
151.10	680.0
151.60	618.0
152.90	2484.0
153.90	1799.0
155.00	4853.0
155.90	4569.0
156.90	991.0
158.00	1175.0
159.00	1645.0
160.00	2224.0
161.00	2624.0
162.10	1040.0
162.90	457.0
164.90	2829.0
165.90	2170.0
167.00	17080.0
168.00	7247.0
169.00	1191.0
169.90	560.0
170.90	906.0
171.90	1274.0
172.90	1619.0
174.00	3040.0
175.00	5858.0
176.00	1018.0
176.90	2811.0
179.00	12465.0
180.10	6358.0
180.90	3190.0
181.80	586.0
183.00	424.0
183.90	785.0
185.00	3466.0
186.00	34456.0
187.00	9585.0
188.00	1170.0
189.00	3905.0
190.00	457.0
191.20	1438.0
192.00	3606.0
193.10	4056.0
194.10	904.0
194.90	449.0
196.00	4549.0
196.90	1756.0
198.00	197312.0
199.00	13938.0
200.00	1370.0
201.90	1152.0
202.90	2137.0
204.00	8513.0
205.00	17376.0
206.10	62792.0
207.00	10016.0
208.00	2568.0
209.10	1369.0
209.90	1632.0
211.10	2324.0
211.80	498.0
213.20	371.0
215.10	703.0
215.90	1389.0
217.00	20440.0
218.10	2614.0
220.00	449.0
221.00	5078.0
222.00	1506.0
223.00	4122.0
224.00	36808.0
225.10	8309.0
226.20	1043.0
227.00	21528.0
228.00	3755.0
229.10	4199.0
229.80	582.0
230.80	1108.0
233.00	338.0
234.00	2030.0
235.10	1208.0
235.80	1389.0
236.90	756.0
237.20	702.0
239.10	1334.0
240.00	737.0
241.00	1062.0
242.00	2091.0
243.10	3083.0
244.10	26400.0
245.10	4222.0
246.00	7671.0
246.90	1399.0
248.90	802.0

Instrument :
BNA_M
ClientSampleId :
DFTPP08

250.90	534.0
252.10	482.0
253.00	902.0
254.10	1153.0
255.00	139648.0
256.10	20368.0
257.00	2097.0
258.00	11767.0
258.90	2647.0
260.20	743.0
264.10	457.0
265.00	5395.0
265.90	689.0
267.00	322.0
271.10	417.0
272.10	301.0
273.00	3311.0
274.00	10844.0
275.00	55376.0
276.00	7651.0
277.10	7419.0
277.90	827.0
280.90	555.0
281.10	594.0
283.00	504.0
284.10	828.0
285.10	1187.0
288.90	511.0
289.90	668.0
291.90	360.0
293.00	785.0
294.90	588.0
296.00	23032.0
296.90	3989.0
301.10	368.0
302.90	2107.0
303.80	1236.0
308.20	461.0
314.10	611.0
315.00	2763.0
315.90	1146.0
321.00	1010.0
321.90	379.0
323.10	6303.0
324.10	733.0
326.80	1156.0
327.90	932.0
331.80	432.0
332.90	534.0
334.00	4699.0
334.90	1349.0
340.90	608.0
346.00	1492.0
352.00	1807.0
353.00	1191.0
354.10	1706.0
355.10	409.0
365.00	9633.0
366.00	1569.0
371.00	498.0
372.10	2287.0
373.00	901.0
382.90	560.0
392.20	423.0
402.00	1157.0
403.00	1470.0
403.70	345.0
421.00	806.0
422.00	1383.0
423.10	7389.0
424.00	1721.0
439.20	306.0
441.10	19616.0
442.10	135104.0
443.10	23536.0
444.00	2142.0

Instrument :
BNA_M
ClientSampleId :
DFTPP08

m/z	Abundance
38.00	1162.0
39.00	5413.0
41.10	1415.0
41.80	485.0
43.00	997.0
43.90	1666.0
49.00	650.0
50.10	21568.0
51.00	59824.0
52.00	2843.0
53.90	414.0
55.00	813.0
56.00	2369.0
57.00	4324.0
61.00	901.0
62.00	869.0
63.00	2835.0
63.90	739.0
64.90	766.0
67.00	611.0
68.00	1122.0
69.00	61336.0
70.10	555.0
73.10	2695.0
74.00	9296.0
75.00	11571.0
76.10	4291.0
77.10	75504.0
78.10	5547.0
79.00	6390.0
79.90	3948.0
81.00	5498.0
82.00	996.0
83.00	1081.0
84.90	1241.0
86.00	2354.0
87.10	424.0
87.90	347.0
90.90	1527.0
92.00	1369.0
93.00	12546.0
93.90	1037.0
94.80	511.0
96.10	633.0
96.80	653.0
98.00	11201.0
98.90	6865.0
100.00	414.0
100.90	2089.0
102.90	1096.0
103.90	2037.0
105.00	2441.0
106.10	335.0
107.00	30064.0
108.00	2594.0
108.90	1668.0
110.00	37120.0
111.00	6678.0
112.00	1153.0
115.00	420.0
116.20	1906.0
117.00	38456.0
118.10	2235.0
120.00	360.0
122.00	2662.0
123.00	3445.0
123.90	1231.0
124.80	1057.0
127.00	71264.0
127.90	5655.0
129.00	31960.0
130.00	3295.0
131.00	1163.0
132.90	457.0
133.70	504.0
134.20	929.0
134.90	2920.0
136.00	1165.0
137.00	2168.0
137.90	393.0
139.90	578.0
140.90	5155.0
142.10	1699.0
143.00	1217.0
145.20	513.0
146.10	1501.0
147.00	3057.0
148.00	7362.0
149.00	1318.0
151.20	899.0
153.10	1287.0

Instrument :
BNA_M
ClientSampleId :
DFTPP08

154.00	1412.0
155.00	4515.0
156.00	4009.0
156.80	1007.0
157.80	894.0
158.10	926.0
159.10	426.0
160.10	1983.0
160.80	1586.0
161.80	627.0
163.60	428.0
165.00	3435.0
166.20	1125.0
167.00	11212.0
168.00	4591.0
169.10	1402.0
171.00	674.0
171.90	1347.0
172.80	1177.0
174.10	2267.0
175.00	3141.0
176.00	1368.0
177.10	2135.0
177.90	1293.0
178.90	8397.0
180.00	3785.0
181.00	3341.0
182.00	782.0
183.90	1210.0
185.00	3931.0
186.00	22488.0
187.10	7566.0
188.20	717.0
189.00	1621.0
190.10	380.0
191.00	1657.0
192.00	3539.0
193.00	2693.0
193.80	554.0
194.90	618.0
196.00	5115.0
197.10	1327.0
198.00	149440.0
199.00	8450.0
199.90	1064.0
200.30	963.0
201.80	945.0
203.00	1452.0
204.00	8152.0
205.00	11858.0
206.00	44312.0
207.00	6997.0
207.90	1713.0
209.00	948.0
209.80	495.0
210.30	597.0
211.10	1607.0
211.80	628.0
215.10	1120.0
215.90	874.0
216.90	15477.0
218.00	2601.0
220.10	510.0
221.00	5377.0
221.90	2413.0
223.10	2445.0
224.10	24320.0
225.10	6125.0
225.90	682.0
227.00	16840.0
228.00	2322.0
229.10	2525.0
230.00	501.0
230.90	1571.0
232.90	429.0
233.90	1140.0
235.10	854.0
236.20	1246.0
237.10	1258.0
237.90	319.0
238.90	1181.0
239.70	745.0
241.00	602.0
242.10	1420.0
243.00	1144.0
244.00	18552.0
244.90	2719.0
245.90	6259.0
247.00	1466.0
249.00	772.0
253.00	577.0
254.00	910.0
255.00	106928.0
256.00	13082.0
257.10	1479.0

Instrument :
BNA_M
ClientSampleId :
DFTPP08

258.00	8728.0
259.00	1173.0
261.00	451.0
264.10	328.0
265.10	3501.0
267.10	371.0
270.90	393.0
271.90	537.0
273.10	2204.0
274.00	9021.0
275.00	45368.0
276.10	5870.0
277.00	5889.0
278.00	1233.0
280.90	570.0
283.20	468.0
284.10	579.0
284.90	961.0
290.00	315.0
293.10	840.0
294.10	544.0
295.20	641.0
296.00	17288.0
297.00	3324.0
301.80	413.0
302.10	575.0
303.10	1720.0
304.10	473.0
314.10	1151.0
314.90	1741.0
316.00	1135.0
316.80	403.0
321.00	924.0
323.10	4987.0
323.90	1151.0
327.00	1571.0
328.00	466.0
333.10	424.0
334.00	3566.0
335.10	857.0
341.00	627.0
346.00	1196.0
347.00	361.0
352.00	2109.0
352.90	1076.0
354.10	1344.0
355.10	336.0
359.90	356.0
365.00	8265.0
366.00	1632.0
372.00	2001.0
373.10	598.0
377.00	445.0
382.80	589.0
384.20	381.0
390.00	379.0
390.90	425.0
401.90	493.0
403.10	1221.0
404.10	553.0
421.00	987.0
422.10	832.0
423.00	7754.0
424.00	1776.0
441.10	22528.0
442.10	127120.0
443.10	25832.0
444.10	2263.0

Instrument :
BNA_M
ClientSampleId :
DFTPP08

m/z	Abundance
39.00	357.0
39.90	409.0
41.10	302.0
42.90	506.0
44.00	939.0
52.90	386.0
55.10	685.0
56.90	646.0
67.20	400.0
68.20	329.0
69.10	371.0
71.10	563.0
73.00	1264.0
78.00	451.0
79.00	640.0
81.00	371.0
82.80	364.0
83.30	307.0
91.00	532.0
93.10	354.0
95.00	875.0
97.20	306.0
105.00	500.0
107.10	715.0
108.10	402.0
147.00	333.0
151.30	383.0
193.00	727.0
207.00	1458.0
208.30	306.0
209.00	342.0
251.00	339.0
281.20	523.0
355.10	446.0

Instrument :

BNA_M

ClientSampleId :

DFTPP08