

Data Path : Z:\HPCHEM1\BNA M\DATA\BM030617\
 Data File : BM009095.D
 Acq On : 07 Mar 2017 00:17
 Operator : SJ/MA
 Sample : I2101-10
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SVOC-GPC2-BLANK

Quant Time: Mar 07 04:28:13 2017
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM02.2-EPA-BM022317.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Mar 07 04:22:08 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.89	152	123130	20.00	ng/ul	0.00
18) Naphthalene-d8	10.69	136	447104	20.00	ng/ul	0.00
35) Acenaphthene-d10	14.53	164	244992	20.00	ng/ul	0.00
61) Phenanthrene-d10	17.27	188	547605	20.00	ng/ul	0.00
75) Chrysene-d12	21.44	240	634213	20.00	ng/ul	0.00
83) Perylene-d12	23.78	264	625817	20.00	ng/ul	0.00

System Monitoring Compounds

3) 1,4-Dioxane-d8	0.00	96	0d	0.00	ng/uL
5) Phenol-d5	0.00	99	0	0.00	ng/ul
7) Bis-(2-Chloroethyl)ether-d	0.00	67	0	0.00	ng/ul
9) 2-Chlorophenol-d4	0.00	132	0	0.00	ng/ul
13) 4-Methylphenol-d8	0.00	113	0	0.00	ng/ul
19) Nitrobenzene-d5	0.00	128	0	0.00	ng/ul
22) 2-Nitrophenol-d4	0.00	143	0	0.00	ng/ul
26) 2,4-Dichlorophenol-d3	0.00	165	0	0.00	ng/ul
29) 4-Chloroaniline-d4	0.00	131	0d	0.00	ng/ul
43) Dimethylphthalate-d6	0.00	166	0	0.00	ng/ul
46) Acenaphthylene-d8	0.00	160	0	0.00	ng/ul
51) 4-Nitrophenol-d4	0.00	143	0	0.00	ng/ul
57) Fluorene-d10	0.00	176	0d	0.00	ng/ul
62) 4,6-Dinitro-2-methylphenol	0.00	200	0	0.00	ng/ul
70) Anthracene-d10	0.00	188	0d	0.00	ng/ul
76) Pyrene-d10	0.00	212	0	0.00	ng/ul
87) Benzo(a)pyrene-d12	0.00	264	0d	0.00	ng/ul

Target Compounds Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA M\DATA\BM030617\
Data File : BM009095.D
Acq On : 07 Mar 2017 00:17
Operator : SJ/MA
Sample : I2101-10
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SVOC-GPC2-BLANK

Quant Time: Mar 07 04:28:13 2017
Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM02.2-EPA-BM022317.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Mar 07 04:22:08 2017
Response via : Initial Calibration

