

m/z	Abundance
37.10	416.0
38.00	2307.0
39.00	10198.0
39.90	1042.0
41.00	1463.0
43.00	688.0
44.00	764.0
45.10	737.0
48.90	1094.0
50.00	41576.0
51.10	147328.0
52.00	7183.0
52.90	850.0
55.00	1113.0
56.00	3817.0
57.00	10707.0
58.10	1391.0
59.00	804.0
60.20	522.0
60.90	2374.0
62.10	2174.0
63.10	7512.0
64.00	713.0
65.00	2178.0
67.10	496.0
67.90	2307.0
69.00	143296.0
70.00	892.0
73.00	2778.0
74.00	15288.0
75.00	25392.0
76.10	8130.0
77.10	160000.0
78.10	10165.0
79.00	11288.0
80.00	8079.0
81.00	12039.0
81.90	3815.0
83.10	2428.0
84.00	634.0
85.10	2544.0
86.00	3784.0
87.10	1775.0
88.10	514.0
89.00	481.0
91.00	3603.0
91.90	4254.0
93.00	20280.0
94.00	528.0
94.90	686.0
96.10	1474.0
97.10	475.0
98.00	16129.0
99.10	11813.0
99.90	2018.0
101.00	6284.0
102.10	465.0
103.00	2289.0
104.00	4456.0
104.90	4505.0
106.00	1610.0
107.00	46432.0
108.00	6880.0
109.00	1401.0
110.00	70784.0
111.00	13096.0
111.90	1450.0
112.90	523.0
115.00	752.0
116.00	2583.0
117.00	48952.0
118.00	3768.0
118.90	416.0
120.00	910.0
121.00	548.0
121.90	3536.0
123.00	4315.0
124.00	2187.0
125.00	1425.0
126.10	816.0
127.00	119744.0
128.00	8248.0
129.00	53576.0
129.90	4736.0
131.00	791.0
131.90	445.0
133.90	2131.0
135.00	4601.0
135.90	2226.0
137.00	3499.0
137.80	446.0

Instrument :
BNA_M
ClientSampleId :
DFTPP43

140.00	399.0
141.00	8513.0
141.90	2617.0
142.90	1380.0
144.00	587.0
145.90	2167.0
147.00	5172.0
148.00	11307.0
148.90	2493.0
149.80	852.0
150.90	638.0
151.40	419.0
151.80	560.0
152.30	407.0
153.00	2274.0
154.10	1613.0
154.90	4969.0
156.00	6151.0
157.00	1364.0
158.00	1667.0
159.20	863.0
159.90	3117.0
161.10	3540.0
161.90	1589.0
163.90	506.0
165.10	3336.0
166.00	2022.0
167.00	16464.0
168.00	10329.0
169.10	1728.0
169.80	696.0
170.90	1306.0
171.90	2264.0
172.90	1993.0
174.00	2774.0
175.00	5064.0
176.00	1196.0
177.00	2276.0
178.10	1722.0
178.90	11996.0
180.00	8886.0
181.00	3885.0
181.90	610.0
184.00	1458.0
185.00	6515.0
186.00	41112.0
187.00	12586.0
187.90	2271.0
188.90	2775.0
189.90	689.0
191.00	1832.0
191.90	4016.0
193.00	4733.0
194.10	1362.0
195.90	5299.0
198.00	254144.0
198.90	19040.0
200.10	1291.0
201.30	1024.0
202.90	3079.0
204.10	9763.0
205.00	19224.0
206.00	72856.0
207.10	10301.0
208.00	3914.0
209.10	805.0
210.20	1680.0
211.00	2566.0
211.70	1182.0
213.00	537.0
214.90	393.0
216.00	1938.0
216.90	21560.0
217.90	1989.0
221.10	9279.0
222.10	2194.0
223.00	5102.0
224.00	43648.0
225.00	8922.0
226.10	1161.0
227.00	22480.0
227.90	3871.0
228.90	4035.0
230.00	388.0
230.90	1240.0
231.80	389.0
233.00	815.0
233.90	1211.0
234.90	1787.0
236.00	1576.0
237.00	1804.0
239.20	696.0
239.90	909.0
241.00	1556.0
242.20	1414.0

Instrument :
BNA_M
ClientSampleId :
DFTPP43

243.10	2457.0
244.00	32728.0
245.00	5065.0
246.00	8169.0
246.90	1759.0
248.00	331.0
249.00	452.0
251.90	525.0
253.00	1317.0
254.10	1881.0
255.00	160576.0
256.00	25312.0
257.00	2527.0
258.00	11463.0
258.90	2758.0
259.90	439.0
261.00	692.0
264.90	4847.0
266.00	735.0
266.80	506.0
270.00	436.0
270.80	565.0
271.80	727.0
273.00	5306.0
273.90	11743.0
275.00	62144.0
276.00	8448.0
277.00	7541.0
278.00	1441.0
278.90	491.0
280.90	641.0
281.90	531.0
283.00	1175.0
283.80	618.0
284.90	1112.0
291.10	329.0
292.00	614.0
293.00	1985.0
293.70	658.0
294.90	485.0
296.00	23488.0
297.00	4193.0
300.90	384.0
301.80	887.0
303.00	2236.0
304.10	913.0
309.90	775.0
312.90	412.0
314.00	961.0
315.00	2289.0
315.90	713.0
320.90	959.0
323.10	5354.0
323.90	782.0
326.60	728.0
327.10	585.0
328.10	792.0
333.00	785.0
334.00	3446.0
334.90	1404.0
341.00	617.0
342.00	343.0
345.90	1358.0
347.00	307.0
352.00	1396.0
353.10	751.0
354.10	1363.0
355.00	567.0
356.10	407.0
365.00	8923.0
365.90	1151.0
370.80	461.0
372.00	2599.0
373.20	461.0
383.10	674.0
384.20	614.0
402.00	1241.0
403.00	1292.0
404.00	429.0
415.20	549.0
421.00	791.0
422.00	1305.0
423.10	6915.0
424.00	3861.0
425.10	441.0
439.40	302.0
441.00	20912.0
442.10	133376.0
443.10	27256.0
444.00	2617.0

Instrument :
BNA_M
ClientSampleId :
DFTPP43

m/z	Abundance
36.90	688.0
38.00	1306.0
39.10	12208.0
40.10	1713.0
41.00	1012.0
43.00	967.0
44.00	665.0
44.90	1053.0
49.10	1113.0
50.00	43096.0
51.00	161344.0
52.00	9496.0
53.00	344.0
55.00	2112.0
55.90	4699.0
57.00	9288.0
58.00	1103.0
61.00	3063.0
62.10	2340.0
63.00	7358.0
63.90	600.0
64.20	596.0
65.00	3559.0
67.00	795.0
68.10	1627.0
69.00	165184.0
70.10	1026.0
71.00	526.0
73.10	4688.0
74.00	18768.0
75.00	23920.0
76.10	10133.0
77.00	168512.0
78.10	12582.0
79.00	12070.0
80.00	7734.0
81.00	11721.0
82.00	3280.0
83.00	1993.0
84.10	741.0
85.00	1719.0
86.00	4015.0
87.00	2855.0
88.10	366.0
91.00	2282.0
92.00	4101.0
93.00	21232.0
93.90	1665.0
95.10	691.0
95.90	1176.0
97.00	981.0
98.00	15293.0
99.00	13599.0
99.90	801.0
100.20	740.0
101.00	6682.0
102.00	463.0
102.90	2190.0
104.10	4288.0
105.00	4929.0
106.00	1023.0
107.00	56536.0
108.00	7601.0
109.00	1378.0
110.00	81240.0
111.00	13656.0
111.90	1600.0
112.90	527.0
114.80	455.0
116.00	2959.0
116.90	62968.0
117.90	4313.0
119.00	796.0
120.10	460.0
121.00	438.0
121.90	3739.0
122.90	6067.0
124.00	2520.0
125.00	3042.0
126.10	919.0
127.00	147008.0
128.00	10678.0
129.00	62672.0
130.10	3926.0
131.00	1150.0
131.90	912.0
133.00	697.0
133.90	2201.0
135.00	5152.0
136.00	2145.0
137.00	3971.0

Instrument :
BNA_M
ClientSampleId :
DFTPP43

137.90	890.0
138.80	417.0
140.10	1323.0
141.00	8037.0
141.90	2599.0
143.00	1962.0
144.00	580.0
144.90	675.0
145.90	1481.0
146.90	5262.0
148.00	12596.0
149.00	2914.0
149.90	406.0
151.00	1074.0
152.20	957.0
153.00	2593.0
154.10	1697.0
155.00	4718.0
156.10	9470.0
156.80	1573.0
157.30	1299.0
158.00	1593.0
159.10	1695.0
160.00	3486.0
161.00	4302.0
161.90	1048.0
162.60	309.0
163.10	345.0
164.00	495.0
165.00	4166.0
166.00	3837.0
167.00	21712.0
168.00	13886.0
169.00	2101.0
170.10	677.0
171.00	1051.0
172.00	1148.0
173.00	2319.0
174.10	3975.0
175.10	6714.0
176.00	2325.0
177.00	4394.0
177.90	1114.0
179.00	15904.0
179.90	9283.0
181.00	3490.0
181.90	1100.0
183.20	579.0
184.00	1142.0
185.00	6957.0
186.00	47224.0
187.00	14773.0
188.00	1892.0
189.00	3917.0
189.90	914.0
191.00	2158.0
192.00	5449.0
193.00	7067.0
194.00	1249.0
194.70	341.0
195.10	512.0
195.90	7944.0
196.90	1719.0
198.00	329024.0
199.00	21312.0
200.00	2315.0
201.30	1678.0
202.10	421.0
202.90	4073.0
204.00	15408.0
205.00	20736.0
206.00	96104.0
207.10	15000.0
207.90	4150.0
209.00	2656.0
210.00	1620.0
211.10	4056.0
212.00	973.0
213.20	608.0
215.00	1527.0
216.00	1817.0
217.00	29904.0
218.00	4094.0
218.90	847.0
220.20	1342.0
221.10	9921.0
221.90	3830.0
223.00	6421.0
224.00	60616.0
225.00	11962.0
226.10	1705.0
227.00	29936.0
227.90	3003.0
229.00	4966.0
229.90	802.0

Instrument :
BNA_M
ClientSampleId :
DFTPP43

231.00	1828.0
232.10	576.0
233.90	1787.0
235.00	1585.0
235.80	664.0
236.10	788.0
237.00	2353.0
238.80	951.0
239.90	1247.0
240.90	551.0
241.30	361.0
242.10	3081.0
243.00	4014.0
244.10	40704.0
245.00	6231.0
245.90	9464.0
247.00	2528.0
248.00	484.0
249.00	1981.0
250.20	566.0
251.10	668.0
252.70	1607.0
254.10	1783.0
255.00	227520.0
256.00	29976.0
257.10	3819.0
258.00	14930.0
259.00	2795.0
261.10	366.0
263.90	459.0
265.00	6711.0
265.90	635.0
266.90	648.0
270.10	575.0
270.60	604.0
271.90	983.0
273.00	5240.0
274.00	17360.0
275.00	89032.0
276.00	12619.0
277.00	10279.0
278.00	936.0
279.00	740.0
281.00	925.0
282.00	636.0
283.00	1432.0
283.90	499.0
285.10	1594.0
286.10	357.0
288.90	326.0
292.00	489.0
293.00	3232.0
293.90	304.0
294.80	643.0
296.00	33344.0
296.90	3902.0
298.10	321.0
301.00	509.0
301.90	538.0
303.00	4960.0
304.00	1058.0
310.10	498.0
313.80	911.0
314.90	3778.0
316.00	2201.0
317.00	505.0
321.00	1900.0
322.10	493.0
323.00	7896.0
323.90	1113.0
327.00	2310.0
328.00	1378.0
331.90	914.0
332.90	949.0
334.00	5866.0
335.00	1605.0
341.00	977.0
341.90	520.0
345.80	1400.0
346.90	614.0
352.00	1791.0
353.00	2103.0
354.10	2360.0
354.90	913.0
365.00	13715.0
366.00	1149.0
371.00	981.0
372.00	3988.0
373.00	1199.0
383.00	1469.0
383.90	484.0
390.10	700.0
390.90	630.0
402.10	1578.0
402.90	2147.0

Instrument :
BNA_M
ClientSampleId :
DFTPP43

404.00	860.0
415.10	342.0
421.10	1471.0
422.00	1983.0
423.10	9854.0
424.10	3999.0
425.00	797.0
439.60	369.0
441.10	37632.0
442.10	217600.0
443.10	41912.0
444.10	4046.0

Instrument :
BNA_M
ClientSampleId :
DFTPP43

m/z	Abundance
37.90	1100.0
39.00	6464.0
40.00	880.0
41.00	944.0
43.10	717.0
44.00	676.0
48.90	1223.0
50.00	27752.0
51.00	97072.0
52.00	4853.0
53.10	391.0
55.10	1692.0
55.90	2921.0
57.10	8010.0
58.00	642.0
59.30	547.0
61.00	1552.0
62.00	1655.0
63.00	4871.0
64.00	792.0
65.00	2211.0
67.10	518.0
68.10	1175.0
69.00	99632.0
70.20	665.0
73.00	3590.0
74.10	10976.0
75.00	17744.0
76.00	5693.0
77.00	111152.0
78.00	7253.0
79.00	9501.0
80.00	5269.0
81.00	10203.0
82.00	1811.0
83.10	1605.0
85.10	1760.0
85.90	2383.0
86.90	970.0
87.90	406.0
91.10	2190.0
92.10	2048.0
93.00	13966.0
94.10	1225.0
96.10	835.0
97.20	385.0
97.90	12481.0
99.00	9416.0
99.90	782.0
101.00	4005.0
102.10	390.0
103.00	1604.0
103.90	2753.0
104.90	3197.0
105.90	695.0
107.00	35872.0
108.00	5238.0
108.90	2082.0
110.00	51504.0
111.00	8492.0
111.90	1118.0
112.90	385.0
116.10	1931.0
117.00	36048.0
118.10	2453.0
118.80	609.0
120.10	790.0
121.00	314.0
121.90	1973.0
123.00	3992.0
124.10	1026.0
125.00	1458.0
126.10	1092.0
127.00	96496.0
128.00	6892.0
129.00	40152.0
130.00	4109.0
131.00	1927.0
133.10	846.0
134.00	1120.0
135.00	4408.0
136.00	1577.0
137.10	2623.0
137.80	780.0
138.90	361.0
140.10	831.0
141.00	6778.0
142.00	2408.0
143.10	1649.0
143.90	739.0
145.10	675.0

Instrument :
BNA_M
ClientSampleId :
DFTPP43

145.90	1868.0
147.00	4293.0
147.90	8342.0
149.10	1628.0
149.70	547.0
150.10	454.0
151.10	959.0
151.80	576.0
152.30	405.0
153.00	1888.0
154.20	551.0
154.90	3293.0
155.90	6144.0
156.90	850.0
157.90	1225.0
158.90	1211.0
160.00	2333.0
160.90	2831.0
161.90	1156.0
163.10	453.0
164.90	2761.0
166.00	1225.0
167.00	16896.0
168.00	8385.0
169.00	1243.0
171.10	924.0
172.00	1552.0
172.90	1344.0
174.00	2051.0
175.00	6973.0
176.10	1223.0
177.00	3224.0
177.90	1074.0
179.00	12695.0
180.00	7527.0
181.00	2897.0
182.20	343.0
182.80	440.0
184.00	1363.0
185.00	5226.0
186.00	34784.0
187.00	8971.0
187.90	1118.0
189.00	1581.0
190.10	357.0
190.80	1359.0
192.00	2841.0
193.00	3504.0
194.00	569.0
195.10	599.0
196.00	5259.0
197.00	2522.0
198.00	228608.0
198.90	15055.0
199.90	1475.0
200.90	408.0
201.40	913.0
201.90	799.0
203.00	2399.0
204.00	9839.0
205.00	16824.0
206.00	65576.0
207.00	10910.0
208.00	3271.0
209.00	1278.0
209.90	876.0
210.90	3838.0
211.80	559.0
214.00	354.0
214.90	1074.0
216.00	1745.0
217.00	19856.0
218.00	2270.0
218.80	626.0
219.90	300.0
220.90	7535.0
221.80	3276.0
222.90	4347.0
224.00	40792.0
225.00	10265.0
226.10	1010.0
227.00	19528.0
227.90	3394.0
229.00	3419.0
230.00	689.0
231.00	1395.0
232.20	343.0
232.80	419.0
234.10	1567.0
235.00	1770.0
235.90	1166.0
237.00	1932.0
237.90	415.0
239.00	1302.0
239.90	426.0

Instrument :
BNA_M
ClientSampleId :
DFTPP43

241.00	1364.0
242.00	2157.0
243.10	3015.0
244.00	33000.0
245.10	4729.0
245.90	6196.0
247.00	1563.0
248.00	618.0
248.90	827.0
251.00	513.0
252.10	763.0
253.20	1177.0
254.10	1816.0
255.00	164544.0
256.00	23352.0
257.10	2360.0
258.00	9580.0
258.80	1395.0
260.10	442.0
261.10	329.0
265.00	4673.0
266.00	707.0
271.00	569.0
271.90	503.0
273.00	4595.0
274.00	15392.0
275.00	66712.0
276.00	9248.0
277.00	6782.0
277.80	1029.0
279.00	682.0
281.10	952.0
282.00	336.0
283.00	924.0
285.10	1483.0
292.10	320.0
293.00	1499.0
295.00	641.0
296.00	25664.0
297.00	2763.0
299.10	501.0
303.00	1654.0
309.10	368.0
309.90	541.0
314.10	1298.0
314.90	2502.0
316.00	866.0
321.00	1057.0
321.90	396.0
323.00	8265.0
324.10	1809.0
327.00	1559.0
327.90	818.0
331.90	490.0
333.00	1114.0
334.10	3451.0
335.00	755.0
341.00	643.0
341.90	459.0
346.00	1669.0
347.10	433.0
352.10	2954.0
353.10	1594.0
354.00	2686.0
355.00	600.0
365.00	11072.0
365.90	1423.0
371.10	349.0
372.00	2343.0
373.00	960.0
376.80	416.0
382.80	734.0
383.00	690.0
390.10	482.0
401.90	1323.0
403.10	2408.0
404.00	559.0
421.00	2003.0
422.00	1768.0
423.10	12596.0
424.10	3510.0
441.10	31456.0
442.10	204160.0
443.10	40824.0
444.10	2838.0

Instrument :
BNA_M
ClientSampleId :
DFTPP43

m/z	Abundance
39.20	377.0
39.90	467.0
40.90	749.0
41.90	395.0
43.10	1107.0
44.00	467.0
50.90	577.0
55.10	506.0
57.20	332.0
58.10	928.0
59.20	301.0
67.00	378.0
69.00	408.0
70.90	346.0
73.10	2524.0
73.90	467.0
75.00	852.0
77.00	455.0
78.20	315.0
79.20	370.0
85.00	341.0
90.90	542.0
92.00	455.0
96.10	601.0
96.80	425.0
105.20	658.0
108.90	416.0
119.10	344.0
119.80	420.0
135.10	975.0
147.00	631.0
162.90	438.0
185.00	445.0
191.00	385.0
193.00	435.0
207.00	2583.0
208.00	662.0
209.00	449.0
255.00	973.0
267.10	485.0
281.00	1180.0


Instrument :
BNA_M
ClientSampleId :
DFTPP43

Data Path : Z:\HPCHEM1\BNA_M\Data\BM030617\
Data File : BM009081.D
Acq On : 06 Mar 2017 15:02
Operator : SJ/MA
Sample : DFTPP43
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DFTPP43

Integration File: RTEINT.P

Method : Z:\HPCHEM1\BNA_M\METHODS\SOM-EPA-SIM-BM030317.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Tue Mar 07 03:37:41 2017



AutoFind: Scans 2483, 2484, 2485; Background Corrected with Scan 2476

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	49.8	134671	PASS
68	69	0.00	2	0.7	934	PASS
69	198	0.00	100	50.1	135629	PASS
70	69	0.00	2	0.6	861	PASS
127	198	10	80	44.7	121082	PASS
197	198	0.00	2	0.5	1413	PASS
198	198	100	100	100.0	270592	PASS
199	198	5	9	6.8	18469	PASS
275	198	10	60	26.8	72629	PASS
365	198	1	100	4.2	11236	PASS
441	443	0.01	100	81.8	30000	PASS
442	198	50	100	68.4	185045	PASS
443	442	15	24	19.8	36664	PASS