

m/z	Abundance
37.20	608.0
38.10	819.0
39.10	7676.0
40.20	859.0
40.90	424.0
43.90	591.0
45.10	420.0
48.90	1496.0
50.10	27168.0
51.00	106544.0
52.00	6820.0
53.20	400.0
55.10	483.0
55.90	3776.0
57.00	6022.0
58.00	476.0
60.90	1370.0
61.80	808.0
62.20	1015.0
63.00	5115.0
64.10	487.0
65.00	1305.0
68.00	1996.0
69.00	97872.0
69.90	557.0
72.90	2182.0
74.00	12217.0
75.00	16448.0
76.10	4996.0
77.10	115248.0
78.10	7090.0
79.00	8389.0
80.10	7052.0
80.90	9130.0
82.00	1902.0
82.90	1118.0
84.00	518.0
85.00	1330.0
85.90	2290.0
86.90	1074.0
88.20	343.0
91.00	2131.0
91.90	1845.0
93.00	13376.0
94.00	1501.0
95.00	336.0
96.00	607.0
98.00	11648.0
99.00	11072.0
100.00	736.0
101.10	2933.0
102.00	527.0
103.10	1736.0
103.90	2422.0
105.00	2560.0
106.00	1079.0
107.00	29688.0
108.10	4585.0
109.10	1279.0
110.00	46016.0
111.00	8750.0
111.90	1341.0
113.00	492.0
116.00	1299.0
116.90	35728.0
117.90	2922.0
119.90	303.0
122.10	2631.0
123.00	3427.0
123.90	1254.0
125.00	970.0
126.10	542.0
127.00	77336.0
128.00	6452.0
129.00	36736.0
129.90	3580.0
131.00	485.0
132.00	542.0
133.10	863.0
134.10	871.0
135.10	3754.0
136.10	607.0
136.90	1475.0
138.00	1012.0
139.80	476.0
141.00	6046.0
142.10	991.0
142.90	879.0
145.90	1279.0
147.00	3017.0
147.90	6956.0

Instrument :
BNA_M
ClientSampleId :
DFTPP44

149.00	1424.0
150.00	408.0
150.80	607.0
151.10	639.0
151.80	417.0
153.00	861.0
154.00	1894.0
155.00	4316.0
156.10	3742.0
157.10	1026.0
157.90	1651.0
159.00	1180.0
159.90	2656.0
160.90	2490.0
161.90	1359.0
165.00	2319.0
166.00	1042.0
167.00	16728.0
168.00	7871.0
168.80	718.0
171.00	589.0
171.80	1486.0
172.90	1544.0
174.00	2169.0
175.00	2568.0
176.20	1089.0
177.10	2051.0
178.00	599.0
179.00	7689.0
180.00	6769.0
180.90	3082.0
181.80	383.0
183.80	619.0
184.20	614.0
185.00	3167.0
186.00	29200.0
187.00	8270.0
188.00	1386.0
188.90	1426.0
189.90	543.0
190.90	1228.0
192.00	2239.0
193.00	2315.0
193.90	382.0
196.10	2967.0
197.00	1046.0
198.00	166720.0
198.90	12946.0
199.90	477.0
200.90	726.0
201.20	625.0
201.80	398.0
203.10	1766.0
204.00	8335.0
205.00	11939.0
206.00	46728.0
207.00	7858.0
208.00	1914.0
209.00	973.0
210.10	1348.0
210.80	1996.0
213.90	316.0
214.70	531.0
216.00	2084.0
216.90	13721.0
217.90	2491.0
220.20	676.0
221.00	6520.0
221.70	2658.0
222.90	3202.0
224.00	29368.0
225.00	6953.0
226.00	955.0
227.00	16009.0
228.00	2053.0
228.90	3074.0
231.10	565.0
231.90	465.0
234.10	1076.0
235.00	1086.0
235.80	378.0
237.00	1853.0
238.10	531.0
238.80	325.0
239.80	346.0
240.90	1003.0
241.90	1497.0
243.00	2171.0
244.00	20960.0
245.00	2906.0
246.00	5469.0
246.90	406.0
248.90	871.0
252.90	501.0
254.00	1020.0

Instrument :
BNA_M
ClientSampleId :
DFTPP44

255.00	103328.0
256.00	16728.0
257.00	1443.0
257.90	6702.0
259.00	936.0
263.80	439.0
264.90	3515.0
273.00	3947.0
274.00	11221.0
275.00	41752.0
276.00	5659.0
277.00	4965.0
277.80	524.0
281.00	603.0
281.80	368.0
283.10	704.0
283.90	556.0
284.90	759.0
292.90	1138.0
293.90	488.0
296.00	15465.0
296.90	1585.0
301.80	461.0
303.00	1057.0
303.90	553.0
313.90	1142.0
315.10	2354.0
316.10	528.0
322.90	3188.0
324.00	959.0
326.90	1444.0
327.80	492.0
332.90	464.0
334.00	2654.0
334.90	865.0
346.00	1151.0
347.00	409.0
351.90	1592.0
352.90	655.0
353.90	696.0
355.10	423.0
365.00	6116.0
365.80	397.0
372.00	1496.0
372.80	339.0
402.00	861.0
403.10	1354.0
404.00	420.0
421.00	950.0
421.90	1454.0
423.00	3715.0
424.10	1453.0
425.00	493.0
441.00	12772.0
442.10	80544.0
443.00	13812.0
444.10	2261.0

Instrument :
BNA_M
ClientSampleId :
DFTPP44

m/z	Abundance
37.00	544.0
38.00	2155.0
39.10	8851.0
39.90	740.0
43.80	780.0
48.80	902.0
50.00	34200.0
51.00	132800.0
52.10	6812.0
53.00	504.0
54.90	463.0
55.20	496.0
56.00	5559.0
57.00	10053.0
57.90	308.0
61.00	1614.0
62.10	2412.0
63.00	5762.0
63.90	320.0
65.00	3570.0
65.80	399.0
66.30	361.0
68.00	2147.0
69.00	123936.0
70.00	1077.0
73.10	2219.0
74.00	13251.0
75.00	20312.0
76.10	7126.0
77.10	139968.0
78.00	9541.0
79.00	10618.0
80.00	6703.0
81.00	9184.0
82.10	2300.0
83.00	2012.0
84.90	1222.0
86.00	3755.0
87.10	1585.0
88.00	783.0
91.00	2083.0
92.10	2347.0
93.00	18544.0
93.90	1012.0
95.10	419.0
96.10	620.0
98.00	14704.0
99.10	10278.0
99.90	2075.0
100.90	5345.0
102.90	1531.0
103.90	3295.0
104.90	3554.0
106.00	816.0
107.00	39784.0
108.00	5509.0
109.00	1557.0
110.00	60152.0
111.00	12842.0
112.00	1418.0
113.10	394.0
116.00	1985.0
117.00	47816.0
118.00	3644.0
122.00	2135.0
123.00	4358.0
124.00	2622.0
125.00	2669.0
125.90	691.0
127.00	102088.0
128.00	7578.0
129.00	46928.0
130.00	4377.0
131.00	936.0
131.90	311.0
132.40	329.0
134.00	808.0
134.90	3981.0
135.80	1268.0
137.00	2976.0
138.10	766.0
140.20	801.0
141.00	5074.0
142.00	2548.0
142.80	1843.0
145.90	1613.0
147.00	4877.0
148.00	11569.0
148.90	2897.0
149.80	998.0
150.80	888.0

Instrument :
BNA_M
ClientSampleId :
DFTPP44

151.70	716.0
152.10	481.0
152.90	1733.0
154.00	1751.0
155.10	4788.0
156.00	4434.0
157.00	1152.0
157.50	496.0
158.00	1208.0
158.90	852.0
160.00	2333.0
161.10	4139.0
162.00	1813.0
163.70	531.0
165.00	2640.0
166.00	3219.0
167.00	15801.0
168.00	8837.0
168.90	1725.0
169.90	1011.0
170.90	878.0
171.80	1062.0
173.00	1455.0
173.90	2673.0
175.10	5205.0
176.20	1433.0
177.10	2353.0
177.80	1032.0
178.90	11259.0
180.00	8327.0
181.00	3311.0
181.90	424.0
182.30	434.0
182.80	366.0
183.90	1184.0
185.00	5185.0
186.00	38416.0
187.10	11321.0
188.00	1266.0
189.00	3376.0
191.00	1368.0
191.90	4023.0
193.00	4326.0
195.10	392.0
196.00	5705.0
198.00	238528.0
198.90	13915.0
200.00	1460.0
201.20	760.0
201.80	630.0
202.90	2353.0
204.10	12476.0
205.00	19880.0
206.00	69768.0
207.00	12050.0
208.00	4507.0
208.90	1354.0
210.10	1564.0
211.10	2468.0
212.00	348.0
214.90	661.0
215.90	2989.0
217.00	17928.0
218.00	3877.0
219.60	539.0
221.00	7967.0
222.00	1783.0
222.90	4456.0
224.00	41392.0
225.10	10417.0
226.10	2434.0
227.00	22168.0
228.10	2545.0
229.00	4599.0
230.10	542.0
231.10	1680.0
234.10	1263.0
235.00	1110.0
236.00	1401.0
237.10	2371.0
239.10	502.0
240.00	536.0
241.00	624.0
242.00	2190.0
243.20	2275.0
244.10	26840.0
245.10	3623.0
246.00	8496.0
246.90	1087.0
248.90	895.0
251.00	366.0
252.00	488.0
253.00	943.0
254.00	886.0
255.00	163584.0

Instrument :
BNA_M
ClientSampleId :
DFTPP44

256.00	22472.0
257.00	2108.0
258.00	9447.0
259.10	2066.0
263.80	399.0
265.10	3853.0
266.10	382.0
271.00	1103.0
272.00	883.0
273.00	4015.0
274.00	12731.0
275.00	63384.0
276.00	8372.0
277.00	6327.0
278.00	1206.0
278.70	398.0
281.00	364.0
282.20	496.0
283.00	871.0
284.00	583.0
285.10	1042.0
286.00	447.0
288.80	350.0
289.90	413.0
292.00	565.0
292.90	1463.0
293.80	413.0
294.90	617.0
296.00	21952.0
296.90	3383.0
300.90	382.0
303.10	1772.0
304.00	670.0
307.90	535.0
309.00	704.0
314.00	861.0
315.00	1957.0
316.10	756.0
321.00	496.0
321.90	327.0
323.10	5040.0
324.10	1778.0
326.90	1424.0
327.80	402.0
331.90	724.0
333.00	693.0
334.00	3563.0
335.10	1122.0
341.10	1497.0
342.20	678.0
346.00	1142.0
352.00	2064.0
353.00	955.0
354.10	1568.0
355.10	487.0
365.00	8105.0
365.80	986.0
371.10	393.0
372.10	3122.0
373.00	994.0
382.90	376.0
383.90	358.0
402.00	1679.0
403.00	1580.0
404.10	505.0
421.00	865.0
422.00	1617.0
423.00	8165.0
424.10	2361.0
425.00	719.0
438.30	327.0
441.10	25352.0
442.10	155904.0
443.00	27952.0
444.00	2714.0

Instrument :
BNA_M
ClientSampleId :
DFTPP44

m/z	Abundance
38.10	1313.0
39.00	5088.0
40.00	1247.0
41.00	425.0
44.00	1241.0
49.10	894.0
50.10	22632.0
51.00	85544.0
52.00	4586.0
55.10	642.0
56.00	2219.0
57.00	5721.0
58.20	341.0
60.90	1290.0
62.10	899.0
63.00	3034.0
63.90	429.0
65.00	1704.0
67.90	956.0
69.00	85200.0
70.10	402.0
73.00	1977.0
74.10	8321.0
75.00	14812.0
76.00	4838.0
77.10	88800.0
78.00	6415.0
79.00	7097.0
80.00	5186.0
81.00	7541.0
82.10	1558.0
82.90	1258.0
85.00	1513.0
85.80	2125.0
87.00	842.0
89.10	306.0
91.00	1267.0
91.90	1620.0
93.00	13588.0
93.90	1095.0
95.00	367.0
95.90	445.0
97.00	356.0
98.00	9733.0
98.90	8571.0
100.00	557.0
101.00	4241.0
103.00	1600.0
103.90	2050.0
105.00	1893.0
106.00	652.0
107.00	29848.0
108.00	3834.0
109.00	591.0
110.00	41904.0
110.90	8314.0
111.80	1309.0
113.00	615.0
115.20	379.0
115.80	1327.0
117.00	30192.0
118.00	1614.0
119.90	307.0
122.00	2280.0
122.90	3162.0
123.90	1317.0
125.00	856.0
127.00	80688.0
128.00	5934.0
129.00	33976.0
129.90	2134.0
131.10	945.0
133.90	1007.0
135.00	2613.0
135.90	1138.0
137.00	2541.0
138.10	338.0
139.90	611.0
141.00	6383.0
141.90	1466.0
142.80	1520.0
146.00	969.0
146.90	4031.0
148.00	9880.0
149.10	1455.0
151.00	1009.0
152.00	580.0
153.10	2019.0
154.00	1304.0
155.00	2327.0
156.00	3489.0

Instrument :
BNA_M
ClientSampleId :
DFTPP44

156.70	480.0
157.70	946.0
158.80	722.0
159.10	681.0
159.90	1951.0
161.00	2251.0
161.90	740.0
165.00	1893.0
166.00	1644.0
167.00	11833.0
168.00	6400.0
168.90	523.0
169.80	590.0
171.00	849.0
171.80	1099.0
173.00	1379.0
174.00	2302.0
175.10	3121.0
176.00	790.0
177.00	2047.0
177.90	1059.0
179.00	9288.0
180.10	6067.0
180.90	2740.0
182.00	617.0
183.80	519.0
185.00	4489.0
186.00	29232.0
187.00	8768.0
187.90	1190.0
188.90	2294.0
190.00	315.0
191.10	1242.0
191.90	1871.0
192.90	3960.0
194.00	1034.0
196.00	3984.0
196.90	1567.0
197.90	192320.0
199.00	12822.0
200.10	1018.0
201.10	1317.0
202.00	516.0
203.10	2173.0
204.00	9592.0
204.90	12382.0
206.00	54160.0
207.00	9290.0
207.90	3251.0
208.90	881.0
209.90	1216.0
210.80	1256.0
211.20	1081.0
215.00	1087.0
216.00	1553.0
217.00	19208.0
218.00	2752.0
221.00	7350.0
221.90	1545.0
223.10	3048.0
224.00	31960.0
225.00	8005.0
226.10	867.0
227.00	14158.0
228.00	2152.0
229.00	3489.0
229.80	423.0
231.00	999.0
234.00	1001.0
235.00	997.0
235.80	718.0
237.00	1817.0
238.90	598.0
240.10	696.0
241.00	1798.0
242.00	2786.0
243.00	1812.0
244.10	22664.0
245.00	3431.0
245.90	6436.0
247.00	1847.0
248.90	1121.0
250.90	489.0
253.00	1414.0
253.80	717.0
255.00	131648.0
256.00	21120.0
257.00	2077.0
257.90	8172.0
258.90	2070.0
265.00	3872.0
265.80	301.0
267.00	310.0
272.10	338.0
273.00	4605.0

Instrument :
BNA_M
ClientSampleId :
DFTPP44

273.90	10909.0
275.00	56328.0
276.00	7405.0
276.90	5246.0
278.10	728.0
281.00	531.0
282.90	1143.0
284.00	794.0
284.80	931.0
290.00	318.0
292.00	701.0
293.10	1552.0
294.00	556.0
294.90	585.0
296.00	17760.0
297.00	2508.0
300.90	317.0
303.10	1635.0
304.10	300.0
307.90	450.0
310.00	342.0
314.10	716.0
315.00	2009.0
315.90	627.0
320.90	493.0
323.10	6531.0
324.10	1055.0
327.00	673.0
327.90	1246.0
333.00	505.0
334.00	4849.0
335.00	771.0
341.00	577.0
346.00	1758.0
352.00	1798.0
353.10	1423.0
354.00	1044.0
355.10	697.0
365.00	10089.0
366.00	1401.0
372.10	2151.0
373.10	644.0
383.00	801.0
402.10	623.0
402.90	1634.0
421.10	906.0
421.80	774.0
423.10	8816.0
424.00	2200.0
441.00	24144.0
442.00	153344.0
443.10	25432.0
444.10	2164.0

Instrument :
BNA_M
ClientSampleId :
DFTPP44

m/z	Abundance
39.90	470.0
44.00	1320.0
73.00	925.0
96.00	430.0
207.10	1869.0
207.80	745.0
209.00	464.0
281.10	562.0


Instrument :
BNA_M
ClientSampleId :
DFTPP44

Data Path : Z:\HPCHEM1\BNA_M\Data\BM030617\
Data File : BM009092.D
Acq On : 06 Mar 2017 22:28
Operator : SJ/MA
Sample : DFTPP44
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DFTPP44

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM022317.M
Title : SVOA CALIBRATION
Last Update : Tue Mar 07 07:10:24 2017



AutoFind: Scans 2482, 2483, 2484; Background Corrected with Scan 2477

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	54.4	108296	PASS
68	69	0.00	2	1.7	1699	PASS
69	198	0.00	100	51.4	102336	PASS
70	69	0.00	2	0.7	678	PASS
127	198	10	80	43.5	86704	PASS
197	198	0.00	2	0.4	871	PASS
198	198	100	100	100.0	199189	PASS
199	198	5	9	6.6	13227	PASS
275	198	10	60	27.0	53821	PASS
365	198	1	100	4.1	8103	PASS
441	443	0.01	100	92.7	20756	PASS
442	198	50	100	65.2	129930	PASS
443	442	15	24	17.2	22398	PASS