


Data Path : Z:\HPCHEM1\BNA\_M\Data\BM030618\  
Data File : BM014491.D  
Acq On : 06 Mar 2018 09:04  
Operator : SJ/JU  
Sample : DFTPP09  
Misc :  
ALS Vial : 1 Sample Multiplier: 1

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP09

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA\_M\METHODS\SOM-EPA-BM021418.M  
Title : SVOA CALIBRATION  
Last Update : Mon Mar 05 23:57:03 2018



AutoFind: Scans 2445, 2446, 2447; Background Corrected with Scan 2440

| Target Mass | Rel. to Mass | Lower Limit% | Upper Limit% | Rel. Abn% | Raw Abn | Result Pass/Fail |
|-------------|--------------|--------------|--------------|-----------|---------|------------------|
| 51          | 198          | 10           | 80           | 55.4      | 112002  | PASS             |
| 68          | 69           | 0.00         | 2            | 1.0       | 1122    | PASS             |
| 69          | 198          | 0.00         | 100          | 54.2      | 109552  | PASS             |
| 70          | 69           | 0.00         | 2            | 0.7       | 794     | PASS             |
| 127         | 198          | 10           | 80           | 53.5      | 108325  | PASS             |
| 197         | 198          | 0.00         | 2            | 0.7       | 1463    | PASS             |
| 198         | 198          | 100          | 100          | 100.0     | 202304  | PASS             |
| 199         | 198          | 5            | 9            | 7.1       | 14388   | PASS             |
| 275         | 198          | 10           | 60           | 27.7      | 56080   | PASS             |
| 365         | 198          | 1            | 100          | 4.4       | 8872    | PASS             |
| 441         | 443          | 0.01         | 100          | 81.0      | 20749   | PASS             |
| 442         | 198          | 50           | 100          | 64.9      | 131272  | PASS             |
| 443         | 442          | 15           | 24           | 19.5      | 25624   | PASS             |

| m/z    | Abundance |
|--------|-----------|
| 37.00  | 941.0     |
| 38.10  | 1799.0    |
| 39.10  | 5908.0    |
| 40.00  | 719.0     |
| 41.10  | 555.0     |
| 44.00  | 575.0     |
| 45.10  | 310.0     |
| 49.00  | 1200.0    |
| 50.10  | 31472.0   |
| 51.00  | 97368.0   |
| 52.10  | 5953.0    |
| 53.00  | 418.0     |
| 55.00  | 1152.0    |
| 56.00  | 2415.0    |
| 57.00  | 5919.0    |
| 61.00  | 925.0     |
| 62.10  | 1309.0    |
| 63.00  | 3841.0    |
| 64.10  | 539.0     |
| 64.90  | 1561.0    |
| 67.00  | 452.0     |
| 68.10  | 1635.0    |
| 69.00  | 90328.0   |
| 69.90  | 749.0     |
| 73.10  | 2238.0    |
| 74.00  | 11066.0   |
| 75.00  | 15786.0   |
| 76.00  | 6016.0    |
| 77.00  | 96024.0   |
| 78.10  | 7254.0    |
| 79.00  | 10914.0   |
| 80.00  | 6783.0    |
| 81.00  | 10462.0   |
| 82.00  | 2076.0    |
| 83.00  | 1642.0    |
| 84.90  | 873.0     |
| 86.00  | 3175.0    |
| 87.10  | 1001.0    |
| 88.00  | 479.0     |
| 91.10  | 1012.0    |
| 91.90  | 2093.0    |
| 93.00  | 15135.0   |
| 94.10  | 1379.0    |
| 94.90  | 587.0     |
| 96.10  | 820.0     |
| 97.10  | 644.0     |
| 98.00  | 13445.0   |
| 99.00  | 8352.0    |
| 100.00 | 903.0     |
| 101.00 | 3900.0    |
| 102.90 | 1271.0    |
| 104.00 | 3194.0    |
| 104.90 | 2690.0    |
| 105.80 | 497.0     |
| 107.00 | 31920.0   |
| 107.90 | 4762.0    |
| 109.20 | 944.0     |
| 110.00 | 46400.0   |
| 111.00 | 8802.0    |
| 112.00 | 832.0     |
| 113.10 | 876.0     |
| 115.90 | 1736.0    |
| 117.00 | 47232.0   |
| 117.90 | 2622.0    |
| 119.00 | 317.0     |
| 122.10 | 3544.0    |
| 123.00 | 3526.0    |
| 124.10 | 1432.0    |
| 124.90 | 1602.0    |
| 125.70 | 638.0     |
| 127.00 | 86712.0   |
| 128.00 | 8029.0    |
| 129.00 | 39192.0   |
| 130.00 | 3016.0    |
| 131.10 | 582.0     |
| 133.90 | 1091.0    |
| 135.00 | 4105.0    |
| 136.10 | 1205.0    |
| 137.10 | 1066.0    |
| 137.80 | 347.0     |
| 139.80 | 1129.0    |
| 141.10 | 5892.0    |
| 142.10 | 1920.0    |
| 143.00 | 1972.0    |
| 145.90 | 1205.0    |
| 146.90 | 3463.0    |
| 148.00 | 9491.0    |
| 149.00 | 2225.0    |
| 150.10 | 547.0     |
| 150.90 | 467.0     |
| 152.10 | 433.0     |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP09

|        |          |
|--------|----------|
| 153.00 | 2531.0   |
| 154.00 | 1205.0   |
| 154.90 | 4055.0   |
| 156.00 | 3273.0   |
| 157.00 | 653.0    |
| 158.00 | 1084.0   |
| 158.90 | 1261.0   |
| 159.90 | 1652.0   |
| 160.90 | 2022.0   |
| 162.10 | 1128.0   |
| 165.00 | 2058.0   |
| 166.10 | 1390.0   |
| 167.00 | 14614.0  |
| 167.90 | 6499.0   |
| 169.00 | 849.0    |
| 170.20 | 303.0    |
| 171.10 | 366.0    |
| 171.80 | 946.0    |
| 172.90 | 1186.0   |
| 174.00 | 3354.0   |
| 175.10 | 3580.0   |
| 175.90 | 1167.0   |
| 177.00 | 1549.0   |
| 177.80 | 606.0    |
| 179.00 | 8444.0   |
| 180.00 | 5515.0   |
| 181.00 | 2517.0   |
| 182.20 | 687.0    |
| 183.00 | 850.0    |
| 183.80 | 792.0    |
| 185.00 | 4740.0   |
| 186.00 | 27832.0  |
| 187.00 | 7260.0   |
| 188.10 | 960.0    |
| 188.90 | 1742.0   |
| 191.00 | 973.0    |
| 192.00 | 1731.0   |
| 192.90 | 2896.0   |
| 193.80 | 439.0    |
| 194.10 | 422.0    |
| 196.00 | 4944.0   |
| 197.10 | 1251.0   |
| 198.00 | 144384.0 |
| 199.00 | 10397.0  |
| 200.00 | 797.0    |
| 200.80 | 374.0    |
| 201.40 | 422.0    |
| 203.10 | 1724.0   |
| 204.00 | 7538.0   |
| 205.10 | 9695.0   |
| 206.10 | 45080.0  |
| 207.00 | 6912.0   |
| 207.90 | 2387.0   |
| 208.80 | 714.0    |
| 210.20 | 1378.0   |
| 210.90 | 1266.0   |
| 211.20 | 1399.0   |
| 211.90 | 424.0    |
| 215.10 | 922.0    |
| 215.90 | 1457.0   |
| 217.00 | 14316.0  |
| 218.00 | 1417.0   |
| 219.10 | 473.0    |
| 220.00 | 477.0    |
| 221.10 | 3955.0   |
| 222.00 | 2106.0   |
| 223.00 | 3107.0   |
| 224.00 | 28768.0  |
| 225.00 | 7719.0   |
| 225.90 | 971.0    |
| 226.90 | 13024.0  |
| 228.10 | 1759.0   |
| 229.00 | 2157.0   |
| 231.10 | 1162.0   |
| 234.10 | 615.0    |
| 234.80 | 1086.0   |
| 235.90 | 600.0    |
| 237.00 | 1643.0   |
| 239.00 | 827.0    |
| 240.20 | 562.0    |
| 240.90 | 1093.0   |
| 242.00 | 1349.0   |
| 243.10 | 1346.0   |
| 244.10 | 16872.0  |
| 245.10 | 2563.0   |
| 245.90 | 4547.0   |
| 246.90 | 1547.0   |
| 249.10 | 615.0    |
| 250.80 | 460.0    |
| 251.80 | 447.0    |
| 253.00 | 401.0    |
| 254.00 | 861.0    |
| 255.00 | 96576.0  |
| 256.00 | 15415.0  |
| 256.90 | 1536.0   |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP09

|        |         |
|--------|---------|
| 258.00 | 7372.0  |
| 258.80 | 568.0   |
| 263.30 | 306.0   |
| 265.00 | 2968.0  |
| 266.10 | 784.0   |
| 266.80 | 537.0   |
| 270.10 | 633.0   |
| 272.10 | 634.0   |
| 272.90 | 2132.0  |
| 274.00 | 7319.0  |
| 275.00 | 36440.0 |
| 276.00 | 6189.0  |
| 277.10 | 4750.0  |
| 278.00 | 1128.0  |
| 282.90 | 702.0   |
| 284.90 | 309.0   |
| 293.10 | 1145.0  |
| 294.00 | 532.0   |
| 296.00 | 14497.0 |
| 297.00 | 2454.0  |
| 302.10 | 411.0   |
| 302.90 | 908.0   |
| 303.80 | 316.0   |
| 313.10 | 326.0   |
| 313.90 | 785.0   |
| 315.00 | 2405.0  |
| 315.90 | 588.0   |
| 320.90 | 671.0   |
| 321.90 | 589.0   |
| 323.10 | 4524.0  |
| 323.90 | 921.0   |
| 327.10 | 1271.0  |
| 332.20 | 521.0   |
| 333.20 | 715.0   |
| 334.00 | 2362.0  |
| 335.30 | 375.0   |
| 341.00 | 628.0   |
| 346.00 | 951.0   |
| 351.90 | 863.0   |
| 352.80 | 614.0   |
| 353.10 | 826.0   |
| 353.90 | 1103.0  |
| 355.30 | 511.0   |
| 364.90 | 5343.0  |
| 366.10 | 735.0   |
| 372.00 | 1706.0  |
| 373.00 | 402.0   |
| 383.10 | 510.0   |
| 402.00 | 404.0   |
| 403.10 | 952.0   |
| 421.00 | 941.0   |
| 422.00 | 601.0   |
| 423.10 | 3022.0  |
| 424.00 | 945.0   |
| 441.10 | 10249.0 |
| 442.10 | 72856.0 |
| 443.10 | 13321.0 |
| 444.20 | 1386.0  |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP09

| m/z    | Abundance |
|--------|-----------|
| 37.10  | 1031.0    |
| 38.00  | 1702.0    |
| 39.00  | 12826.0   |
| 40.00  | 1884.0    |
| 40.90  | 420.0     |
| 44.00  | 1313.0    |
| 45.10  | 568.0     |
| 49.20  | 1351.0    |
| 50.10  | 41248.0   |
| 51.10  | 135488.0  |
| 52.00  | 6399.0    |
| 54.90  | 1389.0    |
| 55.90  | 3539.0    |
| 57.00  | 8687.0    |
| 58.00  | 307.0     |
| 60.90  | 1829.0    |
| 62.10  | 2238.0    |
| 63.00  | 6220.0    |
| 64.00  | 1237.0    |
| 65.00  | 2123.0    |
| 67.00  | 660.0     |
| 67.80  | 2008.0    |
| 69.00  | 136128.0  |
| 70.00  | 815.0     |
| 73.10  | 2347.0    |
| 74.00  | 17680.0   |
| 75.00  | 23736.0   |
| 76.00  | 6133.0    |
| 77.10  | 146624.0  |
| 78.10  | 8200.0    |
| 79.00  | 11695.0   |
| 80.00  | 10159.0   |
| 81.00  | 12276.0   |
| 82.00  | 2554.0    |
| 83.00  | 2022.0    |
| 85.00  | 2701.0    |
| 86.00  | 3154.0    |
| 86.90  | 1702.0    |
| 87.90  | 914.0     |
| 88.90  | 456.0     |
| 91.10  | 2957.0    |
| 92.00  | 2819.0    |
| 93.00  | 25792.0   |
| 93.90  | 943.0     |
| 96.10  | 894.0     |
| 98.00  | 19728.0   |
| 99.00  | 12187.0   |
| 100.00 | 984.0     |
| 101.00 | 5907.0    |
| 103.10 | 1150.0    |
| 103.90 | 5485.0    |
| 105.10 | 4013.0    |
| 105.90 | 1231.0    |
| 107.00 | 51376.0   |
| 108.00 | 7416.0    |
| 109.10 | 1042.0    |
| 110.00 | 72024.0   |
| 111.00 | 12052.0   |
| 112.00 | 1902.0    |
| 113.00 | 736.0     |
| 116.10 | 3349.0    |
| 117.00 | 70216.0   |
| 118.00 | 5063.0    |
| 118.80 | 433.0     |
| 119.90 | 359.0     |
| 121.90 | 3809.0    |
| 123.10 | 6342.0    |
| 124.00 | 2413.0    |
| 125.10 | 1943.0    |
| 126.20 | 1029.0    |
| 127.00 | 136000.0  |
| 128.00 | 9207.0    |
| 129.00 | 63608.0   |
| 130.00 | 5348.0    |
| 130.90 | 1778.0    |
| 131.90 | 431.0     |
| 133.10 | 717.0     |
| 133.90 | 2095.0    |
| 134.90 | 4011.0    |
| 135.90 | 2553.0    |
| 137.10 | 3050.0    |
| 137.80 | 327.0     |
| 138.90 | 646.0     |
| 139.80 | 599.0     |
| 140.20 | 942.0     |
| 141.00 | 10624.0   |
| 142.00 | 3831.0    |
| 143.00 | 2077.0    |
| 145.20 | 558.0     |
| 146.10 | 2461.0    |
| 147.00 | 5002.0    |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP09

|        |          |
|--------|----------|
| 147.90 | 14413.0  |
| 149.00 | 2421.0   |
| 150.00 | 1204.0   |
| 150.90 | 1131.0   |
| 151.70 | 598.0    |
| 153.00 | 3246.0   |
| 154.00 | 2320.0   |
| 155.00 | 5249.0   |
| 156.00 | 7927.0   |
| 156.90 | 1542.0   |
| 157.90 | 1227.0   |
| 159.00 | 1082.0   |
| 159.90 | 2587.0   |
| 160.90 | 4259.0   |
| 161.90 | 1468.0   |
| 162.90 | 773.0    |
| 164.90 | 3242.0   |
| 166.10 | 2572.0   |
| 167.00 | 19888.0  |
| 168.00 | 10909.0  |
| 169.00 | 1591.0   |
| 169.90 | 869.0    |
| 171.00 | 834.0    |
| 172.00 | 2037.0   |
| 173.00 | 1752.0   |
| 174.10 | 3454.0   |
| 175.00 | 6363.0   |
| 176.10 | 884.0    |
| 177.00 | 2440.0   |
| 177.90 | 1460.0   |
| 179.00 | 12763.0  |
| 180.00 | 9931.0   |
| 181.00 | 4218.0   |
| 182.00 | 666.0    |
| 182.50 | 383.0    |
| 184.10 | 1183.0   |
| 185.00 | 7045.0   |
| 186.00 | 45992.0  |
| 187.00 | 11217.0  |
| 187.90 | 923.0    |
| 189.00 | 3286.0   |
| 190.00 | 734.0    |
| 190.90 | 2148.0   |
| 191.90 | 4667.0   |
| 193.00 | 3626.0   |
| 194.10 | 543.0    |
| 195.10 | 788.0    |
| 195.90 | 6581.0   |
| 197.10 | 1509.0   |
| 198.00 | 247872.0 |
| 199.00 | 19040.0  |
| 200.10 | 1448.0   |
| 201.20 | 308.0    |
| 201.80 | 972.0    |
| 202.90 | 2917.0   |
| 204.00 | 13805.0  |
| 205.00 | 19680.0  |
| 206.00 | 73304.0  |
| 207.10 | 10672.0  |
| 208.00 | 4718.0   |
| 208.90 | 920.0    |
| 209.90 | 1223.0   |
| 210.90 | 3179.0   |
| 212.80 | 382.0    |
| 215.00 | 951.0    |
| 216.00 | 2168.0   |
| 217.00 | 23104.0  |
| 217.90 | 3518.0   |
| 219.10 | 365.0    |
| 220.90 | 7723.0   |
| 221.80 | 2485.0   |
| 223.10 | 4349.0   |
| 224.00 | 48112.0  |
| 225.00 | 10357.0  |
| 225.90 | 1078.0   |
| 227.00 | 22144.0  |
| 228.00 | 3779.0   |
| 229.00 | 4839.0   |
| 231.10 | 1674.0   |
| 232.10 | 468.0    |
| 232.80 | 305.0    |
| 234.00 | 1935.0   |
| 234.90 | 1744.0   |
| 236.10 | 1127.0   |
| 237.10 | 1681.0   |
| 237.80 | 466.0    |
| 239.00 | 1922.0   |
| 239.90 | 637.0    |
| 240.90 | 1175.0   |
| 242.10 | 2607.0   |
| 242.90 | 2112.0   |
| 244.10 | 33272.0  |
| 245.00 | 4402.0   |
| 246.00 | 7101.0   |
| 247.00 | 1790.0   |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP09

|        |          |
|--------|----------|
| 247.90 | 672.0    |
| 249.00 | 818.0    |
| 251.10 | 429.0    |
| 253.00 | 804.0    |
| 253.90 | 699.0    |
| 254.10 | 680.0    |
| 255.00 | 166528.0 |
| 256.00 | 25136.0  |
| 256.90 | 1305.0   |
| 258.00 | 11860.0  |
| 259.00 | 1735.0   |
| 259.80 | 469.0    |
| 264.00 | 332.0    |
| 265.00 | 5191.0   |
| 265.90 | 684.0    |
| 271.10 | 1155.0   |
| 272.10 | 1028.0   |
| 273.10 | 5016.0   |
| 274.00 | 10764.0  |
| 275.00 | 65576.0  |
| 276.00 | 11049.0  |
| 277.00 | 8297.0   |
| 277.90 | 1683.0   |
| 280.90 | 403.0    |
| 283.00 | 1199.0   |
| 284.00 | 775.0    |
| 284.90 | 778.0    |
| 285.90 | 320.0    |
| 289.00 | 461.0    |
| 292.10 | 632.0    |
| 292.90 | 1216.0   |
| 294.10 | 605.0    |
| 296.00 | 24800.0  |
| 296.90 | 4574.0   |
| 302.80 | 1237.0   |
| 304.10 | 469.0    |
| 308.00 | 423.0    |
| 309.00 | 326.0    |
| 313.90 | 1021.0   |
| 315.00 | 2715.0   |
| 316.00 | 1186.0   |
| 321.00 | 798.0    |
| 321.80 | 431.0    |
| 323.10 | 5565.0   |
| 324.00 | 1099.0   |
| 327.10 | 1606.0   |
| 328.00 | 805.0    |
| 334.00 | 4284.0   |
| 335.00 | 1236.0   |
| 341.20 | 930.0    |
| 345.90 | 1289.0   |
| 351.90 | 1710.0   |
| 353.20 | 1438.0   |
| 353.90 | 1627.0   |
| 355.20 | 472.0    |
| 365.00 | 10384.0  |
| 365.90 | 866.0    |
| 371.20 | 950.0    |
| 372.00 | 2497.0   |
| 373.10 | 942.0    |
| 383.00 | 1171.0   |
| 402.00 | 923.0    |
| 403.00 | 1346.0   |
| 403.90 | 393.0    |
| 421.00 | 1626.0   |
| 421.90 | 1599.0   |
| 423.10 | 8364.0   |
| 424.00 | 1982.0   |
| 440.00 | 310.0    |
| 441.10 | 23320.0  |
| 442.10 | 145344.0 |
| 443.10 | 27880.0  |
| 444.00 | 2378.0   |

Instrument :

BNA\_M

ClientSampleId :

DFTPP09

| m/z    | Abundance |
|--------|-----------|
| 36.90  | 668.0     |
| 38.00  | 993.0     |
| 39.10  | 6976.0    |
| 40.00  | 763.0     |
| 41.10  | 838.0     |
| 43.90  | 649.0     |
| 49.10  | 971.0     |
| 50.00  | 29832.0   |
| 51.00  | 103152.0  |
| 52.10  | 5741.0    |
| 54.90  | 835.0     |
| 56.00  | 2639.0    |
| 57.00  | 6399.0    |
| 60.90  | 587.0     |
| 61.90  | 1736.0    |
| 63.10  | 4604.0    |
| 64.00  | 836.0     |
| 65.00  | 1716.0    |
| 67.10  | 465.0     |
| 68.20  | 1733.0    |
| 69.00  | 102200.0  |
| 70.10  | 820.0     |
| 73.00  | 1919.0    |
| 74.00  | 15344.0   |
| 75.00  | 18000.0   |
| 76.10  | 6882.0    |
| 77.10  | 118080.0  |
| 78.00  | 7390.0    |
| 78.90  | 9092.0    |
| 80.00  | 8235.0    |
| 81.00  | 9272.0    |
| 82.00  | 2095.0    |
| 82.90  | 1819.0    |
| 84.90  | 1465.0    |
| 86.00  | 3290.0    |
| 87.00  | 766.0     |
| 87.90  | 367.0     |
| 88.90  | 469.0     |
| 90.90  | 2902.0    |
| 92.00  | 2484.0    |
| 93.00  | 21352.0   |
| 94.00  | 1092.0    |
| 96.10  | 1259.0    |
| 97.10  | 330.0     |
| 98.00  | 14266.0   |
| 99.00  | 8627.0    |
| 100.00 | 822.0     |
| 101.00 | 5578.0    |
| 103.00 | 1632.0    |
| 103.90 | 2660.0    |
| 104.90 | 2923.0    |
| 105.90 | 1152.0    |
| 107.00 | 41688.0   |
| 108.00 | 5778.0    |
| 109.00 | 1175.0    |
| 110.00 | 59256.0   |
| 111.00 | 9265.0    |
| 112.00 | 1298.0    |
| 113.00 | 834.0     |
| 116.00 | 3386.0    |
| 117.00 | 55112.0   |
| 118.00 | 4128.0    |
| 119.00 | 337.0     |
| 119.90 | 440.0     |
| 122.00 | 2921.0    |
| 123.00 | 4143.0    |
| 123.90 | 1641.0    |
| 124.80 | 1520.0    |
| 126.10 | 1221.0    |
| 127.00 | 102264.0  |
| 128.10 | 8123.0    |
| 129.00 | 51784.0   |
| 130.00 | 3152.0    |
| 131.00 | 1031.0    |
| 131.90 | 329.0     |
| 132.80 | 326.0     |
| 134.00 | 1944.0    |
| 135.00 | 4013.0    |
| 135.90 | 2571.0    |
| 136.90 | 2553.0    |
| 138.10 | 385.0     |
| 139.00 | 352.0     |
| 140.00 | 1059.0    |
| 141.00 | 8583.0    |
| 141.90 | 2456.0    |
| 143.00 | 1597.0    |
| 144.00 | 318.0     |
| 145.00 | 735.0     |
| 146.00 | 1271.0    |
| 146.90 | 3550.0    |
| 148.00 | 13095.0   |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP09

|        |          |
|--------|----------|
| 149.00 | 2934.0   |
| 150.00 | 1073.0   |
| 151.00 | 930.0    |
| 151.90 | 670.0    |
| 152.90 | 1868.0   |
| 154.00 | 1922.0   |
| 155.00 | 4299.0   |
| 156.00 | 5325.0   |
| 157.00 | 1008.0   |
| 157.90 | 1177.0   |
| 158.10 | 1203.0   |
| 159.00 | 1027.0   |
| 160.00 | 2567.0   |
| 161.00 | 3632.0   |
| 162.10 | 993.0    |
| 163.90 | 312.0    |
| 164.80 | 1824.0   |
| 166.00 | 2124.0   |
| 166.90 | 16680.0  |
| 168.00 | 9152.0   |
| 168.90 | 1274.0   |
| 169.90 | 770.0    |
| 171.10 | 844.0    |
| 172.00 | 1340.0   |
| 173.00 | 1376.0   |
| 174.00 | 2340.0   |
| 175.00 | 5674.0   |
| 176.00 | 1414.0   |
| 176.90 | 3928.0   |
| 177.90 | 804.0    |
| 179.00 | 11177.0  |
| 180.00 | 8182.0   |
| 181.00 | 2658.0   |
| 182.00 | 399.0    |
| 183.20 | 581.0    |
| 184.00 | 1018.0   |
| 185.10 | 5829.0   |
| 186.00 | 39632.0  |
| 187.00 | 10912.0  |
| 188.00 | 1352.0   |
| 188.90 | 3545.0   |
| 190.20 | 422.0    |
| 191.00 | 1693.0   |
| 192.00 | 4590.0   |
| 193.00 | 5701.0   |
| 194.00 | 789.0    |
| 195.10 | 535.0    |
| 195.90 | 5643.0   |
| 197.00 | 1629.0   |
| 198.00 | 214656.0 |
| 198.90 | 13728.0  |
| 199.90 | 718.0    |
| 201.60 | 1502.0   |
| 202.90 | 2323.0   |
| 204.00 | 8469.0   |
| 205.00 | 18256.0  |
| 206.00 | 69024.0  |
| 207.10 | 7811.0   |
| 208.10 | 2961.0   |
| 209.00 | 1704.0   |
| 210.10 | 2407.0   |
| 211.00 | 2586.0   |
| 214.90 | 1326.0   |
| 216.00 | 2062.0   |
| 217.00 | 21072.0  |
| 218.00 | 1596.0   |
| 219.20 | 326.0    |
| 220.10 | 772.0    |
| 221.10 | 6379.0   |
| 222.00 | 2550.0   |
| 223.00 | 4602.0   |
| 224.00 | 42072.0  |
| 225.10 | 8837.0   |
| 226.10 | 705.0    |
| 227.00 | 24472.0  |
| 228.00 | 2573.0   |
| 229.00 | 3677.0   |
| 229.90 | 362.0    |
| 230.90 | 2130.0   |
| 233.00 | 348.0    |
| 234.00 | 1078.0   |
| 235.00 | 1932.0   |
| 235.90 | 823.0    |
| 237.10 | 1778.0   |
| 238.00 | 438.0    |
| 238.90 | 1362.0   |
| 240.10 | 961.0    |
| 240.90 | 1102.0   |
| 242.00 | 2544.0   |
| 243.10 | 2411.0   |
| 244.00 | 30736.0  |
| 245.00 | 4918.0   |
| 246.00 | 8581.0   |
| 247.00 | 2006.0   |
| 247.90 | 651.0    |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP09

|        |          |
|--------|----------|
| 248.80 | 1297.0   |
| 250.90 | 519.0    |
| 252.00 | 504.0    |
| 252.80 | 379.0    |
| 254.10 | 1072.0   |
| 255.00 | 163392.0 |
| 256.00 | 22816.0  |
| 257.00 | 1058.0   |
| 258.00 | 9439.0   |
| 259.00 | 1988.0   |
| 260.00 | 691.0    |
| 261.00 | 336.0    |
| 265.10 | 5165.0   |
| 265.90 | 593.0    |
| 271.10 | 405.0    |
| 272.00 | 660.0    |
| 273.10 | 4915.0   |
| 274.00 | 11839.0  |
| 275.00 | 66224.0  |
| 276.10 | 8592.0   |
| 277.00 | 8434.0   |
| 278.00 | 1027.0   |
| 279.00 | 389.0    |
| 281.00 | 519.0    |
| 283.10 | 919.0    |
| 284.00 | 486.0    |
| 285.00 | 1388.0   |
| 292.00 | 676.0    |
| 293.00 | 1842.0   |
| 293.80 | 680.0    |
| 296.00 | 26304.0  |
| 296.90 | 4041.0   |
| 297.80 | 363.0    |
| 300.90 | 381.0    |
| 303.00 | 2811.0   |
| 304.20 | 717.0    |
| 307.90 | 396.0    |
| 308.90 | 509.0    |
| 309.90 | 312.0    |
| 314.10 | 1313.0   |
| 315.00 | 3486.0   |
| 316.00 | 1634.0   |
| 321.00 | 859.0    |
| 323.10 | 8097.0   |
| 324.10 | 1340.0   |
| 326.70 | 951.0    |
| 327.10 | 1056.0   |
| 328.10 | 1416.0   |
| 333.10 | 438.0    |
| 334.00 | 5384.0   |
| 334.90 | 1027.0   |
| 341.00 | 619.0    |
| 346.00 | 1083.0   |
| 352.20 | 2273.0   |
| 352.80 | 625.0    |
| 353.10 | 620.0    |
| 354.00 | 1609.0   |
| 354.90 | 380.0    |
| 364.10 | 374.0    |
| 365.00 | 10891.0  |
| 366.00 | 2061.0   |
| 370.90 | 676.0    |
| 372.00 | 2634.0   |
| 373.00 | 912.0    |
| 382.90 | 1016.0   |
| 383.90 | 362.0    |
| 389.90 | 365.0    |
| 402.10 | 2160.0   |
| 403.10 | 1860.0   |
| 404.00 | 797.0    |
| 420.90 | 1633.0   |
| 422.00 | 678.0    |
| 423.10 | 10765.0  |
| 424.20 | 2688.0   |
| 441.10 | 28680.0  |
| 442.10 | 175616.0 |
| 443.10 | 35672.0  |
| 444.10 | 3349.0   |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP09

| m/z    | Abundance |
|--------|-----------|
| 38.90  | 337.0     |
| 39.80  | 488.0     |
| 43.10  | 912.0     |
| 43.90  | 506.0     |
| 54.80  | 392.0     |
| 55.30  | 430.0     |
| 72.90  | 927.0     |
| 77.00  | 340.0     |
| 90.90  | 348.0     |
| 93.80  | 325.0     |
| 117.00 | 314.0     |
| 177.00 | 359.0     |
| 207.20 | 1497.0    |
| 281.10 | 631.0     |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP09