

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM030624\
 Data File : BM044534.D
 Acq On : 07 Mar 2024 02:08
 Operator : MA/JU
 Sample : P1676-11
 Misc :
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 E06Z7

Quant Time: Mar 07 02:55:32 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM030524.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Mar 05 23:05:11 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.728	152	281374	20.000	ng/u1	0.00
20) Naphthalene-d8	10.510	136	1254828	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.363	164	814255	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.115	188	1630513	20.000	ng/u1	0.00
79) Chrysene-d12	21.303	240	1241351	20.000	ng/u1	0.00
88) Perylene-d12	23.550	264	1134388	20.000	ng/u1	-0.01

System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.287	96	29102	3.508	ng/uL	0.00
4) Pyridine-d5	3.681	84	438812	17.971	ng/u1	0.00
7) Phenol-d5	6.898	99	581899	20.115	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.075	67	387200	21.669	ng/u1	0.00
11) 2-Chlorophenol-d4	7.263	132	443398	20.857	ng/u1	0.00
15) 4-Methylphenol-d8	8.428	113	477082	20.999	ng/u1	-0.01
21) Nitrobenzene-d5	8.892	128	244204	23.111	ng/u1	0.00
24) 2-Nitrophenol-d4	9.604	143	266280	23.376	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.134	165	443178	22.860	ng/u1	0.00
31) 4-Chloroaniline-d4	10.657	131	704191	21.871	ng/u1	0.00
46) Dimethylphthalate-d6	13.786	166	1575081	24.660	ng/u1	0.00
49) Acenaphthylene-d8	14.057	160	1793128	24.547	ng/u1	0.00
54) 4-Nitrophenol-d4	14.574	143	247444	18.681	ng/u1	0.00
60) Fluorene-d10	15.363	176	1367452	25.476	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.486	200	215368	20.348	ng/u1	0.00
73) Anthracene-d10	17.215	188	2073492	26.605	ng/u1	0.00
81) Pyrene-d10	19.509	212	2318316	30.450	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.409	264	1623861	27.235	ng/u1	-0.01

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM030624\
Data File : BM044534.D
Acq On : 07 Mar 2024 02:08
Operator : MA/JU
Sample : P1676-11
Misc :
ALS Vial : 25 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
E06Z7

Quant Time: Mar 07 02:55:32 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM030524.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Mar 05 23:05:11 2024
Response via : Initial Calibration

